

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2558

Instrument ID: P7

Start Date : 07/16/2025

Run Number: LB136512

End Date : 07/16/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
S0	S0	1055	100		100							
S2	S2	1058	100		98							
S3	S3	1105	99		98							
S4	S4	1108	99		97							
S5	S5	1111	100		97							
S6	S6	1114	103		100							
S7	S7	1117	102		99							
S8	S8	1120	102		90							
ICV01	ICV01	1130	102		97							
LLICV01	LLICV01	1140	102		97							
ICB01	ICB01	1147	101		97							
ICSA01	ICSA01	1221	113		104							
ICSAB01	ICSAB01	1228	116		105							
CCV01	CCV01	1251	111		99							
CCB01	CCB01	1259	110		104							
CRI	CRI	1302	111		105							
ZZZZZZ	ZZZZZZ	1306										
ZZZZZZ	ZZZZZZ	1309										
ZZZZZZ	ZZZZZZ	1312										
ZZZZZZ	ZZZZZZ	1315										
ZZZZZZ	ZZZZZZ	1318										
ZZZZZZ	ZZZZZZ	1322										
ZZZZZZ	ZZZZZZ	1325										
ZZZZZZ	ZZZZZZ	1328										
ZZZZZZ	ZZZZZZ	1337										
CCV02	CCV02	1341	108		98							
CCB02	CCB02	1347	107		102							
ZZZZZZ	ZZZZZZ	1350										
ZZZZZZ	ZZZZZZ	1353										
ZZZZZZ	ZZZZZZ	1356										
ZZZZZZ	ZZZZZZ	1403										
ZZZZZZ	ZZZZZZ	1407										
ZZZZZZ	ZZZZZZ	1410										
ZZZZZZ	ZZZZZZ	1413										

Internal Standard %RI Limit: 30 -120

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2558

Instrument ID: P7

Start Date : 07/16/2025

Run Number: LB136512

End Date : 07/16/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
ZZZZZZ	ZZZZZZ	1416										
ZZZZZZ	ZZZZZZ	1419										
ZZZZZZ	ZZZZZZ	1422										
CCV03	CCV03	1440	108		97							
CCB03	CCB03	1450	105		99							
ZZZZZZ	ZZZZZZ	1454										
ZZZZZZ	ZZZZZZ	1500										
ZZZZZZ	ZZZZZZ	1503										
ZZZZZZ	ZZZZZZ	1506										
ZZZZZZ	ZZZZZZ	1509										
ZZZZZZ	ZZZZZZ	1516										
ZZZZZZ	ZZZZZZ	1519										
ZZZZZZ	ZZZZZZ	1523										
ZZZZZZ	ZZZZZZ	1526										
ZZZZZZ	ZZZZZZ	1529										
CCV04	CCV04	1545	102		91							
CCB04	CCB04	1554	102		94							
ZZZZZZ	ZZZZZZ	1558										
ZZZZZZ	ZZZZZZ	1601										
PB168856BL	PB168856BL	1614	103		96							
PB168856BS	PB168856BS	1618	106		99							
Q2555-02L	OU4-TS-29-07	1627	107		98							
Q2555-02MS	OU4-TS-29-07	1630	104		96							
Q2555-02MSD	OU4-TS-29-07	1633	105		97							
Q2555-02A	OU4-TS-29-07	1636	104		96							
Q2555-02DUP	OU4-TS-29-07	1642	108		98							
ZZZZZZ	ZZZZZZ	1646										
CCV05	CCV05	1702	100		90							
CCB05	CCB05	1708	103		97							
PB168856TB	PB168856TB	1716	104		97							
ZZZZZZ	ZZZZZZ	1719										
Q2558-02	OU4-TS-Denal	1722	109		100							
Q2558-04	OU4-TS-Grill	1725	110		101							
ZZZZZZ	ZZZZZZ	1734										

Internal Standard %RI Limit: 30 -120

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2558

Instrument ID: P7

Start Date : 07/16/2025

Run Number: LB136512

End Date : 07/16/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
CCV06	CCV06	1757	108		96							
CCB06	CCB06	1805	104		96							

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2558

Instrument ID: P7

Start Date : 07/16/2025

Run Number: LB136512

End Date : 07/16/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
S0	S0	1055	100									
S2	S2	1058	99									
S3	S3	1105	98									
S4	S4	1108	101									
S5	S5	1111	98									
S6	S6	1114	100									
S7	S7	1117	98									
S8	S8	1120	92									
ICV01	ICV01	1130	104									
LLICV01	LLICV01	1140	103									
ICB01	ICB01	1147	102									
ICSA01	ICSA01	1221	102									
ICSAB01	ICSAB01	1228	104									
CCV01	CCV01	1251	95									
CCB01	CCB01	1259	106									
CRI	CRI	1302	106									
ZZZZZZ	ZZZZZZ	1306										
ZZZZZZ	ZZZZZZ	1309										
ZZZZZZ	ZZZZZZ	1312										
ZZZZZZ	ZZZZZZ	1315										
ZZZZZZ	ZZZZZZ	1318										
ZZZZZZ	ZZZZZZ	1322										
ZZZZZZ	ZZZZZZ	1325										
ZZZZZZ	ZZZZZZ	1328										
ZZZZZZ	ZZZZZZ	1337										
CCV02	CCV02	1341	97									
CCB02	CCB02	1347	102									
ZZZZZZ	ZZZZZZ	1350										
ZZZZZZ	ZZZZZZ	1353										
ZZZZZZ	ZZZZZZ	1356										
ZZZZZZ	ZZZZZZ	1403										
ZZZZZZ	ZZZZZZ	1407										
ZZZZZZ	ZZZZZZ	1410										
ZZZZZZ	ZZZZZZ	1413										

Internal Standard %RI Limit: 30 -120

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2558

Instrument ID: P7

Start Date : 07/16/2025

Run Number: LB136512

End Date : 07/16/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
ZZZZZZ	ZZZZZZ	1416										
ZZZZZZ	ZZZZZZ	1419										
ZZZZZZ	ZZZZZZ	1422										
CCV03	CCV03	1440	95									
CCB03	CCB03	1450	100									
ZZZZZZ	ZZZZZZ	1454										
ZZZZZZ	ZZZZZZ	1500										
ZZZZZZ	ZZZZZZ	1503										
ZZZZZZ	ZZZZZZ	1506										
ZZZZZZ	ZZZZZZ	1509										
ZZZZZZ	ZZZZZZ	1516										
ZZZZZZ	ZZZZZZ	1519										
ZZZZZZ	ZZZZZZ	1523										
ZZZZZZ	ZZZZZZ	1526										
ZZZZZZ	ZZZZZZ	1529										
CCV04	CCV04	1545	96									
CCB04	CCB04	1554	100									
ZZZZZZ	ZZZZZZ	1558										
ZZZZZZ	ZZZZZZ	1601										
PB168856BL	PB168856BL	1614	100									
PB168856BS	PB168856BS	1618	101									
Q2555-02L	OU4-TS-29-07	1627	100									
Q2555-02MS	OU4-TS-29-07	1630	98									
Q2555-02MSD	OU4-TS-29-07	1633	97									
Q2555-02A	OU4-TS-29-07	1636	98									
Q2555-02DUP	OU4-TS-29-07	1642	98									
ZZZZZZ	ZZZZZZ	1646										
CCV05	CCV05	1702	91									
CCB05	CCB05	1708	101									
PB168856TB	PB168856TB	1716	100									
ZZZZZZ	ZZZZZZ	1719										
Q2558-02	OU4-TS-Denal	1722	104									
Q2558-04	OU4-TS-Grill	1725	102									
ZZZZZZ	ZZZZZZ	1734										

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2558

Instrument ID: P7

Start Date : 07/16/2025

Run Number: LB136512

End Date : 07/16/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
CCV06	CCV06	1757	99									
CCB06	CCB06	1805	100									

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2558

Instrument ID: P7

Start Date : 07/22/2025

Run Number: LB136576

End Date : 07/22/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
S0	S0	1442	100		100							
S2	S2	1448	101		104							
S3	S3	1451	101		104							
S4	S4	1455	102		104							
S5	S5	1458	104		105							
S6	S6	1500	105		106							
S7	S7	1503	109		107							
S8	S8	1506	106		100							
ICV01	ICV01	1639	104		107							
LLICV01	LLICV01	1642	104		106							
ICB01	ICB01	1645	104		105							
ICSA01	ICSA01	1648	110		105							
ICSAB01	ICSAB01	1655	114		112							
CCV01	CCV01	1658	113		109							
CCB01	CCB01	1709	106		107							
CRI	CRI	1712	107		107							
ZZZZZZ	ZZZZZZ	1715										
ZZZZZZ	ZZZZZZ	1719										
ZZZZZZ	ZZZZZZ	1721										
ZZZZZZ	ZZZZZZ	1725										
ZZZZZZ	ZZZZZZ	1728										
ZZZZZZ	ZZZZZZ	1731										
ZZZZZZ	ZZZZZZ	1734										
ZZZZZZ	ZZZZZZ	1737										
ZZZZZZ	ZZZZZZ	1740										
CCV02	CCV02	1748	117		114							
CCB02	CCB02	1804	108		110							
ZZZZZZ	ZZZZZZ	1807										
ZZZZZZ	ZZZZZZ	1811										
ZZZZZZ	ZZZZZZ	1814										
ZZZZZZ	ZZZZZZ	1817										
ZZZZZZ	ZZZZZZ	1820										
ZZZZZZ	ZZZZZZ	1823										
ZZZZZZ	ZZZZZZ	1826										

Internal Standard %RI Limit: 30 -120

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2558

Instrument ID: P7

Start Date : 07/22/2025

Run Number: LB136576

End Date : 07/22/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
ZZZZZZ	ZZZZZZ	1829										
ZZZZZZ	ZZZZZZ	1833										
ZZZZZZ	ZZZZZZ	1836										
CCV03	CCV03	1900	114		110							
CCB03	CCB03	1913	110		112							
ZZZZZZ	ZZZZZZ	1918										
ZZZZZZ	ZZZZZZ	1921										
ZZZZZZ	ZZZZZZ	1924										
ZZZZZZ	ZZZZZZ	1927										
Q2555-02DUP	OU4-TS-29-07	1931	113		110							
Q2555-02L	OU4-TS-29-07	1934	113		112							
Q2555-02MS	OU4-TS-29-07	1937	114		109							
Q2555-02MSD	OU4-TS-29-07	1940	113		111							
Q2555-02A	OU4-TS-29-07	1943	113		111							
ZZZZZZ	ZZZZZZ	1946										
CCV04	CCV04	1953	116		109							
CCB04	CCB04	2003	110		111							
Q2558-02	OU4-TS-Denal	2006	111		112							
ZZZZZZ	ZZZZZZ	2010										
Q2555-02DUP	OU4-TS-29-07	2013	114		111							
ZZZZZZ	ZZZZZZ	2025										
Q2555-02DUP	OU4-TS-29-07	2028	111		110							
Q2555-02L	OU4-TS-29-07	2032	112		108							
Q2555-02MS	OU4-TS-29-07	2035	111		108							
Q2555-02MSD	OU4-TS-29-07	2038	110		108							
Q2555-02A	OU4-TS-29-07	2041	110		107							
CCV05	CCV05	2102	115		110							
CCB05	CCB05	2109	110		112							
ZZZZZZ	ZZZZZZ	2112										
Q2555-02DUP	OU4-TS-29-07	2116	114		113							
Q2555-02L	OU4-TS-29-07	2119	111		111							
Q2555-02MS	OU4-TS-29-07	2122	110		110							
Q2555-02MSD	OU4-TS-29-07	2125	111		108							
Q2555-02A	OU4-TS-29-07	2128	111		109							

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2558

Instrument ID: P7

Start Date : 07/22/2025

Run Number: LB136576

End Date : 07/22/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
ZZZZZZ	ZZZZZZ	2132										
ZZZZZZ	ZZZZZZ	2135										
ZZZZZZ	ZZZZZZ	2138										
ZZZZZZ	ZZZZZZ	2141										
CCV06	CCV06	2204	114		105							
CCB06	CCB06	2211	108		110							

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2558

Instrument ID: P7

Start Date : 07/22/2025

Run Number: LB136576

End Date : 07/22/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
S0	S0	1442	100									
S2	S2	1448	101									
S3	S3	1451	102									
S4	S4	1455	101									
S5	S5	1458	104									
S6	S6	1500	104									
S7	S7	1503	104									
S8	S8	1506	99									
ICV01	ICV01	1639	106									
LLICV01	LLICV01	1642	105									
ICB01	ICB01	1645	105									
ICSA01	ICSA01	1648	105									
ICSAB01	ICSAB01	1655	108									
CCV01	CCV01	1658	104									
CCB01	CCB01	1709	106									
CRI	CRI	1712	108									
ZZZZZZ	ZZZZZZ	1715										
ZZZZZZ	ZZZZZZ	1719										
ZZZZZZ	ZZZZZZ	1721										
ZZZZZZ	ZZZZZZ	1725										
ZZZZZZ	ZZZZZZ	1728										
ZZZZZZ	ZZZZZZ	1731										
ZZZZZZ	ZZZZZZ	1734										
ZZZZZZ	ZZZZZZ	1737										
ZZZZZZ	ZZZZZZ	1740										
CCV02	CCV02	1748	107									
CCB02	CCB02	1804	108									
ZZZZZZ	ZZZZZZ	1807										
ZZZZZZ	ZZZZZZ	1811										
ZZZZZZ	ZZZZZZ	1814										
ZZZZZZ	ZZZZZZ	1817										
ZZZZZZ	ZZZZZZ	1820										
ZZZZZZ	ZZZZZZ	1823										
ZZZZZZ	ZZZZZZ	1826										

Internal Standard %RI Limit: 30 -120

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2558

Instrument ID: P7

Start Date : 07/22/2025

Run Number: LB136576

End Date : 07/22/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
ZZZZZZ	ZZZZZZ	1829										
ZZZZZZ	ZZZZZZ	1833										
ZZZZZZ	ZZZZZZ	1836										
CCV03	CCV03	1900	106									
CCB03	CCB03	1913	108									
ZZZZZZ	ZZZZZZ	1918										
ZZZZZZ	ZZZZZZ	1921										
ZZZZZZ	ZZZZZZ	1924										
ZZZZZZ	ZZZZZZ	1927										
Q2555-02DUP	OU4-TS-29-07	1931	111									
Q2555-02L	OU4-TS-29-07	1934	113									
Q2555-02MS	OU4-TS-29-07	1937	112									
Q2555-02MSD	OU4-TS-29-07	1940	111									
Q2555-02A	OU4-TS-29-07	1943	110									
ZZZZZZ	ZZZZZZ	1946										
CCV04	CCV04	1953	106									
CCB04	CCB04	2003	111									
Q2558-02	OU4-TS-Denal	2006	111									
ZZZZZZ	ZZZZZZ	2010										
Q2555-02DUP	OU4-TS-29-07	2013	111									
ZZZZZZ	ZZZZZZ	2025										
Q2555-02DUP	OU4-TS-29-07	2028	113									
Q2555-02L	OU4-TS-29-07	2032	113									
Q2555-02MS	OU4-TS-29-07	2035	113									
Q2555-02MSD	OU4-TS-29-07	2038	113									
Q2555-02A	OU4-TS-29-07	2041	112									
CCV05	CCV05	2102	105									
CCB05	CCB05	2109	108									
ZZZZZZ	ZZZZZZ	2112										
Q2555-02DUP	OU4-TS-29-07	2116	111									
Q2555-02L	OU4-TS-29-07	2119	112									
Q2555-02MS	OU4-TS-29-07	2122	111									
Q2555-02MSD	OU4-TS-29-07	2125	109									
Q2555-02A	OU4-TS-29-07	2128	110									

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2558

Instrument ID: P7

Start Date : 07/22/2025

Run Number: LB136576

End Date : 07/22/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
ZZZZZZ	ZZZZZZ	2132										
ZZZZZZ	ZZZZZZ	2135										
ZZZZZZ	ZZZZZZ	2138										
ZZZZZZ	ZZZZZZ	2141										
CCV06	CCV06	2204	108									
CCB06	CCB06	2211	111									

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2558

Instrument ID: P7

Start Date : 07/24/2025

Run Number: LB136601

End Date : 07/24/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
S0	S0	1042	100		100							
S2	S2	1049	102		102							
S3	S3	1052	104		103							
S4	S4	1055	104		103							
S5	S5	1058	105		103							
S6	S6	1101	109		104							
S7	S7	1104	108		107							
S8	S8	1107	109		99							
ICV01	ICV01	1125	109		105							
LLICV01	LLICV01	1134	106		104							
ICB01	ICB01	1138	106		105							
ICSA01	ICSA01	1141	111		106							
ICSAB01	ICSAB01	1149	114		106							
CCV01	CCV01	1152	113		106							
CCB01	CCB01	1157	108		106							
CRI	CRI	1201	108		105							
ZZZZZZ	ZZZZZZ	1204										
Q2555-02DUP	OU4-TS-29-07	1207	107		104							
Q2555-02L	OU4-TS-29-07	1211	106		104							
Q2555-02MS	OU4-TS-29-07	1214	108		104							
Q2555-02MSD	OU4-TS-29-07	1217	107		104							
Q2555-02A	OU4-TS-29-07	1220	107		105							
ZZZZZZ	ZZZZZZ	1224										
ZZZZZZ	ZZZZZZ	1231										
ZZZZZZ	ZZZZZZ	1234										
CCV02	CCV02	1239	114		105							
CCB02	CCB02	1245	109		105							

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: ACE

SDG No.: Q2558

Instrument ID: P7

Start Date : 07/24/2025

Run Number: LB136601

End Date : 07/24/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
S0	S0	1042	100									
S2	S2	1049	100									
S3	S3	1052	98									
S4	S4	1055	100									
S5	S5	1058	101									
S6	S6	1101	102									
S7	S7	1104	102									
S8	S8	1107	97									
ICV01	ICV01	1125	105									
LLICV01	LLICV01	1134	103									
ICB01	ICB01	1138	104									
ICSA01	ICSA01	1141	102									
ICSAB01	ICSAB01	1149	102									
CCV01	CCV01	1152	99									
CCB01	CCB01	1157	103									
CRI	CRI	1201	103									
ZZZZZZ	ZZZZZZ	1204										
Q2555-02DUP	OU4-TS-29-07	1207	102									
Q2555-02L	OU4-TS-29-07	1211	102									
Q2555-02MS	OU4-TS-29-07	1214	102									
Q2555-02MSD	OU4-TS-29-07	1217	103									
Q2555-02A	OU4-TS-29-07	1220	103									
ZZZZZZ	ZZZZZZ	1224										
ZZZZZZ	ZZZZZZ	1231										
ZZZZZZ	ZZZZZZ	1234										
CCV02	CCV02	1239	102									
CCB02	CCB02	1245	102									