

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2558
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/17/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031089.D **Time Analyzed:** 10:04

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.643	9.547	9.747	832.110	712.500	16.8
2,4,5-TP(Silvex)	9.348	9.253	9.453	727.860	712.500	2.2
2,4-D	8.463	8.367	8.567	779.000	705.000	10.5
2,4-DB	10.220	10.125	10.325	918.400	712.500	28.9
2,4-DCAA	7.329	7.233	7.433	616.310	750.000	-17.8
Dalapon	2.692	2.598	2.798	596.770	682.500	-12.6
DICAMBA	7.519	7.422	7.622	648.490	705.000	-8.0
DICHLORPROP	8.231	8.135	8.335	664.120	705.000	-5.8
Dinoseb	11.435	11.339	11.539	722.950	705.000	2.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2558
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/17/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031089.D **Time Analyzed:** 10:04

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.359	10.258	10.458	601.660	712.500	-15.6
2,4,5-TP(Silvex)	9.932	9.832	10.032	600.130	712.500	-15.8
2,4-D	9.027	8.927	9.127	582.660	705.000	-17.4
2,4-DB	10.927	10.826	11.026	591.890	712.500	-16.9
2,4-DCAA	7.768	7.669	7.869	612.160	750.000	-18.4
Dalapon	2.704	2.608	2.808	566.740	682.500	-17.0
DICAMBA	7.971	7.871	8.071	583.860	705.000	-17.2
DICHLORPROP	8.691	8.591	8.791	580.540	705.000	-17.7
Dinoseb	11.311	11.210	11.410	580.320	705.000	-17.7