

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2558
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/17/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031100.D **Time Analyzed:** 17:04

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.641	9.547	9.747	819.240	712.500	15.0
2,4,5-TP(Silvex)	9.348	9.253	9.453	741.040	712.500	4.0
2,4-D	8.462	8.367	8.567	777.450	705.000	10.3
2,4-DB	10.218	10.125	10.325	902.960	712.500	26.7
2,4-DCAA	7.329	7.233	7.433	712.900	750.000	-4.9
Dalapon	2.694	2.598	2.798	594.830	682.500	-12.8
DICAMBA	7.518	7.422	7.622	658.300	705.000	-6.6
DICHLORPROP	8.230	8.135	8.335	677.760	705.000	-3.9
Dinoseb	11.434	11.339	11.539	728.780	705.000	3.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2558
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/17/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031100.D **Time Analyzed:** 17:04

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.359	10.258	10.458	612.360	712.500	-14.1
2,4,5-TP(Silvex)	9.932	9.832	10.032	609.110	712.500	-14.5
2,4-D	9.027	8.927	9.127	592.540	705.000	-16.0
2,4-DB	10.926	10.826	11.026	595.160	712.500	-16.5
2,4-DCAA	7.768	7.669	7.869	619.530	750.000	-17.4
Dalapon	2.703	2.608	2.808	569.200	682.500	-16.6
DICAMBA	7.971	7.871	8.071	593.510	705.000	-15.8
DICHLORPROP	8.690	8.591	8.791	581.990	705.000	-17.4
Dinoseb	11.311	11.210	11.410	586.790	705.000	-16.8