

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2558
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/17/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031110.D **Time Analyzed:** 21:13

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.640	9.547	9.747	838.970	712.500	17.8
2,4,5-TP(Silvex)	9.347	9.253	9.453	747.360	712.500	4.9
2,4-D	8.461	8.367	8.567	784.350	705.000	11.3
2,4-DB	10.216	10.125	10.325	922.410	712.500	29.5
2,4-DCAA	7.327	7.233	7.433	643.930	750.000	-14.1
Dalapon	2.693	2.598	2.798	589.210	682.500	-13.7
DICAMBA	7.517	7.422	7.622	668.310	705.000	-5.2
DICHLORPROP	8.229	8.135	8.335	684.280	705.000	-2.9
Dinoseb	11.433	11.339	11.539	719.920	705.000	2.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2558
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/17/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031110.D **Time Analyzed:** 21:13

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.358	10.258	10.458	615.280	712.500	-13.6
2,4,5-TP(Silvex)	9.931	9.832	10.032	611.560	712.500	-14.2
2,4-D	9.027	8.927	9.127	595.250	705.000	-15.6
2,4-DB	10.926	10.826	11.026	603.060	712.500	-15.4
2,4-DCAA	7.768	7.669	7.869	624.760	750.000	-16.7
Dalapon	2.706	2.608	2.808	590.940	682.500	-13.4
DICAMBA	7.970	7.871	8.071	595.700	705.000	-15.5
DICHLORPROP	8.690	8.591	8.791	589.070	705.000	-16.4
Dinoseb	11.310	11.210	11.410	580.610	705.000	-17.6