

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2558
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/23/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031202.D **Time Analyzed:** 19:14

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.630	9.536	9.736	822.390	712.500	15.4
2,4,5-TP(Silvex)	9.337	9.242	9.442	772.900	712.500	8.5
2,4-D	8.452	8.356	8.556	782.140	705.000	10.9
2,4-DB	10.206	10.113	10.313	849.240	712.500	19.2
2,4-DCAA	7.321	7.225	7.425	790.520	750.000	5.4
Dalapon	2.690	2.590	2.790	661.670	682.500	-3.1
DICAMBA	7.510	7.414	7.614	750.580	705.000	6.5
DICHLORPROP	8.221	8.124	8.324	743.100	705.000	5.4
Dinoseb	11.421	11.327	11.527	741.760	705.000	5.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2558
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/23/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031202.D **Time Analyzed:** 19:14

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.352	10.256	10.456	754.190	712.500	5.9
2,4,5-TP(Silvex)	9.926	9.829	10.029	751.460	712.500	5.5
2,4-D	9.021	8.924	9.124	738.570	705.000	4.8
2,4-DB	10.920	10.823	11.023	757.480	712.500	6.3
2,4-DCAA	7.764	7.666	7.866	783.550	750.000	4.5
Dalapon	2.705	2.603	2.803	675.100	682.500	-1.1
DICAMBA	7.967	7.868	8.068	755.420	705.000	7.2
DICHLORPROP	8.685	8.588	8.788	732.000	705.000	3.8
Dinoseb	11.305	11.208	11.408	720.870	705.000	2.3