

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2558
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL07 **Date Analyzed:** 07/24/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031233.D **Time Analyzed:** 18:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.629	9.536	9.736	717.230	712.500	0.7
2,4,5-TP(Silvex)	9.335	9.242	9.442	689.480	712.500	-3.2
2,4-D	8.451	8.356	8.556	729.310	705.000	3.4
2,4-DB	10.207	10.113	10.313	688.910	712.500	-3.3
2,4-DCAA	7.319	7.225	7.425	666.340	750.000	-11.2
Dalapon	2.687	2.590	2.790	581.360	682.500	-14.8
DICAMBA	7.509	7.414	7.614	601.250	705.000	-14.7
DICHLORPROP	8.219	8.124	8.324	687.790	705.000	-2.4
Dinoseb	11.419	11.327	11.527	642.410	705.000	-8.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2558
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL07 **Date Analyzed:** 07/24/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031233.D **Time Analyzed:** 18:39

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.352	10.256	10.456	649.390	712.500	-8.9
2,4,5-TP(Silvex)	9.926	9.829	10.029	671.800	712.500	-5.7
2,4-D	9.021	8.924	9.124	696.610	705.000	-1.2
2,4-DB	10.920	10.823	11.023	652.750	712.500	-8.4
2,4-DCAA	7.764	7.666	7.866	693.770	750.000	-7.5
Dalapon	2.704	2.603	2.803	594.300	682.500	-12.9
DICAMBA	7.966	7.868	8.068	650.840	705.000	-7.7
DICHLORPROP	8.685	8.588	8.788	670.240	705.000	-4.9
Dinoseb	11.305	11.208	11.408	616.950	705.000	-12.5