

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2558
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/11/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073711.D **Time Analyzed:** 10:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.634	5.539	5.739	505.600	500.000	1.1
Aroclor-1016-2	5.656	5.560	5.760	529.710	500.000	5.9
Aroclor-1016-3	5.718	5.622	5.822	529.070	500.000	5.8
Aroclor-1016-4	5.816	5.720	5.920	540.550	500.000	8.1
Aroclor-1016-5	6.108	6.012	6.212	541.320	500.000	8.3
Aroclor-1260-1	7.225	7.129	7.329	515.830	500.000	3.2
Aroclor-1260-2	7.479	7.383	7.583	453.160	500.000	-9.4
Aroclor-1260-3	7.836	7.740	7.940	499.080	500.000	-0.2
Aroclor-1260-4	8.061	7.964	8.164	512.730	500.000	2.5
Aroclor-1260-5	8.377	8.283	8.483	510.400	500.000	2.1
Decachlorobiphenyl	10.171	10.076	10.276	53.340	50.000	6.7
Tetrachloro-m-xylene	4.484	4.388	4.588	54.130	50.000	8.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2558
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/11/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073711.D **Time Analyzed:** 10:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.854	4.758	4.958	482.620	500.000	-3.5
Aroclor-1016-2	4.872	4.775	4.975	481.890	500.000	-3.6
Aroclor-1016-3	5.048	4.951	5.151	495.000	500.000	-1.0
Aroclor-1016-4	5.090	4.993	5.193	489.330	500.000	-2.1
Aroclor-1016-5	5.303	5.207	5.407	515.260	500.000	3.1
Aroclor-1260-1	6.334	6.237	6.437	507.790	500.000	1.6
Aroclor-1260-2	6.522	6.425	6.625	512.860	500.000	2.6
Aroclor-1260-3	6.674	6.577	6.777	484.660	500.000	-3.1
Aroclor-1260-4	7.143	7.046	7.246	500.470	500.000	0.1
Aroclor-1260-5	7.386	7.289	7.489	494.820	500.000	-1.0
Decachlorobiphenyl	8.777	8.680	8.880	55.190	50.000	10.4
Tetrachloro-m-xylene	3.777	3.678	3.878	49.570	50.000	-0.9