

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2558
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/12/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073748.D **Time Analyzed:** 01:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.634	5.539	5.739	508.270	500.000	1.7
Aroclor-1016-2	5.656	5.560	5.760	520.810	500.000	4.2
Aroclor-1016-3	5.718	5.622	5.822	527.420	500.000	5.5
Aroclor-1016-4	5.815	5.720	5.920	537.060	500.000	7.4
Aroclor-1016-5	6.107	6.012	6.212	537.930	500.000	7.6
Aroclor-1260-1	7.224	7.129	7.329	552.860	500.000	10.6
Aroclor-1260-2	7.477	7.383	7.583	460.680	500.000	-7.9
Aroclor-1260-3	7.835	7.740	7.940	514.540	500.000	2.9
Aroclor-1260-4	8.058	7.964	8.164	527.960	500.000	5.6
Aroclor-1260-5	8.376	8.283	8.483	519.500	500.000	3.9
Decachlorobiphenyl	10.169	10.076	10.276	53.590	50.000	7.2
Tetrachloro-m-xylene	4.484	4.388	4.588	53.170	50.000	6.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2558
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/12/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073748.D **Time Analyzed:** 01:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.857	4.758	4.958	515.850	500.000	3.2
Aroclor-1016-2	4.874	4.775	4.975	522.960	500.000	4.6
Aroclor-1016-3	5.050	4.951	5.151	521.790	500.000	4.4
Aroclor-1016-4	5.092	4.993	5.193	515.110	500.000	3.0
Aroclor-1016-5	5.305	5.207	5.407	544.070	500.000	8.8
Aroclor-1260-1	6.335	6.237	6.437	532.730	500.000	6.5
Aroclor-1260-2	6.523	6.425	6.625	554.660	500.000	10.9
Aroclor-1260-3	6.675	6.577	6.777	509.860	500.000	2.0
Aroclor-1260-4	7.144	7.046	7.246	529.100	500.000	5.8
Aroclor-1260-5	7.387	7.289	7.489	523.750	500.000	4.8
Decachlorobiphenyl	8.776	8.680	8.880	58.080	50.000	16.2
Tetrachloro-m-xylene	3.779	3.678	3.878	52.420	50.000	4.8