

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2558
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/14/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073754.D **Time Analyzed:** 09:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.636	5.539	5.739	545.940	500.000	9.2
Aroclor-1016-2	5.657	5.560	5.760	564.210	500.000	12.8
Aroclor-1016-3	5.719	5.622	5.822	557.610	500.000	11.5
Aroclor-1016-4	5.816	5.720	5.920	572.720	500.000	14.5
Aroclor-1016-5	6.108	6.012	6.212	588.820	500.000	17.8
Aroclor-1260-1	7.226	7.129	7.329	546.190	500.000	9.2
Aroclor-1260-2	7.479	7.383	7.583	494.960	500.000	-1.0
Aroclor-1260-3	7.837	7.740	7.940	554.810	500.000	11.0
Aroclor-1260-4	8.060	7.964	8.164	568.090	500.000	13.6
Aroclor-1260-5	8.378	8.283	8.483	570.270	500.000	14.1
Decachlorobiphenyl	10.169	10.076	10.276	58.580	50.000	17.2
Tetrachloro-m-xylene	4.485	4.388	4.588	59.660	50.000	19.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2558
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/14/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073754.D **Time Analyzed:** 09:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.854	4.758	4.958	470.210	500.000	-6.0
Aroclor-1016-2	4.871	4.775	4.975	462.780	500.000	-7.4
Aroclor-1016-3	5.047	4.951	5.151	473.250	500.000	-5.4
Aroclor-1016-4	5.089	4.993	5.193	473.560	500.000	-5.3
Aroclor-1016-5	5.302	5.207	5.407	488.830	500.000	-2.2
Aroclor-1260-1	6.332	6.237	6.437	478.860	500.000	-4.2
Aroclor-1260-2	6.521	6.425	6.625	473.190	500.000	-5.4
Aroclor-1260-3	6.672	6.577	6.777	469.920	500.000	-6.0
Aroclor-1260-4	7.142	7.046	7.246	487.210	500.000	-2.6
Aroclor-1260-5	7.383	7.289	7.489	477.720	500.000	-4.5
Decachlorobiphenyl	8.774	8.680	8.880	52.100	50.000	4.2
Tetrachloro-m-xylene	3.776	3.678	3.878	47.790	50.000	-4.4