

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2558

Client ID : SSTDCCC040

Date Analyzed: 07/16/2025

Lab File ID: BF143117.D

Time Analyzed: 19:39

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	135149	6.969	510164	8.25	254065	10.00
UPPER LIMIT	270298	7.469	1020330	8.751	508130	10.504
LOWER LIMIT	67574.5	6.469	255082	7.751	127033	9.504
EPA SAMPLE NO.						
01 OU4-TS-Denali-070925	148809	6.97	543974	8.25	239712	10.00
02 OU4-TS-Grillo-OG-070925	142532	6.97	490921	8.25	206095	10.00
03 TP-17MS	153802	6.97	579798	8.25	274749	10.00
04 TP-17MSD	154642	6.97	573690	8.25	274130	10.00

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.