

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2558

Client ID : SSTDICCC040

Date Analyzed: 07/17/2025

Lab File ID: BF143144.D

Time Analyzed: 13:03

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	167098	6.963	641413	8.25	331881	10.00
UPPER LIMIT	334196	7.463	1282830	8.745	663762	10.498
LOWER LIMIT	83549	6.463	320707	7.745	165941	9.498
EPA SAMPLE NO.						
01 PB168813BL	146767	6.96	569126	8.24	303459	10.00

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.