

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2558

Client ID: SSTDCCC040

Date Analyzed: 07/16/2025

Lab File ID: BF143117.D

Time Analyzed: 19:39

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	397173	11.492	210714	14.127	255524	15.633
UPPER LIMIT	794346	11.992	421428	14.627	511048	16.133
LOWER LIMIT	198587	10.992	105357	13.627	127762	15.133
EPA SAMPLE NO.						
01 OU4-TS-Denali-070925	333996	11.49	302047	14.13	277642	15.63
02 OU4-TS-Grillo-OG-070925	317891	11.49	291247	14.13	224121	15.64
03 TP-17MS	404136	11.49	268901	14.13	324370	15.64
04 TP-17MSD	404556	11.49	270975	14.13	316202	15.64

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.