

Report of Analysis

Client:	PSEG	Date Collected:	07/10/25
Project:	Aubudon Gas District	Date Received:	07/10/25
Client Sample ID:	AUD-25-0117DL2	SDG No.:	Q2561
Lab Sample ID:	Q2561-05DL2	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	98.7
Sample Wt/Vol:	30.05	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO112252.D	1000	07/11/25 08:20	07/14/25 12:53	PB168809

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	17.0	UD	4.00	17.0	mg/Kg
11104-28-2	Aroclor-1221	17.0	UD	4.10	17.0	mg/Kg
11141-16-5	Aroclor-1232	17.0	UD	3.80	17.0	mg/Kg
53469-21-9	Aroclor-1242	17.0	UD	4.10	17.0	mg/Kg
12672-29-6	Aroclor-1248	17.0	UD	6.00	17.0	mg/Kg
11097-69-1	Aroclor-1254	82.0	D	3.20	17.0	mg/Kg
37324-23-5	Aroclor-1262	17.0	UD	5.10	17.0	mg/Kg
11100-14-4	Aroclor-1268	17.0	UD	3.60	17.0	mg/Kg
11096-82-5	Aroclor-1260	17.0	UD	3.30	17.0	mg/Kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	0	*	30 - 150	%	SPK: 20
2051-24-3	Decachlorobiphenyl	0	*	30 - 150	%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit