



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

### **G Environmental**

**Project Name: Banker**

**Project # N/A**

**Order ID # Q2562**

**Test Name: Anions Group1**

### **A. Number of Samples and Date of Receipt:**

4 Water samples were received on 07/10/2025.

### **B. Parameters:**

According to the Chain of Custody document, the following analyses were requested: Anions Group1. This data package contains results for Anions Group1.

### **C. Analytical Techniques:**

The analysis of Anions Group1 was based on method 300.0.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

Sample 17 BANK was diluted due to high concentrations for Chloride & Sample SUPPLY was diluted due to high concentrations for Chloride & Sample MW25 was diluted due to high concentrations for Chloride & Sample MW1 was diluted due to high concentrations for Chloride.

The Blank Spike met requirements for all compounds.

The Duplicate analysis met criteria for all compounds.

The Matrix Spike (SUPPLYMS) analysis met criteria for all compounds except for Chloride due to matrix interferences.

The Matrix Spike Duplicate (SUPPLYMSD) analysis met criteria for all compounds except for Chloride due to matrix interferences.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

### **E. Additional Comments:**

Sample Q2562-03 was analyzed with 10x straight dilution for Chloride due to high Concentration.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_