

Instrument ID: IC-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136460

Review By	rubina	Review On	7/15/2025 8:26:14 AM
Supervise By	Iwona	Supervise On	7/15/2025 12:00:12 PM
SubDirectory	LB136460	Test	Anions
STD. NAME	STD REF.#		
ICAL Standard	WP113629,WP113630,WP113631,WP113632,WP113633,WP113634,WP113635		
ICV Standard	WP113636		
CCV Standard	WP113914		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP113915		
Chk Standard	WP113637,WP113638		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	STD1	STD1	CAL1	06/23/25 15:17	All standards, samples, and	RM/IZ	OK
2	STD2	STD2	CAL2	06/23/25 15:38	QC are filtered through	RM/IZ	OK
3	STD3	STD3	CAL3	06/23/25 16:00	0.45um, filter lot W3160	RM/IZ	OK
4	STD4	STD4	CAL4	06/23/25 16:21		RM/IZ	OK
5	STD5	STD5	CAL5	06/23/25 16:43		RM/IZ	OK
6	STD6	STD6	CAL6	06/23/25 17:04		RM/IZ	OK
7	STD7	STD7	CAL7	06/23/25 17:25		RM/IZ	OK
8	ICV1	ICV1	ICV	06/23/25 17:47		RM/IZ	OK
9	ICB1	ICB1	ICB	06/23/25 18:08		RM/IZ	OK
10	CCV1	CCV1	CCV	07/11/25 09:09		RM/IZ	OK
11	CCB1	CCB1	CCB	07/11/25 09:52		RM/IZ	OK
12	LB136460BLW	LB136460BLW	MB	07/11/25 10:14		RM/IZ	OK
13	LB136460BSW	LB136460BSW	LCS	07/11/25 10:35		RM/IZ	OK
14	Q2565-01	MOO-25-0192-0193	SAM	07/11/25 10:57		RM/IZ	OK
15	Q2562-01	17 BANK	SAM	07/11/25 11:18	Cl is high	RM/IZ	Dilution
16	Q2562-02	SUPPLY	SAM	07/11/25 11:40	Cl is high	RM/IZ	Dilution
17	Q2562-02MS	SUPPLYMS	MS	07/11/25 12:01	9.5ml of sample, 0.5mL W3092	RM/IZ	OK
18	Q2562-02MSD	SUPPLYMSD	MSD	07/11/25 12:23	9.5ml of sample, 0.5mL W3092	RM/IZ	OK

Instrument ID: IC-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136460

Review By	rubina	Review On	7/15/2025 8:26:14 AM
Supervise By	Iwona	Supervise On	7/15/2025 12:00:12 PM
SubDirectory	LB136460	Test	Anions
STD. NAME	STD REF.#		
ICAL Standard	WP113629,WP113630,WP113631,WP113632,WP113633,WP113634,WP113635		
ICV Standard	WP113636		
CCV Standard	WP113914		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP113915		
Chk Standard	WP113637,WP113638		

19	Q2562-03	MW25	SAM	07/11/25 12:45	10X stright, Still Cl is high	RM/IZ	Dilution
20	Q2562-04	MW1	SAM	07/11/25 13:06	Cl is high	RM/IZ	Dilution
21	CCV2	CCV2	CCV	07/11/25 14:11		RM/IZ	OK
22	CCB2	CCB2	CCB	07/11/25 14:32		RM/IZ	OK
23	Q2562-02DL	SUPPLYDL	SAM	07/11/25 14:54	100X for Cl	RM/IZ	Confirms
24	Q2562-03DL	MW25DL	SAM	07/11/25 15:15	2000X for Cl	RM/IZ	Confirms
25	Q2562-04DL	MW1DL	SAM	07/11/25 15:37	200X for Cl	RM/IZ	Confirms
26	Q2562-01DL	17 BANKDL	SAM	07/11/25 15:58	200X for Cl	RM/IZ	Confirms
27	CCV3	CCV3	CCV	07/11/25 16:20		RM/IZ	OK
28	CCB3	CCB3	CCB	07/11/25 16:41		RM/IZ	OK