

Initial and Continuing Calibration Verification

Client: G Environmental

SDG No.: Q2562

Project: Banker

RunNo.: LB136460

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	9.7	10	97	90-110	06/23/2025
Chloride	mg/L	2.9	3	97	90-110	06/23/2025
Fluoride	mg/L	2	2	100	90-110	06/23/2025
Nitrite	mg/L	2.9	3	97	90-110	06/23/2025
Nitrate	mg/L	2.4	2.5	96	90-110	06/23/2025
Sulfate	mg/L	14.3	15	95	90-110	06/23/2025
Orthophosphate as P	mg/L	4.9	5	98	90-110	06/23/2025
Sample ID: CCV1						
Bromide	mg/L	9.8	10	98	90-110	07/11/2025
Chloride	mg/L	2.9	3	97	90-110	07/11/2025
Fluoride	mg/L	2	2	100	90-110	07/11/2025
Nitrite	mg/L	2.9	3	97	90-110	07/11/2025
Nitrate	mg/L	2.4	2.5	96	90-110	07/11/2025
Sulfate	mg/L	14.6	15	97	90-110	07/11/2025
Orthophosphate as P	mg/L	4.5	5	90	90-110	07/11/2025
Sample ID: CCV2						
Bromide	mg/L	9.7	10	97	90-110	07/11/2025
Chloride	mg/L	3.2	3	107	90-110	07/11/2025
Fluoride	mg/L	2	2	100	90-110	07/11/2025
Nitrite	mg/L	2.9	3	97	90-110	07/11/2025
Nitrate	mg/L	2.3	2.5	92	90-110	07/11/2025
Sulfate	mg/L	14.6	15	97	90-110	07/11/2025
Orthophosphate as P	mg/L	4.9	5	98	90-110	07/11/2025
Sample ID: CCV3						
Bromide	mg/L	9.7	10	97	90-110	07/11/2025
Chloride	mg/L	3.2	3	107	90-110	07/11/2025
Fluoride	mg/L	2	2	100	90-110	07/11/2025
Nitrite	mg/L	2.9	3	97	90-110	07/11/2025
Nitrate	mg/L	2.3	2.5	92	90-110	07/11/2025
Sulfate	mg/L	14.6	15	97	90-110	07/11/2025
Orthophosphate as P	mg/L	4.9	5	98	90-110	07/11/2025