



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q2562 **Order ID:** Q2562
Client: G Environmental **Project ID:** Banker

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : 17 BANK Q2562-01	17 BANK	Water	Sodium	312000		128	500	ug/L
Client ID : SUPPLY Q2562-02	SUPPLY	Water	Sodium	63700	D	640	2500	ug/L
Client ID : MW25 Q2562-03	MW25	Water	Sodium	4690000	D	1280	5000	ug/L
Client ID : MW1 Q2562-04	MW1	Water	Sodium	455000		128	500	ug/L



SAMPLE DATA

Report of Analysis

Client:	G Environmental	Date Collected:	07/09/25
Project:	Banker	Date Received:	07/10/25
Client Sample ID:	17 BANK	SDG No.:	Q2562
Lab Sample ID:	Q2562-01	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-23-5	Sodium	312000	1	128		500	ug/L	07/11/25 08:55	07/16/25 14:10	6020B	3010A

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals ICP-Group1			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits

Report of Analysis

Client:	G Environmental	Date Collected:	07/09/25
Project:	Banker	Date Received:	07/10/25
Client Sample ID:	SUPPLY	SDG No.:	Q2562
Lab Sample ID:	Q2562-02	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-23-5	Sodium	63700	D	5	640	2500	ug/L	07/11/25 08:55	07/16/25 15:16	6020B	3010A

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals ICP-Group1			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits

Report of Analysis

Client:	G Environmental	Date Collected:	07/09/25
Project:	Banker	Date Received:	07/10/25
Client Sample ID:	MW25	SDG No.:	Q2562
Lab Sample ID:	Q2562-03	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-23-5	Sodium	4690000	D	10	1280	5000	ug/L	07/11/25 08:55	07/16/25 16:01	6020B	3010A

Color Before:	Brown	Clarity Before:	Cloudy	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals ICP-Group1			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits

Report of Analysis

Client:	G Environmental	Date Collected:	07/09/25
Project:	Banker	Date Received:	07/10/25
Client Sample ID:	MW1	SDG No.:	Q2562
Lab Sample ID:	Q2562-04	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-23-5	Sodium	455000	1	128		500	ug/L	07/11/25 08:55	07/16/25 15:09	6020B	3010A

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals ICP-Group1			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q2562

Contract: GENV01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Sodium	128	+/-250	U	500	P	07/16/2025	11:47	LB136512

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q2562

Contract: GENV01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Sodium	128	+/-250	U	500	P	07/16/2025	12:59	LB136512
CCB02	Sodium	128	+/-250	U	500	P	07/16/2025	13:47	LB136512
CCB03	Sodium	128	+/-250	U	500	P	07/16/2025	14:50	LB136512
CCB04	Sodium	128	+/-250	U	500	P	07/16/2025	15:54	LB136512
CCB05	Sodium	128	+/-250	U	500	P	07/16/2025	17:08	LB136512
CCB06	Sodium	128	+/-250	U	500	P	07/16/2025	18:05	LB136512

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q2562

Instrument: P7

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168807BL		WATER		Batch Number:	PB168807		Prep Date:	07/11/2025	
	Sodium	128	<250	U	500	P	07/16/2025	14:03	LB136512



METAL CALIBRATION DATA

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q2562

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Sodium	2170	2000	109	90 - 110	P	07/16/2025	11:30	LB136512

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q2562

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Sodium	461	500	92	80 - 120	P	07/16/2025	11:40	LB136512

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

Contract: GENV01

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

SDG No.: Q2562

Lab Code: ACE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Sodium	265000	250000	106	90 - 110	P	07/16/2025	12:51	LB136512
CCV02	Sodium	257000	250000	103	90 - 110	P	07/16/2025	13:41	LB136512
CCV03	Sodium	257000	250000	103	90 - 110	P	07/16/2025	14:40	LB136512
CCV04	Sodium	257000	250000	103	90 - 110	P	07/16/2025	15:45	LB136512
CCV05	Sodium	257000	250000	103	90 - 110	P	07/16/2025	17:02	LB136512
CCV06	Sodium	261000	250000	104	90 - 110	P	07/16/2025	17:57	LB136512



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental

SDG No.: Q2562

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI	Sodium	473	500	95	70 - 130	P	07/16/2025	13:02	LB136512

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: G Environmental

SDG No.: Q2562

Contract: GENV01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P7

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Sodium	101000	100000	101	0	0	07/16/2025	12:21	LB136512
ICSAB01	Sodium	99100	100000	99	0	0	07/16/2025	12:28	LB136512



METAL QC DATA

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>G Environmental</u>	level: <u>low</u>	sdg no.: <u>Q2562</u>
contract: <u>GENV01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q2562-02</u>	client id: <u>SUPPLYMS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2562-02MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Sodium	ug/L	75 - 125	105000	D	63700	D	50000	82		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>G Environmental</u>	level: <u>low</u>	sdg no.: <u>Q2562</u>
contract: <u>GENV01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q2562-02</u>	client id: <u>SUPPLYMSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2562-02MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Sodium	ug/L	75 - 125	105000	D	63700	D	50000	83		P



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals
- 5b -

Client: G Environmental

SDG No.: Q2562

Contract: GENV01

Lab Code: ACE

Matrix:

Level: LOW

Client ID:

Sample ID:

Spiked ID:

Analyte	Units	Acceptance Limit %R	C	Sample Result	C	Spike Added	% Recovery	Qual	M
---------	-------	------------------------	---	------------------	---	----------------	---------------	------	---

.....

- A
- B
- C
- D
- E
- F
- G
- H
- I
- J

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>G Environmental</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q2562</u>
Contract:	<u>GENV01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q2562-02</u>	Client ID:	<u>SUPPLYDUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q2562-02DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Sodium	ug/L	20	63700	D	66600	D	4		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>G Environmental</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q2562</u>
Contract:	<u>GENV01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q2562-02MS</u>	Client ID:	<u>SUPPLYMSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q2562-02MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Sodium	ug/L	20	105000	D	105000	D	0		P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental

SDG No.: Q2562

Contract: GENV01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168807BS Sodium	ug/L	50000	50000		100	80 - 120	P

FORM 8A

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: G Environmental

Contract: GENV01

Lab Code: ACE

SDG No.: Q2562

Instrument ID: P7

Start Date : 07/16/2025

Run Number: LB136512

End Date : 07/16/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 6Li	Q	Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q
S0	S0	1055	100		100		100		100		100	
S2	S2	1058	100		99		100		99		102	
S3	S3	1105	100		97		98		98		100	
S4	S4	1108	98		96		97		97		100	
S5	S5	1111	96		95		97		96		99	
S6	S6	1114	96		97		100		96		102	
S7	S7	1117	93		98		100		97		103	
S8	S8	1120	90		103		104		96		103	
ICV01	ICV01	1130	89		106		108		107		103	
LLICV01	LLICV01	1140	90		108		108		107		102	
ICB01	ICB01	1147	91		105		107		105		102	
ICSA01	ICSA01	1221	83		105		109		105		113	
ICSAB01	ICSAB01	1228	81		102		114		109		118	
CCV01	CCV01	1251	80		104		111		104		111	
CCB01	CCB01	1259	83		105		112		110		111	
CRI	CRI	1302	85		105		113		112		112	
ZZZZZZ	ZZZZZZ	1306										
ZZZZZZ	ZZZZZZ	1309										
ZZZZZZ	ZZZZZZ	1312										
ZZZZZZ	ZZZZZZ	1315										
ZZZZZZ	ZZZZZZ	1318										
ZZZZZZ	ZZZZZZ	1322										
ZZZZZZ	ZZZZZZ	1325										
ZZZZZZ	ZZZZZZ	1328										
ZZZZZZ	ZZZZZZ	1337										
CCV02	CCV02	1341	85		104		110		103		110	
CCB02	CCB02	1347	85		104		110		109		109	
ZZZZZZ	ZZZZZZ	1350										
ZZZZZZ	ZZZZZZ	1353										
ZZZZZZ	ZZZZZZ	1356										
PB168807BL	PB168807BL	1403	86		104		110		109		109	
PB168807BS	PB168807BS	1407	87		107		111		108		111	
Q2562-01	17 BANK	1410	91		114		120		113		118	
Q2562-02	SUPPLY	1413	91		115		118		114		118	
Q2562-02DUP	SUPPLYDUP	1416	91		114		122		116		118	
Q2562-02L	SUPPLYL	1419	85		107		111		109		108	

Internal Standard %RI Limit: 70 - 130

FORM 8A

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: G Environmental

Contract: GENV01

Lab Code: ACE

SDG No.: Q2562

Instrument ID: P7

Start Date : 07/16/2025

Run Number: LB136512

End Date : 07/16/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 6Li	Q	Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q
Q2562-02MS	SUPPLYMS	1422	104		132	*	138	*	130		135	*
CCV03	CCV03	1440	89		108		111		103		108	
CCB03	CCB03	1450	87		107		109		106		106	
Q2562-02MS	SUPPLYMS	1454	107		133	*	136	*	129		132	*
Q2562-02MSD	SUPPLYMSD	1500	92		117		122		113		118	
Q2562-02A	SUPPLYA	1503	99		128		134	*	127		131	*
Q2562-03	MW25	1506	65	*	103		113		94		100	
Q2562-04	MW1	1509	84		115		123		113		114	
Q2562-02	SUPPLY	1516	87		112		113		111		112	
Q2562-02DUP	SUPPLYDUP	1519	89		109		112		108		109	
Q2562-02L	SUPPLYL	1523	91		107		109		106		105	
Q2562-02MS	SUPPLYMS	1526	91		109		111		108		108	
Q2562-02MSD	SUPPLYMSD	1529	91		110		111		108		106	
CCV04	CCV04	1545	88		110		110		101		104	
CCB04	CCB04	1554	88		108		109		107		103	
Q2562-02A	SUPPLYA	1558	93		113		113		109		107	
Q2562-03	MW25	1601	91		112		111		105		107	
ZZZZZZ	ZZZZZZ	1614										
ZZZZZZ	ZZZZZZ	1618										
ZZZZZZ	ZZZZZZ	1627										
ZZZZZZ	ZZZZZZ	1630										
ZZZZZZ	ZZZZZZ	1633										
ZZZZZZ	ZZZZZZ	1636										
ZZZZZZ	ZZZZZZ	1642										
ZZZZZZ	ZZZZZZ	1646										
CCV05	CCV05	1702	84		105		107		98		103	
CCB05	CCB05	1708	88		109		110		108		108	

FORM 8A

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: G Environmental

Contract: GENV01

Lab Code: ACE

SDG No.: Q2562

Instrument ID: P7

Start Date : 07/16/2025

Run Number: LB136512

End Date : 07/16/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q	Element 165Ho	Q
S0	S0	1055	100		100		100		100		100	
S2	S2	1058	99		99		100		99		99	
S3	S3	1105	97		96		97		97		98	
S4	S4	1108	99		98		99		99		101	
S5	S5	1111	97		97		97		99		98	
S6	S6	1114	99		99		100		102		101	
S7	S7	1117	101		101		99		103		101	
S8	S8	1120	107		107		99		103		102	
ICV01	ICV01	1130	112		113		110		107		107	
LLICV01	LLICV01	1140	110		110		110		106		104	
ICB01	ICB01	1147	109		108		108		105		103	
ICSA01	ICSA01	1221	108		110		109		108		108	
ICSAB01	ICSAB01	1228	107		109		109		109		108	
CCV01	CCV01	1251	106		106		104		105		104	
CCB01	CCB01	1259	108		111		114		108		108	
CRI	CRI	1302	111		112		113		110		108	
ZZZZZZ	ZZZZZZ	1306										
ZZZZZZ	ZZZZZZ	1309										
ZZZZZZ	ZZZZZZ	1312										
ZZZZZZ	ZZZZZZ	1315										
ZZZZZZ	ZZZZZZ	1318										
ZZZZZZ	ZZZZZZ	1322										
ZZZZZZ	ZZZZZZ	1325										
ZZZZZZ	ZZZZZZ	1328										
ZZZZZZ	ZZZZZZ	1337										
CCV02	CCV02	1341	109		110		104		108		107	
CCB02	CCB02	1347	107		108		109		106		106	
ZZZZZZ	ZZZZZZ	1350										
ZZZZZZ	ZZZZZZ	1353										
ZZZZZZ	ZZZZZZ	1356										
PB168807BL	PB168807BL	1403	107		108		110		108		108	
PB168807BS	PB168807BS	1407	109		109		109		110		108	
Q2562-01	17 BANK	1410	115		118		114		114		114	
Q2562-02	SUPPLY	1413	117		116		116		115		115	
Q2562-02DUP	SUPPLYDUP	1416	116		116		114		111		111	
Q2562-02L	SUPPLYL	1419	107		109		109		105		105	

Internal Standard %RI Limit: 70 - 130

FORM 8A

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: G Environmental

Contract: GENV01

Lab Code: ACE

SDG No.: Q2562

Instrument ID: P7

Start Date : 07/16/2025

Run Number: LB136512

End Date : 07/16/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q	Element 165Ho	Q
Q2562-02MS	SUPPLYMS	1422	131	*	131	*	126		127		125	
CCV03	CCV03	1440	109		108		103		108		106	
CCB03	CCB03	1450	106		108		108		105		103	
Q2562-02MS	SUPPLYMS	1454	130		131	*	127		127		128	
Q2562-02MSD	SUPPLYMSD	1500	116		117		115		115		114	
Q2562-02A	SUPPLYA	1503	129		128		125		127		125	
Q2562-03	MW25	1506	114		116		98		102		101	
Q2562-04	MW1	1509	119		121		115		114		114	
Q2562-02	SUPPLY	1516	111		111		111		108		106	
Q2562-02DUP	SUPPLYDUP	1519	109		108		109		105		105	
Q2562-02L	SUPPLYL	1523	109		109		110		105		104	
Q2562-02MS	SUPPLYMS	1526	111		110		109		106		106	
Q2562-02MSD	SUPPLYMSD	1529	109		108		107		105		105	
CCV04	CCV04	1545	112		112		105		107		107	
CCB04	CCB04	1554	108		110		108		106		104	
Q2562-02A	SUPPLYA	1558	114		114		113		110		109	
Q2562-03	MW25	1601	111		114		108		109		108	
ZZZZZZ	ZZZZZZ	1614										
ZZZZZZ	ZZZZZZ	1618										
ZZZZZZ	ZZZZZZ	1627										
ZZZZZZ	ZZZZZZ	1630										
ZZZZZZ	ZZZZZZ	1633										
ZZZZZZ	ZZZZZZ	1636										
ZZZZZZ	ZZZZZZ	1642										
ZZZZZZ	ZZZZZZ	1646										
CCV05	CCV05	1702	105		106		99		101		102	
CCB05	CCB05	1708	108		110		110		106		106	

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: G Environmental

Contract: GENV01

Lab Code: ACE

SDG No.: Q2562

Instrument ID: P7

Start Date : 07/16/2025

Run Number: LB136512

End Date : 07/16/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
S0	S0	1055	100		100							
S2	S2	1058	100		98							
S3	S3	1105	99		98							
S4	S4	1108	99		97							
S5	S5	1111	100		97							
S6	S6	1114	103		100							
S7	S7	1117	102		99							
S8	S8	1120	102		90							
ICV01	ICV01	1130	102		97							
LLICV01	LLICV01	1140	102		97							
ICB01	ICB01	1147	101		97							
ICSA01	ICSA01	1221	113		104							
ICSAB01	ICSAB01	1228	116		105							
CCV01	CCV01	1251	111		99							
CCB01	CCB01	1259	110		104							
CRI	CRI	1302	111		105							
ZZZZZZ	ZZZZZZ	1306										
ZZZZZZ	ZZZZZZ	1309										
ZZZZZZ	ZZZZZZ	1312										
ZZZZZZ	ZZZZZZ	1315										
ZZZZZZ	ZZZZZZ	1318										
ZZZZZZ	ZZZZZZ	1322										
ZZZZZZ	ZZZZZZ	1325										
ZZZZZZ	ZZZZZZ	1328										
ZZZZZZ	ZZZZZZ	1337										
CCV02	CCV02	1341	108		98							
CCB02	CCB02	1347	107		102							
ZZZZZZ	ZZZZZZ	1350										
ZZZZZZ	ZZZZZZ	1353										
ZZZZZZ	ZZZZZZ	1356										
PB168807BL	PB168807BL	1403	108		102							
PB168807BS	PB168807BS	1407	110		103							
Q2562-01	17 BANK	1410	116		106							
Q2562-02	SUPPLY	1413	116		107							

Internal Standard %RI Limit: 70 -130

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: G Environmental
Lab Code: ACE
Instrument ID: P7
Run Number: LB136512

Contract: GENV01
SDG No.: Q2562
Start Date : 07/16/2025
End Date : 07/16/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
Q2562-02DUP	SUPPLYDUP	1416	116		107							
Q2562-02L	SUPPLYL	1419	109		101							
Q2562-02MS	SUPPLYMS	1422	134	*	121							
CCV03	CCV03	1440	108		97							
CCB03	CCB03	1450	105		99							
Q2562-02MS	SUPPLYMS	1454	132	*	119							
Q2562-02MSD	SUPPLYMSD	1500	118		107							
Q2562-02A	SUPPLYA	1503	130		120							
Q2562-03	MW25	1506	97		80							
Q2562-04	MW1	1509	114		101							
Q2562-02	SUPPLY	1516	109		99							
Q2562-02DUP	SUPPLYDUP	1519	107		99							
Q2562-02L	SUPPLYL	1523	104		97							
Q2562-02MS	SUPPLYMS	1526	106		97							
Q2562-02MSD	SUPPLYMSD	1529	104		98							
CCV04	CCV04	1545	102		91							
CCB04	CCB04	1554	102		94							
Q2562-02A	SUPPLYA	1558	107		98							
Q2562-03	MW25	1601	106		94							
ZZZZZZ	ZZZZZZ	1614										
ZZZZZZ	ZZZZZZ	1618										
ZZZZZZ	ZZZZZZ	1627										
ZZZZZZ	ZZZZZZ	1630										
ZZZZZZ	ZZZZZZ	1633										
ZZZZZZ	ZZZZZZ	1636										
ZZZZZZ	ZZZZZZ	1642										
ZZZZZZ	ZZZZZZ	1646										
CCV05	CCV05	1702	100		90							
CCB05	CCB05	1708	103		97							

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: G Environmental

Contract: GENV01

Lab Code: ACE

SDG No.: Q2562

Instrument ID: P7

Start Date : 07/16/2025

Run Number: LB136512

End Date : 07/16/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
S0	S0	1055	100									
S2	S2	1058	99									
S3	S3	1105	98									
S4	S4	1108	101									
S5	S5	1111	98									
S6	S6	1114	100									
S7	S7	1117	98									
S8	S8	1120	92									
ICV01	ICV01	1130	104									
LLICV01	LLICV01	1140	103									
ICB01	ICB01	1147	102									
ICSA01	ICSA01	1221	102									
ICSAB01	ICSAB01	1228	104									
CCV01	CCV01	1251	95									
CCB01	CCB01	1259	106									
CRI	CRI	1302	106									
ZZZZZZ	ZZZZZZ	1306										
ZZZZZZ	ZZZZZZ	1309										
ZZZZZZ	ZZZZZZ	1312										
ZZZZZZ	ZZZZZZ	1315										
ZZZZZZ	ZZZZZZ	1318										
ZZZZZZ	ZZZZZZ	1322										
ZZZZZZ	ZZZZZZ	1325										
ZZZZZZ	ZZZZZZ	1328										
ZZZZZZ	ZZZZZZ	1337										
CCV02	CCV02	1341	97									
CCB02	CCB02	1347	102									
ZZZZZZ	ZZZZZZ	1350										
ZZZZZZ	ZZZZZZ	1353										
ZZZZZZ	ZZZZZZ	1356										
PB168807BL	PB168807BL	1403	105									
PB168807BS	PB168807BS	1407	104									
Q2562-01	17 BANK	1410	106									
Q2562-02	SUPPLY	1413	107									

Internal Standard %RI Limit: 70 -130

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: G Environmental

Contract: GENV01

Lab Code: ACE

SDG No.: Q2562

Instrument ID: P7

Start Date : 07/16/2025

Run Number: LB136512

End Date : 07/16/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
Q2562-02DUP	SUPPLYDUP	1416	104									
Q2562-02L	SUPPLYL	1419	100									
Q2562-02MS	SUPPLYMS	1422	117									
CCV03	CCV03	1440	95									
CCB03	CCB03	1450	100									
Q2562-02MS	SUPPLYMS	1454	118									
Q2562-02MSD	SUPPLYMSD	1500	105									
Q2562-02A	SUPPLYA	1503	115									
Q2562-03	MW25	1506	88									
Q2562-04	MW1	1509	104									
Q2562-02	SUPPLY	1516	101									
Q2562-02DUP	SUPPLYDUP	1519	100									
Q2562-02L	SUPPLYL	1523	101									
Q2562-02MS	SUPPLYMS	1526	100									
Q2562-02MSD	SUPPLYMSD	1529	99									
CCV04	CCV04	1545	96									
CCB04	CCB04	1554	100									
Q2562-02A	SUPPLYA	1558	102									
Q2562-03	MW25	1601	99									
ZZZZZZ	ZZZZZZ	1614										
ZZZZZZ	ZZZZZZ	1618										
ZZZZZZ	ZZZZZZ	1627										
ZZZZZZ	ZZZZZZ	1630										
ZZZZZZ	ZZZZZZ	1633										
ZZZZZZ	ZZZZZZ	1636										
ZZZZZZ	ZZZZZZ	1642										
ZZZZZZ	ZZZZZZ	1646										
CCV05	CCV05	1702	91									
CCB05	CCB05	1708	101									

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

SUPPLYL

Lab Name: Alliance **Contract:** GENV01
Lab Code: ACE **Lb No.:** lb136512 **Lab Sample ID :** Q2562-02L **SDG No.:** Q2562
Matrix (soil/water): Water **Level (low/med):** LOW
Concentration Units: ug/L

Analyte	Initial Sample Result (I)		Serial Dilution Result (S)		% Differ- ence	Q	M
Sodium	63700	D	70400	D	10		P

metals
- 14 -
ANALYSIS RUN LOG

Client: G Environmental

Contract: GENV01

Lab code: ACE

Sdg no.: Q2562

Instrument id number: _____

Method: _____

Run number: LB136512

Start date: 07/16/2025

End date: 07/16/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1055	Na
S2	S2	1	1058	Na
S3	S3	1	1105	Na
S4	S4	1	1108	Na
S5	S5	1	1111	Na
S6	S6	1	1114	Na
S7	S7	1	1117	Na
S8	S8	1	1120	Na
ICV01	ICV01	1	1130	Na
LLICV01	LLICV01	1	1140	Na
ICB01	ICB01	1	1147	Na
ICSA01	ICSA01	1	1221	Na
ICSAB01	ICSAB01	1	1228	Na
CCV01	CCV01	1	1251	Na
CCB01	CCB01	1	1259	Na
CRI	CRI	1	1302	Na
CCV02	CCV02	1	1341	Na
CCB02	CCB02	1	1347	Na
PB168807BL	PB168807BL	1	1403	Na
PB168807BS	PB168807BS	1	1407	Na
Q2562-01	17 BANK	1	1410	Na
CCV03	CCV03	1	1440	Na
CCB03	CCB03	1	1450	Na
Q2562-04	MW1	1	1509	Na
Q2562-02	SUPPLY	5	1516	Na
Q2562-02DUP	SUPPLYDUP	5	1519	Na
Q2562-02L	SUPPLYL	25	1523	Na
Q2562-02MS	SUPPLYMS	5	1526	Na
Q2562-02MSD	SUPPLYMSD	5	1529	Na
CCV04	CCV04	1	1545	Na
CCB04	CCB04	1	1554	Na
Q2562-03	MW25	10	1601	Na
CCV05	CCV05	1	1702	Na
CCB05	CCB05	1	1708	Na
CCV06	CCV06	1	1757	Na
CCB06	CCB06	1	1805	Na



METAL PREPARATION & INSTRUMENT DATA

LAB CHRONICLE

OrderID:	Q2562	OrderDate:	7/10/2025 12:19:00 PM
Client:	G Environmental	Project:	Banker
Contact:	Gary Landis	Location:	O33

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2562-01	17 BANK	Water	Metals ICP-Group1	6020B	07/09/25	07/11/25	07/16/25	07/10/25
Q2562-02	SUPPLY	Water	Metals ICP-Group1	6020B	07/09/25	07/11/25	07/16/25	07/10/25
Q2562-03	MW25	Water	Metals ICP-Group1	6020B	07/09/25	07/11/25	07/16/25	07/10/25
Q2562-04	MW1	Water	Metals ICP-Group1	6020B	07/09/25	07/11/25	07/16/25	07/10/25



METAL PREPARATION & ANALYICAL SUMMARY

Metals

- 13 -

SAMPLE PREPARATION SUMMARY

Client: G Environmental

SDG No.: Q2562

Contract: GENV01

Lab Code: ACE

Method: _____

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(mL)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB168807							
PB168807BL	PB168807BL	MB	WATER	07/11/2025	50.0	50.0	
PB168807BS	PB168807BS	LCS	WATER	07/11/2025	50.0	50.0	
Q2562-01	17 BANK	SAM	WATER	07/11/2025	50.0	50.0	
Q2562-02	SUPPLY	SAM	WATER	07/11/2025	50.0	50.0	
Q2562-02DUP	SUPPLYDUP	DUP	WATER	07/11/2025	50.0	50.0	
Q2562-02MS	SUPPLYMS	MS	WATER	07/11/2025	50.0	50.0	
Q2562-02MSD	SUPPLYMSD	MSD	WATER	07/11/2025	50.0	50.0	
Q2562-03	MW25	SAM	WATER	07/11/2025	50.0	50.0	
Q2562-04	MW1	SAM	WATER	07/11/2025	50.0	50.0	

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB136512

Review By	Janvi	Review On	7/17/2025 1:50:29 PM
Supervise By	jaswal	Supervise On	7/18/2025 9:13:11 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86136,MP861437,MP86149,MP86151,MP86153,MP86155,MP86157,MP86160,MP86161		
ICV Standard	MP86162		
CCV Standard	MP86163		
ICSA Standard	MP86164,MP86165		
CRI Standard	MP86160		
LCS Standard			
Chk Standard	MP86167,MP86168		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE	TUNE	07/16/25 09:47		Jaswal	OK
2	S0	S0	CAL1	07/16/25 10:55		Jaswal	OK
3	S2	S2	CAL3	07/16/25 10:58		Jaswal	OK
4	S3	S3	CAL4	07/16/25 11:05		Jaswal	OK
5	S4	S4	CAL5	07/16/25 11:08		Jaswal	OK
6	S5	S5	CAL6	07/16/25 11:11		Jaswal	OK
7	S6	S6	CAL7	07/16/25 11:14		Jaswal	OK
8	S7	S7	CAL8	07/16/25 11:17		Jaswal	OK
9	S8	S8	CAL9	07/16/25 11:20		Jaswal	OK
10	ICV01	ICV01	ICV	07/16/25 11:30		Jaswal	OK
11	LLICV01	LLICV01	LLICV	07/16/25 11:40		Jaswal	OK
12	ICB01	ICB01	ICB	07/16/25 11:47		Jaswal	OK
13	ICSA01	ICSA01	ICSA	07/16/25 12:21		Jaswal	OK
14	ICSAB01	ICSAB01	ICSAB	07/16/25 12:28		Jaswal	OK
15	CCV01	CCV01	CCV	07/16/25 12:51		Jaswal	OK
16	CCB01	CCB01	CCB	07/16/25 12:59		Jaswal	OK
17	CRI	CRI	CRDL	07/16/25 13:02		Jaswal	OK
18	Q2535-02DL	70125-LIQDL	SAM	07/16/25 13:06	Fe,Mn High	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB136512

Review By	Janvi	Review On	7/17/2025 1:50:29 PM
Supervise By	jaswal	Supervise On	7/18/2025 9:13:11 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86136,MP861437,MP86149,MP86151,MP86153,MP86155,MP86157,MP86160,MP86161		
ICV Standard	MP86162		
CCV Standard	MP86163		
ICSA Standard	MP86164,MP86165		
CRI Standard	MP86160		
LCS Standard			
Chk Standard	MP86167,MP86168		

19	Q2551-01	293	SAM	07/16/25 13:09	K,Na High	Jaswal	Dilution
20	Q2550-01DL	3326DL	SAM	07/16/25 13:12		Jaswal	OK
21	Q2550-01DUPDL	3326DUPDL	DUP	07/16/25 13:15		Jaswal	OK
22	Q2550-01LDL	3326LDL	SD	07/16/25 13:18		Jaswal	OK
23	Q2550-01MSDL	3326MSDL	MS	07/16/25 13:22		Jaswal	OK
24	Q2550-01MSDDL	3326MSDDL	MSD	07/16/25 13:25		Jaswal	OK
25	Q2550-01ADL	3326ADL	PS	07/16/25 13:28		Jaswal	OK
26	Q2535-02DL2	70125-LIQDL2	SAM	07/16/25 13:37	25X For Fe,Mn	Jaswal	Confirms
27	CCV02	CCV02	CCV	07/16/25 13:41		Jaswal	OK
28	CCB02	CCB02	CCB	07/16/25 13:47		Jaswal	OK
29	PB168792BL	PB168792BL	MB	07/16/25 13:50		Jaswal	OK
30	PB168792BS	PB168792BS	LCS	07/16/25 13:53		Jaswal	OK
31	Q2551-01DL	293DL	SAM	07/16/25 13:56	5X For K,Na	Jaswal	Confirms
32	PB168807BL	PB168807BL	MB	07/16/25 14:03		Jaswal	OK
33	PB168807BS	PB168807BS	LCS	07/16/25 14:07		Jaswal	OK
34	Q2562-01	17 BANK	SAM	07/16/25 14:10		Jaswal	OK
35	Q2562-02	SUPPLY	SAM	07/16/25 14:13	INT_STD 45Sc(2) Fail	Jaswal	Dilution
36	Q2562-02DUP	SUPPLYDUP	DUP	07/16/25 14:16	INT_STD 45Sc(2) Fail	Jaswal	Dilution
37	Q2562-02L	SUPPLYL	SD	07/16/25 14:19	INT_STD 45Sc(2) Fail	Jaswal	Dilution
38	Q2562-02MS	SUPPLYMS	MS	07/16/25 14:22	INT_STD 45Sc(2) Fail	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB136512

Review By	Janvi	Review On	7/17/2025 1:50:29 PM
Supervise By	jaswal	Supervise On	7/18/2025 9:13:11 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86136,MP861437,MP86149,MP86151,MP86153,MP86155,MP86157,MP86160,MP86161		
ICV Standard	MP86162		
CCV Standard	MP86163		
ICSA Standard	MP86164,MP86165		
CRI Standard	MP86160		
LCS Standard			
Chk Standard	MP86167,MP86168		

39	CCV03	CCV03	CCV	07/16/25 14:40		Jaswal	OK
40	CCB03	CCB03	CCB	07/16/25 14:50		Jaswal	OK
41	Q2562-02MSRE	SUPPLYMS	MS	07/16/25 14:54	Not Use	Jaswal	Not Ok
42	Q2562-02MSD	SUPPLYMSD	MSD	07/16/25 15:00	INT_STD 45Sc(2) Fail	Jaswal	Dilution
43	Q2562-02A	SUPPLYA	PS	07/16/25 15:03		Jaswal	OK
44	Q2562-03	MW25	SAM	07/16/25 15:06	Na High	Jaswal	Dilution
45	Q2562-04	MW1	SAM	07/16/25 15:09		Jaswal	OK
46	Q2562-02DL	SUPPLYDL	SAM	07/16/25 15:16	5X For INT_STD 45Sc(2)	Jaswal	Confirms
47	Q2562-02DUPDL	SUPPLYDUPDL	DUP	07/16/25 15:19	5X For INT_STD 45Sc(2)	Jaswal	Confirms
48	Q2562-02LDL	SUPPLYLDL	SD	07/16/25 15:23	5X For INT_STD 45Sc(2)	Jaswal	Confirms
49	Q2562-02MSDL	SUPPLYMSDL	MS	07/16/25 15:26	5X For INT_STD 45Sc(2)	Jaswal	Confirms
50	Q2562-02MSDDL	SUPPLYMSDDL	MSD	07/16/25 15:29	5X For INT_STD 45Sc(2)	Jaswal	Confirms
51	CCV04	CCV04	CCV	07/16/25 15:45		Jaswal	OK
52	CCB04	CCB04	CCB	07/16/25 15:54		Jaswal	OK
53	Q2562-02ADL	SUPPLYADL	PS	07/16/25 15:58	Not Required	Jaswal	Not Ok
54	Q2562-03DL	MW25DL	SAM	07/16/25 16:01	10X For Na	Jaswal	Confirms
55	PB168856BL	PB168856BL	MB	07/16/25 16:14		Jaswal	OK
56	PB168856BS	PB168856BS	LCS	07/16/25 16:18		Jaswal	OK
57	Q2555-02LDL	OU4-TS-29-070925LDL	SD	07/16/25 16:27	INT_STD 89Y(1&2) Fail	Jaswal	Dilution
58	Q2555-02MSDL	OU4-TS-29-070925MSDL	MS	07/16/25 16:30	INT_STD 89Y(1&2) Fail	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB136512

Review By	Janvi	Review On	7/17/2025 1:50:29 PM
Supervise By	jaswal	Supervise On	7/18/2025 9:13:11 AM

STD. NAME	STD REF.#
ICAL Standard	MP86136,MP861437,MP86149,MP86151,MP86153,MP86155,MP86157,MP86160,MP86161
ICV Standard	MP86162
CCV Standard	MP86163
ICSA Standard	MP86164,MP86165
CRI Standard	MP86160
LCS Standard	
Chk Standard	MP86167,MP86168

59	Q2555-02MSDDL	OU4-TS-29-070925M	MSD	07/16/25 16:33	INT_STD 89Y(1&2) Fail	Jaswal	Dilution
60	Q2555-02ADL	OU4-TS-29-070925A	PS	07/16/25 16:36		Jaswal	OK
61	Q2555-02DUPDL	OU4-TS-29-070925D	DUP	07/16/25 16:42	INT_STD 89Y(1&2) Fail	Jaswal	Dilution
62	Q2555-02DL	OU4-TS-29-070925D	SAM	07/16/25 16:46	INT_STD 89Y(1&2) Fail	Jaswal	Dilution
63	CCV05	CCV05	CCV	07/16/25 17:02		Jaswal	OK
64	CCB05	CCB05	CCB	07/16/25 17:08		Jaswal	OK
65	PB168856TB	PB168856TB	MB	07/16/25 17:16		Jaswal	OK
66	Q2555-04DL	OU4-TS-30-070925D	SAM	07/16/25 17:19	INT_STD 89Y(1&2) Fail	Jaswal	Dilution
67	Q2558-02DL	OU4-TS-Denali-07092	SAM	07/16/25 17:22	INT_STD 89Y(1) Fail	Jaswal	Dilution
68	Q2558-04DL	OU4-TS-Grillo-OG-07	SAM	07/16/25 17:25		Jaswal	OK
69	Q2555-04DL2	OU4-TS-30-070925D	SAM	07/16/25 17:34	Not Required	Jaswal	Not Ok
70	CCV06	CCV06	CCV	07/16/25 17:57		Jaswal	OK
71	CCB06	CCB06	CCB	07/16/25 18:05		Jaswal	OK

SOP ID : M3010A-Digestion-17

SDG No : N/A

Matrix : WATER

Pipette ID: ICP A

Balance ID : N/A

Filter paper ID : N/A

pH Strip ID : M6069

Hood ID : #3

Block ID: 1. HOT BLOCK #1 2. N/A

Start Digest Date: 07/11/2025 Time : 08:55 Temp : 96 °C

End Digest Date: 07/11/2025 Time : 11:05 Temp : 96 °C

Digestion tube ID: M5595

Block thermometer ID: MET-DIG. #1

Dig Technician Signature: *SKS*

Supervisor Signature: *[Signature]*

Temp : 1. 96°C 2. N/A

Standard Name	MLS USED	STD REF. # FROM LOG
Spike Sol 1	0.50	MP86169
Spike Sol 2	1.00	MP86170
Spike Sol 3	1.00	MP86171
Spike Sol 4	1.00	MP86172
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Conc. HNO3	3.00	M6158
1:1 HCL	5.00	MP85156
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

HOT BLOCK#1 CELL#50 96C
MP86171,MP86172*

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
07/11/25 12:05	SKS met dig	SKS (metals lab)
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	pH	Initial Vol (ml)	Final Vol (ml)	Color Before	Color After	Clarity Before	Clarity After	Comment	Prep Pos
PB168807BL	PBW807	<2	50	50	Colorless	Colorless	Clear	Clear	N/A	N/A
PB168807BS	LCS807	<2	50	50	Colorless	Colorless	Clear	Clear	MP86169,MP86170,X*	N/A
Q2562-01	17 BANK	<2	50	50	Colorless	Colorless	Clear	Clear	N/A	N/A
Q2562-02	SUPPLY	<2	50	50	Colorless	Colorless	Clear	Clear	N/A	N/A
Q2562-02MS	SUPPLYMS	<2	50	50	Colorless	Colorless	Clear	Clear	MP86169,MP86170,X*	N/A
Q2562-02MSD	SUPPLYMSD	<2	50	50	Colorless	Colorless	Clear	Clear	MP86169,MP86170,X*	N/A
Q2562-02DUP	SUPPLYDUP	<2	50	50	Colorless	Colorless	Clear	Clear	N/A	N/A
Q2562-03	MW25	<2	50	50	Brown	Colorless	Cloudy	Clear	N/A	N/A
Q2562-04	MW1	<2	50	50	Colorless	Colorless	Clear	Clear	N/A	N/A