

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q2562

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Sodium	2170	2000	109	90 - 110	P	07/16/2025	11:30	LB136512

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q2562

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Sodium	461	500	92	80 - 120	P	07/16/2025	11:40	LB136512

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q2562

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Sodium	265000	250000	106	90 - 110	P	07/16/2025	12:51	LB136512
CCV02	Sodium	257000	250000	103	90 - 110	P	07/16/2025	13:41	LB136512
CCV03	Sodium	257000	250000	103	90 - 110	P	07/16/2025	14:40	LB136512
CCV04	Sodium	257000	250000	103	90 - 110	P	07/16/2025	15:45	LB136512
CCV05	Sodium	257000	250000	103	90 - 110	P	07/16/2025	17:02	LB136512
CCV06	Sodium	261000	250000	104	90 - 110	P	07/16/2025	17:57	LB136512