

metals
- 5a -
MATRIX SPIKE SUMMARY

client: G Environmental **level:** low **sdg no.:** Q2562
contract: GENV01 **lab code:** ACE
matrix: Water **sample id:** Q2562-02 **client id:** SUPPLYMS
Percent Solids for Sample: NA **Spiked ID:** Q2562-02MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Sodium	ug/L	75 - 125	105000	D	63700	D	50000	82		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: G Environmental **level:** low **sdg no.:** Q2562
contract: GENV01 **lab code:** ACE
matrix: Water **sample id:** Q2562-02 **client id:** SUPPLYMSD
Percent Solids for Sample: NA **Spiked ID:** Q2562-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Sodium	ug/L	75 - 125	105000	D	63700	D	50000	83		P