

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q2570

Project: Pre Treatment Plant 2025

RunNo.: LB136445

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Orthophosphate as P	mg/L	0.490	0.50	98	90-110	07/11/2025
Sample ID: CCV1 Orthophosphate as P	mg/L	0.502	0.5	100	90-110	07/11/2025
Sample ID: CCV2 Orthophosphate as P	mg/L	0.492	0.5	98	90-110	07/11/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q2570

Project: Pre Treatment Plant 2025

RunNo.: LB136450

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.504	0.50	101	90-110	07/11/2025
Sample ID: CCV1 Phosphorus, Total	mg/L	0.493	0.50	99	90-110	07/11/2025
Sample ID: CCV2 Phosphorus, Total	mg/L	0.495	0.50	99	90-110	07/11/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q2570

Project: Pre Treatment Plant 2025

RunNo.: LB136459

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.99	1	99	90-110	07/14/2025
Sample ID: CCV1 Ammonia as N	mg/L	1	1	100	90-110	07/14/2025
Sample ID: CCV2 Ammonia as N	mg/L	0.98	1	98	90-110	07/14/2025
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	07/14/2025
Sample ID: CCV4 Ammonia as N	mg/L	1	1	100	90-110	07/14/2025