

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2570

Project: Pre Treatment Plant 2025

Run No.: LB136441

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136441BS							
BOD5	mg/L	198	181		91	1	84.6-115.4	07/11/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2570

Project: Pre Treatment Plant 2025

Run No.: LB136445

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136445BS							
Orthophosphate as P	mg/L	0.5	0.49		98	1	90-110	07/11/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2570

Project: Pre Treatment Plant 2025

Run No.: LB136471

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136471BS							
TSS	mg/L	550	533		97	1	90-110	07/15/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2570

Project: Pre Treatment Plant 2025

Run No.: LB136493

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136493BS							
Oil and Grease	mg/L	20.0	16.8		84	1	78-114	07/16/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2570

Project: Pre Treatment Plant 2025

Run No.: LB136450

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168824BS							
Phosphorus, Total	mg/L	0.50	0.50		100	1	90-110	07/11/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2570

Project: Pre Treatment Plant 2025

Run No.: LB136459

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168835BS							
Ammonia as N	mg/L	1	0.96		96	1	90-110	07/14/2025