

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB136452

Slope : 98.5

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/11/2025	14:20
2	CAL2	1	Water	NA	NA	20.2	7.01	07/11/2025	14:21
3	CAL3	1	Water	NA	NA	20.2	10.02	07/11/2025	14:22
4	ICV	1	Water	NA	NA	20.2	7.01	07/11/2025	14:25
5	CCV1	1	Water	NA	NA	20.2	2.01	07/11/2025	14:33
6	Q2578-03	1	Solid	20.02	20	20.8	11.66	07/11/2025	14:55
7	Q2578-03DUP	1	Solid	20.03	20	20.9	11.68	07/11/2025	14:56
8	Q2578-07	1	Solid	20.02	20	21.3	11.38	07/11/2025	15:00
9	Q2578-11	1	Solid	20.03	20	21.2	11.51	07/11/2025	15:05
10	Q2578-15	1	Solid	20.04	20	21.2	11.69	07/11/2025	15:15
11	Q2578-19	1	Solid	20.02	20	21.3	11.81	07/11/2025	15:25
12	Q2579-03	1	Solid	20.03	20	21.1	11.39	07/11/2025	15:35
13	Q2579-07	1	Solid	20.04	20	21.2	11.91	07/11/2025	15:45
14	Q2586-04	1	Solid	20.02	20	21.9	9.02	07/11/2025	15:56
15	CCV2	1	Water	NA	NA	20.3	12.02	07/11/2025	16:02

WORKLIST(Hardcopy Internal Chain)

136452

WorkList Name : corrosivity q2589

WorkList ID : 190678

Department : Wet-Chemistry

Date : 07-11-2025 13:43:46

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2578-03	WC-A5-01-C	Solid	Corrosivity	Cool 4 deg C	ENTA05	D41	07/10/2025	9045D
Q2578-07	WC-A2-09-C	Solid	Corrosivity	Cool 4 deg C	ENTA05	D41	07/10/2025	9045D
Q2578-11	WC-A2-10-C	Solid	Corrosivity	Cool 4 deg C	ENTA05	D41	07/10/2025	9045D
Q2578-15	WC-A2-11-C	Solid	Corrosivity	Cool 4 deg C	ENTA05	D41	07/10/2025	9045D
Q2578-19	WC-A2-12-C	Solid	Corrosivity	Cool 4 deg C	ENTA05	D41	07/10/2025	9045D
Q2579-03	WC-A2-13-C	Solid	Corrosivity	Cool 4 deg C	ENTA05	D41	07/10/2025	9045D
Q2579-07	WC-A2-14-C	Solid	Corrosivity	Cool 4 deg C	ENTA05	D41	07/10/2025	9045D
Q2586-04	TP-16	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	07/11/2025	9045D

Date/Time 07/11/25 13:50
 Raw Sample Received by: SPWCL
 Raw Sample Relinquished by: RHCWCL

Date/Time 07/11/25 18:00
 Raw Sample Received by: RHCWCL
 Raw Sample Relinquished by: SPWCL