


LB136476

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.4756e-004

C= -1.3138e-002

Rho= 0.9995155

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0
0.0	0.000	0.002	0.002	62	0.000	62	0				1
0.2	0.200	0.218	0.018	935	0.0 %	935					9
2.5	2.500	2.389	-0.111	9703	0.0 %	9703					-4
5.0	5.000	5.040	0.040	20413	0.0 %	20413					1
7.5	7.500	7.703	0.203	31170	0.0 %	31170					3
10.0	10.000	9.847	-0.153	39828	0.0 %	39828					-2

LB136476 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Jul 2025 10:41:48

Type	Sample ID	Extended ID	Conc.	µ Abs.	Units	Type	Std Conc	Method	Date
S	0.0 - 1	5.0	-	62	PPB	Std	0.0000	7470A	15 Jul 2025 11:46:56
S	0.2 - 1	5.2	-	935	PPB	Std	0.2000	7470A	15 Jul 2025 11:49:13
S	2.5 - 1	5.2	-	9703	PPB	Std	2.5000	7470A	15 Jul 2025 11:51:30
S	5.0 - 1	5.2	-	20413	PPB	Std	5.0000	7470A	15 Jul 2025 11:53:47
S	7.5 - 1	5.2	-	31170	PPB	Std	7.5000	7470A	15 Jul 2025 11:56:04
S	10.0 - 1	5.2	-	39828	PPB	Std	10.0000	7470A	15 Jul 2025 12:03:46
U	ICV27 - 1	ICV27	3.8137	15458	PPB	SMPL	-7470A		15 Jul 2025 12:06:37
U	ICB27 - 1	ICB27	-0.0565	-175	PPB	SMPL	-7470A		15 Jul 2025 12:08:53
U	CCV01 - 1	CCV01	5.3080	21494	PPB	SMPL	-7470A		15 Jul 2025 12:11:11
U	CCB01 - 1	CCB01	-0.0718	-237	PPB	SMPL	-7470A		15 Jul 2025 12:13:26
U	CRA - 1	CRA	0.2319	990	PPB	SMPL	-7470A		15 Jul 2025 12:21:13
U	HighStd - 1	HighStd	9.5544	38647	PPB	SMPL	-7470A		15 Jul 2025 12:23:28
U	ChkStd - 1	ChkStd	7.3929	29916	PPB	SMPL	-7470A		15 Jul 2025 12:25:43
U	PB168849BL - 1	PBW	-0.0092	16	PPB	SMPL	-7470A		15 Jul 2025 12:32:58
U	PB168849BS - 1	LCSW	4.0756	16516	PPB	SMPL	-7470A		15 Jul 2025 12:35:15
U	Q2561-07 - 1	AUD-25-0117	-0.0542	-166	PPB	SMPL	-7470A		15 Jul 2025 12:37:31
U	Q2565-06 - 1	MOO-25-0194-0195	-0.0374	-98	PPB	SMPL	-7470A		15 Jul 2025 12:39:51
U	Q2565-07 - 1	MOO-25-0191	-0.0159	-11	PPB	SMPL	-7470A		15 Jul 2025 12:42:07
U	Q2565-08 - 1	MOO-25-0196	-0.0045	35	PPB	SMPL	-7470A		15 Jul 2025 12:44:25
U	Q2565-09 - 1	MOO-25-0180	-0.0097	14	PPB	SMPL	-7470A		15 Jul 2025 12:46:40
U	CCV02 - 1	CCV02	5.1426	20826	PPB	SMPL	-7470A		15 Jul 2025 12:48:55
U	CCB02 - 1	CCB02	-0.0691	-226	PPB	SMPL	-7470A		15 Jul 2025 12:51:10
U	Q2571-04 - 1	TP-18	-0.0141	-4	PPB	SMPL	-7470A		15 Jul 2025 12:53:28
U	Q2571-08 - 1	TP-17	-0.0141	-4	PPB	SMPL	-7470A		15 Jul 2025 12:55:43
U	Q2578-03 - 1	WC-A5-01-C	-0.0074	23	PPB	SMPL	-7470A		15 Jul 2025 12:57:58
U	Q2578-07 - 1	WC-A2-09-C	0.0015	59	PPB	SMPL	-7470A		15 Jul 2025 13:00:14
U	Q2578-11 - 1	WC-A2-10-C	-0.0045	35	PPB	SMPL	-7470A		15 Jul 2025 13:02:29
U	Q2578-15 - 1	WC-A2-11-C	-0.0104	11	PPB	SMPL	-7470A		15 Jul 2025 13:04:46
U	Q2578-19 - 1	WC-A2-12-C	-0.0060	29	PPB	SMPL	-7470A		15 Jul 2025 13:07:02
U	Q2579-03 - 1	WC-A2-13-C	-0.0008	50	PPB	SMPL	-7470A		15 Jul 2025 13:09:19
U	Q2579-07 - 1	WC-A2-14-C	-0.0067	26	PPB	SMPL	-7470A		15 Jul 2025 13:11:36
U	Q2586-04 - 1	TP-16	-0.0035	39	PPB	SMPL	-7470A		15 Jul 2025 13:13:52
U	CCV03 - 1	CCV03	5.1604	20898	PPB	SMPL	-7470A		15 Jul 2025 13:16:07
U	CCB03 - 1	CCB03	-0.0629	-201	PPB	SMPL	-7470A		15 Jul 2025 13:18:24
U	Q2586-04DUP - 1	TP-16DUP	-0.0121	4	PPB	SMPL	-7470A		15 Jul 2025 13:20:41
U	Q2586-04MS - 1	TP-16MS	4.6920	19006	PPB	SMPL	-7470A		15 Jul 2025 13:22:57
U	Q2586-04MSD - 1	TP-16MSD	4.6779	18949	PPB	SMPL	-7470A		15 Jul 2025 13:25:13
U	PB168860BL - 1	PBW	-0.0706	-232	PPB	SMPL	-7470A		15 Jul 2025 13:27:30
U	PB168860BS - 1	LCSW	4.1476	16807	PPB	SMPL	-7470A		15 Jul 2025 13:32:33
U	Q2555-02 - 1	OU4-TS-29-070925	0.0250	154	PPB	SMPL	-7470A		15 Jul 2025 13:34:49
U	Q2555-02DUP - 1	OU4-TS-29-070925DUP	0.0223	143	PPB	SMPL	-7470A		15 Jul 2025 13:39:56
U	Q2555-02MS - 1	OU4-TS-29-070925MS	3.9481	16001	PPB	SMPL	-7470A		15 Jul 2025 13:47:19
U	Q2555-02MSD - 1	OU4-TS-29-070925MSD	4.1046	16633	PPB	SMPL	-7470A		15 Jul 2025 13:49:37
U	Q2555-04 - 1	OU4-TS-30-070925	-0.0250	-48	PPB	SMPL	-7470A		15 Jul 2025 13:50:54
U	CCV04 - 1	CCV04	4.7151	19099	PPB	SMPL	-7470A		15 Jul 2025 13:53:12
U	CCB04 - 1	CCB04	-0.0230	-40	PPB	SMPL	-7470A		15 Jul 2025 13:58:29
U	Q2558-02 - 1	OU4-TS-Denali-070925	0.0889	412	PPB	SMPL	-7470A		15 Jul 2025 14:00:44
U	Q2558-04 - 1	OU4-TS-Grillo-OG-070925	0.0849	396	PPB	SMPL	-7470A		15 Jul 2025 14:03:00
U	PB168803TB - 1	PB168803TB	-0.0181	-20	PPB	SMPL	-7470A		15 Jul 2025 14:09:19
U	PB168860TB - 1	PB168838TB	-0.0228	-39	PPB	SMPL	-7470A		15 Jul 2025 14:11:34
U	Q2586-04LX5 - 1		-0.0156	-10	PPB	SMPL	-7470A		15 Jul 2025 14:13:50
U	Q2586-04A - 1		4.2452	17201	PPB	SMPL	-7470A		15 Jul 2025 14:16:06
U	Q2555-02LX5 - 1		-0.0545	-167	PPB	SMPL	-7470A		15 Jul 2025 14:18:22
U	Q2555-02A - 1		3.8528	15616	PPB	SMPL	-7470A		15 Jul 2025 14:20:41
U	CCV05 - 1	CCV05	4.8180	19515	PPB	SMPL	-7470A		15 Jul 2025 14:22:59
U	CCB05 - 1	CCB05	-0.0364	-94	PPB	SMPL	-7470A		15 Jul 2025 14:27:47