



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT

SDG No.: Q2578

Contract: ENTA05

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.23	0.2	116	70 - 130	CV	07/15/2025	12:21	LB136476
CRI01	Arsenic	20.1	20.0	101	65 - 135	P	07/16/2025	14:22	LB136511
	Barium	100	100	100	65 - 135	P	07/16/2025	14:22	LB136511
	Cadmium	5.84	6.0	97	65 - 135	P	07/16/2025	14:22	LB136511
	Chromium	9.85	10.0	98	65 - 135	P	07/16/2025	14:22	LB136511
	Copper	21.1	20.0	106	65 - 135	P	07/16/2025	14:22	LB136511
	Lead	11.0	12.0	92	65 - 135	P	07/16/2025	14:22	LB136511
	Nickel	39.9	40.0	100	65 - 135	P	07/16/2025	14:22	LB136511
	Selenium	18.3	20.0	92	65 - 135	P	07/16/2025	14:22	LB136511
	Silver	10.4	10.0	104	65 - 135	P	07/16/2025	14:22	LB136511
	Zinc	43.1	40.0	108	65 - 135	P	07/16/2025	14:22	LB136511