



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q2578

Contract: ENTA05

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB27	Mercury	0.076	+/-0.2	U	0.20	CV	07/15/2025	12:08	LB136476

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q2578

Contract: ENTA05

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Mercury	0.076	+/-0.2	U	0.20	CV	07/15/2025	12:13	LB136476
CCB02	Mercury	0.076	+/-0.2	U	0.20	CV	07/15/2025	12:51	LB136476
CCB03	Mercury	0.076	+/-0.2	U	0.20	CV	07/15/2025	13:18	LB136476
CCB04	Mercury	0.076	+/-0.2	U	0.20	CV	07/15/2025	13:58	LB136476
CCB05	Mercury	0.076	+/-0.2	U	0.20	CV	07/15/2025	14:27	LB136476

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q2578

Contract: ENTA05

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	07/16/2025	14:11	LB136511
	Barium	14.6	+/-50	U	100	P	07/16/2025	14:11	LB136511
	Cadmium	0.50	+/-3	U	6.00	P	07/16/2025	14:11	LB136511
	Chromium	2.12	+/-5	U	10.0	P	07/16/2025	14:11	LB136511
	Copper	4.60	+/-10	U	20.0	P	07/16/2025	14:11	LB136511
	Lead	2.30	+/-6	U	12.0	P	07/16/2025	14:11	LB136511
	Nickel	3.06	+/-20	U	40.0	P	07/16/2025	14:11	LB136511
	Selenium	9.64	+/-10	U	20.0	P	07/16/2025	14:11	LB136511
	Silver	1.62	+/-5	U	10.0	P	07/16/2025	14:11	LB136511
	Zinc	16.7	+/-20	U	40.0	P	07/16/2025	14:11	LB136511



- 3a -

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	07/16/2025	14:57	LB136511
	Barium	14.6	+/-50	U	100	P	07/16/2025	14:57	LB136511
	Cadmium	0.50	+/-3	U	6.00	P	07/16/2025	14:57	LB136511
	Chromium	2.12	+/-5	U	10.0	P	07/16/2025	14:57	LB136511
	Copper	4.60	+/-10	U	20.0	P	07/16/2025	14:57	LB136511
	Lead	2.30	+/-6	U	12.0	P	07/16/2025	14:57	LB136511
	Nickel	3.06	+/-20	U	40.0	P	07/16/2025	14:57	LB136511
	Selenium	9.64	+/-10	U	20.0	P	07/16/2025	14:57	LB136511
	Silver	1.62	+/-5	U	10.0	P	07/16/2025	14:57	LB136511
	Zinc	16.7	+/-20	U	40.0	P	07/16/2025	14:57	LB136511
CCB02	Arsenic	5.12	+/-10	U	20.0	P	07/16/2025	15:43	LB136511
	Barium	14.6	+/-50	U	100	P	07/16/2025	15:43	LB136511
	Cadmium	0.50	+/-3	U	6.00	P	07/16/2025	15:43	LB136511
	Chromium	2.12	+/-5	U	10.0	P	07/16/2025	15:43	LB136511
	Copper	4.60	+/-10	U	20.0	P	07/16/2025	15:43	LB136511
	Lead	2.30	+/-6	U	12.0	P	07/16/2025	15:43	LB136511
	Nickel	3.06	+/-20	U	40.0	P	07/16/2025	15:43	LB136511
	Selenium	9.64	+/-10	U	20.0	P	07/16/2025	15:43	LB136511
	Silver	1.62	+/-5	U	10.0	P	07/16/2025	15:43	LB136511
	Zinc	16.7	+/-20	U	40.0	P	07/16/2025	15:43	LB136511
CCB03	Arsenic	5.12	+/-10	U	20.0	P	07/16/2025	16:30	LB136511
	Barium	14.6	+/-50	U	100	P	07/16/2025	16:30	LB136511
	Cadmium	0.50	+/-3	U	6.00	P	07/16/2025	16:30	LB136511
	Chromium	2.12	+/-5	U	10.0	P	07/16/2025	16:30	LB136511
	Copper	4.60	+/-10	U	20.0	P	07/16/2025	16:30	LB136511
	Lead	2.30	+/-6	U	12.0	P	07/16/2025	16:30	LB136511
	Nickel	3.06	+/-20	U	40.0	P	07/16/2025	16:30	LB136511
	Selenium	9.64	+/-10	U	20.0	P	07/16/2025	16:30	LB136511
	Silver	1.62	+/-5	U	10.0	P	07/16/2025	16:30	LB136511
	Zinc	16.7	+/-20	U	40.0	P	07/16/2025	16:30	LB136511
CCB04	Arsenic	5.12	+/-10	U	20.0	P	07/16/2025	17:40	LB136511
	Barium	14.6	+/-50	U	100	P	07/16/2025	17:40	LB136511
	Cadmium	0.50	+/-3	U	6.00	P	07/16/2025	17:40	LB136511
	Chromium	2.12	+/-5	U	10.0	P	07/16/2025	17:40	LB136511
	Copper	4.60	+/-10	U	20.0	P	07/16/2025	17:40	LB136511
	Lead	2.30	+/-6	U	12.0	P	07/16/2025	17:40	LB136511
	Nickel	3.06	+/-20	U	40.0	P	07/16/2025	17:40	LB136511
	Selenium	9.64	+/-10	U	20.0	P	07/16/2025	17:40	LB136511
	Silver	1.62	+/-5	U	10.0	P	07/16/2025	17:40	LB136511

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q2578

Contract: ENTA05

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Zinc	16.7	+/-20	U	40.0	P	07/16/2025	17:40	LB136511
CCB05	Arsenic	5.12	+/-10	U	20.0	P	07/16/2025	18:35	LB136511
	Barium	14.6	+/-50	U	100	P	07/16/2025	18:35	LB136511
	Cadmium	0.50	+/-3	U	6.00	P	07/16/2025	18:35	LB136511
	Chromium	2.12	+/-5	U	10.0	P	07/16/2025	18:35	LB136511
	Copper	4.60	+/-10	U	20.0	P	07/16/2025	18:35	LB136511
	Lead	2.30	+/-6	U	12.0	P	07/16/2025	18:35	LB136511
	Nickel	3.06	+/-20	U	40.0	P	07/16/2025	18:35	LB136511
	Selenium	9.64	+/-10	U	20.0	P	07/16/2025	18:35	LB136511
	Silver	1.62	+/-5	U	10.0	P	07/16/2025	18:35	LB136511
	Zinc	16.7	+/-20	U	40.0	P	07/16/2025	18:35	LB136511
CCB06	Arsenic	5.12	+/-10	U	20.0	P	07/16/2025	18:56	LB136511
	Barium	14.6	+/-50	U	100	P	07/16/2025	18:56	LB136511
	Cadmium	0.50	+/-3	U	6.00	P	07/16/2025	18:56	LB136511
	Chromium	2.12	+/-5	U	10.0	P	07/16/2025	18:56	LB136511
	Copper	4.60	+/-10	U	20.0	P	07/16/2025	18:56	LB136511
	Lead	2.30	+/-6	U	12.0	P	07/16/2025	18:56	LB136511
	Nickel	3.06	+/-20	U	40.0	P	07/16/2025	18:56	LB136511
	Selenium	9.64	+/-10	U	20.0	P	07/16/2025	18:56	LB136511
	Silver	1.62	+/-5	U	10.0	P	07/16/2025	18:56	LB136511
	Zinc	16.7	+/-20	U	40.0	P	07/16/2025	18:56	LB136511