

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: ENTACT

SDG No.: Q2578

Contract: ENTA05

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	9.98			-20	20	07/16/2025	14:26	LB136511
	Barium	2.39	6.0	40	-94	106	07/16/2025	14:26	LB136511
	Cadmium	-0.36	1.0	36	-5	7	07/16/2025	14:26	LB136511
	Chromium	56.1	52.0	108	42	62	07/16/2025	14:26	LB136511
	Copper	-0.33	2.0	16	-18	22	07/16/2025	14:26	LB136511
	Lead	-7.00			-12	12	07/16/2025	14:26	LB136511
	Nickel	18.6	2.0	930	-38	42	07/16/2025	14:26	LB136511
	Selenium	9.85			-20	20	07/16/2025	14:26	LB136511
	Silver	1.35			-10	10	07/16/2025	14:26	LB136511
	Zinc	5.50			-40	40	07/16/2025	14:26	LB136511
ICSAB01	Arsenic	104	104	100	88.4	120	07/16/2025	14:38	LB136511
	Barium	477	537	89	437	637	07/16/2025	14:38	LB136511
	Cadmium	913	972	94	826	1120	07/16/2025	14:38	LB136511
	Chromium	483	542	89	460	624	07/16/2025	14:38	LB136511
	Copper	453	511	89	434	588	07/16/2025	14:38	LB136511
	Lead	38.4	49.0	78	37	61	07/16/2025	14:38	LB136511
	Nickel	924	954	97	810	1100	07/16/2025	14:38	LB136511
	Selenium	60.4	46.0	131	26	66	07/16/2025	14:38	LB136511
	Silver	204	201	102	170	232	07/16/2025	14:38	LB136511
	Zinc	1000	952	105	809	1095	07/16/2025	14:38	LB136511