

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT

SDG No.: Q2578

Contract: ENTA05

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168849BS Mercury	ug/L	4.0	4.08		102	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT
Contract: ENTA05

SDG No.: Q2578
Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168857BS							
Arsenic	ug/L	4000	3960		99	80 - 120	P
Barium	ug/L	1000	1040		104	80 - 120	P
Cadmium	ug/L	1000	965		96	80 - 120	P
Chromium	ug/L	2000	2000		100	80 - 120	P
Copper	ug/L	1500	1580		105	80 - 120	P
Lead	ug/L	5000	4850		97	80 - 120	P
Nickel	ug/L	2500	2510		100	80 - 120	P
Selenium	ug/L	10000	9940		99	80 - 120	P
Silver	ug/L	380	382		100	80 - 120	P
Zinc	ug/L	1000	1060		106	80 - 120	P