

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q2578

Continuing Calib Date: 07/14/2025

Initial Calibration Date(s): 07/08/2025

07/08/2025

Continuing Calib Time: 16:25

Initial Calibration Time(s): 14:06

22:16

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.76	4.76	4.66	4.86	0.00
Aroclor-1242-2 (2)	4.78	4.78	4.68	4.88	0.00
Aroclor-1242-3 (3)	4.83	4.84	4.74	4.94	0.01
Aroclor-1242-4 (4)	4.95	4.96	4.86	5.06	0.01
Aroclor-1242-5 (5)	5.60	5.61	5.51	5.71	0.01
Tetrachloro-m-xylene	3.67	3.67	3.57	3.77	0.00
Decachlorobiphenyl	8.70	8.70	8.60	8.80	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q2578

Continuing Calib Date: 07/14/2025

Initial Calibration Date(s): 07/08/2025

07/08/2025

Continuing Calib Time: 16:25

Initial Calibration Time(s): 14:06

22:16

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.74	4.74	4.64	4.84	0.00
Aroclor-1242-2 (2)	4.76	4.76	4.66	4.86	0.00
Aroclor-1242-3 (3)	4.93	4.94	4.84	5.04	0.01
Aroclor-1242-4 (4)	5.02	5.02	4.92	5.12	0.00
Aroclor-1242-5 (5)	5.54	5.54	5.44	5.64	0.00
Tetrachloro-m-xylene	3.67	3.67	3.57	3.77	0.01
Decachlorobiphenyl	8.65	8.65	8.55	8.75	0.00