

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2578
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/14/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112259.D **Time Analyzed:** 16:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.757	4.661	4.861	516.900	500.000	3.4
Aroclor-1016-2	4.776	4.680	4.880	525.820	500.000	5.2
Aroclor-1016-3	4.833	4.736	4.936	507.700	500.000	1.5
Aroclor-1016-4	4.953	4.856	5.056	524.090	500.000	4.8
Aroclor-1016-5	5.209	5.113	5.313	542.000	500.000	8.4
Aroclor-1260-1	6.247	6.151	6.351	499.820	500.000	0.0
Aroclor-1260-2	6.437	6.341	6.541	500.070	500.000	0.0
Aroclor-1260-3	6.803	6.707	6.907	500.250	500.000	0.1
Aroclor-1260-4	7.062	6.967	7.167	515.850	500.000	3.2
Aroclor-1260-5	7.305	7.209	7.409	518.100	500.000	3.6
Decachlorobiphenyl	8.698	8.603	8.803	51.880	50.000	3.8
Tetrachloro-m-xylene	3.671	3.573	3.773	55.190	50.000	10.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2578
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/14/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112259.D **Time Analyzed:** 16:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.741	4.661	4.861	495.640	500.000	-0.9
Aroclor-1016-2	4.760	4.679	4.879	495.110	500.000	-1.0
Aroclor-1016-3	4.935	4.854	5.054	490.430	500.000	-1.9
Aroclor-1016-4	4.978	4.897	5.097	483.780	500.000	-3.2
Aroclor-1016-5	5.189	5.109	5.309	518.230	500.000	3.6
Aroclor-1260-1	6.218	6.138	6.338	490.910	500.000	-1.8
Aroclor-1260-2	6.406	6.326	6.526	500.980	500.000	0.2
Aroclor-1260-3	6.557	6.478	6.678	502.560	500.000	0.5
Aroclor-1260-4	7.028	6.948	7.148	473.910	500.000	-5.2
Aroclor-1260-5	7.269	7.190	7.390	485.600	500.000	-2.9
Decachlorobiphenyl	8.646	8.568	8.768	49.630	50.000	-0.7
Tetrachloro-m-xylene	3.666	3.584	3.784	52.280	50.000	4.6