

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2578
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/14/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112270.D **Time Analyzed:** 20:16

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.758	4.661	4.861	566.300	500.000	13.3
Aroclor-1016-2	4.776	4.680	4.880	570.420	500.000	14.1
Aroclor-1016-3	4.833	4.736	4.936	552.280	500.000	10.5
Aroclor-1016-4	4.953	4.856	5.056	566.420	500.000	13.3
Aroclor-1016-5	5.210	5.113	5.313	597.680	500.000	19.5
Aroclor-1260-1	6.247	6.151	6.351	449.180	500.000	-10.2
Aroclor-1260-2	6.437	6.341	6.541	439.890	500.000	-12.0
Aroclor-1260-3	6.803	6.707	6.907	558.730	500.000	11.7
Aroclor-1260-4	7.062	6.967	7.167	544.230	500.000	8.8
Aroclor-1260-5	7.304	7.209	7.409	444.700	500.000	-11.1
Decachlorobiphenyl	8.697	8.603	8.803	43.550	50.000	-12.9
Tetrachloro-m-xylene	3.670	3.573	3.773	59.970	50.000	19.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2578
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/14/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112270.D **Time Analyzed:** 20:16

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.738	4.661	4.861	524.110	500.000	4.8
Aroclor-1016-2	4.758	4.679	4.879	548.630	500.000	9.7
Aroclor-1016-3	4.932	4.854	5.054	544.810	500.000	9.0
Aroclor-1016-4	4.975	4.897	5.097	530.620	500.000	6.1
Aroclor-1016-5	5.187	5.109	5.309	550.020	500.000	10.0
Aroclor-1260-1	6.216	6.138	6.338	507.230	500.000	1.4
Aroclor-1260-2	6.404	6.326	6.526	505.440	500.000	1.1
Aroclor-1260-3	6.555	6.478	6.678	496.400	500.000	-0.7
Aroclor-1260-4	7.026	6.948	7.148	460.080	500.000	-8.0
Aroclor-1260-5	7.267	7.190	7.390	466.320	500.000	-6.7
Decachlorobiphenyl	8.645	8.568	8.768	47.560	50.000	-4.9
Tetrachloro-m-xylene	3.663	3.584	3.784	57.170	50.000	14.3