

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2578
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/14/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073768.D **Time Analyzed:** 14:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.638	5.539	5.739	444.570	500.000	-11.1
Aroclor-1016-2	5.659	5.560	5.760	465.050	500.000	-7.0
Aroclor-1016-3	5.721	5.622	5.822	465.930	500.000	-6.8
Aroclor-1016-4	5.818	5.720	5.920	478.780	500.000	-4.2
Aroclor-1016-5	6.110	6.012	6.212	488.440	500.000	-2.3
Aroclor-1260-1	7.227	7.129	7.329	460.120	500.000	-8.0
Aroclor-1260-2	7.480	7.383	7.583	405.330	500.000	-18.9
Aroclor-1260-3	7.838	7.740	7.940	446.530	500.000	-10.7
Aroclor-1260-4	8.062	7.964	8.164	456.350	500.000	-8.7
Aroclor-1260-5	8.379	8.283	8.483	463.970	500.000	-7.2
Decachlorobiphenyl	10.171	10.076	10.276	48.630	50.000	-2.7
Tetrachloro-m-xylene	4.487	4.388	4.588	49.030	50.000	-1.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2578
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/14/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073768.D **Time Analyzed:** 14:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.856	4.758	4.958	447.770	500.000	-10.4
Aroclor-1016-2	4.874	4.775	4.975	445.900	500.000	-10.8
Aroclor-1016-3	5.050	4.951	5.151	446.930	500.000	-10.6
Aroclor-1016-4	5.092	4.993	5.193	443.960	500.000	-11.2
Aroclor-1016-5	5.305	5.207	5.407	472.050	500.000	-5.6
Aroclor-1260-1	6.335	6.237	6.437	440.220	500.000	-12.0
Aroclor-1260-2	6.523	6.425	6.625	442.420	500.000	-11.5
Aroclor-1260-3	6.675	6.577	6.777	432.130	500.000	-13.6
Aroclor-1260-4	7.144	7.046	7.246	455.880	500.000	-8.8
Aroclor-1260-5	7.386	7.289	7.489	449.420	500.000	-10.1
Decachlorobiphenyl	8.777	8.680	8.880	49.350	50.000	-1.3
Tetrachloro-m-xylene	3.779	3.678	3.878	44.490	50.000	-11.0