

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2578
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/14/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073783.D **Time Analyzed:** 19:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.638	5.539	5.739	429.480	500.000	-14.1
Aroclor-1016-2	5.660	5.560	5.760	456.850	500.000	-8.6
Aroclor-1016-3	5.722	5.622	5.822	450.490	500.000	-9.9
Aroclor-1016-4	5.820	5.720	5.920	455.610	500.000	-8.9
Aroclor-1016-5	6.111	6.012	6.212	469.800	500.000	-6.0
Aroclor-1260-1	7.228	7.129	7.329	462.440	500.000	-7.5
Aroclor-1260-2	7.480	7.383	7.583	404.550	500.000	-19.1
Aroclor-1260-3	7.839	7.740	7.940	444.780	500.000	-11.0
Aroclor-1260-4	8.063	7.964	8.164	467.350	500.000	-6.5
Aroclor-1260-5	8.380	8.283	8.483	461.010	500.000	-7.8
Decachlorobiphenyl	10.173	10.076	10.276	48.020	50.000	-4.0
Tetrachloro-m-xylene	4.488	4.388	4.588	47.760	50.000	-4.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2578
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/14/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073783.D **Time Analyzed:** 19:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.856	4.758	4.958	453.210	500.000	-9.4
Aroclor-1016-2	4.874	4.775	4.975	449.210	500.000	-10.2
Aroclor-1016-3	5.050	4.951	5.151	458.590	500.000	-8.3
Aroclor-1016-4	5.092	4.993	5.193	456.700	500.000	-8.7
Aroclor-1016-5	5.305	5.207	5.407	488.120	500.000	-2.4
Aroclor-1260-1	6.334	6.237	6.437	446.200	500.000	-10.8
Aroclor-1260-2	6.524	6.425	6.625	445.950	500.000	-10.8
Aroclor-1260-3	6.675	6.577	6.777	447.270	500.000	-10.5
Aroclor-1260-4	7.144	7.046	7.246	466.470	500.000	-6.7
Aroclor-1260-5	7.387	7.289	7.489	466.520	500.000	-6.7
Decachlorobiphenyl	8.778	8.680	8.880	50.860	50.000	1.7
Tetrachloro-m-xylene	3.779	3.678	3.878	45.150	50.000	-9.7