

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS071125

Review By	Abdul	Review On	7/14/2025 8:48:45 AM
Supervise By	mohammad	Supervise On	7/15/2025 1:41:05 AM
SubDirectory	PS071125	HP Acquire Method	HP Processing Method ps061825 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560
CCC	PP24559
Internal Standard/PEM	
ICV/I.BLK	PP24562
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS031004.D	11 Jul 2025 15:11		AR\AJ	Ok
2	I.BLK	I.BLK	PS031005.D	11 Jul 2025 15:35		AR\AJ	Ok
3	HSTDICC200	HSTDICC200	PS031006.D	11 Jul 2025 16:00		AR\AJ	Ok,M
4	HSTDICC500	HSTDICC500	PS031007.D	11 Jul 2025 16:24		AR\AJ	Ok,M
5	HSTDICC750	HSTDICC750	PS031008.D	11 Jul 2025 16:48		AR\AJ	Ok
6	HSTDICC1000	HSTDICC1000	PS031009.D	11 Jul 2025 17:12		AR\AJ	Ok
7	HSTDICC1500	HSTDICC1500	PS031010.D	11 Jul 2025 17:36		AR\AJ	Ok,M
8	HSTDICV750	ICVPS071125	PS031011.D	11 Jul 2025 18:00		AR\AJ	Ok
9	I.BLK	I.BLK	PS031012.D	11 Jul 2025 18:25		AR\AJ	Ok,M
10	HSTDCCC750	HSTDCCC750	PS031013.D	11 Jul 2025 18:49		AR\AJ	Ok
11	Q2517-01RE	TP-14RE	PS031014.D	11 Jul 2025 20:01	Surrogate low in 2nd column	AR\AJ	Confirms
12	Q2514-10RE	TP-90RE	PS031015.D	11 Jul 2025 20:25	Surrogate low in 1st column	AR\AJ	Confirms
13	Q2493-01MS	WC-11MS	PS031016.D	11 Jul 2025 20:49	F Flag in comp#1 , Comp#14 not detected,Comp#9,10 recovery fail	AR\AJ	Not Ok
14	Q2493-01MSD	WC-11MSD	PS031017.D	11 Jul 2025 21:14	F Flag in comp#1 , Comp#14 not detected,Comp#9,10 recovery fail	AR\AJ	Not Ok
15	I.BLK	I.BLK	PS031018.D	11 Jul 2025 21:38		AR\AJ	Ok
16	HSTDCCC750	HSTDCCC750	PS031019.D	11 Jul 2025 22:02		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS071525

Review By	yogesh	Review On	7/15/2025 3:24:40 PM
Supervise By	mohammad	Supervise On	7/18/2025 3:10:15 AM
SubDirectory	PS071525	HP Acquire Method	HP Processing Method ps071125 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC	PP24559		
Internal Standard/PEM			
ICV/I.BLK	PP24562		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS031029.D	15 Jul 2025 10:57		AR\AJ	Ok
2	I.BLK	I.BLK	PS031030.D	15 Jul 2025 17:08		AR\AJ	Ok
3	HSTDCCC750	HSTDCCC750	PS031031.D	15 Jul 2025 17:57		AR\AJ	Ok,M
4	PB168844BL	PB168844BL	PS031032.D	15 Jul 2025 18:21		AR\AJ	Ok
5	PB168844BS	PB168844BS	PS031033.D	15 Jul 2025 18:45		AR\AJ	Ok,M
6	PB168728TB	PB168728TB	PS031034.D	15 Jul 2025 19:09		AR\AJ	Ok
7	PB168803TB	PB168803TB	PS031035.D	15 Jul 2025 19:33		AR\AJ	Ok
8	Q2489-01	G4(0-6)	PS031036.D	15 Jul 2025 19:58		AR\AJ	Ok
9	Q2489-01MS	G4(0-6)MS	PS031037.D	15 Jul 2025 20:22		AR\AJ	Ok,M
10	Q2489-01MSD	G4(0-6)MSD	PS031038.D	15 Jul 2025 20:46		AR\AJ	Ok,M
11	Q2489-02	G4(6-12)	PS031039.D	15 Jul 2025 21:10		AR\AJ	Ok
12	Q2489-03	G3(0-6)	PS031040.D	15 Jul 2025 21:34		AR\AJ	Ok,M
13	Q2489-04	G3(6-12)	PS031041.D	15 Jul 2025 21:58		AR\AJ	Ok,M
14	Q2489-05	G2(0-6)	PS031042.D	15 Jul 2025 22:23		AR\AJ	Ok,M
15	Q2489-06	G2(6-12)	PS031043.D	15 Jul 2025 22:47	Surrogate high in 1st column	AR\AJ	ReRun
16	Q2489-07	G1(0-6)	PS031044.D	15 Jul 2025 23:11		AR\AJ	Ok
17	Q2489-08	G1(6-12)	PS031045.D	15 Jul 2025 23:35		AR\AJ	Ok
18	I.BLK	I.BLK	PS031046.D	15 Jul 2025 23:59		AR\AJ	Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS071525

Review By	yogesh	Review On	7/15/2025 3:24:40 PM
Supervise By	mohammad	Supervise On	7/18/2025 3:10:15 AM
SubDirectory	PS071525	HP Acquire Method	HP Processing Method ps071125 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC	PP24559		
Internal Standard/PEM			
ICV/I.BLK	PP24562		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	HSTDCCC750	HSTDCCC750	PS031047.D	16 Jul 2025 00:48		AR\AJ	Ok,M
20	PB168825BL	PB168825BL	PS031048.D	16 Jul 2025 01:36		AR\AJ	Ok,M
21	PB168825BS	PB168825BS	PS031049.D	16 Jul 2025 02:00		AR\AJ	Ok,M
22	PB168825TB	PB168825TB	PS031050.D	16 Jul 2025 02:24		AR\AJ	Ok
23	Q2489-01	G4(0-6)	PS031051.D	16 Jul 2025 02:48		AR\AJ	Ok
24	Q2489-01MS	G4(0-6)MS	PS031052.D	16 Jul 2025 03:13		AR\AJ	Ok,M
25	Q2489-01MSD	G4(0-6)MSD	PS031053.D	16 Jul 2025 03:37		AR\AJ	Ok,M
26	Q2489-02	G4(6-12)	PS031054.D	16 Jul 2025 04:01		AR\AJ	Ok
27	Q2489-03	G3(0-6)	PS031055.D	16 Jul 2025 04:25	Surrogate high in both column	AR\AJ	Not Ok
28	Q2489-04	G3(6-12)	PS031056.D	16 Jul 2025 04:49		AR\AJ	Ok
29	Q2489-05	G2(0-6)	PS031057.D	16 Jul 2025 05:13		AR\AJ	Ok
30	Q2489-06	G2(6-12)	PS031058.D	16 Jul 2025 05:38		AR\AJ	Ok
31	Q2489-07	G1(0-6)	PS031059.D	16 Jul 2025 06:02		AR\AJ	Ok,M
32	Q2489-08	G1(6-12)	PS031060.D	16 Jul 2025 06:26		AR\AJ	Ok
33	I.BLK	I.BLK	PS031061.D	16 Jul 2025 06:50		AR\AJ	Ok
34	HSTDCCC750	HSTDCCC750	PS031062.D	16 Jul 2025 07:14		AR\AJ	Ok,M
35	Q2578-03	WC-A5-01-C	PS031063.D	16 Jul 2025 10:34		AR\AJ	Ok
36	Q2578-07	WC-A2-09-C	PS031064.D	16 Jul 2025 10:58	Surrogate fail in 1st col	AR\AJ	ReRun
37	Q2578-11	WC-A2-10-C	PS031065.D	16 Jul 2025 11:22		AR\AJ	Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS071525

Review By	yogesh	Review On	7/15/2025 3:24:40 PM
Supervise By	mohammad	Supervise On	7/18/2025 3:10:15 AM
SubDirectory	PS071525	HP Acquire Method	HP Processing Method ps071125 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC Internal Standard/PEM	PP24559		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24562		

38	Q2578-15	WC-A2-11-C	PS031066.D	16 Jul 2025 13:03		AR\AJ	Ok,M
39	Q2578-19	WC-A2-12-C	PS031067.D	16 Jul 2025 13:27		AR\AJ	Ok
40	Q2579-03	WC-A2-13-C	PS031068.D	16 Jul 2025 13:51		AR\AJ	Ok
41	Q2579-07	WC-A2-14-C	PS031069.D	16 Jul 2025 14:15		AR\AJ	Ok
42	Q2578-07RE	WC-A2-09-CRE	PS031070.D	16 Jul 2025 14:40	Surrogate fail in 1st col	AR\AJ	Confirms
43	I.BLK	I.BLK	PS031071.D	16 Jul 2025 15:04		AR\AJ	Ok
44	HSTDCCC750	HSTDCCC750	PS031072.D	16 Jul 2025 15:28		AR\AJ	Ok,M

M : Manual Integration