

Surrogate Summary

SDG No.: Q2578

Client: ENTACT

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD089537.D	PIBLK-PD089537.D	Decachlorobiphen	1	20	21.2	106		30 (57)	150 (171)
		Tetrachloro-m-xyl	1	20	18.5	92		30 (61)	150 (148)
		Decachlorobiphen	2	20	20.5	103		30 (57)	150 (171)
		Tetrachloro-m-xyl	2	20	19.9	100		30 (61)	150 (148)
I.BLK-PD089559.D	PIBLK-PD089559.D	Decachlorobiphen	1	20	21.8	109		30 (57)	150 (171)
		Tetrachloro-m-xyl	1	20	19.3	96		30 (61)	150 (148)
		Decachlorobiphen	2	20	22.1	111		30 (57)	150 (171)
		Tetrachloro-m-xyl	2	20	21.4	107		30 (61)	150 (148)
PB168867BL	PB168867BL	Decachlorobiphen	1	20	20.9	105		30 (57)	150 (171)
		Tetrachloro-m-xyl	1	20	17.8	89		30 (61)	150 (148)
		Decachlorobiphen	2	20	20.7	104		30 (57)	150 (171)
		Tetrachloro-m-xyl	2	20	19.2	96		30 (61)	150 (148)
PB168867BS	PB168867BS	Decachlorobiphen	1	20	20.7	103		30 (57)	150 (171)
		Tetrachloro-m-xyl	1	20	17.4	87		30 (61)	150 (148)
		Decachlorobiphen	2	20	20.8	104		30 (57)	150 (171)
		Tetrachloro-m-xyl	2	20	20.0	100		30 (61)	150 (148)
PB168803TB	PB168803TB	Decachlorobiphen	1	20	20.8	104		30 (57)	150 (171)
		Tetrachloro-m-xyl	1	20	17.5	88		30 (61)	150 (148)
		Decachlorobiphen	2	20	20.7	103		30 (57)	150 (171)
		Tetrachloro-m-xyl	2	20	19.4	97		30 (61)	150 (148)
Q2578-03	WC-A5-01-C	Decachlorobiphen	1	20	19.0	95		30 (57)	150 (171)
		Tetrachloro-m-xyl	1	20	14.3	71		30 (61)	150 (148)
		Decachlorobiphen	2	20	18.7	93		30 (57)	150 (171)
		Tetrachloro-m-xyl	2	20	14.3	72		30 (61)	150 (148)
Q2578-03MS	WC-A5-01-CMS	Decachlorobiphen	1	20	20.9	105		30 (57)	150 (171)
		Tetrachloro-m-xyl	1	20	15.5	77		30 (61)	150 (148)
		Decachlorobiphen	2	20	20.7	104		30 (57)	150 (171)
		Tetrachloro-m-xyl	2	20	15.4	77		30 (61)	150 (148)
Q2578-03MSD	WC-A5-01-CMSD	Decachlorobiphen	1	20	20.9	105		30 (57)	150 (171)
		Tetrachloro-m-xyl	1	20	15.4	77		30 (61)	150 (148)
		Decachlorobiphen	2	20	20.8	104		30 (57)	150 (171)
		Tetrachloro-m-xyl	2	20	15.9	80		30 (61)	150 (148)
Q2578-07	WC-A2-09-C	Decachlorobiphen	1	20	15.4	77		30 (57)	150 (171)
		Tetrachloro-m-xyl	1	20	16.1	80		30 (61)	150 (148)
		Decachlorobiphen	2	20	15.3	76		30 (57)	150 (171)
		Tetrachloro-m-xyl	2	20	17.6	88		30 (61)	150 (148)
Q2578-11	WC-A2-10-C	Decachlorobiphen	1	20	15.8	79		30 (57)	150 (171)
		Tetrachloro-m-xyl	1	20	14.8	74		30 (61)	150 (148)
		Decachlorobiphen	2	20	15.9	79		30 (57)	150 (171)
		Tetrachloro-m-xyl	2	20	15.8	79		30 (61)	150 (148)
Q2578-15	WC-A2-11-C	Decachlorobiphen	1	20	19.5	97		30 (57)	150 (171)

Surrogate Summary

SDG No.: Q2578

Client: ENTACT

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q2578-15	WC-A2-11-C	Tetrachloro-m-xyl	1	20	15.6	78		30 (61)	150 (148)
		Decachlorobiphen	2	20	19.8	99		30 (57)	150 (171)
		Tetrachloro-m-xyl	2	20	16.7	83		30 (61)	150 (148)
Q2578-19	WC-A2-12-C	Decachlorobiphen	1	20	19.5	98		30 (57)	150 (171)
		Tetrachloro-m-xyl	1	20	15.8	79		30 (61)	150 (148)
		Decachlorobiphen	2	20	19.5	98		30 (57)	150 (171)
		Tetrachloro-m-xyl	2	20	16.7	83		30 (61)	150 (148)
		Decachlorobiphen	1	20	21.8	109		30 (57)	150 (171)
I.BLK-PD089574.D	PIBLK-PD089574.D	Tetrachloro-m-xyl	1	20	19.1	95		30 (61)	150 (148)
		Decachlorobiphen	2	20	22.3	112		30 (57)	150 (171)
		Tetrachloro-m-xyl	2	20	21.0	105		30 (61)	150 (148)
		Decachlorobiphen	1	20	21.8	109		30 (57)	150 (171)