

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071425\
 Data File : BM050429.D
 Acq On : 14 Jul 2025 16:53
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 14 17:29:18 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM070925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 08 18:32:25 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	110	0.00
2	1,4-Dioxane	40.000	38.676	3.3	110	0.00
3	Pyridine	40.000	39.934	0.2	112	0.00
4	n-Nitrosodimethylamine	40.000	40.722	-1.8	114	0.00
5 S	2-Fluorophenol	80.000	80.393	-0.5	111	0.00
6	Aniline	40.000	40.237	-0.6	110	0.00
7 S	Phenol-d6	80.000	82.395	-3.0	113	0.00
8	2-Chlorophenol	40.000	40.070	-0.2	110	0.00
9	Benzaldehyde	40.000	44.877	-12.2	119	0.00
10 C	Phenol	40.000	40.591	-1.5	113	0.00
11	bis(2-Chloroethyl)ether	40.000	39.256	1.9	110	0.00
12	1,3-Dichlorobenzene	40.000	39.084	2.3	110	0.00
13 C	1,4-Dichlorobenzene	40.000	39.004	2.5	111	0.00
14	1,2-Dichlorobenzene	40.000	39.116	2.2	111	0.00
15	Benzyl Alcohol	40.000	40.162	-0.4	111	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.024	2.4	110	0.00
17	2-Methylphenol	40.000	40.241	-0.6	110	0.00
18	Hexachloroethane	40.000	39.443	1.4	111	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.504	-3.8	111	0.00
20	3+4-Methylphenols	40.000	39.839	0.4	109	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	109	0.00
22	Acetophenone	40.000	39.815	0.5	110	0.00
23 S	Nitrobenzene-d5	80.000	81.413	-1.8	111	0.00
24	Nitrobenzene	40.000	40.131	-0.3	111	0.00
25	Isophorone	40.000	39.716	0.7	107	0.00
26 C	2-Nitrophenol	40.000	42.742	-6.9	113	0.00
27	2,4-Dimethylphenol	40.000	39.483	1.3	109	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.943	2.6	107	0.00
29 C	2,4-Dichlorophenol	40.000	40.196	-0.5	109	0.00
30	1,2,4-Trichlorobenzene	40.000	38.342	4.1	107	0.00
31	Naphthalene	40.000	38.781	3.0	108	0.00
32	Benzoic acid	40.000	39.787	0.5	119	0.00
33	4-Chloroaniline	40.000	40.145	-0.4	109	0.00
34 C	Hexachlorobutadiene	40.000	37.580	6.1	105	0.00
35	Caprolactam	40.000	40.432	-1.1	108	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.774	-1.9	109	0.00
37	2-Methylnaphthalene	40.000	39.410	1.5	108	0.00
38	1-Methylnaphthalene	40.000	39.418	1.5	108	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	107	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.323	4.2	105	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.245	4.4	105	-0.01
42 S	2,4,6-Tribromophenol	80.000	83.942	-4.9	110	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.394	-1.0	108	0.00
44	2,4,5-Trichlorophenol	40.000	40.188	-0.5	107	0.00
45 S	2-Fluorobiphenyl	80.000	77.961	2.5	106	0.00
46	1,1'-Biphenyl	40.000	39.087	2.3	108	0.00
47	2-Chloronaphthalene	40.000	39.380	1.5	109	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071425\
 Data File : BM050429.D
 Acq On : 14 Jul 2025 16:53
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 14 17:29:18 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM070925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 08 18:32:25 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.062	-7.7	109	0.00
49	Acenaphthylene	40.000	40.151	-0.4	108	0.00
50	Dimethylphthalate	40.000	39.254	1.9	107	0.00
51	2,6-Dinitrotoluene	40.000	42.250	-5.6	110	0.00
52 C	Acenaphthene	40.000	39.774	0.6	109	0.00
53	3-Nitroaniline	40.000	42.864	-7.2	110	0.00
54 P	2,4-Dinitrophenol	40.000	38.955	2.6	118	0.00
55	Dibenzofuran	40.000	39.114	2.2	107	0.00
56 P	4-Nitrophenol	40.000	41.709	-4.3	110	0.00
57	2,4-Dinitrotoluene	40.000	39.178	2.1	110	0.00
58	Fluorene	40.000	39.804	0.5	108	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.793	-2.0	108	0.00
60	Diethylphthalate	40.000	39.771	0.6	107	-0.01
61	4-Chlorophenyl-phenylether	40.000	39.180	2.1	106	0.00
62	4-Nitroaniline	40.000	40.267	-0.7	111	0.00
63	Azobenzene	40.000	40.674	-1.7	109	-0.01
64 I	Phenanthrene-d10	20.000	20.000	0.0	109	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.794	3.0	117	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.314	1.7	108	-0.01
67	4-Bromophenyl-phenylether	40.000	39.033	2.4	108	-0.01
68	Hexachlorobenzene	40.000	38.647	3.4	107	0.00
69	Atrazine	40.000	40.910	-2.3	108	-0.01
70 C	Pentachlorophenol	40.000	40.653	-1.6	112	0.00
71	Phenanthrene	40.000	38.896	2.8	109	0.00
72	Anthracene	40.000	39.846	0.4	110	-0.01
73	Carbazole	40.000	40.769	-1.9	112	0.00
74	Di-n-butylphthalate	40.000	41.275	-3.2	110	-0.01
75 C	Fluoranthene	40.000	40.753	-1.9	111	-0.01
76 I	Chrysene-d12	20.000	20.000	0.0	116	-0.01
77	Benzidine	40.000	23.465	41.3#	63	-0.01
78	Pyrene	40.000	37.888	5.3	112	-0.01
79 S	Terphenyl-d14	80.000	83.219	-4.0	109	-0.01
80	Butylbenzylphthalate	40.000	42.873	-7.2	119	-0.02
81	Benzo(a)anthracene	40.000	39.905	0.2	118	-0.01
82	3,3'-Dichlorobenzidine	40.000	38.750	3.1	115	-0.01
83	Chrysene	40.000	39.487	1.3	117	-0.01
84	Bis(2-ethylhexyl)phthalate	40.000	43.607	-9.0	120	-0.01
85 c	Di-n-octyl phthalate	40.000	43.719	-9.3	126	-0.02
86 I	Perylene-d12	20.000	20.000	0.0	122	-0.02
87	Indeno(1,2,3-cd)pyrene	40.000	40.924	-2.3	125	-0.02
88	Benzo(b)fluoranthene	40.000	39.820	0.4	121	-0.02
89	Benzo(k)fluoranthene	40.000	38.728	3.2	119	-0.02
90 C	Benzo(a)pyrene	40.000	40.414	-1.0	123	-0.02
91	Dibenzo(a,h)anthracene	40.000	40.853	-2.1	125	-0.04
92	Benzo(g,h,i)perylene	40.000	40.712	-1.8	125	-0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071425\
Data File : BM050429.D
Acq On : 14 Jul 2025 16:53
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Jul 14 17:29:18 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM070925.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 08 18:32:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0