

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2578

Client ID : SSTDCCC040

Date Analyzed: 07/14/2025

Lab File ID: BM050429.D

Time Analyzed: 16:53

Instrument ID: BNA_M

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	427187	7.863	1630120	10.66	1036890	14.49
UPPER LIMIT	854374	8.363	3260240	11.157	2073780	14.986
LOWER LIMIT	213594	7.363	815060	10.157	518445	13.986
EPA SAMPLE NO.						
01 WC-A5-01-C	406303	7.86	1510570	10.66	1001060	14.49
02 WC-A2-09-C	411418	7.86	1469590	10.66	901307	14.49
03 WC-A2-10-C	396203	7.86	1451300	10.66	964408	14.49

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.