284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	ENTACT	Date Collected:	07/10/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	07/10/25
Client Sample ID:	WC-A2-10-G	SDG No.:	Q2578
Lab Sample ID:	Q2578-09	Matrix:	TCLP
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	TCLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :	SW5035		

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX046978.D	1	07/14/25 16:49	VX071425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	0.00026	U	0.00026	0.0050	mg/L
75-35-4	1,1-Dichloroethene	0.00023	U	0.00023	0.0050	mg/L
78-93-3	2-Butanone	0.00098	U	0.00098	0.025	mg/L
56-23-5	Carbon Tetrachloride	0.00025	U	0.00025	0.0050	mg/L
67-66-3	Chloroform	0.00025	U	0.00025	0.0050	mg/L
71-43-2	Benzene	0.00015	U	0.00015	0.0050	mg/L
107-06-2	1,2-Dichloroethane	0.00022	U	0.00022	0.0050	mg/L
79-01-6	Trichloroethene	0.000090	U	0.000090	0.0050	mg/L
127-18-4	Tetrachloroethene	0.00023	U	0.00023	0.0050	mg/L
108-90-7	Chlorobenzene	0.00012	U	0.00012	0.0050	mg/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.1		70 (74) - 130 (125)	106%	SPK: 50
1868-53-7	Dibromofluoromethane	0.010	*	70 (75) - 130 (124)	0%	SPK: 50
2037-26-5	Toluene-d8	48.1		70 (86) - 130 (113)	96%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.7		70 (77) - 130 (121)	103%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	383000	5.568			
540-36-3	1,4-Difluorobenzene	656000	6.769			
3114-55-4	Chlorobenzene-d5	592000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	310000	12.018			

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products