

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ENTA05
 Lab Code: ACE SDG NO.: Q2578
 Lab File ID: VX046961.D Date Analyzed: 07/14/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:16
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	220043	12.018				
UPPER LIMIT	440086	12.518				
LOWER LIMIT	110022	11.518				
EPA SAMPLE NO.						
WC-A5-01-G	270125	12.02				
WC-A2-09-G	272001	12.02				
WC-A2-10-G	309717	12.02				
WC-A2-11-G	268456	12.02				
WC-A2-12-G	295784	12.02				
VX0714WBL01	263407	12.02				
VX0714WBS01	241384	12.02				
VX0714WBSD01	230311	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.