

## Cover Page

**Order ID :** Q2579

**Project ID :** 540 Degraw St, Brooklyn, NY - E9309

**Client :** ENTACT

**Lab Sample Number**

Q2579-01  
Q2579-02  
Q2579-03  
Q2579-04  
Q2579-05  
Q2579-06  
Q2579-07  
Q2579-08

**Client Sample Number**

WC-A2-13-G  
WC-A2-13-C  
WC-A2-13-C  
WC-A2-13-C  
WC-A2-14-G  
WC-A2-14-C  
WC-A2-14-C  
WC-A2-14-C

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : \_\_\_\_\_

Date: 7/19/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

## DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following “ Results Qualifiers” are used:

|           |                                                                                                                                                                                                                                                                                                                                                                               |
|-----------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <b>J</b>  | Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).                                                                                                                                                                                  |
| <b>U</b>  | Indicates the analyte was analyzed for, but not detected.                                                                                                                                                                                                                                                                                                                     |
| <b>ND</b> | Indicates the analyte was analyzed for, but not detected                                                                                                                                                                                                                                                                                                                      |
| <b>E</b>  | Indicates the reported value is estimated because of the presence of interference                                                                                                                                                                                                                                                                                             |
| <b>M</b>  | Indicates Duplicate injection precision not met.                                                                                                                                                                                                                                                                                                                              |
| <b>N</b>  | Indicates the spiked sample recovery is not within control limits.                                                                                                                                                                                                                                                                                                            |
| <b>S</b>  | Indicates the reported value was determined by the Method of Standard Addition (MSA).                                                                                                                                                                                                                                                                                         |
| <b>*</b>  | Indicates that the duplicate analysis is not within control limits.                                                                                                                                                                                                                                                                                                           |
| <b>+</b>  | Indicates the correlation coefficient for the MSA is less than 0.995.                                                                                                                                                                                                                                                                                                         |
| <b>D</b>  | Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.                                                                                                                                                                                                                                       |
| <b>M</b>  | Method qualifiers<br>“P” for ICP instrument<br>“PM” for ICP when Microwave Digestion is used<br>“CV” for Manual Cold Vapor AA<br>“AV” for automated Cold Vapor AA<br>“CA” for MIDI-Distillation Spectrophotometric<br>“AS” for Semi -Automated Spectrophotometric<br>“C” for Manual Spectrophotometric<br>“T” for Titrimetric<br>“NR” for analyte not required to be analyzed |
| <b>OR</b> | Indicates the analyte’s concentration exceeds the calibrated range of the instrument for that specific analysis.                                                                                                                                                                                                                                                              |
| <b>Q</b>  | Indicates the LCS did not meet the control limits requirements                                                                                                                                                                                                                                                                                                                |
| <b>H</b>  | Sample Analysis Out Of Hold Time                                                                                                                                                                                                                                                                                                                                              |

## APPENDIX A

### QA REVIEW GENERAL DOCUMENTATION

Project #: Q2579

Completed

For thorough review, the report must have the following:

#### GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

#### COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

#### CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

#### ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: MAHESH PATEL

Date: 07/19/2025

## LAB CHRONICLE

|                 |                 |                   |                                     |
|-----------------|-----------------|-------------------|-------------------------------------|
| <b>OrderID:</b> | Q2579           | <b>OrderDate:</b> | 7/10/2025 3:31:00 PM                |
| <b>Client:</b>  | ENTACT          | <b>Project:</b>   | 540 Degraw St, Brooklyn, NY - E9309 |
| <b>Contact:</b> | Austin Farmerie | <b>Location:</b>  | D41                                 |

| LabID    | ClientID   | Matrix | Test                | Method     | Sample Date       | Prep Date | Anal Date         | Received |
|----------|------------|--------|---------------------|------------|-------------------|-----------|-------------------|----------|
| Q2579-02 | WC-A2-13-C | SOIL   |                     |            | 07/10/25<br>12:00 |           |                   | 07/10/25 |
|          |            |        | Oil and Grease      | 9071B      |                   |           | 07/15/25<br>11:45 |          |
|          |            |        | Paint Filter        | 9095B      |                   |           | 07/14/25<br>10:35 |          |
|          |            |        | pH                  | 9045D      |                   |           | 07/11/25<br>15:35 |          |
|          |            |        | TS                  | SM2540 B   |                   |           | 07/11/25<br>11:00 |          |
|          |            |        | TVS                 | 160.4      |                   |           | 07/11/25<br>16:00 |          |
| Q2579-03 | WC-A2-13-C | SOIL   |                     |            | 07/10/25<br>12:00 |           |                   | 07/10/25 |
|          |            |        | Corrosivity         | 9045D      |                   |           | 07/11/25<br>15:35 |          |
|          |            |        | Ignitability        | 1030       |                   |           | 07/14/25<br>09:15 |          |
|          |            |        | Reactive Cyanide    | 9012B      |                   | 07/11/25  | 07/11/25<br>13:13 |          |
|          |            |        | Reactive Sulfide    | 9034       |                   | 07/11/25  | 07/11/25<br>15:47 |          |
| Q2579-04 | WC-A2-13-C | WATER  |                     |            | 07/10/25<br>12:00 |           |                   | 07/10/25 |
|          |            |        | ASTM Ammonia        | SM4500-NH3 |                   | 07/14/25  | 07/14/25<br>16:28 |          |
|          |            |        | ASTM COD            | SM5220 D   |                   |           | 07/15/25<br>12:32 |          |
|          |            |        | ASTM Oil and Grease | 1664A      |                   |           | 07/14/25<br>13:00 |          |



## LAB CHRONICLE

|                  |            |          |                     |            |          |                   |                   |
|------------------|------------|----------|---------------------|------------|----------|-------------------|-------------------|
| Q2579-06         | WC-A2-14-C | SOIL     | ASTM TS             | SM2540 B   |          | 07/14/25<br>12:00 | 07/10/25<br>12:00 |
|                  |            |          | Oil and Grease      | 9071B      |          | 07/15/25<br>11:45 |                   |
|                  |            |          | Paint Filter        | 9095B      |          | 07/14/25<br>10:42 |                   |
|                  |            |          | pH                  | 9045D      |          | 07/11/25<br>15:45 |                   |
|                  |            |          | TS                  | SM2540 B   |          | 07/11/25<br>11:00 |                   |
|                  |            |          | TVS                 | 160.4      |          | 07/11/25<br>16:00 |                   |
|                  |            |          | Q2579-07            | WC-A2-14-C | SOIL     |                   |                   |
| Corrosivity      | 9045D      |          |                     |            |          | 07/11/25<br>15:45 |                   |
| Ignitability     | 1030       |          |                     |            |          | 07/14/25<br>09:22 |                   |
| Reactive Cyanide | 9012B      | 07/11/25 |                     |            |          | 07/11/25<br>13:13 |                   |
| Reactive Sulfide | 9034       | 07/11/25 |                     |            |          | 07/11/25<br>15:49 |                   |
| Q2579-08         | WC-A2-14-C | WATER    |                     |            |          |                   | 07/10/25<br>12:00 |
|                  |            |          | ASTM Ammonia        | SM4500-NH3 | 07/14/25 | 07/14/25<br>16:28 |                   |
|                  |            |          | ASTM COD            | SM5220 D   |          | 07/15/25<br>12:33 |                   |
|                  |            |          | ASTM Oil and Grease | 1664A      |          | 07/14/25<br>13:00 |                   |
|                  |            |          | ASTM TS             | SM2540 B   |          | 07/14/25<br>12:00 |                   |



# SAMPLE DATA

## Report of Analysis

|                   |                                     |                 |                |
|-------------------|-------------------------------------|-----------------|----------------|
| Client:           | ENTACT                              | Date Collected: | 07/10/25 12:00 |
| Project:          | 540 Degraw St, Brooklyn, NY - E9309 | Date Received:  | 07/10/25       |
| Client Sample ID: | WC-A2-13-C                          | SDG No.:        | Q2579          |
| Lab Sample ID:    | Q2579-02                            | Matrix:         | SOIL           |
|                   |                                     | % Solid:        | 72.1           |

| Parameter      | Conc. | Qua. | DF | MDL  | LOQ / CRQL | Units(Dry Weight) | Prep Date | Date Ana.      | Ana Met.     |
|----------------|-------|------|----|------|------------|-------------------|-----------|----------------|--------------|
| Oil and Grease | 374   |      | 1  | 8.04 | 34.6       | mg/Kg             |           | 07/15/25 11:45 | SW9071B      |
| Paint Filter   | 1.00  | U    | 1  | 1.00 | 1.00       | ml/100gm          |           | 07/14/25 10:35 | 9095B        |
| pH             | 11.4  | H    | 1  | 0    | 0          | pH                |           | 07/11/25 15:35 | 9045D        |
| TS             | 71.7  |      | 1  | 1.00 | 5.00       | %                 |           | 07/11/25 11:00 | SM 2540 B-20 |
| TVS            | 4.60  | J    | 1  | 1.00 | 10.0       | %                 |           | 07/11/25 16:00 | 160.4        |

Comments: pH result reported at temperature 21.1 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

## Report of Analysis

|                   |                                     |                 |                |
|-------------------|-------------------------------------|-----------------|----------------|
| Client:           | ENTACT                              | Date Collected: | 07/10/25 12:00 |
| Project:          | 540 Degraw St, Brooklyn, NY - E9309 | Date Received:  | 07/10/25       |
| Client Sample ID: | WC-A2-13-C                          | SDG No.:        | Q2579          |
| Lab Sample ID:    | Q2579-03                            | Matrix:         | SOIL           |
|                   |                                     | % Solid:        | 100            |

| Parameter        | Conc.  | Qua. | DF | MDL    | LOQ / CRQL | Units | Prep Date      | Date Ana.      | Ana Met. |
|------------------|--------|------|----|--------|------------|-------|----------------|----------------|----------|
| Corrosivity      | 11.4   | H    | 1  | 0      | 0          | pH    |                | 07/11/25 15:35 | 9045D    |
| Ignitability     | NO     |      | 1  | 0      | 0          | oC    |                | 07/14/25 09:15 | 1030     |
| Reactive Cyanide | 0.0083 | U    | 1  | 0.0083 | 0.050      | mg/Kg | 07/11/25 08:45 | 07/11/25 13:13 | 9012B    |
| Reactive Sulfide | 6.36   | J    | 1  | 0.20   | 10.0       | mg/Kg | 07/11/25 12:45 | 07/11/25 15:47 | 9034     |

Comments: pH result reported at temperature 21.1 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

## Report of Analysis

|                   |                                     |                 |                |
|-------------------|-------------------------------------|-----------------|----------------|
| Client:           | ENTACT                              | Date Collected: | 07/10/25 12:00 |
| Project:          | 540 Degraw St, Brooklyn, NY - E9309 | Date Received:  | 07/10/25       |
| Client Sample ID: | WC-A2-13-C                          | SDG No.:        | Q2579          |
| Lab Sample ID:    | Q2579-04                            | Matrix:         | WATER          |
|                   |                                     | % Solid:        | 0              |

| Parameter           | Conc. | Qua. | DF | MDL   | LOQ / CRQL | Units | Prep Date      | Date Ana.      | Ana Met.                          |
|---------------------|-------|------|----|-------|------------|-------|----------------|----------------|-----------------------------------|
| ASTM Ammonia        | 0.40  |      | 1  | 0.030 | 0.10       | mg/L  | 07/14/25 12:50 | 07/14/25 16:28 | SM 4500-NH3<br>B plus NH3<br>G-11 |
| ASTM COD            | 46.9  |      | 1  | 1.50  | 10.0       | mg/L  |                | 07/15/25 12:32 | SM 5220 D-11                      |
| ASTM Oil and Grease | 0.50  | J    | 1  | 0.29  | 5.00       | mg/L  |                | 07/14/25 13:00 | SW1664A                           |
| ASTM TS             | 528   |      | 1  | 1.00  | 5.00       | mg/L  |                | 07/14/25 12:00 | SM 2540 B-20                      |

Comments:

U = Not Detected  
LOQ = Limit of Quantitation  
MDL = Method Detection Limit  
LOD = Limit of Detection  
D = Dilution  
Q = indicates LCS control criteria did not meet requirements  
H = Sample Analysis Out Of Hold Time

J = Estimated Value  
B = Analyte Found in Associated Method Blank  
\* = indicates the duplicate analysis is not within control limits.  
E = Indicates the reported value is estimated because of the presence of interference.  
OR = Over Range  
N = Spiked sample recovery not within control limits

## Report of Analysis

|                   |                                     |                 |                |
|-------------------|-------------------------------------|-----------------|----------------|
| Client:           | ENTACT                              | Date Collected: | 07/10/25 12:00 |
| Project:          | 540 Degraw St, Brooklyn, NY - E9309 | Date Received:  | 07/10/25       |
| Client Sample ID: | WC-A2-14-C                          | SDG No.:        | Q2579          |
| Lab Sample ID:    | Q2579-06                            | Matrix:         | SOIL           |
|                   |                                     | % Solid:        | 82.5           |

| Parameter      | Conc. | Qua. | DF | MDL  | LOQ / CRQL | Units(Dry Weight) | Prep Date | Date Ana.      | Ana Met.     |
|----------------|-------|------|----|------|------------|-------------------|-----------|----------------|--------------|
| Oil and Grease | 127   |      | 1  | 7.03 | 30.3       | mg/Kg             |           | 07/15/25 11:45 | SW9071B      |
| Paint Filter   | 1.00  | U    | 1  | 1.00 | 1.00       | ml/100gm          |           | 07/14/25 10:42 | 9095B        |
| pH             | 11.9  | H    | 1  | 0    | 0          | pH                |           | 07/11/25 15:45 | 9045D        |
| TS             | 82.2  |      | 1  | 1.00 | 5.00       | %                 |           | 07/11/25 11:00 | SM 2540 B-20 |
| TVS            | 1.80  | J    | 1  | 1.00 | 10.0       | %                 |           | 07/11/25 16:00 | 160.4        |

Comments: pH result reported at temperature 21.2 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

## Report of Analysis

|                   |                                     |                 |                |
|-------------------|-------------------------------------|-----------------|----------------|
| Client:           | ENTACT                              | Date Collected: | 07/10/25 12:00 |
| Project:          | 540 Degraw St, Brooklyn, NY - E9309 | Date Received:  | 07/10/25       |
| Client Sample ID: | WC-A2-14-C                          | SDG No.:        | Q2579          |
| Lab Sample ID:    | Q2579-07                            | Matrix:         | SOIL           |
|                   |                                     | % Solid:        | 100            |

| Parameter        | Conc.  | Qua. | DF | MDL    | LOQ / CRQL | Units | Prep Date      | Date Ana.      | Ana Met. |
|------------------|--------|------|----|--------|------------|-------|----------------|----------------|----------|
| Corrosivity      | 11.9   | H    | 1  | 0      | 0          | pH    |                | 07/11/25 15:45 | 9045D    |
| Ignitability     | NO     |      | 1  | 0      | 0          | oC    |                | 07/14/25 09:22 | 1030     |
| Reactive Cyanide | 0.0083 | U    | 1  | 0.0083 | 0.050      | mg/Kg | 07/11/25 08:45 | 07/11/25 13:13 | 9012B    |
| Reactive Sulfide | 4.75   | J    | 1  | 0.20   | 10.0       | mg/Kg | 07/11/25 12:45 | 07/11/25 15:49 | 9034     |

Comments: pH result reported at temperature 212 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

## Report of Analysis

|                   |                                     |                 |                |
|-------------------|-------------------------------------|-----------------|----------------|
| Client:           | ENTACT                              | Date Collected: | 07/10/25 12:00 |
| Project:          | 540 Degraw St, Brooklyn, NY - E9309 | Date Received:  | 07/10/25       |
| Client Sample ID: | WC-A2-14-C                          | SDG No.:        | Q2579          |
| Lab Sample ID:    | Q2579-08                            | Matrix:         | WATER          |
|                   |                                     | % Solid:        | 0              |

| Parameter           | Conc. | Qua. | DF | MDL   | LOQ / CRQL | Units | Prep Date      | Date Ana.      | Ana Met.                          |
|---------------------|-------|------|----|-------|------------|-------|----------------|----------------|-----------------------------------|
| ASTM Ammonia        | 0.26  |      | 1  | 0.030 | 0.10       | mg/L  | 07/14/25 12:50 | 07/14/25 16:28 | SM 4500-NH3<br>B plus NH3<br>G-11 |
| ASTM COD            | 5.26  | J    | 1  | 1.50  | 10.0       | mg/L  |                | 07/15/25 12:33 | SM 5220 D-11                      |
| ASTM Oil and Grease | 0.30  | J    | 1  | 0.29  | 5.00       | mg/L  |                | 07/14/25 13:00 | SW1664A                           |
| ASTM TS             | 878   |      | 1  | 1.00  | 5.00       | mg/L  |                | 07/14/25 12:00 | SM 2540 B-20                      |

Comments: \_\_\_\_\_

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits





# QC RESULT SUMMARY

## Initial and Continuing Calibration Verification

**Client:** ENTACT

**SDG No.:** Q2579

**Project:** 540 Degraw St, Brooklyn, NY - E9309

**RunNo.:** LB136439

| Analyte                                    | Units | Result | True Value | % Recovery | Acceptance Window (%R) | Analysis Date |
|--------------------------------------------|-------|--------|------------|------------|------------------------|---------------|
| Sample ID: <b>ICV1</b><br>Reactive Cyanide | mg/L  | 0.098  | 0.099      | 99         | 85-115                 | 07/11/2025    |
| Sample ID: <b>CCV1</b><br>Reactive Cyanide | mg/L  | 0.25   | 0.25       | 100        | 90-110                 | 07/11/2025    |
| Sample ID: <b>CCV2</b><br>Reactive Cyanide | mg/L  | 0.25   | 0.25       | 100        | 90-110                 | 07/11/2025    |
| Sample ID: <b>CCV3</b><br>Reactive Cyanide | mg/L  | 0.25   | 0.25       | 100        | 90-110                 | 07/11/2025    |
| Sample ID: <b>CCV4</b><br>Reactive Cyanide | mg/L  | 0.25   | 0.25       | 100        | 90-110                 | 07/11/2025    |

## Initial and Continuing Calibration Verification

**Client:** ENTACT

**SDG No.:** Q2579

**Project:** 540 Degraw St, Brooklyn, NY - E9309

**RunNo.:** LB136451

| Analyte    |      | Units | Result | True Value | % Recovery | Acceptance Window (%R) | Analysis Date |
|------------|------|-------|--------|------------|------------|------------------------|---------------|
| Sample ID: | ICV  |       |        |            |            |                        |               |
| pH         |      | pH    | 7.01   | 7          | 100        | 90-110                 | 07/11/2025    |
| Sample ID: | CCV1 |       |        |            |            |                        |               |
| pH         |      | pH    | 2.01   | 2.00       | 101        | 90-110                 | 07/11/2025    |
| Sample ID: | CCV2 |       |        |            |            |                        |               |
| pH         |      | pH    | 12.02  | 12.00      | 100        | 90-110                 | 07/11/2025    |

## Initial and Continuing Calibration Verification

**Client:** ENTACT

**SDG No.:** Q2579

**Project:** 540 Degraw St, Brooklyn, NY - E9309

**RunNo.:** LB136452

| Analyte     |      | Units | Result | True Value | %<br>Recovery | Acceptance<br>Window (%R) | Analysis<br>Date |
|-------------|------|-------|--------|------------|---------------|---------------------------|------------------|
| Sample ID:  | ICV  |       |        |            |               |                           |                  |
| Corrosivity |      | pH    | 7.01   | 7          | 100           | 90-110                    | 07/11/2025       |
| Sample ID:  | CCV1 |       |        |            |               |                           |                  |
| Corrosivity |      | pH    | 2.01   | 2.00       | 101           | 90-110                    | 07/11/2025       |
| Sample ID:  | CCV2 |       |        |            |               |                           |                  |
| Corrosivity |      | pH    | 12.02  | 12.00      | 100           | 90-110                    | 07/11/2025       |

## Initial and Continuing Calibration Verification

**Client:** ENTACT

**SDG No.:** Q2579

**Project:** 540 Degraw St, Brooklyn, NY - E9309

**RunNo.:** LB136468

| Analyte                                | Units | Result | True Value | % Recovery | Acceptance Window (%R) | Analysis Date |
|----------------------------------------|-------|--------|------------|------------|------------------------|---------------|
| Sample ID: <b>ICV1</b><br>ASTM Ammonia | mg/L  | 0.99   | 1          | 99         | 90-110                 | 07/14/2025    |
| Sample ID: <b>CCV1</b><br>ASTM Ammonia | mg/L  | 0.98   | 1          | 98         | 90-110                 | 07/14/2025    |
| Sample ID: <b>CCV2</b><br>ASTM Ammonia | mg/L  | 1      | 1          | 100        | 90-110                 | 07/14/2025    |
| Sample ID: <b>CCV3</b><br>ASTM Ammonia | mg/L  | 0.99   | 1          | 99         | 90-110                 | 07/14/2025    |

## Initial and Continuing Calibration Verification

**Client:** ENTACT

**SDG No.:** Q2579

**Project:** 540 Degraw St, Brooklyn, NY - E9309

**RunNo.:** LB136478

| Analyte                | Units | Result | True Value | % Recovery | Acceptance Window (%R) | Analysis Date |
|------------------------|-------|--------|------------|------------|------------------------|---------------|
| Sample ID: <b>ICV</b>  |       |        |            |            |                        |               |
| ASTM COD               | mg/L  | 50.962 | 50         | 102        | 95-105                 | 05/28/2025    |
| Sample ID: <b>CCV1</b> |       |        |            |            |                        |               |
| ASTM COD               | mg/L  | 48.931 | 50         | 98         | 95-105                 | 07/15/2025    |
| Sample ID: <b>CCV2</b> |       |        |            |            |                        |               |
| ASTM COD               | mg/L  | 50.962 | 50         | 102        | 95-105                 | 07/15/2025    |
| Sample ID: <b>CCV3</b> |       |        |            |            |                        |               |
| ASTM COD               | mg/L  | 49.946 | 50         | 100        | 95-105                 | 07/15/2025    |

### Initial and Continuing Calibration Blank Summary

**Client:** ENTACT

**SDG No.:** Q2579

**Project:** 540 Degraw St, Brooklyn, NY - E9309

**RunNo.:** LB136439

| Analyte                                    | Units | Result   | Acceptance Limits | Conc Qual | MDL     | RDL   | Analysis Date |
|--------------------------------------------|-------|----------|-------------------|-----------|---------|-------|---------------|
| Sample ID: <b>ICB1</b><br>Reactive Cyanide | mg/L  | < 0.0025 | 0.0025            | U         | 0.00096 | 0.005 | 07/11/2025    |
| Sample ID: <b>CCB1</b><br>Reactive Cyanide | mg/L  | < 0.0025 | 0.0025            | U         | 0.00096 | 0.005 | 07/11/2025    |
| Sample ID: <b>CCB2</b><br>Reactive Cyanide | mg/L  | < 0.0025 | 0.0025            | U         | 0.00096 | 0.005 | 07/11/2025    |
| Sample ID: <b>CCB3</b><br>Reactive Cyanide | mg/L  | < 0.0025 | 0.0025            | U         | 0.00096 | 0.005 | 07/11/2025    |
| Sample ID: <b>CCB4</b><br>Reactive Cyanide | mg/L  | < 0.0025 | 0.0025            | U         | 0.00096 | 0.005 | 07/11/2025    |

### Initial and Continuing Calibration Blank Summary

**Client:** ENTACT

**SDG No.:** Q2579

**Project:** 540 Degraw St, Brooklyn, NY - E9309

**RunNo.:** LB136468

| Analyte                                | Units | Result   | Acceptance Limits | Conc Qual | MDL   | RDL | Analysis Date |
|----------------------------------------|-------|----------|-------------------|-----------|-------|-----|---------------|
| Sample ID: <b>ICB1</b><br>ASTM Ammonia | mg/L  | < 0.0500 | 0.0500            | U         | 0.030 | 0.1 | 07/14/2025    |
| Sample ID: <b>CCB1</b><br>ASTM Ammonia | mg/L  | < 0.0500 | 0.0500            | U         | 0.030 | 0.1 | 07/14/2025    |
| Sample ID: <b>CCB2</b><br>ASTM Ammonia | mg/L  | < 0.0500 | 0.0500            | U         | 0.030 | 0.1 | 07/14/2025    |
| Sample ID: <b>CCB3</b><br>ASTM Ammonia | mg/L  | < 0.0500 | 0.0500            | U         | 0.030 | 0.1 | 07/14/2025    |



### Initial and Continuing Calibration Blank Summary

**Client:** ENTACT

**SDG No.:** Q2579

**Project:** 540 Degraw St, Brooklyn, NY - E9309

**RunNo.:** LB136478

| Analyte                            | Units | Result   | Acceptance Limits | Conc Qual | MDL  | RDL | Analysis Date |
|------------------------------------|-------|----------|-------------------|-----------|------|-----|---------------|
| Sample ID: <b>ICB</b><br>ASTM COD  | mg/L  | < 5.0000 | 5.0000            | U         | 1.50 | 10  | 05/28/2025    |
| Sample ID: <b>CCB1</b><br>ASTM COD | mg/L  | < 5.0000 | 5.0000            | U         | 1.50 | 10  | 07/15/2025    |
| Sample ID: <b>CCB2</b><br>ASTM COD | mg/L  | < 5.0000 | 5.0000            | U         | 1.50 | 10  | 07/15/2025    |
| Sample ID: <b>CCB3</b><br>ASTM COD | mg/L  | < 5.0000 | 5.0000            | U         | 1.50 | 10  | 07/15/2025    |

## Preparation Blank Summary

**Client:** ENTACT

**SDG No.:** Q2579

**Project:** 540 Degraw St, Brooklyn, NY - E9309

| Analyte                                             | Units | Result    | Acceptance Limits | Conc Qual | MDL    | RDL  | Analysis Date |
|-----------------------------------------------------|-------|-----------|-------------------|-----------|--------|------|---------------|
| Sample ID: <b>LB136448BL</b><br>TS                  | %     | < 2.5000  | 2.5000            | U         | 1      | 5    | 07/11/2025    |
| Sample ID: <b>LB136449BL</b><br>TVS                 | %     | < 5.0000  | 5.0000            | U         | 1      | 10   | 07/11/2025    |
| Sample ID: <b>LB136455BL</b><br>ASTM TS             | mg/L  | < 2.5000  | 2.5000            | U         | 1      | 5    | 07/14/2025    |
| Sample ID: <b>LB136456BL</b><br>ASTM Oil and Grease | mg/L  | < 2.5000  | 2.5000            | U         | 0.29   | 5.0  | 07/14/2025    |
| Sample ID: <b>LB136474BL</b><br>Oil and Grease      | mg/Kg | < 12.5000 | 12.5000           | U         | 5.8    | 25   | 07/15/2025    |
| Sample ID: <b>LB136478BL</b><br>ASTM COD            | mg/L  | < 5.0000  | 5.0000            | U         | 1.5    | 10.0 | 07/15/2025    |
| Sample ID: <b>PB168800BL</b><br>Reactive Sulfide    | mg/Kg | < 5.0000  | 5.0000            | U         | 0.201  | 10   | 07/11/2025    |
| Sample ID: <b>PB168801BL</b><br>Reactive Cyanide    | mg/Kg | < 0.0250  | 0.0250            | U         | 0.0084 | 0.05 | 07/11/2025    |
| Sample ID: <b>PB168836BL</b><br>ASTM Ammonia        | mg/L  | < 0.0500  | 0.0500            | U         | 0.03   | 0.1  | 07/14/2025    |

## Matrix Spike Summary

|                   |                                     |                                         |          |
|-------------------|-------------------------------------|-----------------------------------------|----------|
| <b>Client:</b>    | ENTACT                              | <b>SDG No.:</b>                         | Q2579    |
| <b>Project:</b>   | 540 Degraw St, Brooklyn, NY - E9309 | <b>Sample ID:</b>                       | Q2578-04 |
| <b>Client ID:</b> | WC-A5-01-CMS                        | <b>Percent Solids for Spike Sample:</b> | 0        |

| Analyte             | Units | Acceptance<br>Limit %R | Spiked<br>Result | Conc.<br>Qualifier | Sample<br>Result | Conc.<br>Qualifier | Spike<br>Added | Dilution<br>Factor | %<br>Rec | Qual | Analysis<br>Date |
|---------------------|-------|------------------------|------------------|--------------------|------------------|--------------------|----------------|--------------------|----------|------|------------------|
| ASTM COD            | mg/L  | 75-125                 | 98.7             |                    | 48.9             |                    | 50.0           | 1                  | 100      |      | 07/15/2025       |
| ASTM Ammonia        | mg/L  | 75-125                 | 1.10             |                    | 0.23             |                    | 1              | 1                  | 87       |      | 07/14/2025       |
| ASTM Oil and Grease | mg/L  | 78-114                 | 20.4             |                    | 0.30             | J                  | 20.0           | 1                  | 101      |      | 07/14/2025       |

## Matrix Spike Summary

|                   |                                     |                                         |          |
|-------------------|-------------------------------------|-----------------------------------------|----------|
| <b>Client:</b>    | ENTACT                              | <b>SDG No.:</b>                         | Q2579    |
| <b>Project:</b>   | 540 Degraw St, Brooklyn, NY - E9309 | <b>Sample ID:</b>                       | Q2578-04 |
| <b>Client ID:</b> | WC-A5-01-CMSD                       | <b>Percent Solids for Spike Sample:</b> | 0        |

| Analyte             | Units | Acceptance<br>Limit %R | Spiked<br>Result | Conc.<br>Qualifier | Sample<br>Result | Conc.<br>Qualifier | Spike<br>Added | Dilution<br>Factor | %<br>Rec | Qual | Analysis<br>Date |
|---------------------|-------|------------------------|------------------|--------------------|------------------|--------------------|----------------|--------------------|----------|------|------------------|
| ASTM COD            | mg/L  | 75-125                 | 97.7             |                    | 48.9             |                    | 50.0           | 1                  | 98       |      | 07/15/2025       |
| ASTM Ammonia        | mg/L  | 75-125                 | 1.30             |                    | 0.23             |                    | 1              | 1                  | 107      |      | 07/14/2025       |
| ASTM Oil and Grease | mg/L  | 78-114                 | 20.5             |                    | 0.30             | J                  | 20.0           | 1                  | 101      |      | 07/14/2025       |

## Matrix Spike Summary

|                   |                                     |                                         |          |
|-------------------|-------------------------------------|-----------------------------------------|----------|
| <b>Client:</b>    | ENTACT                              | <b>SDG No.:</b>                         | Q2579    |
| <b>Project:</b>   | 540 Degraw St, Brooklyn, NY - E9309 | <b>Sample ID:</b>                       | Q2579-06 |
| <b>Client ID:</b> | WC-A2-14-CMS                        | <b>Percent Solids for Spike Sample:</b> | 82.5     |

| Analyte        | Units | Acceptance<br>Limit %R | Spiked<br>Result | Conc.<br>Qualifier | Sample<br>Result | Conc.<br>Qualifier | Spike<br>Added | Dilution<br>Factor | %<br>Rec | Qual | Analysis<br>Date |
|----------------|-------|------------------------|------------------|--------------------|------------------|--------------------|----------------|--------------------|----------|------|------------------|
| Oil and Grease | mg/Kg | 75-125                 | 200              |                    | 127              |                    | 121            | 1                  | 60       | *    | 07/15/2025       |

## Matrix Spike Summary

|                   |                                     |                                         |          |
|-------------------|-------------------------------------|-----------------------------------------|----------|
| <b>Client:</b>    | ENTACT                              | <b>SDG No.:</b>                         | Q2579    |
| <b>Project:</b>   | 540 Degraw St, Brooklyn, NY - E9309 | <b>Sample ID:</b>                       | Q2579-06 |
| <b>Client ID:</b> | WC-A2-14-CMSD                       | <b>Percent Solids for Spike Sample:</b> | 82.5     |

| Analyte        | Units | Acceptance<br>Limit %R | Spiked<br>Result | Conc.<br>Qualifier | Sample<br>Result | Conc.<br>Qualifier | Spike<br>Added | Dilution<br>Factor | %<br>Rec | Qual | Analysis<br>Date |
|----------------|-------|------------------------|------------------|--------------------|------------------|--------------------|----------------|--------------------|----------|------|------------------|
| Oil and Grease | mg/Kg | 75-125                 | 181              |                    | 127              |                    | 121            | 1                  | 45       | *    | 07/15/2025       |

## Duplicate Sample Summary

|                                                                                                            |                                                                                                    |
|------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------|
| <b>Client:</b> ENTACT<br><b>Project:</b> 540 Degraw St, Brooklyn, NY - E9309<br><b>Client ID:</b> TP-15DUP | <b>SDG No.:</b> Q2579<br><b>Sample ID:</b> Q2519-04<br><b>Percent Solids for Spike Sample:</b> 100 |
|------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------|

| Analyte          | Units | Acceptance Limit | Sample Result | Conc. Qualifier | Duplicate Result | Conc. Qualifier | Dilution Factor | RPD/AD | Qual | Analysis Date |
|------------------|-------|------------------|---------------|-----------------|------------------|-----------------|-----------------|--------|------|---------------|
| Reactive Cyanide | mg/Kg | +/-20            | 0.0083        | U               | 0.0083           | U               | 1               | 0      |      | 07/11/2025    |
| Reactive Sulfide | mg/Kg | +/-20            | 3.16          | J               | 3.16             | J               | 1               | 0      |      | 07/11/2025    |

## Duplicate Sample Summary

|                                                                                                                 |                                                                                                     |
|-----------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------|
| <b>Client:</b> ENTACT<br><b>Project:</b> 540 Degraw St, Brooklyn, NY - E9309<br><b>Client ID:</b> WC-A5-01-CDUP | <b>SDG No.:</b> Q2579<br><b>Sample ID:</b> Q2578-02<br><b>Percent Solids for Spike Sample:</b> 84.9 |
|-----------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------|

| Analyte | Units | Acceptance Limit | Sample Result | Conc. Qualifier | Duplicate Result | Conc. Qualifier | Dilution Factor | RPD/AD | Qual | Analysis Date |
|---------|-------|------------------|---------------|-----------------|------------------|-----------------|-----------------|--------|------|---------------|
| TS      | %     | +/-5             | 84.8          |                 | 84.8             |                 | 1               | 0      |      | 07/11/2025    |
| TVS     | %     | +/-5             | 2.90          | J               | 2.80             | J               | 1               | 3.51   |      | 07/11/2025    |



## Duplicate Sample Summary

|                                                                                                                 |                                                                                                    |
|-----------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------|
| <b>Client:</b> ENTACT<br><b>Project:</b> 540 Degraw St, Brooklyn, NY - E9309<br><b>Client ID:</b> WC-A5-01-CDUP | <b>SDG No.:</b> Q2579<br><b>Sample ID:</b> Q2578-02<br><b>Percent Solids for Spike Sample:</b> 100 |
|-----------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------|

| Analyte      | Units    | Acceptance Limit | Sample Result | Conc. Qualifier | Duplicate Result | Conc. Qualifier | Dilution Factor | RPD/AD | Qual | Analysis Date |
|--------------|----------|------------------|---------------|-----------------|------------------|-----------------|-----------------|--------|------|---------------|
| pH           | pH       | +/-20            | 11.7          |                 | 11.7             |                 | 1               | 0.17   |      | 07/11/2025    |
| Paint Filter | ml/100gm | +/-20            | 1.00          | U               | 1.00             | U               | 1               | 0      |      | 07/14/2025    |

## Duplicate Sample Summary

|                                                                                                                 |                                                                                                    |
|-----------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------|
| <b>Client:</b> ENTACT<br><b>Project:</b> 540 Degraw St, Brooklyn, NY - E9309<br><b>Client ID:</b> WC-A5-01-CDUP | <b>SDG No.:</b> Q2579<br><b>Sample ID:</b> Q2578-03<br><b>Percent Solids for Spike Sample:</b> 100 |
|-----------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------|

| Analyte      | Units | Acceptance Limit | Sample Result | Conc. Qualifier | Duplicate Result | Conc. Qualifier | Dilution Factor | RPD/AD | Qual | Analysis Date |
|--------------|-------|------------------|---------------|-----------------|------------------|-----------------|-----------------|--------|------|---------------|
| Corrosivity  | pH    | +/-20            | 11.7          |                 | 11.7             |                 | 1               | 0.17   |      | 07/11/2025    |
| Ignitability | oC    | +/-20            | NO            |                 | NO               |                 | 1               | 0      |      | 07/14/2025    |

## Duplicate Sample Summary

|                                                                                                                 |                                                                                                  |
|-----------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------------------|
| <b>Client:</b> ENTACT<br><b>Project:</b> 540 Degraw St, Brooklyn, NY - E9309<br><b>Client ID:</b> WC-A5-01-CDUP | <b>SDG No.:</b> Q2579<br><b>Sample ID:</b> Q2578-04<br><b>Percent Solids for Spike Sample:</b> 0 |
|-----------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------------------|

| Analyte             | Units | Acceptance Limit | Sample Result | Conc. Qualifier | Duplicate Result | Conc. Qualifier | Dilution Factor | RPD/AD | Qual | Analysis Date |
|---------------------|-------|------------------|---------------|-----------------|------------------|-----------------|-----------------|--------|------|---------------|
| ASTM TS             | mg/L  | +/-5             | 658           |                 | 647              |                 | 1               | 1.69   |      | 07/14/2025    |
| ASTM Oil and Grease | mg/L  | +/-18            | 0.30          | J               | 0.30             | J               | 1               | 0      |      | 07/14/2025    |
| ASTM Ammonia        | mg/L  | +/-20            | 0.23          |                 | 0.23             |                 | 1               | 0      |      | 07/14/2025    |
| ASTM COD            | mg/L  | +/-20            | 48.9          |                 | 49.9             |                 | 1               | 2.02   |      | 07/15/2025    |

## Duplicate Sample Summary

|                   |                                     |                                         |          |
|-------------------|-------------------------------------|-----------------------------------------|----------|
| <b>Client:</b>    | ENTACT                              | <b>SDG No.:</b>                         | Q2579    |
| <b>Project:</b>   | 540 Degraw St, Brooklyn, NY - E9309 | <b>Sample ID:</b>                       | Q2578-04 |
| <b>Client ID:</b> | WC-A5-01-CMSD                       | <b>Percent Solids for Spike Sample:</b> | 0        |

| Analyte             | Units | Acceptance Limit | Sample Result | Conc. Qualifier | Duplicate Result | Conc. Qualifier | Dilution Factor | RPD/AD | Qual | Analysis Date |
|---------------------|-------|------------------|---------------|-----------------|------------------|-----------------|-----------------|--------|------|---------------|
| ASTM Oil and Grease | mg/L  | +/-18            | 20.4          |                 | 20.5             |                 | 1               | 0.49   |      | 07/14/2025    |
| ASTM Ammonia        | mg/L  | +/-20            | 1.10          |                 | 1.30             |                 | 1               | 17     |      | 07/14/2025    |
| ASTM COD            | mg/L  | +/-20            | 98.7          |                 | 97.7             |                 | 1               | 1.02   |      | 07/15/2025    |

## Duplicate Sample Summary

|                                                                                                                 |                                                                                                     |
|-----------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------|
| <b>Client:</b> ENTACT<br><b>Project:</b> 540 Degraw St, Brooklyn, NY - E9309<br><b>Client ID:</b> WC-A2-14-CDUP | <b>SDG No.:</b> Q2579<br><b>Sample ID:</b> Q2579-06<br><b>Percent Solids for Spike Sample:</b> 82.5 |
|-----------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------|

| Analyte        | Units | Acceptance<br>Limit | Sample<br>Result | Conc.<br>Qualifier | Duplicate<br>Result | Conc.<br>Qualifier | Dilution<br>Factor | RPD/<br>AD | Qual | Analysis<br>Date |
|----------------|-------|---------------------|------------------|--------------------|---------------------|--------------------|--------------------|------------|------|------------------|
| Oil and Grease | mg/Kg | +/-20               | 127              |                    | 127                 |                    | 1                  | 0.06       |      | 07/15/2025       |

### Duplicate Sample Summary

|                   |                                     |                                         |          |
|-------------------|-------------------------------------|-----------------------------------------|----------|
| <b>Client:</b>    | ENTACT                              | <b>SDG No.:</b>                         | Q2579    |
| <b>Project:</b>   | 540 Degraw St, Brooklyn, NY - E9309 | <b>Sample ID:</b>                       | Q2579-06 |
| <b>Client ID:</b> | WC-A2-14-CMSD                       | <b>Percent Solids for Spike Sample:</b> | 82.5     |

| Analyte        | Units | Acceptance<br>Limit | Sample<br>Result | Conc.<br>Qualifier | Duplicate<br>Result | Conc.<br>Qualifier | Dilution<br>Factor | RPD/<br>AD | Qual | Analysis<br>Date |
|----------------|-------|---------------------|------------------|--------------------|---------------------|--------------------|--------------------|------------|------|------------------|
| Oil and Grease | mg/Kg | +/-20               | 200              |                    | 181                 |                    | 1                  | 10         |      | 07/15/2025       |

### Laboratory Control Sample Summary

**Client:** ENTACT

**SDG No.:** Q2579

**Project:** 540 Degraw St, Brooklyn, NY - E9309

**Run No.:** LB136456

| Analyte             | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|---------------------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID           | LB136456BS |            |        |                 |            |                 |                     |               |
| ASTM Oil and Grease | mg/L       | 20.0       | 16.7   |                 | 84         | 1               | 78-114              | 07/14/2025    |

### Laboratory Control Sample Summary

**Client:** ENTACT

**SDG No.:** Q2579

**Project:** 540 Degraw St, Brooklyn, NY - E9309

**Run No.:** LB136474

| Analyte        | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|----------------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID      | LB136474BS |            |        |                 |            |                 |                     |               |
| Oil and Grease | mg/Kg      | 100        | 94.9   |                 | 95         | 1               | 80-120              | 07/15/2025    |



### Laboratory Control Sample Summary

**Client:** ENTACT

**SDG No.:** Q2579

**Project:** 540 Degraw St, Brooklyn, NY - E9309

**Run No.:** LB136478

| Analyte   | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|-----------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID | LB136478BS |            |        |                 |            |                 |                     |               |
| ASTM COD  | mg/L       | 50         | 51.0   |                 | 102        | 1               | 90-110              | 07/15/2025    |

### Laboratory Control Sample Summary

**Client:** ENTACT

**SDG No.:** Q2579

**Project:** 540 Degraw St, Brooklyn, NY - E9309

**Run No.:** LB136468

| Analyte      | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|--------------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID    | PB168836BS |            |        |                 |            |                 |                     |               |
| ASTM Ammonia | mg/L       | 1          | 1.00   |                 | 100        | 1               | 90-110              | 07/14/2025    |



# RAW DATA

LB136439

Test results

AquaKem 7.2AQ1

Page:

Alliance Technical Group  
284 Sheffield Street, Mountainside, NJ 07092

7/11/2025 13:32

Reviewed by : RM

Instrument ID : Konelab

Test: Total CN

| Sample Id   | Result  | Dil. 1 + | Response | Errors |
|-------------|---------|----------|----------|--------|
| ICV1        | 97.530  | 0.0      | 0.079    |        |
| ICB1        | -0.062  | 0.0      | 0.001    |        |
| CCV1        | 246.971 | 0.0      | 0.197    |        |
| CCB1        | -0.290  | 0.0      | 0.001    |        |
| PB168801BL  | -0.556  | 0.0      | 0.001    |        |
| Q2519-01    | -0.523  | 0.0      | 0.001    |        |
| Q2519-04DUP | -0.594  | 0.0      | 0.001    |        |
| Q2523-01    | -0.520  | 0.0      | 0.001    |        |
| Q2526-05    | -0.309  | 0.0      | 0.001    |        |
| Q2526-06    | -0.397  | 0.0      | 0.001    |        |
| Q2556-01    | -0.457  | 0.0      | 0.001    |        |
| Q2565-02    | -0.490  | 0.0      | 0.001    |        |
| Q2565-03    | -0.617  | 0.0      | 0.001    |        |
| Q2565-04    | -0.482  | 0.0      | 0.001    |        |
| CCV2        | 246.625 | 0.0      | 0.197    |        |
| CCB2        | -0.202  | 0.0      | 0.001    |        |
| Q2565-05    | -0.247  | 0.0      | 0.001    |        |
| Q2571-04    | -0.371  | 0.0      | 0.001    |        |
| Q2571-08    | -0.532  | 0.0      | 0.001    |        |
| Q2578-03    | -0.375  | 0.0      | 0.001    |        |
| Q2578-07    | -0.543  | 0.0      | 0.001    |        |
| Q2578-11    | -0.309  | 0.0      | 0.001    |        |
| Q2578-15    | -0.464  | 0.0      | 0.001    |        |
| Q2578-19    | -0.422  | 0.0      | 0.001    |        |
| Q2579-03    | -0.427  | 0.0      | 0.001    |        |
| Q2579-07    | -0.243  | 0.0      | 0.001    |        |
| CCV3        | 248.058 | 0.0      | 0.198    |        |
| CCB3        | 0.005   | 0.0      | 0.001    |        |
| PB168777BL  | 0.001   | 0.0      | 0.001    |        |
| Q2545-01    | -0.418  | 0.0      | 0.001    |        |
| Q2545-01DUP | 0.190   | 0.0      | 0.002    |        |
| Q2550-01    | 0.093   | 0.0      | 0.002    |        |
| Q2551-01    | 0.122   | 0.0      | 0.002    |        |
| Q2565-01    | 0.207   | 0.0      | 0.002    |        |
| CCV4        | 251.464 | 0.0      | 0.201    |        |
| CCB4        | 0.174   | 0.0      | 0.002    |        |
| N           | 36      |          |          |        |
| Mean        | 30.044  |          |          |        |
| SD          | 79.9298 |          |          |        |
| CV%         | 266.04  |          |          |        |

Aquakem v. 7.2AQ1

Results from time period:

Fri Jul 11 12:51:15 2025

Fri Jul 11 13:21:13 2025

| Sample Id   | Sam/Ctr/c/ | Test short name | Test type | Result   | Result unit | Result date and time | Stat |
|-------------|------------|-----------------|-----------|----------|-------------|----------------------|------|
| 0.0PPBCN    | A          | Total CN        | P         | -0.9536  | µg/l        | 7/11/2025 8:54:36    |      |
| 5.0PPBCN    | A          | Total CN        | P         | 4.1163   | µg/l        | 7/11/2025 8:54:37    |      |
| 10PPBCN     | A          | Total CN        | P         | 9.0972   | µg/l        | 7/11/2025 8:54:38    |      |
| 50PPBCN     | A          | Total CN        | P         | 49.9481  | µg/l        | 7/11/2025 8:54:39    |      |
| 100PPBCN    | A          | Total CN        | P         | 100.6483 | µg/l        | 7/11/2025 8:54:40    |      |
| 250PPBCN    | A          | Total CN        | P         | 254.4826 | µg/l        | 7/11/2025 8:54:41    |      |
| 500PPBCN    | A          | Total CN        | P         | 497.6612 | µg/l        | 7/11/2025 8:54:42    |      |
| ICV1        | S          | Total CN        | P         | 97.5303  | µg/l        | 7/11/2025 12:51:16   |      |
| ICB1        | S          | Total CN        | P         | -0.0623  | µg/l        | 7/11/2025 12:51:17   |      |
| CCV1        | S          | Total CN        | P         | 246.971  | µg/l        | 7/11/2025 12:51:20   |      |
| CCB1        | S          | Total CN        | P         | -0.2903  | µg/l        | 7/11/2025 12:51:21   |      |
| PB168801BL  | S          | Total CN        | P         | -0.5562  | µg/l        | 7/11/2025 12:51:24   |      |
| Q2519-01    | S          | Total CN        | P         | -0.5232  | µg/l        | 7/11/2025 12:51:25   |      |
| Q2519-04DUP | S          | Total CN        | P         | -0.594   | µg/l        | 7/11/2025 12:58:50   |      |
| Q2523-01    | S          | Total CN        | P         | -0.5198  | µg/l        | 7/11/2025 12:58:51   |      |
| Q2526-05    | S          | Total CN        | P         | -0.3094  | µg/l        | 7/11/2025 12:58:52   |      |
| Q2526-06    | S          | Total CN        | P         | -0.3968  | µg/l        | 7/11/2025 12:58:53   |      |
| Q2556-01    | S          | Total CN        | P         | -0.4571  | µg/l        | 7/11/2025 12:58:54   |      |
| Q2565-02    | S          | Total CN        | P         | -0.4898  | µg/l        | 7/11/2025 12:58:55   |      |
| Q2565-03    | S          | Total CN        | P         | -0.6171  | µg/l        | 7/11/2025 12:58:56   |      |
| Q2565-04    | S          | Total CN        | P         | -0.4819  | µg/l        | 7/11/2025 12:58:57   |      |
| CCV2        | S          | Total CN        | P         | 246.6253 | µg/l        | 7/11/2025 12:59:00   |      |
| CCB2        | S          | Total CN        | P         | -0.2024  | µg/l        | 7/11/2025 13:06:27   |      |
| Q2565-05    | S          | Total CN        | P         | -0.2467  | µg/l        | 7/11/2025 13:06:28   |      |
| Q2571-04    | S          | Total CN        | P         | -0.3713  | µg/l        | 7/11/2025 13:06:29   |      |
| Q2571-08    | S          | Total CN        | P         | -0.5316  | µg/l        | 7/11/2025 13:06:30   |      |
| Q2578-03    | S          | Total CN        | P         | -0.375   | µg/l        | 7/11/2025 13:06:31   |      |
| Q2578-07    | S          | Total CN        | P         | -0.5433  | µg/l        | 7/11/2025 13:06:32   |      |
| Q2578-11    | S          | Total CN        | P         | -0.3092  | µg/l        | 7/11/2025 13:06:33   |      |
| Q2578-15    | S          | Total CN        | P         | -0.4644  | µg/l        | 7/11/2025 13:06:34   |      |
| Q2578-19    | S          | Total CN        | P         | -0.4218  | µg/l        | 7/11/2025 13:06:35   |      |
| Q2579-03    | S          | Total CN        | P         | -0.4266  | µg/l        | 7/11/2025 13:13:57   |      |
| Q2579-07    | S          | Total CN        | P         | -0.2425  | µg/l        | 7/11/2025 13:13:58   |      |
| CCV3        | S          | Total CN        | P         | 248.0581 | µg/l        | 7/11/2025 13:14:01   |      |
| CCB3        | S          | Total CN        | P         | 0.0046   | µg/l        | 7/11/2025 13:14:03   |      |
| PB168777BL  | S          | Total CN        | P         | 0.0012   | µg/l        | 7/11/2025 13:14:06   |      |
| Q2545-01    | S          | Total CN        | P         | -0.4177  | µg/l        | 7/11/2025 13:14:07   |      |
| Q2545-01DUP | S          | Total CN        | P         | 0.1904   | µg/l        | 7/11/2025 13:21:04   |      |
| Q2550-01    | S          | Total CN        | P         | 0.0932   | µg/l        | 7/11/2025 13:21:05   |      |
| Q2551-01    | S          | Total CN        | P         | 0.1218   | µg/l        | 7/11/2025 13:21:06   |      |
| Q2565-01    | S          | Total CN        | P         | 0.2075   | µg/l        | 7/11/2025 13:21:07   |      |
| CCV4        | S          | Total CN        | P         | 251.4644 | µg/l        | 7/11/2025 13:21:10   |      |
| CCB4        | S          | Total CN        | P         | 0.1738   | µg/l        | 7/11/2025 13:21:13   |      |

Calibration results

Aquakem 7.2AQ1

Page: 1

Alliance Technical Group  
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

7/11/2025 8:55


Test Total CN

Accepted 7/11/2025 8:55

Factor 1261  
Bias 0.001

Coeff. of det. 0.999861

Errors



|   | Calibrator | Response | Calc. con. | Conc.    | Errors |
|---|------------|----------|------------|----------|--------|
| 1 | 0.0PPBCN   | 0.001    | -0.9536    | 0.0000   | -      |
| 2 | 5.0PPBCN   | 0.005    | 4.1163     | 5.0000   | -17.7  |
| 3 | 10PPBCN    | 0.009    | 9.0972     | 10.0000  | -9.0   |
| 4 | 50PPBCN    | 0.041    | 49.9481    | 50.0000  | -0.1   |
| 5 | 100PPBCN   | 0.081    | 100.6483   | 100.0000 | 0.6    |
| 6 | 250PPBCN   | 0.203    | 254.4826   | 250.0000 | 1.8    |
| 7 | 500PPBCN   | 0.396    | 497.6612   | 500.0000 | -0.5   |

07/11/2025  
RM

Analysis Method: 9034

ANALYST: rubina

Parameter: Reactive Sulfide

SUPERVISOR REVIEW BY: Iwona

Run Number: LB136443

Constant: 16000

Normality1: 0.025

Normality2: 0.025

| Reagent/Standard                   | Lot/Log # |
|------------------------------------|-----------|
| SODIUM THIOSULFATE, 0.025N, 4LITRE | W3105     |
| IODINE SOLUTION .025N 1L           | W3213     |
| Starch Solution, 4L                | W3149     |

| Seq | Lab ID      | True Value (mg/l) | DF | Initial Weight (g) | Final Volume (ml) | T1 (ml) | T2 Initial | T2 Final | T2 Diff. (ml) | T1 - T2 Diff (mL) | Value Corrected With Blank | Result (ppm) | Anal Date  | Anal Time |
|-----|-------------|-------------------|----|--------------------|-------------------|---------|------------|----------|---------------|-------------------|----------------------------|--------------|------------|-----------|
| 1   | PB168800BL  |                   | 1  | 5.00               | 50                | 2.00    | 0.00       | 1.94     | 1.94          | 0.06              | 0.00                       | 0.00         | 07/11/2025 | 15:00     |
| 2   | Q2519-04    |                   | 1  | 5.06               | 50                | 2.00    | 0.00       | 1.90     | 1.90          | 0.10              | 0.04                       | 3.16         | 07/11/2025 | 15:03     |
| 3   | Q2519-04DUP |                   | 1  | 5.06               | 50                | 2.00    | 0.00       | 1.90     | 1.90          | 0.10              | 0.04                       | 3.16         | 07/11/2025 | 15:06     |
| 4   | Q2523-01    |                   | 1  | 5.03               | 50                | 2.00    | 0.00       | 1.86     | 1.86          | 0.14              | 0.08                       | 6.36         | 07/11/2025 | 15:09     |
| 5   | Q2526-05    |                   | 1  | 5.03               | 50                | 2.00    | 0.00       | 1.90     | 1.90          | 0.10              | 0.04                       | 3.18         | 07/11/2025 | 15:12     |
| 6   | Q2526-06    |                   | 1  | 5.07               | 50                | 2.00    | 0.00       | 1.86     | 1.86          | 0.14              | 0.08                       | 6.31         | 07/11/2025 | 15:14     |
| 7   | Q2556-01    |                   | 1  | 5.01               | 50                | 2.00    | 0.00       | 1.88     | 1.88          | 0.12              | 0.06                       | 4.79         | 07/11/2025 | 15:17     |
| 8   | Q2565-02    |                   | 1  | 5.07               | 50                | 2.00    | 0.00       | 1.90     | 1.90          | 0.10              | 0.04                       | 3.16         | 07/11/2025 | 15:20     |
| 9   | Q2565-03    |                   | 1  | 5.06               | 50                | 2.00    | 0.00       | 1.90     | 1.90          | 0.10              | 0.04                       | 3.16         | 07/11/2025 | 15:22     |
| 10  | Q2565-04    |                   | 1  | 5.02               | 50                | 2.00    | 0.00       | 1.86     | 1.86          | 0.14              | 0.08                       | 6.37         | 07/11/2025 | 15:24     |
| 11  | Q2565-05    |                   | 1  | 5.06               | 50                | 2.00    | 0.00       | 1.88     | 1.88          | 0.12              | 0.06                       | 4.74         | 07/11/2025 | 15:27     |
| 12  | Q2571-04    |                   | 1  | 5.02               | 50                | 2.00    | 0.00       | 1.90     | 1.90          | 0.10              | 0.04                       | 3.19         | 07/11/2025 | 15:30     |

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 \* Normality1) - ((T1 - Value Corrected With Blank) \* Normality2)) \* Constant / Initial Volume

Analysis Method: 9034

ANALYST: rubina

Parameter: Reactive Sulfide

SUPERVISOR REVIEW BY: Iwona

Run Number: LB136443

Constant: 16000

Normality1: 0.025

Normality2: 0.025

| Reagent/Standard                   | Lot/Log # |
|------------------------------------|-----------|
| SODIUM THIOSULFATE, 0.025N, 4LITRE | W3105     |
| IODINE SOLUTION .025N 1L           | W3213     |
| Starch Solution, 4L                | W3149     |

| Seq | Lab ID   | True Value (mg/l) | DF | Initial Weight (g) | Final Volume (ml) | T1 (ml) | T2 Initial | T2 Final | T2 Diff. (ml) | T1 - T2 Diff (mL) | Value Corrected With Blank | Result (ppm) | Anal Date  | Anal Time |
|-----|----------|-------------------|----|--------------------|-------------------|---------|------------|----------|---------------|-------------------|----------------------------|--------------|------------|-----------|
| 13  | Q2571-08 |                   | 1  | 5.01               | 50                | 2.00    | 0.00       | 1.88     | 1.88          | 0.12              | 0.06                       | 4.79         | 07/11/2025 | 15:32     |
| 14  | Q2578-03 |                   | 1  | 5.08               | 50                | 2.00    | 0.00       | 1.86     | 1.86          | 0.14              | 0.08                       | 6.30         | 07/11/2025 | 15:34     |
| 15  | Q2578-07 |                   | 1  | 5.01               | 50                | 2.00    | 0.00       | 1.86     | 1.86          | 0.14              | 0.08                       | 6.39         | 07/11/2025 | 15:36     |
| 16  | Q2578-11 |                   | 1  | 5.05               | 50                | 2.00    | 0.00       | 1.86     | 1.86          | 0.14              | 0.08                       | 6.34         | 07/11/2025 | 15:39     |
| 17  | Q2578-15 |                   | 1  | 5.06               | 50                | 2.00    | 0.00       | 1.90     | 1.90          | 0.10              | 0.04                       | 3.16         | 07/11/2025 | 15:42     |
| 18  | Q2578-19 |                   | 1  | 5.06               | 50                | 2.00    | 0.00       | 1.90     | 1.90          | 0.10              | 0.04                       | 3.16         | 07/11/2025 | 15:45     |
| 19  | Q2579-03 |                   | 1  | 5.03               | 50                | 2.00    | 0.00       | 1.86     | 1.86          | 0.14              | 0.08                       | 6.36         | 07/11/2025 | 15:47     |
| 20  | Q2579-07 |                   | 1  | 5.05               | 50                | 2.00    | 0.00       | 1.88     | 1.88          | 0.12              | 0.06                       | 4.75         | 07/11/2025 | 15:49     |

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 \* Normality1) - ((T1 - Value Corrected With Blank) \* Normality2)) \* Constant / Initial Volume



**TOTAL SOLIDS - SM2540B**

**SUPERVISOR:** Iwona

**ANALYST:** jignesh

**Date:** 07/11/2025

**Run Number:** LB136448

**BalanceID:** WC SC-6

**OvenID:** WC OVEN-1

**ThermometerID:** WET OVEN#1

**TEMP1 IN:** 104 °C 07/11/2025 11:00 **TEMP1 OUT:** 104 °C 07/11/2025 12:00  
**TEMP2 IN:** 104 °C 07/11/2025 12:30 **TEMP2 OUT:** 103 °C 07/11/2025 13:00  
**TEMP3 IN:** 104 °C 07/11/2025 16:00 **TEMP3 OUT:** 103 °C 07/14/2025 07:00  
**TEMP4 IN:** 104 °C 07/14/2025 07:30 **TEMP4 OUT:** 103 °C 07/14/2025 08:00

| Dish # | Lab ID      | Client ID     | Empty Dish Weight (g) | Final Empty Dish Weight (g) | Dish + Sample Weight (g) | Original weight 1st Dish+Sample weight after Drying @103-@105°C (g) | Constant weight 2nd Dish+Sample weight after Drying @103-@105°C (g) | Final Constant weight Final Dish+Sample weight after Drying @103-@105°C (g) | Weight (g) | Result % |
|--------|-------------|---------------|-----------------------|-----------------------------|--------------------------|---------------------------------------------------------------------|---------------------------------------------------------------------|-----------------------------------------------------------------------------|------------|----------|
| 1      | LB136448BL  | LB136448BL    | 65.2317               | 65.2317                     | 65.2317                  | 65.2317                                                             | 65.2317                                                             | 65.2317                                                                     | 0.0000     | 0        |
| 2      | Q2578-02    | WC-A5-01-C    | 78.4185               | 78.4185                     | 99.9115                  | 96.6456                                                             | 96.6456                                                             | 96.6456                                                                     | 18.2271    | 84.8     |
| 3      | Q2578-02DUP | WC-A5-01-CDUP | 82.3636               | 82.3636                     | 102.0585                 | 99.0655                                                             | 99.0655                                                             | 99.0655                                                                     | 16.7019    | 84.8     |
| 4      | Q2578-06    | WC-A2-09-C    | 78.4792               | 78.4792                     | 99.5985                  | 94.1689                                                             | 94.1689                                                             | 94.1689                                                                     | 15.6897    | 74.3     |
| 5      | Q2578-10    | WC-A2-10-C    | 53.5697               | 53.5697                     | 76.1793                  | 69.7568                                                             | 69.7568                                                             | 69.7568                                                                     | 16.1871    | 71.6     |
| 6      | Q2578-14    | WC-A2-11-C    | 88.4181               | 88.4181                     | 110.5384                 | 106.9513                                                            | 106.9513                                                            | 106.9510                                                                    | 18.5332    | 83.8     |
| 7      | Q2578-18    | WC-A2-12-C    | 90.0686               | 90.0686                     | 113.2805                 | 105.3776                                                            | 105.3776                                                            | 105.3780                                                                    | 15.3090    | 66       |
| 8      | Q2579-02    | WC-A2-13-C    | 87.3613               | 87.3613                     | 110.9608                 | 104.2760                                                            | 104.2760                                                            | 104.2760                                                                    | 16.9147    | 71.7     |
| 9      | Q2579-06    | WC-A2-14-C    | 96.9996               | 96.9996                     | 117.7224                 | 114.0235                                                            | 114.0235                                                            | 114.0240                                                                    | 17.0239    | 82.2     |

A = Final Empty Dish Weight (g)

B = Dish + Sample Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result \%} = (C - A) * 100 / (B - A)$$

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS S Q2589      WorkList ID : 190680      Department : Wet-Chemistry      Date : 07-11-2025 13:45:01

| Sample   | Customer Sample | Matrix | Test | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method   |
|----------|-----------------|--------|------|--------------|----------|-----------------------------|--------------|----------|
| Q2578-02 | WC-A5-01-C      | Solid  | TS   | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | SM2540 B |
| Q2578-06 | WC-A2-09-C      | Solid  | TS   | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | SM2540 B |
| Q2578-10 | WC-A2-10-C      | Solid  | TS   | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | SM2540 B |
| Q2578-14 | WC-A2-11-C      | Solid  | TS   | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | SM2540 B |
| Q2578-18 | WC-A2-12-C      | Solid  | TS   | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | SM2540 B |
| Q2579-02 | WC-A2-13-C      | Solid  | TS   | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | SM2540 B |
| Q2579-06 | WC-A2-14-C      | Solid  | TS   | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | SM2540 B |

Date/Time 07-11-25 14:00  
 Raw Sample Received by: CP SM  
 Raw Sample Relinquished by: CP SM

Date/Time 07-11-25 16:30  
 Raw Sample Received by: CP SM  
 Raw Sample Relinquished by: CP SM

**TOTAL VOLATILE SOLIDS 160.4**

**TEMP1 IN:** 104 °C 07/11/2025 16:00    **TEMP1 OUT:** 103 °C 07/14/2025 07:00  
**TEMP2 IN:** 104 °C 07/14/2025 07:30    **TEMP2 OUT:** 103 °C 07/14/2025 08:00  
**TEMP3 IN:** 550 °C 07/14/2025 08:30    **TEMP3 OUT:** 550 °C 07/14/2025 10:00  
**TEMP4 IN:** 540 °C 07/14/2025 10:30    **TEMP4 OUT:** 550 °C 07/14/2025 12:00

**Run Number:** LB136449

**SUPERVISOR:** Iwona

**ANALYST:** jignesh

**BalanceID:** WC SC-6

**OvenID:** WC OVEN-1

| Dish # | Lab ID      | Empty Dish Weight (g) | Final Empty Dish Weight (g) | Empty Dish + Sample Weight (g) | 1st Dish + SampleWt Drying @103-@105°C (g) | Final Dish + SampleWt Drying @103-@105°C (g) | Dish + Samplewt Drying @550 (±50) °C (g) | Final Dish + Samplewt Drying @550 (±50) °C (g) | Weight Diff (g) | Result (%) |
|--------|-------------|-----------------------|-----------------------------|--------------------------------|--------------------------------------------|----------------------------------------------|------------------------------------------|------------------------------------------------|-----------------|------------|
| 1      | LB136449BL  | 65.2317               | 65.2317                     | 65.2317                        | 65.2317                                    | 65.2317                                      | 65.2317                                  | 65.2317                                        | 0.0000          | 0          |
| 2      | Q2578-02    | 78.4185               | 78.4185                     | 99.9115                        | 96.6456                                    | 96.6456                                      | 96.1234                                  | 96.1234                                        | 0.5222          | 2.9        |
| 3      | Q2578-02DUP | 82.3636               | 82.3636                     | 102.0585                       | 99.0655                                    | 99.0655                                      | 98.6000                                  | 98.6000                                        | 0.4655          | 2.8        |
| 4      | Q2578-06    | 78.4792               | 78.4792                     | 99.5985                        | 94.1689                                    | 94.1689                                      | 93.5319                                  | 93.5319                                        | 0.6370          | 4.1        |
| 5      | Q2578-10    | 53.5697               | 53.5697                     | 76.1793                        | 69.7568                                    | 69.7568                                      | 68.7883                                  | 68.7883                                        | 0.9685          | 6          |
| 6      | Q2578-14    | 88.4181               | 88.4181                     | 110.5384                       | 106.9513                                   | 106.9513                                     | 106.5734                                 | 106.5734                                       | 0.3779          | 2          |
| 7      | Q2578-18    | 90.0686               | 90.0686                     | 113.2805                       | 105.3776                                   | 105.3776                                     | 104.7636                                 | 104.7636                                       | 0.6140          | 4          |
| 8      | Q2579-02    | 87.3613               | 87.3613                     | 110.9608                       | 104.2760                                   | 104.276                                      | 103.4899                                 | 103.4899                                       | 0.7861          | 4.6        |
| 9      | Q2579-06    | 96.9996               | 96.9996                     | 117.7224                       | 114.0235                                   | 114.0235                                     | 113.7102                                 | 113.7102                                       | 0.3133          | 1.8        |

**A =** Sample Weight (g)

**B =** Final Dish + Samplewt Drying @550 (±50) °C (g)

**C =** Final Dish + SampleWt Drying @103-@105°C (g)

**D =** Weight (g)

**E =** Final Empty Dish Weight (g)

**F =** Final Dish + SampleWt Drying @103-@105°C (g)

**Weight D =** C - B

**Result % =**  $\frac{D}{F - E} \star 100$

# WORKLIST(Hardcopy Internal Chain)

136449

WorkList Name : TVS Q2579 S      WorkList ID : 190881      Department : Wet-Chemistry      Date : 07-11-2025 13:45:52

| Sample   | Customer Sample | Matrix | Test | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method |
|----------|-----------------|--------|------|--------------|----------|-----------------------------|--------------|--------|
| Q2578-02 | WC-A5-01-C      | Solid  | TVS  | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 160.4  |
| Q2578-06 | WC-A2-09-C      | Solid  | TVS  | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 160.4  |
| Q2578-10 | WC-A2-10-C      | Solid  | TVS  | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 160.4  |
| Q2578-14 | WC-A2-11-C      | Solid  | TVS  | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 160.4  |
| Q2578-18 | WC-A2-12-C      | Solid  | TVS  | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 160.4  |
| Q2579-02 | WC-A2-13-C      | Solid  | TVS  | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 160.4  |
| Q2579-06 | WC-A2-14-C      | Solid  | TVS  | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 160.4  |

Date/Time 07-11-25 14:00  
 Raw Sample Received by: SP WCI  
 Raw Sample Relinquished by: CP SM

Date/Time 07-11-25  
 Raw Sample Received by: CP SM  
 Raw Sample Relinquished by: SP WCI

## Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136451

Slope : 98.5

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

| Calibration Standards           | Chemtech Log# |
|---------------------------------|---------------|
| PH 4 BUFFER SOLUTION            | W3178         |
| BUFFER PH 7.00 GREEN 1PINT PK6  | W3093         |
| PH 10.01 BUFFER, COLOR CD 475ML | W3191         |
| buffer solution pH 7 yellow     | W3217         |
| Buffer Solution, PH2 (500ml)    | W3161         |
| pH 12.00 Buffer                 | W3200         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.2             | 4.01        | 07/11/2025 | 14:20     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.2             | 7.01        | 07/11/2025 | 14:21     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.2             | 10.02       | 07/11/2025 | 14:22     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.2             | 7.01        | 07/11/2025 | 14:25     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.2             | 2.01        | 07/11/2025 | 14:33     |
| 6   | Q2578-02    | 1  | Solid  | 20.02       | 20          | 20.8             | 11.66       | 07/11/2025 | 14:55     |
| 7   | Q2578-02DUP | 1  | Solid  | 20.03       | 20          | 20.9             | 11.68       | 07/11/2025 | 14:56     |
| 8   | Q2578-06    | 1  | Solid  | 20.02       | 20          | 21.3             | 11.38       | 07/11/2025 | 15:00     |
| 9   | Q2578-10    | 1  | Solid  | 20.03       | 20          | 21.2             | 11.51       | 07/11/2025 | 15:05     |
| 10  | Q2578-14    | 1  | Solid  | 20.04       | 20          | 21.2             | 11.69       | 07/11/2025 | 15:15     |
| 11  | Q2578-18    | 1  | Solid  | 20.02       | 20          | 21.3             | 11.81       | 07/11/2025 | 15:25     |
| 12  | Q2579-02    | 1  | Solid  | 20.03       | 20          | 21.1             | 11.39       | 07/11/2025 | 15:35     |
| 13  | Q2579-06    | 1  | Solid  | 20.04       | 20          | 21.2             | 11.91       | 07/11/2025 | 15:45     |
| 14  | Q2589-01    | 1  | Solid  | 20.02       | 20          | 21.9             | 8.22        | 07/11/2025 | 16:00     |
| 15  | CCV2        | 1  | Water  | NA          | NA          | 20.3             | 12.02       | 07/11/2025 | 16:02     |

# WORKLIST(Hardcopy Internal Chain)

87736451

WorkList Name : PH S Q2589      WorkList ID : 190679      Department : Wet-Chemistry      Date : 07-11-2025 13:44:18

| Sample   | Customer Sample | Matrix | Test | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method |
|----------|-----------------|--------|------|--------------|----------|-----------------------------|--------------|--------|
| Q2578-02 | WC-A5-01-C      | Solid  | pH   | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9045D  |
| Q2578-06 | WC-A2-09-C      | Solid  | pH   | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9045D  |
| Q2578-10 | WC-A2-10-C      | Solid  | pH   | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9045D  |
| Q2578-14 | WC-A2-11-C      | Solid  | pH   | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9045D  |
| Q2578-18 | WC-A2-12-C      | Solid  | pH   | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9045D  |
| Q2579-02 | WC-A2-13-C      | Solid  | pH   | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9045D  |
| Q2579-06 | WC-A2-14-C      | Solid  | pH   | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9045D  |
| Q2589-01 | AU-06-071125    | Solid  | pH   | Cool 4 deg C | PSEG05   | D31                         | 07/11/2025   | 9045D  |

Date/Time 07-11-25 13:50

Raw Sample Received by: RM Caw

Raw Sample Relinquished by: RM Caw

Date/Time 07-11-25

Raw Sample Received by: RM Caw

Raw Sample Relinquished by: RM Caw

18:00

RM Caw

RM Caw

## Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB136452

Slope : 98.5

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

| Calibration Standards           | Chemtech Log# |
|---------------------------------|---------------|
| PH 4 BUFFER SOLUTION            | W3178         |
| BUFFER PH 7.00 GREEN 1PINT PK6  | W3093         |
| PH 10.01 BUFFER, COLOR CD 475ML | W3191         |
| buffer solution pH 7 yellow     | W3217         |
| Buffer Solution, PH2 (500ml)    | W3161         |
| pH 12.00 Buffer                 | W3200         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.2             | 4.01        | 07/11/2025 | 14:20     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.2             | 7.01        | 07/11/2025 | 14:21     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.2             | 10.02       | 07/11/2025 | 14:22     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.2             | 7.01        | 07/11/2025 | 14:25     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.2             | 2.01        | 07/11/2025 | 14:33     |
| 6   | Q2578-03    | 1  | Solid  | 20.02       | 20          | 20.8             | 11.66       | 07/11/2025 | 14:55     |
| 7   | Q2578-03DUP | 1  | Solid  | 20.03       | 20          | 20.9             | 11.68       | 07/11/2025 | 14:56     |
| 8   | Q2578-07    | 1  | Solid  | 20.02       | 20          | 21.3             | 11.38       | 07/11/2025 | 15:00     |
| 9   | Q2578-11    | 1  | Solid  | 20.03       | 20          | 21.2             | 11.51       | 07/11/2025 | 15:05     |
| 10  | Q2578-15    | 1  | Solid  | 20.04       | 20          | 21.2             | 11.69       | 07/11/2025 | 15:15     |
| 11  | Q2578-19    | 1  | Solid  | 20.02       | 20          | 21.3             | 11.81       | 07/11/2025 | 15:25     |
| 12  | Q2579-03    | 1  | Solid  | 20.03       | 20          | 21.1             | 11.39       | 07/11/2025 | 15:35     |
| 13  | Q2579-07    | 1  | Solid  | 20.04       | 20          | 21.2             | 11.91       | 07/11/2025 | 15:45     |
| 14  | Q2586-04    | 1  | Solid  | 20.02       | 20          | 21.9             | 9.02        | 07/11/2025 | 15:56     |
| 15  | CCV2        | 1  | Water  | NA          | NA          | 20.3             | 12.02       | 07/11/2025 | 16:02     |

# WORKLIST(Hardcopy Internal Chain)

136452

WorkList Name : corrosivity q2589

WorkList ID : 190678

Department : Wet-Chemistry

Date : 07-11-2025 13:43:46

| Sample   | Customer Sample | Matrix | Test        | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method |
|----------|-----------------|--------|-------------|--------------|----------|-----------------------------|--------------|--------|
| Q2578-03 | WC-A5-01-C      | Solid  | Corrosivity | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9045D  |
| Q2578-07 | WC-A2-09-C      | Solid  | Corrosivity | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9045D  |
| Q2578-11 | WC-A2-10-C      | Solid  | Corrosivity | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9045D  |
| Q2578-15 | WC-A2-11-C      | Solid  | Corrosivity | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9045D  |
| Q2578-19 | WC-A2-12-C      | Solid  | Corrosivity | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9045D  |
| Q2579-03 | WC-A2-13-C      | Solid  | Corrosivity | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9045D  |
| Q2579-07 | WC-A2-14-C      | Solid  | Corrosivity | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9045D  |
| Q2586-04 | TP-16           | Solid  | Corrosivity | Cool 4 deg C | PSEG03   | D41                         | 07/11/2025   | 9045D  |

Date/Time 07/11/25 13:50  
 Raw Sample Received by: SPWCL  
 Raw Sample Relinquished by: RHCWCL

Date/Time 07/11/25 18:00  
 Raw Sample Received by: RHCWCL  
 Raw Sample Relinquished by: SPWCL



**TOTAL SOLIDS - SM2540B**

**SUPERVISOR:** Iwona

**ANALYST:** jignesh

**Date:** 07/14/2025

**Run Number:** LB136455

**BalanceID:** WC SC-6

**OvenID:** WC OVEN-1

**ThermometerID:** WET OVEN#1

**TEMP1 IN:** 103 °C 07/14/2025 12:00 **TEMP1 OUT:** 103 °C 07/14/2025 13:00  
**TEMP2 IN:** 104 °C 07/14/2025 13:30 **TEMP2 OUT:** 104 °C 07/14/2025 14:30  
**TEMP3 IN:** 104 °C 07/14/2025 15:40 **TEMP3 OUT:** 103 °C 07/15/2025 07:30  
**TEMP4 IN:** 104 °C 07/15/2025 08:00 **TEMP4 OUT:** 104 °C 07/15/2025 09:35

| Dish # | Lab ID      | Client ID     | Empty Dish Weight (g) | Final Empty Dish Weight (g) | Sample Vol (ml) | Original weight 1st Dish+Sample weight after Drying @103-@105°C (g) | Constant weight 2nd Dish+Sample weight after Drying @103-@105°C (g) | Final Constant weight Final Dish+Sample weight after Drying @103-@105°C (g) | Weight (g) | Result (mg/L) |
|--------|-------------|---------------|-----------------------|-----------------------------|-----------------|---------------------------------------------------------------------|---------------------------------------------------------------------|-----------------------------------------------------------------------------|------------|---------------|
| 1      | LB136455BL  | LB136455BL    | 85.9631               | 85.9631                     | 100             | 85.9631                                                             | 85.9631                                                             | 85.9631                                                                     | 0.0000     | 0             |
| 2      | Q2578-04    | WC-A5-01-C    | 156.5811              | 156.5811                    | 100             | 156.6469                                                            | 156.6469                                                            | 156.6470                                                                    | 0.0658     | 658           |
| 3      | Q2578-04DUP | WC-A5-01-CDUP | 142.7876              | 142.7876                    | 100             | 142.8523                                                            | 142.8523                                                            | 142.8520                                                                    | 0.0647     | 647           |
| 4      | Q2578-08    | WC-A2-09-C    | 152.1814              | 152.1814                    | 100             | 152.2348                                                            | 152.2348                                                            | 152.2350                                                                    | 0.0534     | 534           |
| 5      | Q2578-12    | WC-A2-10-C    | 106.1470              | 106.1470                    | 100             | 106.2342                                                            | 106.2342                                                            | 106.2340                                                                    | 0.0872     | 872           |
| 6      | Q2578-16    | WC-A2-11-C    | 110.2228              | 110.2228                    | 100             | 110.2946                                                            | 110.2946                                                            | 110.2950                                                                    | 0.0718     | 718           |
| 7      | Q2578-20    | WC-A2-12-C    | 70.8490               | 70.8490                     | 100             | 70.9515                                                             | 70.9515                                                             | 70.9515                                                                     | 0.1025     | 1025          |
| 8      | Q2579-04    | WC-A2-13-C    | 76.7959               | 76.7959                     | 100             | 76.8487                                                             | 76.8487                                                             | 76.8487                                                                     | 0.0528     | 528           |
| 9      | Q2579-08    | WC-A2-14-C    | 69.0157               | 69.0157                     | 100             | 69.1035                                                             | 69.1035                                                             | 69.1035                                                                     | 0.0878     | 878           |

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

# WORKLIST(Hardcopy Internal Chain)

W7196455

WorkList Name : astm ts q2579

WorkList ID : 190707

Department : Wet-Chemistry

Date : 07-14-2025 10:50:23

| Sample   | Customer Sample | Matrix | Test    | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method   |
|----------|-----------------|--------|---------|--------------|----------|-----------------------------|--------------|----------|
| Q2578-04 | WC-A5-01-C      | Solid  | ASTM TS | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | SM2540 B |
| Q2578-08 | WC-A2-09-C      | Solid  | ASTM TS | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | SM2540 B |
| Q2578-12 | WC-A2-10-C      | Solid  | ASTM TS | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | SM2540 B |
| Q2578-16 | WC-A2-11-C      | Solid  | ASTM TS | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | SM2540 B |
| Q2578-20 | WC-A2-12-C      | Solid  | ASTM TS | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | SM2540 B |
| Q2579-04 | WC-A2-13-C      | Solid  | ASTM TS | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | SM2540 B |
| Q2579-08 | WC-A2-14-C      | Solid  | ASTM TS | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | SM2540 B |

Date/Time 07/14/25 12:30

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 07/14/25 16:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

## Extraction and Analytical Summary Report

**Analysis Method:** 1664A  
**Test:** ASTM Oil and Grease  
**Run Number:** LB136456  
**Analysis Date:** 07/14/2025  
**BalanceID:** WC SC-6  
**OvenID:** EXT OVEN-3

**ANALYST:** jignesh  
**REVIEWED BY:** Iwona  
**Extraction Date:** 07/14/2025  
**Extraction IN Time:** 11:15  
**Extraction OUT Time:** 12:00  
**Thermometer ID:** EXT OVEN#3

| Dish # | Lab ID      | Client ID     | Matrix | pH  | Sample Weight (g) | Final Volume (mL) | Empty Dish Weight (g) | Final Empty Dish Weight (g) | Silica Gel Weight (g) | Weight After Drying (g) | Final Weight After Drying (g) | Change Weight (g) | Result in ppm |
|--------|-------------|---------------|--------|-----|-------------------|-------------------|-----------------------|-----------------------------|-----------------------|-------------------------|-------------------------------|-------------------|---------------|
| 1      | LB136456BL  | LB136456BL    | WATER  | 1.3 | 1000              | 100               | 2.5983                | 2.5983                      | 0                     | 2.5984                  | 2.5984                        | 0.0001            | 0.1           |
| 2      | LB136456BS  | LB136456BS    | WATER  | 1.3 | 1000              | 100               | 3.0214                | 3.0214                      | 0                     | 3.0381                  | 3.0381                        | 0.0167            | 16.7          |
| 3      | Q2578-04    | WC-A5-01-C    | WATER  | 1.6 | 1000              | 100               | 2.7415                | 2.7415                      | 0                     | 2.7418                  | 2.7418                        | 0.0003            | 0.3           |
| 4      | Q2578-04DUP | WC-A5-01-CDUP | WATER  | 1.6 | 1000              | 100               | 3.1526                | 3.1526                      | 0                     | 3.1529                  | 3.1529                        | 0.0003            | 0.3           |
| 5      | Q2578-04MS  | WC-A5-01-C    | WATER  | 1.6 | 1000              | 100               | 3.0188                | 3.0188                      | 0                     | 3.0392                  | 3.0392                        | 0.0204            | 20.4          |
| 6      | Q2578-04MSD | WC-A5-01-C    | WATER  | 1.6 | 1000              | 100               | 2.1411                | 2.1411                      | 0                     | 2.1616                  | 2.1616                        | 0.0205            | 20.5          |
| 7      | Q2578-08    | WC-A2-09-C    | WATER  | 1.6 | 1000              | 100               | 3.0299                | 3.0299                      | 0                     | 3.0302                  | 3.0302                        | 0.0003            | 0.3           |
| 8      | Q2578-12    | WC-A2-10-C    | WATER  | 1.6 | 1000              | 100               | 2.4074                | 2.4074                      | 0                     | 2.4079                  | 2.4079                        | 0.0005            | 0.5           |
| 9      | Q2578-16    | WC-A2-11-C    | WATER  | 1.6 | 1000              | 100               | 3.1966                | 3.1966                      | 0                     | 3.1970                  | 3.1970                        | 0.0004            | 0.4           |
| 10     | Q2578-20    | WC-A2-12-C    | WATER  | 1.6 | 1000              | 100               | 2.7044                | 2.7044                      | 0                     | 2.7046                  | 2.7046                        | 0.0002            | 0.2           |
| 11     | Q2579-04    | WC-A2-13-C    | WATER  | 1.6 | 1000              | 100               | 3.1691                | 3.1691                      | 0                     | 3.1696                  | 3.1696                        | 0.0005            | 0.5           |
| 12     | Q2579-08    | WC-A2-14-C    | WATER  | 1.6 | 1000              | 100               | 2.4855                | 2.4855                      | 0                     | 2.4858                  | 2.4858                        | 0.0003            | 0.3           |



# WORKLIST(Hardcopy Internal Chain)

VB 136456

WorkList Name : astm oil & grease q2589 WorkList ID : 190708 Department : Wet-Chemistry Date : 07-14-2025 10:51:24

| Sample   | Customer Sample | Matrix | Test                | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method |
|----------|-----------------|--------|---------------------|--------------|----------|-----------------------------|--------------|--------|
| Q2578-04 | WC-A5-01-C      | Solid  | ASTM Oil and Grease | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 1664A  |
| Q2578-08 | WC-A2-09-C      | Solid  | ASTM Oil and Grease | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 1664A  |
| Q2578-12 | WC-A2-10-C      | Solid  | ASTM Oil and Grease | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 1664A  |
| Q2578-16 | WC-A2-11-C      | Solid  | ASTM Oil and Grease | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 1664A  |
| Q2578-20 | WC-A2-12-C      | Solid  | ASTM Oil and Grease | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 1664A  |
| Q2579-04 | WC-A2-13-C      | Solid  | ASTM Oil and Grease | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 1664A  |
| Q2579-08 | WC-A2-14-C      | Solid  | ASTM Oil and Grease | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 1664A  |

Date/Time 07/14/25 11:10

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 07/14/25

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

17130

[Signature]

[Signature]

## Analytical Summary Report

Analysis Method: 1030  
Parameter: Ignitability  
Run Number: LB136462

Reviewed By: Eman

Supervisor Review By: Iwona

| Seq | LabID       | ClientID      | DF | matrix | Result Status | Burning Rate | Anal Date  | Anal Time |
|-----|-------------|---------------|----|--------|---------------|--------------|------------|-----------|
| 1   | Q2578-03    | WC-A5-01-C    | 1  | Solid  | NO            | 0.00         | 07/14/2025 | 08:30     |
| 2   | Q2578-03DUP | WC-A5-01-CDUP | 1  | Solid  | NO            | 0.00         | 07/14/2025 | 08:37     |
| 3   | Q2578-07    | WC-A2-09-C    | 1  | Solid  | NO            | 0.00         | 07/14/2025 | 08:45     |
| 4   | Q2578-11    | WC-A2-10-C    | 1  | Solid  | NO            | 0.00         | 07/14/2025 | 08:52     |
| 5   | Q2578-15    | WC-A2-11-C    | 1  | Solid  | NO            | 0.00         | 07/14/2025 | 09:00     |
| 6   | Q2578-19    | WC-A2-12-C    | 1  | Solid  | NO            | 0.00         | 07/14/2025 | 09:08     |
| 7   | Q2579-03    | WC-A2-13-C    | 1  | Solid  | NO            | 0.00         | 07/14/2025 | 09:15     |
| 8   | Q2579-07    | WC-A2-14-C    | 1  | Solid  | NO            | 0.00         | 07/14/2025 | 09:22     |
| 9   | Q2586-01    | TP-16         | 1  | Solid  | NO            | 0.00         | 07/14/2025 | 09:30     |
| 10  | Q2586-04    | TP-16         | 1  | Solid  | NO            | 0.00         | 07/14/2025 | 09:37     |

$$\text{Burning Rate} = \frac{\text{Length (mm)}}{\text{Total Time (sec)}}$$

# WORKLIST(Hardcopy Internal Chain)

1b136462

WorkList Name : IGN-071425      WorkList ID : 190686      Department : Wet-Chemistry      Date : 07-14-2025 08:07:39

| Sample   | Customer Sample | Matrix | Test         | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method |
|----------|-----------------|--------|--------------|--------------|----------|-----------------------------|--------------|--------|
| Q2578-03 | WC-A5-01-C      | Solid  | Ignitability | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 1030   |
| Q2578-07 | WC-A2-09-C      | Solid  | Ignitability | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 1030   |
| Q2578-11 | WC-A2-10-C      | Solid  | Ignitability | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 1030   |
| Q2578-15 | WC-A2-11-C      | Solid  | Ignitability | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 1030   |
| Q2578-19 | WC-A2-12-C      | Solid  | Ignitability | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 1030   |
| Q2579-03 | WC-A2-13-C      | Solid  | Ignitability | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 1030   |
| Q2579-07 | WC-A2-14-C      | Solid  | Ignitability | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 1030   |
| Q2586-01 | TP-16           | Solid  | Ignitability | Cool 4 deg C | PSEG03   | D41                         | 07/11/2025   | 1030   |
| Q2586-04 | TP-16           | Solid  | Ignitability | Cool 4 deg C | PSEG03   | D41                         | 07/11/2025   | 1030   |

Date/Time 07/14/25 08:20  
Raw Sample Received by: EM(WC)  
Raw Sample Relinquished by: EM(WC)

Date/Time 07/14/25 11:15  
Raw Sample Received by: EM(WC)  
Raw Sample Relinquished by: EM(WC)

## Analytical Summary Report

**Analysis Method:** 9095B

**Parameter:** Paint Filter

**Run Number:** LB136463

**Reviewed By:** Eman

**Supervisor Review By:** Iwona

**BalanceID:** WC SC-7

| Seq | LabID       | ClientID      | Dilution | Weight (g) | Inst. Conc (ml/100g) | Anal Date  | Anal Time |
|-----|-------------|---------------|----------|------------|----------------------|------------|-----------|
| 1   | Q2578-02    | WC-A5-01-C    | 1        | 100.02     | 0.00                 | 07/14/2025 | 09:50     |
| 2   | Q2578-02DUP | WC-A5-01-CDUP | 1        | 100.05     | 0.00                 | 07/14/2025 | 09:57     |
| 3   | Q2578-06    | WC-A2-09-C    | 1        | 100.03     | 0.00                 | 07/14/2025 | 10:05     |
| 4   | Q2578-10    | WC-A2-10-C    | 1        | 100.01     | 0.00                 | 07/14/2025 | 10:12     |
| 5   | Q2578-14    | WC-A2-11-C    | 1        | 100.06     | 0.00                 | 07/14/2025 | 10:20     |
| 6   | Q2578-18    | WC-A2-12-C    | 1        | 100.04     | 0.00                 | 07/14/2025 | 10:27     |
| 7   | Q2579-02    | WC-A2-13-C    | 1        | 100.02     | 0.00                 | 07/14/2025 | 10:35     |
| 8   | Q2579-06    | WC-A2-14-C    | 1        | 100.01     | 0.00                 | 07/14/2025 | 10:42     |
| 9   | Q2586-01    | TP-16         | 1        | 100.06     | 0.00                 | 07/14/2025 | 10:50     |
| 10  | Q2590-01    | NB-7-071125   | 1        | 100.03     | 0.00                 | 07/14/2025 | 10:58     |



# WORKLIST(Hardcopy Internal Chain)

16136462

WorkList Name : PF-071425

WorkList ID : 190687

Department : Wet-Chemistry

Date : 07-14-2025 08:07:45

| Sample   | Customer Sample | Matrix | Test         | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method |
|----------|-----------------|--------|--------------|--------------|----------|-----------------------------|--------------|--------|
| Q2578-02 | WC-A5-01-C      | Solid  | Paint Filter | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9095B  |
| Q2578-06 | WC-A2-09-C      | Solid  | Paint Filter | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9095B  |
| Q2578-10 | WC-A2-10-C      | Solid  | Paint Filter | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9095B  |
| Q2578-14 | WC-A2-11-C      | Solid  | Paint Filter | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9095B  |
| Q2578-18 | WC-A2-12-C      | Solid  | Paint Filter | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9095B  |
| Q2579-02 | WC-A2-13-C      | Solid  | Paint Filter | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9095B  |
| Q2579-06 | WC-A2-14-C      | Solid  | Paint Filter | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9095B  |
| Q2586-01 | TP-16           | Solid  | Paint Filter | Cool 4 deg C | PSEG03   | D41                         | 07/11/2025   | 9095B  |
| Q2590-01 | NB-07-071125    | Solid  | Paint Filter | Cool 4 deg C | PSEG05   | D51                         | 07/11/2025   | 9095B  |

Date/Time 07/14/25 08:20  
 Raw Sample Received by: EM(wce)  
 Raw Sample Relinquished by: 28(wce)

Date/Time 07/14/25 11:15  
 Raw Sample Received by: 28(wce)  
 Raw Sample Relinquished by: EM(wce)

0136468

Test results

Aquakem 7.2AQ1

Page: 1

Alliance Technical Group  
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

7/14/2025 16:39

Test: Ammonia-N

| Sample Id   | Result | Dil. 1 + | Response | Errors |
|-------------|--------|----------|----------|--------|
| ICV1        | 0.986  | 0.0      | 0.209    |        |
| ICB1        | 0.002  | 0.0      | 0.018    |        |
| CCV1        | 0.980  | 0.0      | 0.208    |        |
| CCB1        | 0.001  | 0.0      | 0.017    |        |
| PB168836BL  | 0.003  | 0.0      | 0.018    |        |
| PB168836BS  | 1.002  | 0.0      | 0.212    |        |
| PB168836TB  | 0.003  | 0.0      | 0.018    |        |
| Q2578-04    | 0.233  | 0.0      | 0.062    |        |
| Q2578-04DUP | 0.229  | 0.0      | 0.062    |        |
| Q2578-04MS  | 1.146  | 0.0      | 0.240    |        |
| Q2578-04MSD | 1.290  | 0.0      | 0.268    |        |
| Q2578-08    | 0.377  | 0.0      | 0.090    |        |
| Q2578-12    | 0.929  | 0.0      | 0.198    |        |
| Q2578-16    | 0.149  | 0.0      | 0.046    |        |
| CCV2        | 1.034  | 0.0      | 0.218    |        |
| CCB2        | 0.008  | 0.0      | 0.019    |        |
| Q2578-20    | 0.831  | 0.0      | 0.179    |        |
| Q2579-04    | 0.400  | 0.0      | 0.095    |        |
| Q2579-08    | 0.260  | 0.0      | 0.068    |        |
| CCV3        | 0.986  | 0.0      | 0.209    |        |
| CCB3        | 0.010  | 0.0      | 0.019    |        |

N 21  
Mean 0.517  
SD 0.4687  
CV% 90.64

07/14/2025  
RM

Aquakem v. 7.2AQ1

Results from time period:

Mon Jul 14 16:06:34 2025

Mon Jul 14 16:32:54 2025

| Sample Id   | Sam/Ctr/c/ | Test short r | Test type | Result | Result unit | Result date and time | Stat |
|-------------|------------|--------------|-----------|--------|-------------|----------------------|------|
| 0.0PPM      | A          | Ammonia-† P  |           | 0.0074 | mg/l        | 7/14/2025 10:59:54   |      |
| 0.1PPM      | A          | Ammonia-† P  |           | 0.1037 | mg/l        | 7/14/2025 10:59:55   |      |
| 0.2PPM      | A          | Ammonia-† P  |           | 0.1989 | mg/l        | 7/14/2025 10:59:56   |      |
| 0.4PPM      | A          | Ammonia-† P  |           | 0.3879 | mg/l        | 7/14/2025 10:59:57   |      |
| 1.0PPM      | A          | Ammonia-† P  |           | 0.9878 | mg/l        | 7/14/2025 10:59:58   |      |
| 1.3PPM      | A          | Ammonia-† P  |           | 1.3509 | mg/l        | 7/14/2025 10:59:59   |      |
| 2.0PPM      | A          | Ammonia-† P  |           | 1.9967 | mg/l        | 7/14/2025 11:00:00   |      |
| ICV1        | S          | Ammonia-† P  |           | 0.9865 | mg/l        | 7/14/2025 16:06:34   |      |
| ICB1        | S          | Ammonia-† P  |           | 0.0021 | mg/l        | 7/14/2025 16:06:37   |      |
| CCV1        | S          | Ammonia-† P  |           | 0.9795 | mg/l        | 7/14/2025 16:06:38   |      |
| CCB1        | S          | Ammonia-† P  |           | 0.0007 | mg/l        | 7/14/2025 16:06:41   |      |
| PB168836BL  | S          | Ammonia-† P  |           | 0.0027 | mg/l        | 7/14/2025 16:06:42   |      |
| PB168836BS  | S          | Ammonia-† P  |           | 1.0025 | mg/l        | 7/14/2025 16:06:45   |      |
| PB168836TB  | S          | Ammonia-† P  |           | 0.0032 | mg/l        | 7/14/2025 16:17:19   |      |
| Q2578-04    | S          | Ammonia-† P  |           | 0.2331 | mg/l        | 7/14/2025 16:17:21   |      |
| Q2578-04DUP | S          | Ammonia-† P  |           | 0.2295 | mg/l        | 7/14/2025 16:17:23   |      |
| Q2578-04MS  | S          | Ammonia-† P  |           | 1.1462 | mg/l        | 7/14/2025 16:17:24   |      |
| Q2578-04MSD | S          | Ammonia-† P  |           | 1.2896 | mg/l        | 7/14/2025 16:17:25   |      |
| Q2578-08    | S          | Ammonia-† P  |           | 0.3767 | mg/l        | 7/14/2025 16:17:26   |      |
| Q2578-12    | S          | Ammonia-† P  |           | 0.9288 | mg/l        | 7/14/2025 16:17:27   |      |
| Q2578-16    | S          | Ammonia-† P  |           | 0.1492 | mg/l        | 7/14/2025 16:17:28   |      |
| CCV2        | S          | Ammonia-† P  |           | 1.034  | mg/l        | 7/14/2025 16:28:03   |      |
| CCB2        | S          | Ammonia-† P  |           | 0.0076 | mg/l        | 7/14/2025 16:28:05   |      |
| Q2578-20    | S          | Ammonia-† P  |           | 0.8307 | mg/l        | 7/14/2025 16:28:06   |      |
| Q2579-04    | S          | Ammonia-† P  |           | 0.4004 | mg/l        | 7/14/2025 16:28:07   |      |
| Q2579-08    | S          | Ammonia-† P  |           | 0.2598 | mg/l        | 7/14/2025 16:28:08   |      |
| CCV3        | S          | Ammonia-† P  |           | 0.9858 | mg/l        | 7/14/2025 16:28:10   |      |
| CCB3        | S          | Ammonia-† P  |           | 0.0098 | mg/l        | 7/14/2025 16:32:53   |      |

Calibration results

Aquakem 7.2AQ1

Page:

Alliance Technical Group

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

7/14/2025 11:07

Test Ammonia-N


Accepted 7/14/2025 11:07

Factor 5.141

Bias 0.017

Coeff. of det. 0.999798

Errors



|   | Calibrator | Response | Calc. con. | Conc.  | Errors |
|---|------------|----------|------------|--------|--------|
| 1 | 0.00PPM    | 0.019    | 0.0074     | 0.0000 | -      |
| 2 | NH3-2PPM   | 0.037    | 0.1037     | 0.1000 | 3.7    |
| 3 | NH3-2PPM   | 0.056    | 0.1989     | 0.2000 | -0.6   |
| 4 | NH3-2PPM   | 0.093    | 0.3879     | 0.4000 | -3.0   |
| 5 | NH3-2PPM   | 0.209    | 0.9878     | 1.0000 | -1.2   |
| 6 | NH3-2PPM   | 0.280    | 1.3509     | 1.3333 | 3.9    |
| 7 | NH3-2PPM   | 0.406    | 1.9967     | 2.0000 | -0.2   |

07/14/2025  
 RM

## Extraction and Analytical Summary Report

**Analysis Method:** 9071B  
**Test:** Oil and Grease  
**Run Number:** LB136474  
**Analysis Date:** 07/15/2025  
**BalanceID:** WC SC-6  
**OvenID:** EXT OVEN-3

**ANALYST:** jignesh  
**REVIEWED BY:** Iwona  
**Extraction Date:** 07/15/2025  
**Extraction IN Time:** 09:00  
**Extraction OUT Time:** 09:50  
**Thermometer ID:** EXT OVEN#3

| Dish # | Lab ID      | Client ID     | Matrix | pH | Sample Weight (g) | Final Volume (mL) | Empty Dish Weight (g) | Final Empty Dish Weight (g) | Silica Gel Weight (g) | Weight After Drying (g) | Final Weight After Drying (g) | Change Weight (g) | Result in ppm |
|--------|-------------|---------------|--------|----|-------------------|-------------------|-----------------------|-----------------------------|-----------------------|-------------------------|-------------------------------|-------------------|---------------|
| 1      | LB136474BL  | LB136474BL    | SOLID  |    | 20.02             | 100               | 2.9601                | 2.9601                      | 0                     | 2.9602                  | 2.9602                        | 0.0001            | 5             |
| 2      | LB136474BS  | LB136474BS    | SOLID  |    | 20.03             | 100               | 2.7414                | 2.7414                      | 0                     | 2.7433                  | 2.7433                        | 0.0019            | 94.86         |
| 3      | Q2578-02    | WC-A5-01-C    | SOLID  |    | 20.02             | 100               | 2.9981                | 2.9981                      | 0                     | 3.0153                  | 3.0153                        | 0.0172            | 859.14        |
| 4      | Q2578-06    | WC-A2-09-C    | SOLID  |    | 20.03             | 100               | 3.0005                | 3.0005                      | 0                     | 3.0086                  | 3.0086                        | 0.0081            | 404.39        |
| 5      | Q2578-10    | WC-A2-10-C    | SOLID  |    | 20.04             | 100               | 3.0036                | 3.0036                      | 0                     | 3.0059                  | 3.0059                        | 0.0023            | 114.77        |
| 6      | Q2578-14    | WC-A2-11-C    | SOLID  |    | 20.02             | 100               | 3.0571                | 3.0571                      | 0                     | 3.0618                  | 3.0618                        | 0.0047            | 234.77        |
| 7      | Q2578-18    | WC-A2-12-C    | SOLID  |    | 20.03             | 100               | 3.0929                | 3.0929                      | 0                     | 3.0933                  | 3.0933                        | 0.0004            | 19.97         |
| 8      | Q2579-02    | WC-A2-13-C    | SOLID  |    | 20.04             | 100               | 3.0662                | 3.0662                      | 0                     | 3.0716                  | 3.0716                        | 0.0054            | 269.46        |
| 9      | Q2579-06    | WC-A2-14-C    | SOLID  |    | 20.03             | 100               | 3.0835                | 3.0835                      | 0                     | 3.0856                  | 3.0856                        | 0.0021            | 104.84        |
| 10     | Q2579-06DUP | WC-A2-14-CDUP | SOLID  |    | 20.02             | 100               | 3.0252                | 3.0252                      | 0                     | 3.0273                  | 3.0273                        | 0.0021            | 104.9         |
| 11     | Q2579-06MS  | WC-A2-14-C    | SOLID  |    | 20.03             | 100               | 3.0268                | 3.0268                      | 0                     | 3.0301                  | 3.0301                        | 0.0033            | 164.75        |
| 12     | Q2579-06MSD | WC-A2-14-C    | SOLID  |    | 20.04             | 100               | 3.0530                | 3.0530                      | 0                     | 3.0560                  | 3.0560                        | 0.0030            | 149.7         |

QC Batch# LB136474

**Test:** Oil and Grease

**Analysis Date:** 07/15/2025

### Chemicals Used:

| Chemical Name  | Chemical Lot # |
|----------------|----------------|
| HEXANE         | W3204          |
| pH Paper 0-14  | NA             |
| Sodium Sulfate | EP2625         |
| 1:1 HCL        | NA             |
| Silica Gel     | NA             |
| Sand           | W3186          |

### Standards Used:

| Standard Name | Amount Used | Standard Lot # |
|---------------|-------------|----------------|
| LCSS          | 1.00 ML     | WP112785       |
| LCSSD         | NA          | NA             |
| MS/MSD        | 1.00 ML     | WP112786       |

### BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-6

## Before Analysis

0.0020 gram Balance: 0.0018 (0.0018-0.0022) In OVEN TEMP1 : 71 °C Dessicator Time In1 : 12:31

1.0000 gram Balance: 1.0004 (0.9950-1.0050) In Time1: 11:45

Bal Check Time: 09:10 Out OVEN TEMP1: 71 °C Dessicator Time Out1: 13:05

Out Time1: 12:30

## After Analysis

0.0020 gram Balance: 0.0019 (0.0018-0.0022) In OVEN TEMP2 : 70 °C Dessicator Time In2 : 14:31

1.0000 gram Balance: 1.0003 (0.9950-1.0050) In Time2: 13:37

Bal Check Time: 15:15 Out OVEN TEMP2: 70 °C Dessicator Time Out2: 15:10

Out Time2: 14:30

# WORKLIST(Hardcopy Internal Chain)

W 136474

WorkList Name : OIL & GREASE Q2589      WorkList ID : 190715      Department : Wet-Chemistry      Date : 07-15-2025 08:17:44

| Sample   | Customer Sample | Matrix | Test           | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method |
|----------|-----------------|--------|----------------|--------------|----------|-----------------------------|--------------|--------|
| Q2578-02 | WC-A5-01-C      | Solid  | Oil and Grease | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9071B  |
| Q2578-06 | WC-A2-09-C      | Solid  | Oil and Grease | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9071B  |
| Q2578-10 | WC-A2-10-C      | Solid  | Oil and Grease | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9071B  |
| Q2578-14 | WC-A2-11-C      | Solid  | Oil and Grease | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9071B  |
| Q2578-18 | WC-A2-12-C      | Solid  | Oil and Grease | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9071B  |
| Q2579-02 | WC-A2-13-C      | Solid  | Oil and Grease | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9071B  |
| Q2579-06 | WC-A2-14-C      | Solid  | Oil and Grease | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9071B  |

07/15/25 08:30

Date/Time  
Raw Sample Received by: AP  
Raw Sample Relinquished by: AP

Date/Time 07/15/25  
Raw Sample Received by: AP  
Raw Sample Relinquished by: AP

15.0

## Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: ASTM COD

SUPERVISOR REVIEW BY: jignesh

Run Number: LB136478

| Reagent/Standard                        | Lot/Log # |
|-----------------------------------------|-----------|
| COD calibration std. 150 ppm            | WP113238  |
| COD calibration std. 100 ppm            | WP113237  |
| COD calibration std. 50 ppm             | WP113235  |
| COD calibration std. 10 ppm             | WP113234  |
| COD calibration std. 0 ppm              | WP113233  |
| COD ICV-LCS std, 50ppm                  | WP113240  |
| COD calibration std. 75 ppm             | WP113236  |
| COD Digestion Vials Low Level 0-150Mg/L | W3129     |
| COD CCV std, 50ppm                      | WP113940  |
| COD ICV-LCS std, 50ppm                  | WP113941  |
| RL CHECK                                | WP113942  |

|                   |                      |                 |
|-------------------|----------------------|-----------------|
| Temp In (C): 148  | Date In: 07/15/2025  | Time In: 09:05  |
| Temp Out (C): 151 | Date Out: 07/15/2025 | Time Out: 11:05 |

Intercept: 0.8179

Slope: 0.9847

Regression: 0.9995

| Seq | Lab ID | TrueValue (mg/l) | DF | MATRIX | Reading | Result (mg/l) | %D    | Anal Date  | Anal Time |
|-----|--------|------------------|----|--------|---------|---------------|-------|------------|-----------|
| 1   | CAL1   | 0                | 1  | Water  | 0.000   | -0.831        |       | 05/28/2025 | 13:10     |
| 2   | CAL2   | 10               | 1  | Water  | 9.000   | 8.309         | -16.9 | 05/28/2025 | 13:10     |
| 3   | CAL3   | 50               | 1  | Water  | 52.000  | 51.977        | 4     | 05/28/2025 | 13:11     |
| 4   | CAL4   | 75               | 1  | Water  | 77.000  | 77.366        | 3.2   | 05/28/2025 | 13:11     |
| 5   | CAL5   | 100              | 1  | Water  | 99.000  | 99.708        | -0.3  | 05/28/2025 | 13:12     |
| 6   | CAL6   | 150              | 1  | Water  | 147.000 | 148.453       | -1    | 05/28/2025 | 13:12     |



## Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: ASTM COD

SUPERVISOR REVIEW BY: jignesh

Run Number: LB136478

| Seq | Lab ID      | True Value (mg/l) | Initial Weight (g) | Final Vol (ml) | DF | MATRIX | Reading | Result | AnalDate   | AnalTime |
|-----|-------------|-------------------|--------------------|----------------|----|--------|---------|--------|------------|----------|
| 1   | ICV         | 50                | NA                 | NA             | 1  | Water  | 51.000  | 50.962 | 05/28/2025 | 13:13    |
| 2   | ICB         |                   | NA                 | NA             | 1  | Water  | 0.000   | -0.831 | 05/28/2025 | 13:13    |
| 3   | CCV1        | 50                | NA                 | NA             | 1  | Water  | 49.000  | 48.931 | 07/15/2025 | 12:25    |
| 4   | CCB1        |                   | NA                 | NA             | 1  | Water  | 1.000   | 0.185  | 07/15/2025 | 12:25    |
| 5   | RL Check    | 10                | NA                 | NA             | 1  | Water  | 9.000   | 8.309  | 07/15/2025 | 12:16    |
| 6   | LB136478BL  |                   | NA                 | NA             | 1  | Water  | 1.000   | 0.185  | 07/15/2025 | 12:26    |
| 7   | LB136478BS  | 50                | NA                 | NA             | 1  | Water  | 51.000  | 50.962 | 07/15/2025 | 12:27    |
| 8   | Q2578-04    |                   | NA                 | NA             | 1  | Water  | 49.000  | 48.931 | 07/15/2025 | 12:27    |
| 9   | Q2578-04DUP |                   | NA                 | NA             | 1  | Water  | 50.000  | 49.946 | 07/15/2025 | 12:28    |
| 10  | Q2578-04MS  | 50                | NA                 | NA             | 1  | Water  | 98.000  | 98.692 | 07/15/2025 | 12:28    |
| 11  | Q2578-04MSD | 50                | NA                 | NA             | 1  | Water  | 97.000  | 97.677 | 07/15/2025 | 12:29    |
| 12  | Q2578-08    |                   | NA                 | NA             | 1  | Water  | 51.000  | 50.962 | 07/15/2025 | 12:29    |
| 13  | Q2578-12    |                   | NA                 | NA             | 1  | Water  | 46.000  | 45.884 | 07/15/2025 | 12:30    |
| 14  | Q2578-16    |                   | NA                 | NA             | 1  | Water  | 15.000  | 14.402 | 07/15/2025 | 12:30    |
| 15  | Q2578-20    |                   | NA                 | NA             | 1  | Water  | 20.000  | 19.480 | 07/15/2025 | 12:31    |
| 16  | CCV2        | 50                | NA                 | NA             | 1  | Water  | 51.000  | 50.962 | 07/15/2025 | 12:31    |
| 17  | CCB2        |                   | NA                 | NA             | 1  | Water  | 1.000   | 0.185  | 07/15/2025 | 12:32    |
| 18  | Q2579-04    |                   | NA                 | NA             | 1  | Water  | 47.000  | 46.900 | 07/15/2025 | 12:32    |
| 19  | Q2579-08    |                   | NA                 | NA             | 1  | Water  | 6.000   | 5.263  | 07/15/2025 | 12:33    |
| 20  | CCV3        | 50                | NA                 | NA             | 1  | Water  | 50.000  | 49.946 | 07/15/2025 | 12:33    |
| 21  | CCB3        |                   | NA                 | NA             | 1  | Water  | 1.000   | 0.185  | 07/15/2025 | 12:34    |

# WORKLIST(Hardcopy Internal Chain)

2136478

WorkList Name : ASTM COD-071425      WorkList ID : 190692      Department : Wet-Chemistry      Date : 07-14-2025 08:55:41

| Sample   | Customer Sample | Matrix | Test     | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method   |
|----------|-----------------|--------|----------|--------------|----------|-----------------------------|--------------|----------|
| Q2578-04 | WC-A5-01-C      | Solid  | ASTM COD | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | SM5220 D |
| Q2578-08 | WC-A2-09-C      | Solid  | ASTM COD | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | SM5220 D |
| Q2578-12 | WC-A2-10-C      | Solid  | ASTM COD | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | SM5220 D |
| Q2578-16 | WC-A2-11-C      | Solid  | ASTM COD | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | SM5220 D |
| Q2578-20 | WC-A2-12-C      | Solid  | ASTM COD | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | SM5220 D |
| Q2579-04 | WC-A2-13-C      | Solid  | ASTM COD | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | SM5220 D |
| Q2579-08 | WC-A2-14-C      | Solid  | ASTM COD | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | SM5220 D |

Date/Time 07/15/25 08:40  
Raw Sample Received by: 12/20/25  
Raw Sample Relinquished by: 08/10/25

Date/Time NA  
Raw Sample Received by: 1  
Raw Sample Relinquished by: 1

SOP ID : M9030B-Sulfide-13

SDG No : N/A

Matrix : SOIL

Pipette ID : WC

Balance ID : WC SC-7

Hood ID : HOOD#1

Block ID : MC-1,MC-2

Weigh By : RM

Start Digest Date: 07/11/2025 Time : 12:45 Temp : N/A

End Digest Date: 07/11/2025 Time : 14:15 Temp : N/A

Digestion tube ID : M5595

Filter paper ID : N/A

pH Meter ID : N/A

Block Thermometer ID : N/A

Prep Technician Signature: RM

Supervisor Signature: 12

| Standard Name | MLS USED | STD REF. # FROM LOG |
|---------------|----------|---------------------|
| PBS003        | 50.0ML   | W3112               |
| N/A           | N/A      | N/A                 |
| N/A           | N/A      | N/A                 |
| N/A           | N/A      | N/A                 |
| N/A           | N/A      | N/A                 |
| N/A           | N/A      | N/A                 |

| Chemical Used     | ML/SAMPLE USED | Lot Number |
|-------------------|----------------|------------|
| 0.5M ZINC ACETATE | 5.0ML          | WP113086   |
| FORMALDEHYDE      | 2.0ML          | W2725      |
| N/A               | N/A            | N/A        |
| N/A               | N/A            | N/A        |
| N/A               | N/A            | N/A        |
| N/A               | N/A            | N/A        |
| N/A               | N/A            | N/A        |
| N/A               | N/A            | N/A        |
| N/A               | N/A            | N/A        |
| N/A               | N/A            | N/A        |
| N/A               | N/A            | N/A        |

Extraction Conformance/Non-Conformance Comments:

N/A

07/11/2025 RM

| Date / Time | Prepped Sample Relinquished By/Location | Received By/Location |
|-------------|-----------------------------------------|----------------------|
|             |                                         |                      |
|             | Preparation Group                       | Analysis Group       |

| Lab Sample ID | Client Sample ID | Initial Weight (g) | Final Vol (ml) | pH  | Sulfide | Oxidizing | Nitrate/Nitrite | Comment | Prep Pos |
|---------------|------------------|--------------------|----------------|-----|---------|-----------|-----------------|---------|----------|
| PB168800BL    | PBS800           | 5.00               | 50             | N/A | N/A     | N/A       | N/A             | N/A     | N/A      |
| Q2519-04DUP   | TP-15DUP         | 5.06               | 50             | N/A | N/A     | N/A       | N/A             | N/A     | N/A      |
| Q2519-04      | TP-15            | 5.06               | 50             | N/A | N/A     | N/A       | N/A             | N/A     | N/A      |
| Q2523-01      | MOO-25-0190      | 5.03               | 50             | N/A | N/A     | N/A       | N/A             | N/A     | N/A      |
| Q2526-05      | LAW-25-0101      | 5.03               | 50             | N/A | N/A     | N/A       | N/A             | N/A     | N/A      |
| Q2526-06      | LAW-25-0102      | 5.07               | 50             | N/A | N/A     | N/A       | N/A             | N/A     | N/A      |
| Q2556-01      | RT3997           | 5.01               | 50             | N/A | N/A     | N/A       | N/A             | N/A     | N/A      |
| Q2565-02      | MOO-25-0194-0195 | 5.07               | 50             | N/A | N/A     | N/A       | N/A             | N/A     | N/A      |
| Q2565-03      | MOO-25-0191      | 5.06               | 50             | N/A | N/A     | N/A       | N/A             | N/A     | N/A      |
| Q2565-04      | MOO-25-0196      | 5.02               | 50             | N/A | N/A     | N/A       | N/A             | N/A     | N/A      |
| Q2565-05      | MOO-25-0180      | 5.06               | 50             | N/A | N/A     | N/A       | N/A             | N/A     | N/A      |
| Q2571-04      | TP-18            | 5.02               | 50             | N/A | N/A     | N/A       | N/A             | N/A     | N/A      |
| Q2571-08      | TP-17            | 5.01               | 50             | N/A | N/A     | N/A       | N/A             | N/A     | N/A      |
| Q2578-03      | WC-A5-01-C       | 5.08               | 50             | N/A | N/A     | N/A       | N/A             | N/A     | N/A      |
| Q2578-07      | WC-A2-09-C       | 5.01               | 50             | N/A | N/A     | N/A       | N/A             | N/A     | N/A      |
| Q2578-11      | WC-A2-10-C       | 5.05               | 50             | N/A | N/A     | N/A       | N/A             | N/A     | N/A      |
| Q2578-15      | WC-A2-11-C       | 5.06               | 50             | N/A | N/A     | N/A       | N/A             | N/A     | N/A      |
| Q2578-19      | WC-A2-12-C       | 5.06               | 50             | N/A | N/A     | N/A       | N/A             | N/A     | N/A      |
| Q2579-03      | WC-A2-13-C       | 5.03               | 50             | N/A | N/A     | N/A       | N/A             | N/A     | N/A      |
| Q2579-07      | WC-A2-14-C       | 5.05               | 50             | N/A | N/A     | N/A       | N/A             | N/A     | N/A      |

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : RSUL-SOIL-7-11

WorkList ID : 190669

Department : Distillation

Date : 07-11-2025 08:02:37

| Sample   | Customer Sample  | Matrix | Test             | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method |
|----------|------------------|--------|------------------|--------------|----------|-----------------------------|--------------|--------|
| Q2519-04 | TP-15            | Solid  | Reactive Sulfide | Cool 4 deg C | PSEG03   | O11                         | 07/08/2025   | 9034   |
| Q2523-01 | MOO-25-0190      | Solid  | Reactive Sulfide | Cool 4 deg C | PSEG03   | O21                         | 07/08/2025   | 9034   |
| Q2526-05 | LAW-25-0101      | Solid  | Reactive Sulfide | Cool 4 deg C | PSEG03   | O13                         | 07/08/2025   | 9034   |
| Q2526-06 | LAW-25-0102      | Solid  | Reactive Sulfide | Cool 4 deg C | PSEG03   | O13                         | 07/08/2025   | 9034   |
| Q2556-01 | RT3997           | Solid  | Reactive Sulfide | Cool 4 deg C | PSEG03   | --Sele                      | 07/10/2025   | 9034   |
| Q2565-02 | MOO-25-0194-0195 | Solid  | Reactive Sulfide | Cool 4 deg C | PSEG03   | O11                         | 07/10/2025   | 9034   |
| Q2565-03 | MOO-25-0191      | Solid  | Reactive Sulfide | Cool 4 deg C | PSEG03   | O11                         | 07/10/2025   | 9034   |
| Q2565-04 | MOO-25-0196      | Solid  | Reactive Sulfide | Cool 4 deg C | PSEG03   | O11                         | 07/10/2025   | 9034   |
| Q2565-05 | MOO-25-0180      | Solid  | Reactive Sulfide | Cool 4 deg C | PSEG03   | O11                         | 07/10/2025   | 9034   |
| Q2571-04 | TP-18            | Solid  | Reactive Sulfide | Cool 4 deg C | PSEG03   | O11                         | 07/10/2025   | 9034   |
| Q2571-08 | TP-17            | Solid  | Reactive Sulfide | Cool 4 deg C | PSEG03   | O22                         | 07/10/2025   | 9034   |
| Q2578-03 | WC-A5-01-C       | Solid  | Reactive Sulfide | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9034   |
| Q2578-07 | WC-A2-09-C       | Solid  | Reactive Sulfide | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9034   |
| Q2578-11 | WC-A2-10-C       | Solid  | Reactive Sulfide | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9034   |
| Q2578-15 | WC-A2-11-C       | Solid  | Reactive Sulfide | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9034   |
| Q2578-19 | WC-A2-12-C       | Solid  | Reactive Sulfide | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9034   |
| Q2579-03 | WC-A2-13-C       | Solid  | Reactive Sulfide | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9034   |
| Q2579-07 | WC-A2-14-C       | Solid  | Reactive Sulfide | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9034   |

Date/Time 07/11/2025 08:10  
Raw Sample Received by: RH (wg)  
Raw Sample Relinquished by: JPC (wg)

Date/Time 07/11/2025 11:40  
Raw Sample Received by: JPC (wg)  
Raw Sample Relinquished by: RH (wg)

SOP ID : M9012B-Total, Amenable and Reactive Cyanide-21

SDG No : N/A

Start Digest Date: 07/11/2025 Time : 08:45 Temp : N/A

Matrix : SOIL

End Digest Date: 07/11/2025 Time : 10:15 Temp : N/A

Pipette ID : N/A

Balance ID : WC SC-7

Hood ID : HOOD#1

Digestion tube ID : M5595

Block Thermometer ID : N/A

Block ID : MC-1,MC-2

Filter paper ID : N/A

Prep Technician Signature: RM

Weigh By : RM

pH Meter ID : N/A

Supervisor Signature: 12

| Standardized Name | MLS USED | STD REF. # FROM LOG |
|-------------------|----------|---------------------|
| PBS003            | 50.0ML   | W3112               |
| N/A               | N/A      | N/A                 |
| N/A               | N/A      | N/A                 |
| N/A               | N/A      | N/A                 |
| N/A               | N/A      | N/A                 |

| Chemical Used | ML/SAMPLE USED | Lot Number |
|---------------|----------------|------------|
| 0.25N NaOH    | 50.0ML         | WP113836   |
| N/A           | N/A            | N/A        |
| N/A           | N/A            | N/A        |
| N/A           | N/A            | N/A        |
| N/A           | N/A            | N/A        |
| N/A           | N/A            | N/A        |
| N/A           | N/A            | N/A        |
| N/A           | N/A            | N/A        |
| N/A           | N/A            | N/A        |
| N/A           | N/A            | N/A        |
| N/A           | N/A            | N/A        |

| LAB SAMPLE ID | CLIENT SAMPLE ID | Comment |
|---------------|------------------|---------|
|               |                  |         |

## Extraction Conformance/Non-Conformance Comments:

N/A

| Date / Time      | Prepped Sample Relinquished By/Location | Received By/Location |
|------------------|-----------------------------------------|----------------------|
| 07/11/2025 10:25 | RM (WC)                                 | RM (WC)              |
|                  | Preparation Group                       | Analysis Group       |

| Lab Sample ID | Client Sample ID | Initial Weight (g) | Final Vol (ml) | pH  | Sulfide | Oxidizing | Nitrate/ Nitrite | Comment | Prep Pos |
|---------------|------------------|--------------------|----------------|-----|---------|-----------|------------------|---------|----------|
| PB168801BL    | PBS801           | 5.00               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q2519-04DUP   | TP-15DUP         | 5.07               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q2519-04      | TP-15            | 5.07               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q2523-01      | MOO-25-0190      | 5.01               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q2526-05      | LAW-25-0101      | 5.04               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q2526-06      | LAW-25-0102      | 5.04               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q2556-01      | RT3997           | 5.01               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q2565-02      | MOO-25-0194-0195 | 5.07               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q2565-03      | MOO-25-0191      | 5.06               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q2565-04      | MOO-25-0196      | 5.03               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q2565-05      | MOO-25-0180      | 5.01               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q2571-04      | TP-18            | 5.04               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q2571-08      | TP-17            | 5.07               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q2578-03      | WC-A5-01-C       | 5.06               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q2578-07      | WC-A2-09-C       | 5.01               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q2578-11      | WC-A2-10-C       | 5.05               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q2578-15      | WC-A2-11-C       | 5.04               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q2578-19      | WC-A2-12-C       | 5.01               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q2579-03      | WC-A2-13-C       | 5.05               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |
| Q2579-07      | WC-A2-14-C       | 5.03               | 50             | N/A | N/A     | N/A       | N/A              | N/A     | N/A      |

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : RCN-SOIL-7-11

WorkList ID : 190670

Department : Distillation

Date : 07-11-2025 08:01:45

| Sample   | Customer Sample  | Matrix | Test             | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method |
|----------|------------------|--------|------------------|--------------|----------|-----------------------------|--------------|--------|
| Q2519-04 | TP-15            | Solid  | Reactive Cyanide | Cool 4 deg C | PSEG03   | O11                         | 07/08/2025   | 9012B  |
| Q2523-01 | MOO-25-0190      | Solid  | Reactive Cyanide | Cool 4 deg C | PSEG03   | O21                         | 07/08/2025   | 9012B  |
| Q2526-05 | LAW-25-0101      | Solid  | Reactive Cyanide | Cool 4 deg C | PSEG03   | O13                         | 07/08/2025   | 9012B  |
| Q2526-06 | LAW-25-0102      | Solid  | Reactive Cyanide | Cool 4 deg C | PSEG03   | O13                         | 07/08/2025   | 9012B  |
| Q2556-01 | RT3997           | Solid  | Reactive Cyanide | Cool 4 deg C | PSEG03   | --Sele                      | 07/10/2025   | 9012B  |
| Q2565-02 | MOO-25-0194-0195 | Solid  | Reactive Cyanide | Cool 4 deg C | PSEG03   | O11                         | 07/10/2025   | 9012B  |
| Q2565-03 | MOO-25-0191      | Solid  | Reactive Cyanide | Cool 4 deg C | PSEG03   | O11                         | 07/10/2025   | 9012B  |
| Q2565-04 | MOO-25-0196      | Solid  | Reactive Cyanide | Cool 4 deg C | PSEG03   | O11                         | 07/10/2025   | 9012B  |
| Q2565-05 | MOO-25-0180      | Solid  | Reactive Cyanide | Cool 4 deg C | PSEG03   | O11                         | 07/10/2025   | 9012B  |
| Q2571-04 | TP-18            | Solid  | Reactive Cyanide | Cool 4 deg C | PSEG03   | O11                         | 07/10/2025   | 9012B  |
| Q2571-08 | TP-17            | Solid  | Reactive Cyanide | Cool 4 deg C | PSEG03   | O22                         | 07/10/2025   | 9012B  |
| Q2578-03 | WC-A5-01-C       | Solid  | Reactive Cyanide | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9012B  |
| Q2578-07 | WC-A2-09-C       | Solid  | Reactive Cyanide | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9012B  |
| Q2578-11 | WC-A2-10-C       | Solid  | Reactive Cyanide | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9012B  |
| Q2578-15 | WC-A2-11-C       | Solid  | Reactive Cyanide | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9012B  |
| Q2578-19 | WC-A2-12-C       | Solid  | Reactive Cyanide | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9012B  |
| Q2579-03 | WC-A2-13-C       | Solid  | Reactive Cyanide | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9012B  |
| Q2579-07 | WC-A2-14-C       | Solid  | Reactive Cyanide | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | 9012B  |

Date/Time 07/11/2025 08:10  
 Raw Sample Received by: RM WLS  
 Raw Sample Relinquished by: [Signature]

Date/Time 07/11/2025 11:00  
 Raw Sample Received by: [Signature]  
 Raw Sample Relinquished by: RM WLS



SOP ID : MSM4500-NH3 B,G-Ammonia-18

SDG No : N/A

Matrix : WATER

Pipette ID : WC

Balance ID : N/A

Hood ID : HOOD#2

Block ID : WC-DIST-BLOCK-1

Weigh By : N/A

Start Digest Date: 07/14/2025 Time : 12:50 Temp : 150 °C

End Digest Date: 07/14/2025 Time : 13:50 Temp : 160 °C

*11 batch*  
07/14/2025 14:10 150°C  
07/14/2025 15:10 160°C

Digestion tube ID : M5595

Block Thermometer ID : WC CYANIDE

Filter paper ID : N/A

Prep Technician Signature: *RM*

pH Meter ID : N/A

Supervisor Signature: *12*

| Standard Name     | MLS USED | STD REF. # FROM LOG |
|-------------------|----------|---------------------|
| LCSW              | 1.0ML    | WP113889            |
| MS/MSD SPIKE SOL. | 1.0ML    | WP113888            |
| PBW               | 50.0ML   | W3112               |
| RL CHECK          | N/A      | AS PER PB168835     |
| N/A               | N/A      | N/A                 |

| Chemical Used    | ML/SAMPLE USED | Lot Number |
|------------------|----------------|------------|
| BORATE BUFFER    | 2.5ML          | WP113886   |
| NAOH 6N          | 0.5-2.0ML      | WP113887   |
| H2SO4 0.04N      | 5.0ML          | WP112828   |
| pH strip-Ammonia | N/A            | W3133      |
| KI-starch paper  | N/A            | W3155      |
| N/A              | N/A            | N/A        |
| N/A              | N/A            | N/A        |
| N/A              | N/A            | N/A        |
| N/A              | N/A            | N/A        |
| N/A              | N/A            | N/A        |
| N/A              | N/A            | N/A        |

Extraction Conformance/Non-Conformance Comments:

N/A

| Date / Time      | Prepped Sample Relinquished By/Location | Received By/Location |
|------------------|-----------------------------------------|----------------------|
| 07/14/2025 15:20 | <i>RM W31</i>                           | <i>RM W31</i>        |
|                  | Preparation Group                       | Analysis Group       |

| Lab Sample ID | Client Sample ID | Initial Vol (ml) | Final Vol (ml) | pH | Sulfide | Oxidizing | Nitrate/ Nitrite | Comment                        | Prep Pos |
|---------------|------------------|------------------|----------------|----|---------|-----------|------------------|--------------------------------|----------|
| PB168836BL    | PBW836           | 50               | 50             | <2 | N/A     | Negative  | N/A              | AFTER ADDING 6N NAOH PH IS 9.5 | N/A      |
| PB168836BS    | LCS836           | 50               | 50             | <2 | N/A     | Negative  | N/A              | AFTER ADDING 6N NAOH PH IS 9.5 | N/A      |
| PB168836TB    | LEB836           | 50               | 50             | <2 | N/A     | Negative  | N/A              | AFTER ADDING 6N NAOH PH IS 9.5 | N/A      |
| Q2578-04DUP   | WC-A5-01-CDUP    | 50               | 50             | <2 | N/A     | Negative  | N/A              | AFTER ADDING 6N NAOH PH IS 9.5 | N/A      |
| Q2578-04MS    | WC-A5-01-CMS     | 50               | 50             | <2 | N/A     | Negative  | N/A              | AFTER ADDING 6N NAOH PH IS 9.5 | N/A      |
| Q2578-04MSD   | WC-A5-01-CMSD    | 50               | 50             | <2 | N/A     | Negative  | N/A              | AFTER ADDING 6N NAOH PH IS 9.5 | N/A      |
| Q2578-04      | WC-A5-01-C       | 50               | 50             | <2 | N/A     | Negative  | N/A              | AFTER ADDING 6N NAOH PH IS 9.5 | N/A      |
| Q2578-08      | WC-A2-09-C       | 50               | 50             | <2 | N/A     | Negative  | N/A              | AFTER ADDING 6N NAOH PH IS 9.5 | N/A      |
| Q2578-12      | WC-A2-10-C       | 50               | 50             | <2 | N/A     | Negative  | N/A              | AFTER ADDING 6N NAOH PH IS 9.5 | N/A      |
| Q2578-16      | WC-A2-11-C       | 50               | 50             | <2 | N/A     | Negative  | N/A              | AFTER ADDING 6N NAOH PH IS 9.5 | N/A      |
| Q2578-20      | WC-A2-12-C       | 50               | 50             | <2 | N/A     | Negative  | N/A              | AFTER ADDING 6N NAOH PH IS 9.5 | N/A      |
| Q2579-04      | WC-A2-13-C       | 50               | 50             | <2 | N/A     | Negative  | N/A              | AFTER ADDING 6N NAOH PH IS 9.5 | N/A      |
| Q2579-08      | WC-A2-14-C       | 50               | 50             | <2 | N/A     | Negative  | N/A              | AFTER ADDING 6N NAOH PH IS 9.5 | N/A      |

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : astm-ammonia-07-14

WorkList ID : 190699

Department : Distillation

Date : 07-14-2025 11:02:08

| Sample   | Customer Sample | Matrix | Test         | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method     |
|----------|-----------------|--------|--------------|--------------|----------|-----------------------------|--------------|------------|
| Q2578-04 | WC-A5-01-C      | Solid  | ASTM Ammonia | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | SM4500-NH3 |
| Q2578-08 | WC-A2-09-C      | Solid  | ASTM Ammonia | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | SM4500-NH3 |
| Q2578-12 | WC-A2-10-C      | Solid  | ASTM Ammonia | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | SM4500-NH3 |
| Q2578-16 | WC-A2-11-C      | Solid  | ASTM Ammonia | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | SM4500-NH3 |
| Q2578-20 | WC-A2-12-C      | Solid  | ASTM Ammonia | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | SM4500-NH3 |
| Q2579-04 | WC-A2-13-C      | Solid  | ASTM Ammonia | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | SM4500-NH3 |
| Q2579-08 | WC-A2-14-C      | Solid  | ASTM Ammonia | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | SM4500-NH3 |

Date/Time 07/14/2025 11:05  
 Raw Sample Received by: RY (wsc)  
 Raw Sample Relinquished by: [Signature]

Date/Time N/A  
 Raw Sample Received by: ''  
 Raw Sample Relinquished by: ''

**Instrument ID:** KONELAB

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136439**

|                  |                                                                |              |                       |
|------------------|----------------------------------------------------------------|--------------|-----------------------|
| Review By        | rubina                                                         | Review On    | 7/14/2025 10:30:26 AM |
| Supervise By     | Iwona                                                          | Supervise On | 7/14/2025 10:33:59 AM |
| SubDirectory     | LB136439                                                       | Test         | Reactive Cyanide      |
| <b>STD. NAME</b> | <b>STD REF.#</b>                                               |              |                       |
| ICAL Standard    | WP113905,WP113906,WP113907,WP113908,WP113909,WP113910,WP113911 |              |                       |
| ICV Standard     | WP113912                                                       |              |                       |
| CCV Standard     | WP113906                                                       |              |                       |
| ICSA Standard    | N/A                                                            |              |                       |
| CRI Standard     | N/A                                                            |              |                       |
| LCS Standard     | N/A                                                            |              |                       |
| Chk Standard     | WP112643,WP112900,WP113913                                     |              |                       |

| Sr# | SampleId    | ClientID    | QcType | Date           | Comment | Operator | Status |
|-----|-------------|-------------|--------|----------------|---------|----------|--------|
| 1   | 0.0PPBCN    | 0.0PPBCN    | CAL1   | 07/11/25 08:54 |         | rubina   | OK     |
| 2   | 5.0PPBCN    | 5.0PPBCN    | CAL2   | 07/11/25 08:54 |         | rubina   | OK     |
| 3   | 10PPBCN     | 10PPBCN     | CAL3   | 07/11/25 08:54 |         | rubina   | OK     |
| 4   | 50PPBCN     | 50PPBCN     | CAL4   | 07/11/25 08:54 |         | rubina   | OK     |
| 5   | 100PPBCN    | 100PPBCN    | CAL5   | 07/11/25 08:54 |         | rubina   | OK     |
| 6   | 250PPBCN    | 250PPBCN    | CAL6   | 07/11/25 08:54 |         | rubina   | OK     |
| 7   | 500PPBCN    | 500PPBCN    | CAL7   | 07/11/25 08:54 |         | rubina   | OK     |
| 8   | ICV1        | ICV1        | ICV    | 07/11/25 12:51 |         | rubina   | OK     |
| 9   | ICB1        | ICB1        | ICB    | 07/11/25 12:51 |         | rubina   | OK     |
| 10  | CCV1        | CCV1        | CCV    | 07/11/25 12:51 |         | rubina   | OK     |
| 11  | CCB1        | CCB1        | CCB    | 07/11/25 12:51 |         | rubina   | OK     |
| 12  | PB168801BL  | PB168801BL  | MB     | 07/11/25 12:51 |         | rubina   | OK     |
| 13  | Q2519-04    | TP-15       | SAM    | 07/11/25 12:51 |         | rubina   | OK     |
| 14  | Q2519-04DUP | TP-15DUP    | DUP    | 07/11/25 12:58 |         | rubina   | OK     |
| 15  | Q2523-01    | MOO-25-0190 | SAM    | 07/11/25 12:58 |         | rubina   | OK     |
| 16  | Q2526-05    | LAW-25-0101 | SAM    | 07/11/25 12:58 |         | rubina   | OK     |
| 17  | Q2526-06    | LAW-25-0102 | SAM    | 07/11/25 12:58 |         | rubina   | OK     |
| 18  | Q2556-01    | RT3997      | SAM    | 07/11/25 12:58 |         | rubina   | OK     |

Instrument ID: KONELAB

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136439**

|                  |                                                                |              |                       |
|------------------|----------------------------------------------------------------|--------------|-----------------------|
| Review By        | rubina                                                         | Review On    | 7/14/2025 10:30:26 AM |
| Supervise By     | Iwona                                                          | Supervise On | 7/14/2025 10:33:59 AM |
| SubDirectory     | LB136439                                                       | Test         | Reactive Cyanide      |
| <b>STD. NAME</b> | <b>STD REF.#</b>                                               |              |                       |
| ICAL Standard    | WP113905,WP113906,WP113907,WP113908,WP113909,WP113910,WP113911 |              |                       |
| ICV Standard     | WP113912                                                       |              |                       |
| CCV Standard     | WP113906                                                       |              |                       |
| ICSA Standard    | N/A                                                            |              |                       |
| CRI Standard     | N/A                                                            |              |                       |
| LCS Standard     | N/A                                                            |              |                       |
| Chk Standard     | WP112643,WP112900,WP113913                                     |              |                       |

|    |             |                  |     |                |  |        |    |
|----|-------------|------------------|-----|----------------|--|--------|----|
| 19 | Q2565-02    | MOO-25-0194-0195 | SAM | 07/11/25 12:58 |  | rubina | OK |
| 20 | Q2565-03    | MOO-25-0191      | SAM | 07/11/25 12:58 |  | rubina | OK |
| 21 | Q2565-04    | MOO-25-0196      | SAM | 07/11/25 12:58 |  | rubina | OK |
| 22 | CCV2        | CCV2             | CCV | 07/11/25 12:59 |  | rubina | OK |
| 23 | CCB2        | CCB2             | CCB | 07/11/25 13:06 |  | rubina | OK |
| 24 | Q2565-05    | MOO-25-0180      | SAM | 07/11/25 13:06 |  | rubina | OK |
| 25 | Q2571-04    | TP-18            | SAM | 07/11/25 13:06 |  | rubina | OK |
| 26 | Q2571-08    | TP-17            | SAM | 07/11/25 13:06 |  | rubina | OK |
| 27 | Q2578-03    | WC-A5-01-C       | SAM | 07/11/25 13:06 |  | rubina | OK |
| 28 | Q2578-07    | WC-A2-09-C       | SAM | 07/11/25 13:06 |  | rubina | OK |
| 29 | Q2578-11    | WC-A2-10-C       | SAM | 07/11/25 13:06 |  | rubina | OK |
| 30 | Q2578-15    | WC-A2-11-C       | SAM | 07/11/25 13:06 |  | rubina | OK |
| 31 | Q2578-19    | WC-A2-12-C       | SAM | 07/11/25 13:06 |  | rubina | OK |
| 32 | Q2579-03    | WC-A2-13-C       | SAM | 07/11/25 13:13 |  | rubina | OK |
| 33 | Q2579-07    | WC-A2-14-C       | SAM | 07/11/25 13:13 |  | rubina | OK |
| 34 | CCV3        | CCV3             | CCV | 07/11/25 13:14 |  | rubina | OK |
| 35 | CCB3        | CCB3             | CCB | 07/11/25 13:14 |  | rubina | OK |
| 36 | PB168777BL  | PB168777BL       | MB  | 07/11/25 13:14 |  | rubina | OK |
| 37 | Q2545-01    | 3918             | SAM | 07/11/25 13:14 |  | rubina | OK |
| 38 | Q2545-01DUP | 3918DUP          | DUP | 07/11/25 13:21 |  | rubina | OK |

Instrument ID: KONELAB

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136439**

|                  |                                                                |              |                       |
|------------------|----------------------------------------------------------------|--------------|-----------------------|
| Review By        | rubina                                                         | Review On    | 7/14/2025 10:30:26 AM |
| Supervise By     | Iwona                                                          | Supervise On | 7/14/2025 10:33:59 AM |
| SubDirectory     | LB136439                                                       | Test         | Reactive Cyanide      |
| <b>STD. NAME</b> | <b>STD REF.#</b>                                               |              |                       |
| ICAL Standard    | WP113905,WP113906,WP113907,WP113908,WP113909,WP113910,WP113911 |              |                       |
| ICV Standard     | WP113912                                                       |              |                       |
| CCV Standard     | WP113906                                                       |              |                       |
| ICSA Standard    | N/A                                                            |              |                       |
| CRI Standard     | N/A                                                            |              |                       |
| LCS Standard     | N/A                                                            |              |                       |
| Chk Standard     | WP112643,WP112900,WP113913                                     |              |                       |

|    |          |                  |     |                |  |        |    |
|----|----------|------------------|-----|----------------|--|--------|----|
| 39 | Q2550-01 | 3326             | SAM | 07/11/25 13:21 |  | rubina | OK |
| 40 | Q2551-01 | 293              | SAM | 07/11/25 13:21 |  | rubina | OK |
| 41 | Q2565-01 | MOO-25-0192-0193 | SAM | 07/11/25 13:21 |  | rubina | OK |
| 42 | CCV4     | CCV4             | CCV | 07/11/25 13:21 |  | rubina | OK |
| 43 | CCB4     | CCB4             | CCB | 07/11/25 13:21 |  | rubina | OK |

**Instrument ID:** TITRAMETRIC

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136443**

|                  |                   |              |                      |
|------------------|-------------------|--------------|----------------------|
| Review By        | rubina            | Review On    | 7/11/2025 3:53:07 PM |
| Supervise By     | Iwona             | Supervise On | 7/11/2025 4:23:51 PM |
| SubDirectory     | LB136443          | Test         | Reactive Sulfide     |
| <b>STD. NAME</b> | <b>STD REF.#</b>  |              |                      |
| ICAL Standard    | N/A               |              |                      |
| ICV Standard     | N/A               |              |                      |
| CCV Standard     | N/A               |              |                      |
| ICSA Standard    | N/A               |              |                      |
| CRI Standard     | N/A               |              |                      |
| LCS Standard     | N/A               |              |                      |
| Chk Standard     | W3105,W3213,W3149 |              |                      |

| Sr# | SampleId    | ClientID         | QcType | Date           | Comment | Operator | Status |
|-----|-------------|------------------|--------|----------------|---------|----------|--------|
| 1   | PB168800BL  | PB168800BL       | MB     | 07/11/25 15:00 |         | rubina   | OK     |
| 2   | Q2519-04    | TP-15            | SAM    | 07/11/25 15:03 |         | rubina   | OK     |
| 3   | Q2519-04DUP | TP-15DUP         | DUP    | 07/11/25 15:06 |         | rubina   | OK     |
| 4   | Q2523-01    | MOO-25-0190      | SAM    | 07/11/25 15:09 |         | rubina   | OK     |
| 5   | Q2526-05    | LAW-25-0101      | SAM    | 07/11/25 15:12 |         | rubina   | OK     |
| 6   | Q2526-06    | LAW-25-0102      | SAM    | 07/11/25 15:14 |         | rubina   | OK     |
| 7   | Q2556-01    | RT3997           | SAM    | 07/11/25 15:17 |         | rubina   | OK     |
| 8   | Q2565-02    | MOO-25-0194-0195 | SAM    | 07/11/25 15:20 |         | rubina   | OK     |
| 9   | Q2565-03    | MOO-25-0191      | SAM    | 07/11/25 15:22 |         | rubina   | OK     |
| 10  | Q2565-04    | MOO-25-0196      | SAM    | 07/11/25 15:24 |         | rubina   | OK     |
| 11  | Q2565-05    | MOO-25-0180      | SAM    | 07/11/25 15:27 |         | rubina   | OK     |
| 12  | Q2571-04    | TP-18            | SAM    | 07/11/25 15:30 |         | rubina   | OK     |
| 13  | Q2571-08    | TP-17            | SAM    | 07/11/25 15:32 |         | rubina   | OK     |
| 14  | Q2578-03    | WC-A5-01-C       | SAM    | 07/11/25 15:34 |         | rubina   | OK     |
| 15  | Q2578-07    | WC-A2-09-C       | SAM    | 07/11/25 15:36 |         | rubina   | OK     |
| 16  | Q2578-11    | WC-A2-10-C       | SAM    | 07/11/25 15:39 |         | rubina   | OK     |
| 17  | Q2578-15    | WC-A2-11-C       | SAM    | 07/11/25 15:42 |         | rubina   | OK     |
| 18  | Q2578-19    | WC-A2-12-C       | SAM    | 07/11/25 15:45 |         | rubina   | OK     |

**Instrument ID:** TITRAMETRIC

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136443**

|                  |                   |              |                      |
|------------------|-------------------|--------------|----------------------|
| Review By        | rubina            | Review On    | 7/11/2025 3:53:07 PM |
| Supervise By     | Iwona             | Supervise On | 7/11/2025 4:23:51 PM |
| SubDirectory     | LB136443          | Test         | Reactive Sulfide     |
| <b>STD. NAME</b> | <b>STD REF.#</b>  |              |                      |
| ICAL Standard    | N/A               |              |                      |
| ICV Standard     | N/A               |              |                      |
| CCV Standard     | N/A               |              |                      |
| ICSA Standard    | N/A               |              |                      |
| CRI Standard     | N/A               |              |                      |
| LCS Standard     | N/A               |              |                      |
| Chk Standard     | W3105,W3213,W3149 |              |                      |

|    |          |            |     |                |  |        |    |
|----|----------|------------|-----|----------------|--|--------|----|
| 19 | Q2579-03 | WC-A2-13-C | SAM | 07/11/25 15:47 |  | rubina | OK |
| 20 | Q2579-07 | WC-A2-14-C | SAM | 07/11/25 15:49 |  | rubina | OK |



**Instrument ID:** WC SC-3

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136448**

|                  |                  |              |                       |
|------------------|------------------|--------------|-----------------------|
| Review By        | jignesh          | Review On    | 7/14/2025 10:27:10 AM |
| Supervise By     | Iwona            | Supervise On | 7/14/2025 11:02:45 AM |
| SubDirectory     | LB136448         | Test         | TS                    |
| <b>STD. NAME</b> | <b>STD REF.#</b> |              |                       |
| ICAL Standard    | N/A              |              |                       |
| ICV Standard     | N/A              |              |                       |
| CCV Standard     | N/A              |              |                       |
| ICSA Standard    | N/A              |              |                       |
| CRI Standard     | N/A              |              |                       |
| LCS Standard     | N/A              |              |                       |
| Chk Standard     | N/A              |              |                       |

| Sr# | SampleID    | ClientID      | QcType | Date           | Comment | Operator | Status |
|-----|-------------|---------------|--------|----------------|---------|----------|--------|
| 1   | LB136448BL  | LB136448BL    | MB     | 07/11/25 11:00 |         | jignesh  | OK     |
| 2   | Q2578-02    | WC-A5-01-C    | SAM    | 07/11/25 11:00 |         | jignesh  | OK     |
| 3   | Q2578-02DUP | WC-A5-01-CDUP | DUP    | 07/11/25 11:00 |         | jignesh  | OK     |
| 4   | Q2578-06    | WC-A2-09-C    | SAM    | 07/11/25 11:00 |         | jignesh  | OK     |
| 5   | Q2578-10    | WC-A2-10-C    | SAM    | 07/11/25 11:00 |         | jignesh  | OK     |
| 6   | Q2578-14    | WC-A2-11-C    | SAM    | 07/11/25 11:00 |         | jignesh  | OK     |
| 7   | Q2578-18    | WC-A2-12-C    | SAM    | 07/11/25 11:00 |         | jignesh  | OK     |
| 8   | Q2579-02    | WC-A2-13-C    | SAM    | 07/11/25 11:00 |         | jignesh  | OK     |
| 9   | Q2579-06    | WC-A2-14-C    | SAM    | 07/11/25 11:00 |         | jignesh  | OK     |

**Instrument ID:** WC SC-3

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136449**

|                  |                  |              |                       |
|------------------|------------------|--------------|-----------------------|
| Review By        | jignesh          | Review On    | 7/14/2025 10:46:28 AM |
| Supervise By     | Iwona            | Supervise On | 7/14/2025 11:00:45 AM |
| SubDirectory     | LB136449         | Test         | TVS                   |
| <b>STD. NAME</b> | <b>STD REF.#</b> |              |                       |
| ICAL Standard    | N/A              |              |                       |
| ICV Standard     | N/A              |              |                       |
| CCV Standard     | N/A              |              |                       |
| ICSA Standard    | N/A              |              |                       |
| CRI Standard     | N/A              |              |                       |
| LCS Standard     | N/A              |              |                       |
| Chk Standard     | N/A              |              |                       |

| Sr# | SampleID    | ClientID      | QcType | Date           | Comment | Operator | Status |
|-----|-------------|---------------|--------|----------------|---------|----------|--------|
| 1   | LB136449BL  | LB136449BL    | MB     | 07/11/25 16:00 |         | jignesh  | OK     |
| 2   | Q2578-02    | WC-A5-01-C    | SAM    | 07/11/25 16:00 |         | jignesh  | OK     |
| 3   | Q2578-02DUP | WC-A5-01-CDUP | DUP    | 07/11/25 16:00 |         | jignesh  | OK     |
| 4   | Q2578-06    | WC-A2-09-C    | SAM    | 07/11/25 16:00 |         | jignesh  | OK     |
| 5   | Q2578-10    | WC-A2-10-C    | SAM    | 07/11/25 16:00 |         | jignesh  | OK     |
| 6   | Q2578-14    | WC-A2-11-C    | SAM    | 07/11/25 16:00 |         | jignesh  | OK     |
| 7   | Q2578-18    | WC-A2-12-C    | SAM    | 07/11/25 16:00 |         | jignesh  | OK     |
| 8   | Q2579-02    | WC-A2-13-C    | SAM    | 07/11/25 16:00 |         | jignesh  | OK     |
| 9   | Q2579-06    | WC-A2-14-C    | SAM    | 07/11/25 16:00 |         | jignesh  | OK     |

**Instrument ID:** WC PH METER-1

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136451**

|                  |                                     |              |                       |
|------------------|-------------------------------------|--------------|-----------------------|
| Review By        | jignesh                             | Review On    | 7/11/2025 4:23:19 PM  |
| Supervise By     | Iwona                               | Supervise On | 7/14/2025 10:56:31 AM |
| SubDirectory     | LB136451                            | Test         | pH                    |
| <b>STD. NAME</b> | <b>STD REF.#</b>                    |              |                       |
| ICAL Standard    | N/A                                 |              |                       |
| ICV Standard     | N/A                                 |              |                       |
| CCV Standard     | N/A                                 |              |                       |
| ICSA Standard    | N/A                                 |              |                       |
| CRI Standard     | N/A                                 |              |                       |
| LCS Standard     | N/A                                 |              |                       |
| Chk Standard     | W3178,W3093,W3191,W3217,W3161,W3200 |              |                       |

| Sr# | SampleId    | ClientID      | QcType | Date           | Comment | Operator | Status |
|-----|-------------|---------------|--------|----------------|---------|----------|--------|
| 1   | CAL1        | CAL1          | CAL    | 07/11/25 14:20 |         | Jignesh  | OK     |
| 2   | CAL2        | CAL2          | CAL    | 07/11/25 14:21 |         | Jignesh  | OK     |
| 3   | CAL3        | CAL3          | CAL    | 07/11/25 14:22 |         | Jignesh  | OK     |
| 4   | ICV         | ICV           | ICV    | 07/11/25 14:25 |         | Jignesh  | OK     |
| 5   | CCV1        | CCV1          | CCV    | 07/11/25 14:33 |         | Jignesh  | OK     |
| 6   | Q2578-02    | WC-A5-01-C    | SAM    | 07/11/25 14:55 |         | Jignesh  | OK     |
| 7   | Q2578-02DUP | WC-A5-01-CDUP | DUP    | 07/11/25 14:56 |         | Jignesh  | OK     |
| 8   | Q2578-06    | WC-A2-09-C    | SAM    | 07/11/25 15:00 |         | Jignesh  | OK     |
| 9   | Q2578-10    | WC-A2-10-C    | SAM    | 07/11/25 15:05 |         | Jignesh  | OK     |
| 10  | Q2578-14    | WC-A2-11-C    | SAM    | 07/11/25 15:15 |         | Jignesh  | OK     |
| 11  | Q2578-18    | WC-A2-12-C    | SAM    | 07/11/25 15:25 |         | Jignesh  | OK     |
| 12  | Q2579-02    | WC-A2-13-C    | SAM    | 07/11/25 15:35 |         | Jignesh  | OK     |
| 13  | Q2579-06    | WC-A2-14-C    | SAM    | 07/11/25 15:45 |         | Jignesh  | OK     |
| 14  | Q2589-01    | AU-06-071125  | SAM    | 07/11/25 16:00 |         | Jignesh  | OK     |
| 15  | CCV2        | CCV2          | CCV    | 07/11/25 16:02 |         | Jignesh  | OK     |

**Instrument ID:** WC PH METER-1

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136452**

|                  |                                     |              |                       |
|------------------|-------------------------------------|--------------|-----------------------|
| Review By        | jignesh                             | Review On    | 7/11/2025 4:30:58 PM  |
| Supervise By     | Iwona                               | Supervise On | 7/14/2025 10:56:16 AM |
| SubDirectory     | LB136452                            | Test         | Corrosivity           |
| <b>STD. NAME</b> | <b>STD REF.#</b>                    |              |                       |
| ICAL Standard    | N/A                                 |              |                       |
| ICV Standard     | N/A                                 |              |                       |
| CCV Standard     | N/A                                 |              |                       |
| ICSA Standard    | N/A                                 |              |                       |
| CRI Standard     | N/A                                 |              |                       |
| LCS Standard     | N/A                                 |              |                       |
| Chk Standard     | W3178,W3093,W3191,W3217,W3161,W3200 |              |                       |

| Sr# | SampleId    | ClientID      | QcType | Date           | Comment | Operator | Status |
|-----|-------------|---------------|--------|----------------|---------|----------|--------|
| 1   | CAL1        | CAL1          | CAL    | 07/11/25 14:20 |         | Jignesh  | OK     |
| 2   | CAL2        | CAL2          | CAL    | 07/11/25 14:21 |         | Jignesh  | OK     |
| 3   | CAL3        | CAL3          | CAL    | 07/11/25 14:22 |         | Jignesh  | OK     |
| 4   | ICV         | ICV           | ICV    | 07/11/25 14:25 |         | Jignesh  | OK     |
| 5   | CCV1        | CCV1          | CCV    | 07/11/25 14:33 |         | Jignesh  | OK     |
| 6   | Q2578-03    | WC-A5-01-C    | SAM    | 07/11/25 14:55 |         | Jignesh  | OK     |
| 7   | Q2578-03DUP | WC-A5-01-CDUP | DUP    | 07/11/25 14:56 |         | Jignesh  | OK     |
| 8   | Q2578-07    | WC-A2-09-C    | SAM    | 07/11/25 15:00 |         | Jignesh  | OK     |
| 9   | Q2578-11    | WC-A2-10-C    | SAM    | 07/11/25 15:05 |         | Jignesh  | OK     |
| 10  | Q2578-15    | WC-A2-11-C    | SAM    | 07/11/25 15:15 |         | Jignesh  | OK     |
| 11  | Q2578-19    | WC-A2-12-C    | SAM    | 07/11/25 15:25 |         | Jignesh  | OK     |
| 12  | Q2579-03    | WC-A2-13-C    | SAM    | 07/11/25 15:35 |         | Jignesh  | OK     |
| 13  | Q2579-07    | WC-A2-14-C    | SAM    | 07/11/25 15:45 |         | Jignesh  | OK     |
| 14  | Q2586-04    | TP-16         | SAM    | 07/11/25 15:56 |         | Jignesh  | OK     |
| 15  | CCV2        | CCV2          | CCV    | 07/11/25 16:02 |         | Jignesh  | OK     |

**Instrument ID:** WC SC-3

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136455**

|                  |                  |              |                       |
|------------------|------------------|--------------|-----------------------|
| Review By        | jignesh          | Review On    | 7/15/2025 10:28:04 AM |
| Supervise By     | Iwona            | Supervise On | 7/15/2025 1:22:21 PM  |
| SubDirectory     | LB136455         | Test         | ASTM TS               |
| <b>STD. NAME</b> | <b>STD REF.#</b> |              |                       |
| ICAL Standard    | N/A              |              |                       |
| ICV Standard     | N/A              |              |                       |
| CCV Standard     | N/A              |              |                       |
| ICSA Standard    | N/A              |              |                       |
| CRI Standard     | N/A              |              |                       |
| LCS Standard     | N/A              |              |                       |
| Chk Standard     | N/A              |              |                       |

| Sr# | SampleID    | ClientID      | QcType | Date           | Comment | Operator | Status |
|-----|-------------|---------------|--------|----------------|---------|----------|--------|
| 1   | LB136455BL  | LB136455BL    | MB     | 07/14/25 12:00 |         | jignesh  | OK     |
| 2   | Q2578-04    | WC-A5-01-C    | SAM    | 07/14/25 12:00 |         | jignesh  | OK     |
| 3   | Q2578-04DUP | WC-A5-01-CDUP | DUP    | 07/14/25 12:00 |         | jignesh  | OK     |
| 4   | Q2578-08    | WC-A2-09-C    | SAM    | 07/14/25 12:00 |         | jignesh  | OK     |
| 5   | Q2578-12    | WC-A2-10-C    | SAM    | 07/14/25 12:00 |         | jignesh  | OK     |
| 6   | Q2578-16    | WC-A2-11-C    | SAM    | 07/14/25 12:00 |         | jignesh  | OK     |
| 7   | Q2578-20    | WC-A2-12-C    | SAM    | 07/14/25 12:00 |         | jignesh  | OK     |
| 8   | Q2579-04    | WC-A2-13-C    | SAM    | 07/14/25 12:00 |         | jignesh  | OK     |
| 9   | Q2579-08    | WC-A2-14-C    | SAM    | 07/14/25 12:00 |         | jignesh  | OK     |

**Instrument ID:** WC SC-3

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136456**

|                  |                                                        |              |                       |
|------------------|--------------------------------------------------------|--------------|-----------------------|
| Review By        | jignesh                                                | Review On    | 7/14/2025 11:54:45 AM |
| Supervise By     | Iwona                                                  | Supervise On | 7/15/2025 1:21:55 PM  |
| SubDirectory     | LB136456                                               | Test         | ASTM Oil and Grease   |
| <b>STD. NAME</b> | <b>STD REF.#</b>                                       |              |                       |
| ICAL Standard    | N/A                                                    |              |                       |
| ICV Standard     | N/A                                                    |              |                       |
| CCV Standard     | N/A                                                    |              |                       |
| ICSA Standard    | N/A                                                    |              |                       |
| CRI Standard     | N/A                                                    |              |                       |
| LCS Standard     | N/A                                                    |              |                       |
| Chk Standard     | W3204,M6069,EP2624,WP112782,NA,NA,WP112783,NA,WO112784 |              |                       |

| Sr# | SampleId    | ClientID      | QcType | Date           | Comment | Operator | Status |
|-----|-------------|---------------|--------|----------------|---------|----------|--------|
| 1   | LB136456BL  | LB136456BL    | MB     | 07/14/25 13:00 |         | jignesh  | OK     |
| 2   | LB136456BS  | LB136456BS    | LCS    | 07/14/25 13:00 |         | jignesh  | OK     |
| 3   | Q2578-04    | WC-A5-01-C    | SAM    | 07/14/25 13:00 |         | jignesh  | OK     |
| 4   | Q2578-04DUP | WC-A5-01-CDUP | DUP    | 07/14/25 13:00 |         | jignesh  | OK     |
| 5   | Q2578-04MS  | WC-A5-01-CMS  | MS     | 07/14/25 13:00 |         | jignesh  | OK     |
| 6   | Q2578-04MSD | WC-A5-01-CMSD | MSD    | 07/14/25 13:00 |         | jignesh  | OK     |
| 7   | Q2578-08    | WC-A2-09-C    | SAM    | 07/14/25 13:00 |         | jignesh  | OK     |
| 8   | Q2578-12    | WC-A2-10-C    | SAM    | 07/14/25 13:00 |         | jignesh  | OK     |
| 9   | Q2578-16    | WC-A2-11-C    | SAM    | 07/14/25 13:00 |         | jignesh  | OK     |
| 10  | Q2578-20    | WC-A2-12-C    | SAM    | 07/14/25 13:00 |         | jignesh  | OK     |
| 11  | Q2579-04    | WC-A2-13-C    | SAM    | 07/14/25 13:00 |         | jignesh  | OK     |
| 12  | Q2579-08    | WC-A2-14-C    | SAM    | 07/14/25 13:00 |         | jignesh  | OK     |

**Instrument ID:** FLAME

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136462**

|                  |                  |              |                      |
|------------------|------------------|--------------|----------------------|
| Review By        | Eman             | Review On    | 7/14/2025 4:26:48 PM |
| Supervise By     | Iwona            | Supervise On | 7/14/2025 4:51:34 PM |
| SubDirectory     | LB136462         | Test         | Ignitability         |
| <b>STD. NAME</b> | <b>STD REF.#</b> |              |                      |
| ICAL Standard    | N/A              |              |                      |
| ICV Standard     | N/A              |              |                      |
| CCV Standard     | N/A              |              |                      |
| ICSA Standard    | N/A              |              |                      |
| CRI Standard     | N/A              |              |                      |
| LCS Standard     | N/A              |              |                      |
| Chk Standard     | N/A              |              |                      |

| Sr# | SampleId    | ClientID      | QcType | Date           | Comment | Operator | Status |
|-----|-------------|---------------|--------|----------------|---------|----------|--------|
| 1   | Q2578-03    | WC-A5-01-C    | SAM    | 07/14/25 08:30 |         | Eman     | OK     |
| 2   | Q2578-03DUP | WC-A5-01-CDUP | DUP    | 07/14/25 08:37 |         | Eman     | OK     |
| 3   | Q2578-07    | WC-A2-09-C    | SAM    | 07/14/25 08:45 |         | Eman     | OK     |
| 4   | Q2578-11    | WC-A2-10-C    | SAM    | 07/14/25 08:52 |         | Eman     | OK     |
| 5   | Q2578-15    | WC-A2-11-C    | SAM    | 07/14/25 09:00 |         | Eman     | OK     |
| 6   | Q2578-19    | WC-A2-12-C    | SAM    | 07/14/25 09:08 |         | Eman     | OK     |
| 7   | Q2579-03    | WC-A2-13-C    | SAM    | 07/14/25 09:15 |         | Eman     | OK     |
| 8   | Q2579-07    | WC-A2-14-C    | SAM    | 07/14/25 09:22 |         | Eman     | OK     |
| 9   | Q2586-01    | TP-16         | SAM    | 07/14/25 09:30 |         | Eman     | OK     |
| 10  | Q2586-04    | TP-16         | SAM    | 07/14/25 09:37 |         | Eman     | OK     |

**Instrument ID:** FILTER/GRAVIMETRIC

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136463**

|                  |                  |              |                      |
|------------------|------------------|--------------|----------------------|
| Review By        | Eman             | Review On    | 7/14/2025 4:26:57 PM |
| Supervise By     | Iwona            | Supervise On | 7/14/2025 4:50:18 PM |
| SubDirectory     | LB136463         | Test         | Paint Filter         |
| <b>STD. NAME</b> | <b>STD REF.#</b> |              |                      |
| ICAL Standard    | N/A              |              |                      |
| ICV Standard     | N/A              |              |                      |
| CCV Standard     | N/A              |              |                      |
| ICSA Standard    | N/A              |              |                      |
| CRI Standard     | N/A              |              |                      |
| LCS Standard     | N/A              |              |                      |
| Chk Standard     | N/A              |              |                      |

| Sr# | SampleId    | ClientID      | QcType | Date           | Comment | Operator | Status |
|-----|-------------|---------------|--------|----------------|---------|----------|--------|
| 1   | Q2578-02    | WC-A5-01-C    | SAM    | 07/14/25 09:50 |         | Eman     | OK     |
| 2   | Q2578-02DUP | WC-A5-01-CDUP | DUP    | 07/14/25 09:57 |         | Eman     | OK     |
| 3   | Q2578-06    | WC-A2-09-C    | SAM    | 07/14/25 10:05 |         | Eman     | OK     |
| 4   | Q2578-10    | WC-A2-10-C    | SAM    | 07/14/25 10:12 |         | Eman     | OK     |
| 5   | Q2578-14    | WC-A2-11-C    | SAM    | 07/14/25 10:20 |         | Eman     | OK     |
| 6   | Q2578-18    | WC-A2-12-C    | SAM    | 07/14/25 10:27 |         | Eman     | OK     |
| 7   | Q2579-02    | WC-A2-13-C    | SAM    | 07/14/25 10:35 |         | Eman     | OK     |
| 8   | Q2579-06    | WC-A2-14-C    | SAM    | 07/14/25 10:42 |         | Eman     | OK     |
| 9   | Q2586-01    | TP-16         | SAM    | 07/14/25 10:50 |         | Eman     | OK     |
| 10  | Q2590-01    | NB-07-071125  | SAM    | 07/14/25 10:58 |         | Eman     | OK     |



**Instrument ID:** KONELAB

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136468**

|                  |                                     |              |                       |
|------------------|-------------------------------------|--------------|-----------------------|
| Review By        | rubina                              | Review On    | 7/15/2025 11:07:50 AM |
| Supervise By     | Iwona                               | Supervise On | 7/15/2025 11:58:23 AM |
| SubDirectory     | LB136468                            | Test         | ASTM Ammonia          |
| <b>STD. NAME</b> | <b>STD REF.#</b>                    |              |                       |
| ICAL Standard    | WP113926                            |              |                       |
| ICV Standard     | WP113928                            |              |                       |
| CCV Standard     | WP113927                            |              |                       |
| ICSA Standard    | N/A                                 |              |                       |
| CRI Standard     | N/A                                 |              |                       |
| LCS Standard     | WP113889                            |              |                       |
| Chk Standard     | WP113852,WP111745,WP113929,WP111660 |              |                       |

| Sr# | SampleId    | ClientID      | QcType | Date           | Comment | Operator | Status |
|-----|-------------|---------------|--------|----------------|---------|----------|--------|
| 1   | 0.0PPM      | 0.0PPM        | CAL1   | 07/14/25 10:59 |         | rubina   | OK     |
| 2   | 0.1PPM      | 0.1PPM        | CAL2   | 07/14/25 10:59 |         | rubina   | OK     |
| 3   | 0.2PPM      | 0.2PPM        | CAL3   | 07/14/25 10:59 |         | rubina   | OK     |
| 4   | 0.4PPM      | 0.4PPM        | CAL4   | 07/14/25 10:59 |         | rubina   | OK     |
| 5   | 1.0PPM      | 1.0PPM        | CAL5   | 07/14/25 10:59 |         | rubina   | OK     |
| 6   | 1.3PPM      | 1.3PPM        | CAL6   | 07/14/25 10:59 |         | rubina   | OK     |
| 7   | 2.0PPM      | 2.0PPM        | CAL7   | 07/14/25 11:00 |         | rubina   | OK     |
| 8   | ICV1        | ICV1          | ICV    | 07/14/25 16:06 |         | rubina   | OK     |
| 9   | ICB1        | ICB1          | ICB    | 07/14/25 16:06 |         | rubina   | OK     |
| 10  | CCV1        | CCV1          | CCV    | 07/14/25 16:06 |         | rubina   | OK     |
| 11  | CCB1        | CCB1          | CCB    | 07/14/25 16:06 |         | rubina   | OK     |
| 12  | PB168836BL  | PB168836BL    | MB     | 07/14/25 16:06 |         | rubina   | OK     |
| 13  | PB168836BS  | PB168836BS    | LCS    | 07/14/25 16:06 |         | rubina   | OK     |
| 14  | PB168836TB  | PB168836TB    | MB     | 07/14/25 16:17 |         | rubina   | OK     |
| 15  | Q2578-04    | WC-A5-01-C    | SAM    | 07/14/25 16:17 |         | rubina   | OK     |
| 16  | Q2578-04DUP | WC-A5-01-CDUP | DUP    | 07/14/25 16:17 |         | rubina   | OK     |
| 17  | Q2578-04MS  | WC-A5-01-CMS  | MS     | 07/14/25 16:17 |         | rubina   | OK     |
| 18  | Q2578-04MSD | WC-A5-01-CMSD | MSD    | 07/14/25 16:17 |         | rubina   | OK     |

Instrument ID: KONELAB

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136468**

|                  |                                     |              |                       |
|------------------|-------------------------------------|--------------|-----------------------|
| Review By        | rubina                              | Review On    | 7/15/2025 11:07:50 AM |
| Supervise By     | Iwona                               | Supervise On | 7/15/2025 11:58:23 AM |
| SubDirectory     | LB136468                            | Test         | ASTM Ammonia          |
| <b>STD. NAME</b> | <b>STD REF.#</b>                    |              |                       |
| ICAL Standard    | WP113926                            |              |                       |
| ICV Standard     | WP113928                            |              |                       |
| CCV Standard     | WP113927                            |              |                       |
| ICSA Standard    | N/A                                 |              |                       |
| CRI Standard     | N/A                                 |              |                       |
| LCS Standard     | WP113889                            |              |                       |
| Chk Standard     | WP113852,WP111745,WP113929,WP111660 |              |                       |

|    |          |            |     |                |  |        |    |
|----|----------|------------|-----|----------------|--|--------|----|
| 19 | Q2578-08 | WC-A2-09-C | SAM | 07/14/25 16:17 |  | rubina | OK |
| 20 | Q2578-12 | WC-A2-10-C | SAM | 07/14/25 16:17 |  | rubina | OK |
| 21 | Q2578-16 | WC-A2-11-C | SAM | 07/14/25 16:17 |  | rubina | OK |
| 22 | CCV2     | CCV2       | CCV | 07/14/25 16:28 |  | rubina | OK |
| 23 | CCB2     | CCB2       | CCB | 07/14/25 16:28 |  | rubina | OK |
| 24 | Q2578-20 | WC-A2-12-C | SAM | 07/14/25 16:28 |  | rubina | OK |
| 25 | Q2579-04 | WC-A2-13-C | SAM | 07/14/25 16:28 |  | rubina | OK |
| 26 | Q2579-08 | WC-A2-14-C | SAM | 07/14/25 16:28 |  | rubina | OK |
| 27 | CCV3     | CCV3       | CCV | 07/14/25 16:28 |  | rubina | OK |
| 28 | CCB3     | CCB3       | CCB | 07/14/25 16:32 |  | rubina | OK |

**Instrument ID:** WC SC-3

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136474**

|                  |                                                  |              |                       |
|------------------|--------------------------------------------------|--------------|-----------------------|
| Review By        | jignesh                                          | Review On    | 7/16/2025 10:59:13 AM |
| Supervise By     | Iwona                                            | Supervise On | 7/16/2025 1:34:02 PM  |
| SubDirectory     | LB136474                                         | Test         | Oil and Grease        |
| <b>STD. NAME</b> | <b>STD REF.#</b>                                 |              |                       |
| ICAL Standard    | N/A                                              |              |                       |
| ICV Standard     | N/A                                              |              |                       |
| CCV Standard     | N/A                                              |              |                       |
| ICSA Standard    | N/A                                              |              |                       |
| CRI Standard     | N/A                                              |              |                       |
| LCS Standard     | N/A                                              |              |                       |
| Chk Standard     | W3204,NA,EP2625,NA,NA,W3186,WP112785,NA,WP112786 |              |                       |

| Sr# | SampleID    | ClientID      | QcType | Date           | Comment | Operator | Status |
|-----|-------------|---------------|--------|----------------|---------|----------|--------|
| 1   | LB136474BL  | LB136474BL    | MB     | 07/15/25 11:45 |         | jignesh  | OK     |
| 2   | LB136474BS  | LB136474BS    | LCS    | 07/15/25 11:45 |         | jignesh  | OK     |
| 3   | Q2578-02    | WC-A5-01-C    | SAM    | 07/15/25 11:45 |         | jignesh  | OK     |
| 4   | Q2578-06    | WC-A2-09-C    | SAM    | 07/15/25 11:45 |         | jignesh  | OK     |
| 5   | Q2578-10    | WC-A2-10-C    | SAM    | 07/15/25 11:45 |         | jignesh  | OK     |
| 6   | Q2578-14    | WC-A2-11-C    | SAM    | 07/15/25 11:45 |         | jignesh  | OK     |
| 7   | Q2578-18    | WC-A2-12-C    | SAM    | 07/15/25 11:45 |         | jignesh  | OK     |
| 8   | Q2579-02    | WC-A2-13-C    | SAM    | 07/15/25 11:45 |         | jignesh  | OK     |
| 9   | Q2579-06    | WC-A2-14-C    | SAM    | 07/15/25 11:45 |         | jignesh  | OK     |
| 10  | Q2579-06DUP | WC-A2-14-CDUP | DUP    | 07/15/25 11:45 |         | jignesh  | OK     |
| 11  | Q2579-06MS  | WC-A2-14-CMS  | MS     | 07/15/25 11:45 |         | jignesh  | OK     |
| 12  | Q2579-06MSD | WC-A2-14-CMSD | MSD    | 07/15/25 11:45 |         | jignesh  | OK     |

**Instrument ID:** SPECTROPHOTOMETER-2

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136478**

|                  |                                                                                            |              |                       |
|------------------|--------------------------------------------------------------------------------------------|--------------|-----------------------|
| Review By        | Iwona                                                                                      | Review On    | 7/15/2025 12:45:43 PM |
| Supervise By     | jignesh                                                                                    | Supervise On | 7/15/2025 12:48:42 PM |
| SubDirectory     | LB136478                                                                                   | Test         | ASTM COD              |
| <b>STD. NAME</b> | <b>STD REF.#</b>                                                                           |              |                       |
| ICAL Standard    | N/A                                                                                        |              |                       |
| ICV Standard     | N/A                                                                                        |              |                       |
| CCV Standard     | N/A                                                                                        |              |                       |
| ICSA Standard    | N/A                                                                                        |              |                       |
| CRI Standard     | N/A                                                                                        |              |                       |
| LCS Standard     | N/A                                                                                        |              |                       |
| Chk Standard     | WP113238,WP113237,WP113235,WP113234,WP113233,WP113240,WP113236,W3129,WP113940,WP113941,WP1 |              |                       |

| Sr# | SampleId    | ClientID      | QcType | Date           | Comment | Operator | Status |
|-----|-------------|---------------|--------|----------------|---------|----------|--------|
| 1   | CAL1        | CAL1          | CAL    | 05/28/25 13:10 |         | Iwona    | OK     |
| 2   | CAL2        | CAL2          | CAL    | 05/28/25 13:10 |         | Iwona    | OK     |
| 3   | CAL3        | CAL3          | CAL    | 05/28/25 13:11 |         | Iwona    | OK     |
| 4   | CAL4        | CAL4          | CAL    | 05/28/25 13:11 |         | Iwona    | OK     |
| 5   | CAL5        | CAL5          | CAL    | 05/28/25 13:12 |         | Iwona    | OK     |
| 6   | CAL6        | CAL6          | CAL    | 05/28/25 13:12 |         | Iwona    | OK     |
| 7   | ICV         | ICV           | ICV    | 05/28/25 13:13 |         | Iwona    | OK     |
| 8   | ICB         | ICB           | ICB    | 05/28/25 13:13 |         | Iwona    | OK     |
| 9   | RL Check    | RL Check      | RL     | 07/15/25 12:16 |         | Iwona    | OK     |
| 10  | CCV1        | CCV1          | CCV    | 07/15/25 12:25 |         | Iwona    | OK     |
| 11  | CCB1        | CCB1          | CCB    | 07/15/25 12:25 |         | Iwona    | OK     |
| 12  | LB136478BL  | LB136478BL    | MB     | 07/15/25 12:26 |         | Iwona    | OK     |
| 13  | LB136478BS  | LB136478BS    | LCS    | 07/15/25 12:27 |         | Iwona    | OK     |
| 14  | Q2578-04    | WC-A5-01-C    | SAM    | 07/15/25 12:27 |         | Iwona    | OK     |
| 15  | Q2578-04DUP | WC-A5-01-CDUP | DUP    | 07/15/25 12:28 |         | Iwona    | OK     |
| 16  | Q2578-04MS  | WC-A5-01-CMS  | MS     | 07/15/25 12:28 |         | Iwona    | OK     |
| 17  | Q2578-04MSD | WC-A5-01-CMSD | MSD    | 07/15/25 12:29 |         | Iwona    | OK     |
| 18  | Q2578-08    | WC-A2-09-C    | SAM    | 07/15/25 12:29 |         | Iwona    | OK     |

**Instrument ID:** SPECTROPHOTOMETER-2

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136478**

|                  |                                                                                            |              |                       |
|------------------|--------------------------------------------------------------------------------------------|--------------|-----------------------|
| Review By        | Iwona                                                                                      | Review On    | 7/15/2025 12:45:43 PM |
| Supervise By     | jignesh                                                                                    | Supervise On | 7/15/2025 12:48:42 PM |
| SubDirectory     | LB136478                                                                                   | Test         | ASTM COD              |
| <b>STD. NAME</b> | <b>STD REF.#</b>                                                                           |              |                       |
| ICAL Standard    | N/A                                                                                        |              |                       |
| ICV Standard     | N/A                                                                                        |              |                       |
| CCV Standard     | N/A                                                                                        |              |                       |
| ICSA Standard    | N/A                                                                                        |              |                       |
| CRI Standard     | N/A                                                                                        |              |                       |
| LCS Standard     | N/A                                                                                        |              |                       |
| Chk Standard     | WP113238,WP113237,WP113235,WP113234,WP113233,WP113240,WP113236,W3129,WP113940,WP113941,WP1 |              |                       |

|    |          |            |     |                |  |       |    |
|----|----------|------------|-----|----------------|--|-------|----|
| 19 | Q2578-12 | WC-A2-10-C | SAM | 07/15/25 12:30 |  | Iwona | OK |
| 20 | Q2578-16 | WC-A2-11-C | SAM | 07/15/25 12:30 |  | Iwona | OK |
| 21 | Q2578-20 | WC-A2-12-C | SAM | 07/15/25 12:31 |  | Iwona | OK |
| 22 | CCV2     | CCV2       | CCV | 07/15/25 12:31 |  | Iwona | OK |
| 23 | CCB2     | CCB2       | CCB | 07/15/25 12:32 |  | Iwona | OK |
| 24 | Q2579-04 | WC-A2-13-C | SAM | 07/15/25 12:32 |  | Iwona | OK |
| 25 | Q2579-08 | WC-A2-14-C | SAM | 07/15/25 12:33 |  | Iwona | OK |
| 26 | CCV3     | CCV3       | CCV | 07/15/25 12:33 |  | Iwona | OK |
| 27 | CCB3     | CCB3       | CCB | 07/15/25 12:34 |  | Iwona | OK |

## Prep Standard - Chemical Standard Summary

**Order ID :** Q2579

**Test :** ASTM Ammonia,ASTM COD,ASTM Oil and Grease,ASTM TS,Corrosivity,Ignitability,Oil and Grease,Paint Filter,Percent Solids,pH,Reactive Cyanide,Reactive Sulfide,TS,TVS

**Prepbatch ID :** PB168800,PB168801,PB168836,

**Sequence ID/Qc Batch ID:** LB136439,LB136443,LB136448,LB136449,LB136451,LB136452,LB136455,LB136456,LB136462,LB136463

### Standard ID :

EP2624,EP2625,WP111660,WP111745,WP112611,WP112612,WP112643,WP112782,WP112783,WP112785,WP112786,WP112828,WP112900,WP113086,WP113231,WP113232,WP113233,WP113234,WP113235,WP113236,WP113237,WP113238,WP113240,WP113836,WP113838,WP113852,WP113885,WP113886,WP113887,WP113888,WP113889,WP113904,WP113905,WP113906,WP113907,WP113908,WP113909,WP113910,WP113911,WP113912,WP113913,WP113926,WP113927,WP113928,WP113929,WP113938,WP113939,WP113940,WP113941,WP113942,

### Chemical ID :

AS PER

PB168835,E3551,E3917,M6041,M6069,M6151,W2663,W2666,W2668,W2725,W2784,W2817,W2871,W2926,W3009,W3019,W3082,W3093,W3105,W3112,W3113,W3129,W3132,W3133,W3139,W3149,W3155,W3161,W3169,W3174,W3178,W3186,W3191,W3195,W3196,W3200,W3201,W3203,W3204,W3213,W3214,W3217,W3219,W3224,WO 112784,



| <u>Recipe ID</u>                                                           | <u>NAME</u>          | <u>NO.</u>             | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u>                  | <u>PipetteID</u> | <u>Supervised By</u>                |
|----------------------------------------------------------------------------|----------------------|------------------------|------------------|------------------------|--------------------|---------------------------------|------------------|-------------------------------------|
| 3923                                                                       | Baked Sodium Sulfate | <a href="#">EP2624</a> | 06/26/2025       | 12/04/2025             | RUPESHKUMAR SHAH   | Extraction_SCALE_2<br>(EX-SC-2) | None             | Riteshkumar Patel<br><br>06/26/2025 |
| <b><u>FROM</u></b> 4000.00000gram of E3551 = Final Quantity: 4000.000 gram |                      |                        |                  |                        |                    |                                 |                  |                                     |

| <u>Recipe ID</u>                                                           | <u>NAME</u>          | <u>NO.</u>             | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u>                  | <u>PipetteID</u> | <u>Supervised By</u>            |
|----------------------------------------------------------------------------|----------------------|------------------------|------------------|------------------------|--------------------|---------------------------------|------------------|---------------------------------|
| 3923                                                                       | Baked Sodium Sulfate | <a href="#">EP2625</a> | 07/15/2025       | 12/04/2025             | RUPESHKUMAR SHAH   | Extraction_SCALE_2<br>(EX-SC-2) | None             | Riteshkumar Patel<br>07/15/2025 |
| <b><u>FROM</u></b> 4000.00000gram of E3551 = Final Quantity: 4000.000 gram |                      |                        |                  |                        |                    |                                 |                  |                                 |



| <u>Recipe ID</u>                                                                                                     | <u>NAME</u>             | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u>          | <u>PipetteID</u> | <u>Supervised By</u>       |
|----------------------------------------------------------------------------------------------------------------------|-------------------------|--------------------------|------------------|------------------------|--------------------|-------------------------|------------------|----------------------------|
| 635                                                                                                                  | EDTA BUFFER FOR AMMONIA | <a href="#">WP111660</a> | 01/28/2025       | 07/28/2025             | Rubina Mughal      | WETCHEM_SCALE_8 (WCS-7) | None             | Iwona Zarych<br>01/28/2025 |
| <b><u>FROM</u></b> 5.50000gram of W3113 + 50.00000gram of W3132 + 950.00000ml of W3112 = Final Quantity: 1000.000 ml |                         |                          |                  |                        |                    |                         |                  |                            |

| <u>Recipe ID</u>                                                                          | <u>NAME</u>                     | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u>       |
|-------------------------------------------------------------------------------------------|---------------------------------|--------------------------|------------------|------------------------|--------------------|----------------|------------------|----------------------------|
| 289                                                                                       | Sodium Hypochlorite for Ammonia | <a href="#">WP111745</a> | 02/03/2025       | 07/31/2025             | Rubina Mughal      | None           | None             | Iwona Zarych<br>02/03/2025 |
| <b><u>FROM</u></b> 50.00000ml of W3112 + 50.00000ml of W3174 = Final Quantity: 100.000 ml |                                 |                          |                  |                        |                    |                |                  |                            |





| <u>Recipe ID</u>                                                                             | <u>NAME</u>                   | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u>                   | <u>PipetteID</u> | <u>Supervised By</u>           |
|----------------------------------------------------------------------------------------------|-------------------------------|--------------------------|------------------|------------------------|--------------------|----------------------------------|------------------|--------------------------------|
| 153                                                                                          | Ammonia Stock Std. (1000 ppm) | <a href="#">WP112611</a> | 04/07/2025       | 10/07/2025             | Rubina Mughal      | WETCHEM_S<br>CALE_8 (WC<br>SC-7) | None             | Iwona Zarych<br><br>04/07/2025 |
| <b><u>FROM</u></b> 3.81900gram of W3196 + 996.18100ml of W3112 = Final Quantity: 1000.000 ml |                               |                          |                  |                        |                    |                                  |                  |                                |

| <u>Recipe ID</u> | <u>NAME</u>                                                               | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u>                   | <u>PipetteID</u> | <u>Supervised By</u>           |
|------------------|---------------------------------------------------------------------------|--------------------------|------------------|------------------------|--------------------|----------------------------------|------------------|--------------------------------|
| 1895             | Ammonia Stock Std,<br>1000PPM-SS                                          | <a href="#">WP112612</a> | 04/07/2025       | 10/07/2025             | Rubina Mughal      | WETCHEM_S<br>CALE_8 (WC<br>SC-7) | None             | Iwona Zarych<br><br>04/07/2025 |
| <u>FROM</u>      | 3.81900gram of W3195 + 996.18100ml of W3112 = Final Quantity: 1000.000 ml |                          |                  |                        |                    |                                  |                  |                                |



| <u>Recipe ID</u> | <u>NAME</u>                                                                 | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u>                   | <u>PipetteID</u> | <u>Supervised By</u>           |
|------------------|-----------------------------------------------------------------------------|--------------------------|------------------|------------------------|--------------------|----------------------------------|------------------|--------------------------------|
| 539              | CN BUFFER                                                                   | <a href="#">WP112643</a> | 04/09/2025       | 10/09/2025             | Niha Farheen Shaik | WETCHEM_S<br>CALE_5 (WC<br>SC-5) | None             | Iwona Zarych<br><br>04/09/2025 |
| <u>FROM</u>      | 138.00000gram of W2668 + 862.00000ml of W3112 = Final Quantity: 1000.000 ml |                          |                  |                        |                    |                                  |                  |                                |

| <u>Recipe ID</u>                                                                         | <u>NAME</u> | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u>       |
|------------------------------------------------------------------------------------------|-------------|--------------------------|------------------|------------------------|--------------------|----------------|------------------|----------------------------|
| 229                                                                                      | 1:1 HCL     | <a href="#">WP112782</a> | 04/22/2025       | 08/18/2025             | Jignesh Parikh     | None           | None             | Iwona Zarych<br>04/22/2025 |
| <b><u>FROM</u></b> 500.00000ml of M6151 + 500.00000ml of W3112 = Final Quantity: 1.000 L |             |                          |                  |                        |                    |                |                  |                            |



| <u>Recipe ID</u> | <u>NAME</u>                                                                                       | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u>          | <u>PipetteID</u> | <u>Supervised By</u>       |
|------------------|---------------------------------------------------------------------------------------------------|--------------------------|------------------|------------------------|--------------------|-------------------------|------------------|----------------------------|
| 2470             | 1664A SPIKING SOLN                                                                                | <a href="#">WP112783</a> | 04/22/2025       | 10/03/2025             | Jignesh Parikh     | WETCHEM_SCALE_8 (WCS-7) | None             | Iwona Zarych<br>04/22/2025 |
| <u>FROM</u>      | 1000.00000ml of E3917 + 4.00000gram of W2817 + 4.00000gram of W2871 = Final Quantity: 1000.000 ml |                          |                  |                        |                    |                         |                  |                            |

[illegible]



| <u>Recipe ID</u> | <u>NAME</u>                                                                                      | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u>            | <u>PipetteID</u> | <u>Supervised By</u>       |
|------------------|--------------------------------------------------------------------------------------------------|--------------------------|------------------|------------------------|--------------------|---------------------------|------------------|----------------------------|
| 3873             | Spiking solution for 9071B - SS                                                                  | <a href="#">WP112786</a> | 04/22/2025       | 10/03/2025             | Jignesh Parikh     | WETCHEM_SCALE_8 (WC SC-7) | None             | Iwona Zarych<br>04/22/2025 |
| <u>FROM</u>      | 1.00000gram of W3009 + 1.00000gram of W3082 + 1000.00000L of E3917 = Final Quantity: 1000.000 ml |                          |                  |                        |                    |                           |                  |                            |

| <u>Recipe ID</u> | <u>NAME</u>                                                             | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>          | <u>Supervised By</u>       |
|------------------|-------------------------------------------------------------------------|--------------------------|------------------|------------------------|--------------------|----------------|---------------------------|----------------------------|
| 1597             | 0.04 N H2SO4                                                            | <a href="#">WP112828</a> | 04/25/2025       | 10/25/2025             | Rubina Mughal      | None           | WETCHEM_PIPETTE_3<br>(WC) | Iwona Zarych<br>04/25/2025 |
| <u>FROM</u>      | 1.00000ml of M6041 + 999.00000ml of W3112 = Final Quantity: 1000.000 ml |                          |                  |                        |                    |                |                           |                            |



| <u>Recipe ID</u> | <u>NAME</u>                                                                                                           | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u>                   | <u>PipetteID</u>   | <u>Supervised By</u>           |
|------------------|-----------------------------------------------------------------------------------------------------------------------|--------------------------|------------------|------------------------|--------------------|----------------------------------|--------------------|--------------------------------|
| 607              | PYRIDINE-BARBITURIC ACID                                                                                              | <a href="#">WP112900</a> | 05/01/2025       | 08/18/2025             | Rubina Mughal      | WETCHEM_S<br>CALE_8 (WC<br>SC-7) | Glass<br>Pipette-A | Iwona Zarych<br><br>05/01/2025 |
| <u>FROM</u>      | 145.00000ml of W3112 + 15.00000gram of W3203 + 15.00000ml of M6151 + 75.00000ml of W3019 = Final Quantity: 250.000 ml |                          |                  |                        |                    |                                  |                    |                                |

| <u>Recipe ID</u>                                                                                                 | <u>NAME</u>       | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u>            | <u>PipetteID</u> | <u>Supervised By</u>       |
|------------------------------------------------------------------------------------------------------------------|-------------------|--------------------------|------------------|------------------------|--------------------|---------------------------|------------------|----------------------------|
| 160                                                                                                              | 0.5M ZINC ACETATE | <a href="#">WP113086</a> | 05/15/2025       | 08/18/2025             | Rubina Mughal      | WETCHEM_SCALE_8 (WC SC-7) | None             | Iwona Zarych<br>05/15/2025 |
| <b><u>FROM</u></b> 0.88900L of W3112 + 1.00000ml of M6151 + 110.00000gram of W2926 = Final Quantity: 1000.000 ml |                   |                          |                  |                        |                    |                           |                  |                            |



| <u>Recipe ID</u>                                                                            | <u>NAME</u>            | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u>          | <u>PipetteID</u> | <u>Supervised By</u>         |
|---------------------------------------------------------------------------------------------|------------------------|--------------------------|------------------|------------------------|--------------------|-------------------------|------------------|------------------------------|
| 2456                                                                                        | COD Stock std, 1000ppm | <a href="#">WP113231</a> | 05/28/2025       | 06/04/2025             | Iwona Zarych       | WETCHEM_SCALE_5 (WCS-5) | None             | Jignesh Parikh<br>05/28/2025 |
| <b><u>FROM</u></b> 0.08500gram of W2784 + 100.00000ml of W3112 = Final Quantity: 100.000 ml |                        |                          |                  |                        |                    |                         |                  |                              |

| <u>Recipe ID</u> | <u>NAME</u>                                                              | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u>                   | <u>PipetteID</u> | <u>Supervised By</u>             |
|------------------|--------------------------------------------------------------------------|--------------------------|------------------|------------------------|--------------------|----------------------------------|------------------|----------------------------------|
| 2457             | COD Stock std-SS, 1000ppm                                                | <a href="#">WP113232</a> | 05/28/2025       | 06/04/2025             | Iwona Zarych       | WETCHEM_S<br>CALE_5 (WC<br>SC-5) | None             | Jignesh Parikh<br><br>05/28/2025 |
| <u>FROM</u>      | 0.08500gram of W3169 + 100.00000ml of W3112 = Final Quantity: 100.000 ml |                          |                  |                        |                    |                                  |                  |                                  |

## Wet Chemistry STANDARD PREPARATION LOG

| <u>Recipe ID</u> | <u>NAME</u>                | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u> |
|------------------|----------------------------|--------------------------|------------------|------------------------|--------------------|----------------|------------------|----------------------|
| 139              | COD calibration std. 0 ppm | <a href="#">WP113233</a> | 05/28/2025       | 06/04/2025             | Iwona Zarych       | None           | None             | Jignesh Parikh       |
|                  |                            |                          |                  |                        |                    |                |                  | 05/28/2025           |

**FROM** 10.00000ml of W3112 = Final Quantity: 10.000 ml

| <u>Recipe ID</u> | <u>NAME</u>                 | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>  | <u>Supervised By</u> |
|------------------|-----------------------------|--------------------------|------------------|------------------------|--------------------|----------------|-------------------|----------------------|
| 138              | COD calibration std. 10 ppm | <a href="#">WP113234</a> | 05/28/2025       | 06/04/2025             | Iwona Zarych       | None           | WETCHEM_FIPETTE_3 | Jignesh Parikh       |
|                  |                             |                          |                  |                        |                    |                | (WC)              | 05/28/2025           |

**FROM** 9.90000ml of W3112 + 0.10000ml of WP113231 = Final Quantity: 10.000 ml



| <u>Recipe ID</u> | <u>NAME</u>                                                            | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>          | <u>Supervised By</u>         |
|------------------|------------------------------------------------------------------------|--------------------------|------------------|------------------------|--------------------|----------------|---------------------------|------------------------------|
| 137              | COD calibration std. 50 ppm                                            | <a href="#">WP113235</a> | 05/28/2025       | 06/04/2025             | Iwona Zarych       | None           | WETCHEM_PIPETTE_3<br>(WC) | Jignesh Parikh<br>05/28/2025 |
| <u>FROM</u>      | 9.50000ml of W3112 + 0.50000ml of WP113231 = Final Quantity: 10.000 ml |                          |                  |                        |                    |                |                           |                              |

| <u>Recipe ID</u> | <u>NAME</u>                                                            | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>          | <u>Supervised By</u>         |
|------------------|------------------------------------------------------------------------|--------------------------|------------------|------------------------|--------------------|----------------|---------------------------|------------------------------|
| 4161             | COD calibration std. 75 ppm                                            | <a href="#">WP113236</a> | 05/28/2025       | 06/04/2025             | Iwona Zarych       | None           | WETCHEM_PIPETTE_3<br>(WC) | Jignesh Parikh<br>05/28/2025 |
| <u>FROM</u>      | 9.25000ml of W3112 + 0.75000ml of WP113231 = Final Quantity: 10.000 ml |                          |                  |                        |                    |                |                           |                              |





| <u>Recipe ID</u> | <u>NAME</u>                                                            | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>          | <u>Supervised By</u>         |
|------------------|------------------------------------------------------------------------|--------------------------|------------------|------------------------|--------------------|----------------|---------------------------|------------------------------|
| 136              | COD calibration std. 100 ppm                                           | <a href="#">WP113237</a> | 05/28/2025       | 06/04/2025             | Iwona Zarych       | None           | WETCHEM_PIPETTE_3<br>(WC) | Jignesh Parikh<br>05/28/2025 |
| <u>FROM</u>      | 9.00000ml of W3112 + 1.00000ml of WP113231 = Final Quantity: 10.000 ml |                          |                  |                        |                    |                |                           |                              |

| <u>Recipe ID</u> | <u>NAME</u>                                                            | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>          | <u>Supervised By</u>         |
|------------------|------------------------------------------------------------------------|--------------------------|------------------|------------------------|--------------------|----------------|---------------------------|------------------------------|
| 135              | COD calibration std. 150 ppm                                           | <a href="#">WP113238</a> | 05/28/2025       | 06/04/2025             | Iwona Zarych       | None           | WETCHEM_PIPETTE_3<br>(WC) | Jignesh Parikh<br>05/28/2025 |
| <u>FROM</u>      | 8.50000ml of W3112 + 1.50000ml of WP113231 = Final Quantity: 10.000 ml |                          |                  |                        |                    |                |                           |                              |

## Wet Chemistry STANDARD PREPARATION LOG

| <u>Recipe ID</u> | <u>NAME</u>            | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>              | <u>Supervised By</u>             |
|------------------|------------------------|--------------------------|------------------|------------------------|--------------------|----------------|-------------------------------|----------------------------------|
| 2459             | COD ICV-LCS std, 50ppm | <a href="#">WP113240</a> | 05/28/2025       | 06/04/2025             | Iwona Zarych       | None           | WETCHEM_F<br>IPETTE_3<br>(WC) | Jignesh Parikh<br><br>05/28/2025 |

**FROM** 9.50000ml of W3112 + 0.50000ml of WP113232 = Final Quantity: 10.000 ml

| <u>Recipe ID</u> | <u>NAME</u>                                | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u>                   | <u>PipetteID</u> | <u>Supervised By</u>           |
|------------------|--------------------------------------------|--------------------------|------------------|------------------------|--------------------|----------------------------------|------------------|--------------------------------|
| 11               | Sodium hydroxide absorbing solution 0.25 N | <a href="#">WP113836</a> | 07/08/2025       | 12/31/2025             | Rubina Mughal      | WETCHEM_S<br>CALE_8 (WC<br>SC-7) | None             | Iwona Zarych<br><br>07/08/2025 |

**FROM** 21.00000L of W3112 + 210.00000gram of W3113 = Final Quantity: 21.000 L



| <u>Recipe ID</u> | <u>NAME</u>                                                               | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>          | <u>Supervised By</u>           |
|------------------|---------------------------------------------------------------------------|--------------------------|------------------|------------------------|--------------------|----------------|---------------------------|--------------------------------|
| 3371             | Cyanide LCS Spike Solution,<br>5PPM                                       | <a href="#">WP113838</a> | 07/08/2025       | 12/24/2025             | Rubina Mughal      | None           | WETCHEM_PIPETTE_3<br>(WC) | Iwona Zarych<br><br>07/08/2025 |
| <u>FROM</u>      | 1.00000ml of W3224 + 199.00000ml of WP113836 = Final Quantity: 200.000 ml |                          |                  |                        |                    |                |                           |                                |

| <u>Recipe ID</u>                                                                           | <u>NAME</u>                          | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u>          | <u>PipetteID</u> | <u>Supervised By</u>       |
|--------------------------------------------------------------------------------------------|--------------------------------------|--------------------------|------------------|------------------------|--------------------|-------------------------|------------------|----------------------------|
| 740                                                                                        | sodium nitroferricyanide for ammonia | <a href="#">WP113852</a> | 07/09/2025       | 08/09/2025             | Rubina Mughal      | WETCHEM_SCALE_5 (WCS-5) | None             | Iwona Zarych<br>07/09/2025 |
| <b><u>FROM</u></b> 0.05000gram of W2666 + 99.95000ml of W3112 = Final Quantity: 100.000 ml |                                      |                          |                  |                        |                    |                         |                  |                            |

## Wet Chemistry STANDARD PREPARATION LOG

| <u>Recipe ID</u> | <u>NAME</u> | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u>            | <u>PipetteID</u> | <u>Supervised By</u>       |
|------------------|-------------|--------------------------|------------------|------------------------|--------------------|---------------------------|------------------|----------------------------|
| 1796             | NaOH, 0.1N  | <a href="#">WP113885</a> | 07/10/2025       | 12/31/2025             | Rubina Mughal      | WETCHEM_SCALE_8 (WC SC-7) | None             | Iwona Zarych<br>07/10/2025 |

**FROM** 4.00000gram of W3113 + 996.00000ml of W3112 = Final Quantity: 1000.000 ml

| <u>Recipe ID</u> | <u>NAME</u>   | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u>            | <u>PipetteID</u> | <u>Supervised By</u>       |
|------------------|---------------|--------------------------|------------------|------------------------|--------------------|---------------------------|------------------|----------------------------|
| 1494             | BORATE BUFFER | <a href="#">WP113886</a> | 07/10/2025       | 12/31/2025             | Rubina Mughal      | WETCHEM_SCALE_8 (WC SC-7) | None             | Iwona Zarych<br>07/10/2025 |

**FROM** 0.90250L of W3112 + 9.50000gram of W3201 + 88.00000ml of WP113885 = Final Quantity: 1.000 L



| <u>Recipe ID</u>                                                                               | <u>NAME</u>       | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u>                   | <u>PipetteID</u> | <u>Supervised By</u>           |
|------------------------------------------------------------------------------------------------|-------------------|--------------------------|------------------|------------------------|--------------------|----------------------------------|------------------|--------------------------------|
| 1471                                                                                           | NaOH Solution, 6N | <a href="#">WP113887</a> | 07/10/2025       | 12/31/2025             | Rubina Mughal      | WETCHEM_S<br>CALE_8 (WC<br>SC-7) | None             | Iwona Zarych<br><br>07/10/2025 |
| <b><u>FROM</u></b> 240.00000gram of W3113 + 760.00000ml of W3112 = Final Quantity: 1000.000 ml |                   |                          |                  |                        |                    |                                  |                  |                                |

| <u>Recipe ID</u>                                                                            | <u>NAME</u>                     | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>          | <u>Supervised By</u>       |
|---------------------------------------------------------------------------------------------|---------------------------------|--------------------------|------------------|------------------------|--------------------|----------------|---------------------------|----------------------------|
| 1322                                                                                        | Ammonia Intermediate Std, 50PPM | <a href="#">WP113888</a> | 07/10/2025       | 08/10/2025             | Rubina Mughal      | None           | WETCHEM_PIPETTE_3<br>(WC) | Iwona Zarych<br>07/10/2025 |
| <b><u>FROM</u></b> 95.00000ml of W3112 + 5.00000ml of WP112611 = Final Quantity: 100.000 ml |                                 |                          |                  |                        |                    |                |                           |                            |



| <u>Recipe ID</u>                                                                            | <u>NAME</u>                                      | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>          | <u>Supervised By</u>       |
|---------------------------------------------------------------------------------------------|--------------------------------------------------|--------------------------|------------------|------------------------|--------------------|----------------|---------------------------|----------------------------|
| 1639                                                                                        | Ammonia Intermediate<br>Std-Second source, 50PPM | <a href="#">WP113889</a> | 07/10/2025       | 08/10/2025             | Rubina Mughal      | None           | WETCHEM_PIPETTE_3<br>(WC) | Iwona Zarych<br>07/10/2025 |
| <b><u>FROM</u></b> 95.00000ml of W3112 + 5.00000ml of WP112612 = Final Quantity: 100.000 ml |                                                  |                          |                  |                        |                    |                |                           |                            |

| <u>Recipe ID</u>                                                                                   | <u>NAME</u>                            | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>  | <u>Supervised By</u> |
|----------------------------------------------------------------------------------------------------|----------------------------------------|--------------------------|------------------|------------------------|--------------------|----------------|-------------------|----------------------|
| 3456                                                                                               | Cyanide Intermediate Working Std, 5PPM | <a href="#">WP113904</a> | 07/11/2025       | 07/12/2025             | Rubina Mughal      | None           | WETCHEM_PIPETTE_3 | Iwona Zarych         |
| <b>FROM</b> 0.25000ml of W3214 + 49.75000ml of WP113836 = Final Quantity: 50.000 ml<br><div></div> |                                        |                          |                  |                        |                    |                |                   |                      |



| <u>Recipe ID</u>                                                                                          | <u>NAME</u>                  | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>  | <u>Supervised By</u> |
|-----------------------------------------------------------------------------------------------------------|------------------------------|--------------------------|------------------|------------------------|--------------------|----------------|-------------------|----------------------|
| 4                                                                                                         | Calibration standard 500 ppb | <a href="#">WP113905</a> | 07/11/2025       | 07/12/2025             | Rubina Mughal      | None           | WETCHEM_PIPETTE_3 | Iwona Zarych         |
| <p>(WC)</p> <p><b>FROM</b> 45.00000ml of WP113836 + 5.00000ml of WP113904 = Final Quantity: 50.000 ml</p> |                              |                          |                  |                        |                    |                |                   |                      |

| <u>Recipe ID</u>                                                                                          | <u>NAME</u>                         | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>  | <u>Supervised By</u> |
|-----------------------------------------------------------------------------------------------------------|-------------------------------------|--------------------------|------------------|------------------------|--------------------|----------------|-------------------|----------------------|
| 3761                                                                                                      | Calibration-CCV CN Standard 250 ppb | <a href="#">WP113906</a> | 07/11/2025       | 07/12/2025             | Rubina Mughal      | None           | WETCHEM_PIPETTE_3 | Iwona Zarych         |
| <p><b><u>FROM</u></b>      2.50000ml of WP113904 + 47.50000ml of WP113836 = Final Quantity: 50.000 ml</p> |                                     |                          |                  |                        |                    |                |                   |                      |



| <u>Recipe ID</u>                                                                                      | <u>NAME</u>                  | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>  | <u>Supervised By</u> |
|-------------------------------------------------------------------------------------------------------|------------------------------|--------------------------|------------------|------------------------|--------------------|----------------|-------------------|----------------------|
| 6                                                                                                     | Calibration Standard 100 ppb | <a href="#">WP113907</a> | 07/11/2025       | 07/12/2025             | Rubina Mughal      | None           | WETCHEM_PIPETTE_3 | Iwona Zarych         |
| <b>FROM</b> 1.00000ml of WP113904 + 49.00000ml of WP113836 = Final Quantity: 50.000 ml<br><div></div> |                              |                          |                  |                        |                    |                |                   |                      |

| <u>Recipe ID</u>                                                                              | <u>NAME</u>                 | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>          | <u>Supervised By</u>       |
|-----------------------------------------------------------------------------------------------|-----------------------------|--------------------------|------------------|------------------------|--------------------|----------------|---------------------------|----------------------------|
| 7                                                                                             | Calibration Standard 50 ppb | <a href="#">WP113908</a> | 07/11/2025       | 07/12/2025             | Rubina Mughal      | None           | WETCHEM_PIPETTE_3<br>(WC) | Iwona Zarych<br>07/11/2025 |
| <b><u>FROM</u></b> 0.50000ml of WP113904 + 49.50000ml of WP113836 = Final Quantity: 50.000 ml |                             |                          |                  |                        |                    |                |                           |                            |





| <u>Recipe ID</u>                                                                                      | <u>NAME</u>                 | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>  | <u>Supervised By</u> |
|-------------------------------------------------------------------------------------------------------|-----------------------------|--------------------------|------------------|------------------------|--------------------|----------------|-------------------|----------------------|
| 8                                                                                                     | Calibration Standard 10 ppb | <a href="#">WP113909</a> | 07/11/2025       | 07/12/2025             | Rubina Mughal      | None           | WETCHEM_PIPETTE_3 | Iwona Zarych         |
| <b>FROM</b> 1.00000ml of WP113905 + 49.00000ml of WP113836 = Final Quantity: 50.000 ml<br><div></div> |                             |                          |                  |                        |                    |                |                   |                      |

| <u>Recipe ID</u>                                                                                      | <u>NAME</u>                | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>  | <u>Supervised By</u> |
|-------------------------------------------------------------------------------------------------------|----------------------------|--------------------------|------------------|------------------------|--------------------|----------------|-------------------|----------------------|
| 9                                                                                                     | Calibration Standard 5 ppb | <a href="#">WP113910</a> | 07/11/2025       | 07/12/2025             | Rubina Mughal      | None           | WETCHEM_PIPETTE_3 | Iwona Zarych         |
| <b>FROM</b> 0.50000ml of WP113905 + 49.50000ml of WP113836 = Final Quantity: 50.000 ml<br><div></div> |                            |                          |                  |                        |                    |                |                   |                      |

## Wet Chemistry STANDARD PREPARATION LOG

| <u>Recipe ID</u> | <u>NAME</u>              | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u> |
|------------------|--------------------------|--------------------------|------------------|------------------------|--------------------|----------------|------------------|----------------------|
| 167              | 0 ppb CN calibration std | <a href="#">WP113911</a> | 07/11/2025       | 07/12/2025             | Rubina Mughal      | None           | None             | Iwona Zarych         |
|                  |                          |                          |                  |                        |                    |                |                  | 07/11/2025           |

**FROM** 50.00000ml of WP113836 = Final Quantity: 50.000 ml

| <u>Recipe ID</u> | <u>NAME</u>          | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>      | <u>Supervised By</u> |
|------------------|----------------------|--------------------------|------------------|------------------------|--------------------|----------------|-----------------------|----------------------|
| 2168             | RCN ICV STD, 100 PPB | <a href="#">WP113912</a> | 07/11/2025       | 07/12/2025             | Rubina Mughal      | None           | WETCHEM_F<br>IPETTE_3 | Iwona Zarych         |
|                  |                      |                          |                  |                        |                    |                | (WC)                  | 07/11/2025           |

**FROM** 1.00000ml of WP113838 + 49.00000ml of WP113836 = Final Quantity: 50.000 ml



| <u>Recipe ID</u> | <u>NAME</u>                                                            | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u>          | <u>PipetteID</u> | <u>Supervised By</u>       |
|------------------|------------------------------------------------------------------------|--------------------------|------------------|------------------------|--------------------|-------------------------|------------------|----------------------------|
| 1582             | Chloramine T solution, 0.014M                                          | <a href="#">WP113913</a> | 07/11/2025       | 07/12/2025             | Rubina Mughal      | WETCHEM_SCALE_5 (WCS-5) | Glass Pipette-A  | Iwona Zarych<br>07/11/2025 |
| <u>FROM</u>      | 0.08000gram of W3139 + 20.00000ml of W3112 = Final Quantity: 20.000 ml |                          |                  |                        |                    |                         |                  |                            |

| <u>Recipe ID</u>                                                                                   | <u>NAME</u>                      | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>  | <u>Supervised By</u> |
|----------------------------------------------------------------------------------------------------|----------------------------------|--------------------------|------------------|------------------------|--------------------|----------------|-------------------|----------------------|
| 275                                                                                                | Ammonia Calibration Std. (2 ppm) | <a href="#">WP113926</a> | 07/14/2025       | 07/15/2025             | Rubina Mughal      | None           | WETCHEM_PIPETTE_3 | Iwona Zarych         |
| <b>FROM</b> 48.00000ml of W3112 + 2.00000ml of WP113888 = Final Quantity: 50.000 ml<br><div></div> |                                  |                          |                  |                        |                    |                |                   |                      |



| <u>Recipe ID</u>                                                                                   | <u>NAME</u>              | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>  | <u>Supervised By</u> |
|----------------------------------------------------------------------------------------------------|--------------------------|--------------------------|------------------|------------------------|--------------------|----------------|-------------------|----------------------|
| 285                                                                                                | Ammonia CCV Std. (1 ppm) | <a href="#">WP113927</a> | 07/14/2025       | 07/15/2025             | Rubina Mughal      | None           | WETCHEM_PIPETTE_3 | Iwona Zarych         |
| <b>FROM</b> 49.00000ml of W3112 + 1.00000ml of WP113888 = Final Quantity: 50.000 ml<br><div></div> |                          |                          |                  |                        |                    |                |                   |                      |

| <u>Recipe ID</u>                                                                           | <u>NAME</u>              | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>          | <u>Supervised By</u>       |
|--------------------------------------------------------------------------------------------|--------------------------|--------------------------|------------------|------------------------|--------------------|----------------|---------------------------|----------------------------|
| 286                                                                                        | Ammonia ICV Std. (1 ppm) | <a href="#">WP113928</a> | 07/14/2025       | 07/15/2025             | Rubina Mughal      | None           | WETCHEM_PIPETTE_3<br>(WC) | Iwona Zarych<br>07/15/2025 |
| <b><u>FROM</u></b> 49.00000ml of W3112 + 1.00000ml of WP113889 = Final Quantity: 50.000 ml |                          |                          |                  |                        |                    |                |                           |                            |



| <u>Recipe ID</u>                                                                                                  | <u>NAME</u>                | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u>            | <u>PipetteID</u> | <u>Supervised By</u>       |
|-------------------------------------------------------------------------------------------------------------------|----------------------------|--------------------------|------------------|------------------------|--------------------|---------------------------|------------------|----------------------------|
| 290                                                                                                               | Phenol reagent for Ammonia | <a href="#">WP113929</a> | 07/14/2025       | 12/31/2025             | Rubina Mughal      | WETCHEM_SCALE_8 (WC SC-7) | None             | Iwona Zarych<br>07/15/2025 |
| <b><u>FROM</u></b> 3.20000gram of W3113 + 8.30000gram of W2663 + 88.80000ml of W3112 = Final Quantity: 100.000 ml |                            |                          |                  |                        |                    |                           |                  |                            |

| <u>Recipe ID</u>                                                                            | <u>NAME</u>            | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u>          | <u>PipetteID</u> | <u>Supervised By</u> |
|---------------------------------------------------------------------------------------------|------------------------|--------------------------|------------------|------------------------|--------------------|-------------------------|------------------|----------------------|
| 2456                                                                                        | COD Stock std, 1000ppm | <a href="#">WP113938</a> | 07/15/2025       | 07/22/2025             | Iwona Zarych       | WETCHEM_SCALE_7 (WCS-6) | None             | Jignesh Parikh       |
| <b><u>FROM</u></b> 0.08500gram of W3219 + 100.00000ml of W3112 = Final Quantity: 100.000 ml |                        |                          |                  |                        |                    |                         |                  |                      |



| <u>Recipe ID</u>                                                                            | <u>NAME</u>               | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u>          | <u>PipetteID</u> | <u>Supervised By</u>         |
|---------------------------------------------------------------------------------------------|---------------------------|--------------------------|------------------|------------------------|--------------------|-------------------------|------------------|------------------------------|
| 2457                                                                                        | COD Stock std-SS, 1000ppm | <a href="#">WP113939</a> | 07/15/2025       | 07/22/2025             | Iwona Zarych       | WETCHEM_SCALE_7 (WCS-6) | None             | Jignesh Parikh<br>07/15/2025 |
| <b><u>FROM</u></b> 0.08500gram of W3169 + 100.00000ml of W3112 = Final Quantity: 100.000 ml |                           |                          |                  |                        |                    |                         |                  |                              |

| <u>Recipe ID</u>                                                                          | <u>NAME</u>        | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>          | <u>Supervised By</u>         |
|-------------------------------------------------------------------------------------------|--------------------|--------------------------|------------------|------------------------|--------------------|----------------|---------------------------|------------------------------|
| 2458                                                                                      | COD CCV std, 50ppm | <a href="#">WP113940</a> | 07/15/2025       | 07/22/2025             | Iwona Zarych       | None           | WETCHEM_PIPETTE_3<br>(WC) | Jignesh Parikh<br>07/15/2025 |
| <b><u>FROM</u></b> 9.50000ml of W3112 + 0.50000ml of WP113938 = Final Quantity: 10.000 ml |                    |                          |                  |                        |                    |                |                           |                              |



| <u>Recipe ID</u> | <u>NAME</u>                                                            | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>          | <u>Supervised By</u>         |
|------------------|------------------------------------------------------------------------|--------------------------|------------------|------------------------|--------------------|----------------|---------------------------|------------------------------|
| 2459             | COD ICV-LCS std, 50ppm                                                 | <a href="#">WP113941</a> | 07/15/2025       | 07/22/2025             | Iwona Zarych       | None           | WETCHEM_PIPETTE_3<br>(WC) | Jignesh Parikh<br>07/15/2025 |
| <u>FROM</u>      | 9.50000ml of W3112 + 0.50000ml of WP113939 = Final Quantity: 10.000 ml |                          |                  |                        |                    |                |                           |                              |

| <u>Recipe ID</u>                                                                          | <u>NAME</u> | <u>NO.</u>               | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u>          | <u>Supervised By</u>         |
|-------------------------------------------------------------------------------------------|-------------|--------------------------|------------------|------------------------|--------------------|----------------|---------------------------|------------------------------|
| 4162                                                                                      | RL CHECK    | <a href="#">WP113942</a> | 07/15/2025       | 07/22/2025             | Iwona Zarych       | None           | WETCHEM_PIPETTE_3<br>(WC) | Jignesh Parikh<br>07/15/2025 |
| <b><u>FROM</u></b> 9.90000ml of W3112 + 0.10000ml of WP113938 = Final Quantity: 10.000 ml |             |                          |                  |                        |                    |                |                           |                              |

## CHEMICAL RECEIPT LOG BOOK

| Supplier                    | ItemCode / ItemName                                    | Lot #  | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|--------------------------------------------------------|--------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1 | 313201 | 12/04/2025      | 01/03/2024 / Rajesh     | 07/20/2023 / Rajesh         | E3551          |

| Supplier         | ItemCode / ItemName                        | Lot #      | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|------------------|--------------------------------------------|------------|-----------------|-------------------------|-----------------------------|----------------|
| Seidler Chemical | BA-9254-03 / Acetone, Ultra Resi (cs/4x4L) | 24H2762008 | 10/03/2025      | 04/03/2025 / Rajesh     | 03/31/2025 / Rajesh         | E3917          |

| Supplier         | ItemCode / ItemName                                     | Lot #      | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|------------------|---------------------------------------------------------|------------|-----------------|-------------------------|-----------------------------|----------------|
| Seidler Chemical | BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L) | 23D2462010 | 03/20/2028      | 08/16/2024 / mohan      | 08/16/2024 / mohan          | M6041          |

| Supplier                    | ItemCode / ItemName                          | Lot #   | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|----------------------------------------------|---------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | 140440 / TEST PAPERS,PH,0-2.5,.2SENSI, 100PK | 80A0441 | 02/29/2028      | 09/03/2024 / jignesh    | 08/19/2024 / Jaswal         | M6069          |

| Supplier         | ItemCode / ItemName                                         | Lot #      | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|------------------|-------------------------------------------------------------|------------|-----------------|-------------------------|-----------------------------|----------------|
| Seidler Chemical | BA-9530-33 / Hydrochloric Acid, Instra-Analyzed (cs/6x2.5L) | 22G2862015 | 08/18/2025      | 02/18/2025 / Sagar      | 01/15/2025 / Sagar          | M6151          |

| Supplier                    | ItemCode / ItemName          | Lot #   | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|------------------------------|---------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | P1060-10 / PHENOL, ACS, 500G | 2HD0179 | 01/27/2030      | 01/27/2020 / apatel     | 01/27/2020 / apatel         | W2663          |



## CHEMICAL RECEIPT LOG BOOK

| Supplier                    | ItemCode / ItemName                   | Lot #   | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|---------------------------------------|---------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | 87683 / Sodium Nitroferricyanide 250g | W12F013 | 02/10/2030      | 02/10/2020 / apatel     | 02/10/2020 / apatel         | W2666          |

| Supplier                    | ItemCode / ItemName                                         | Lot #      | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|-------------------------------------------------------------|------------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | J3818-5 / SODIUM PHOSPHATE, MONOBAS/HYD, CRYST, ACS, 2.5 KG | 0000225799 | 12/03/2025      | 04/05/2021 / Alexander  | 02/10/2020 / apatel         | W2668          |

| Supplier                    | ItemCode / ItemName                        | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|--------------------------------------------|-------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | EMD-FX0410-5 / FORMALDEHYDE SOLUTION 450ML | 60045 | 06/30/2025      | 08/19/2024 / lwona      | 06/22/2020 / apatel         | W2725          |

| Supplier                    | ItemCode / ItemName                              | Lot #  | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|--------------------------------------------------|--------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | P243-500 / Potassium Hydrogen Phthalate, 500 gms | 201089 | 06/30/2025      | 12/23/2020 / apatel     | 12/16/2020 / apatel         | W2784          |

| Supplier                    | ItemCode / ItemName               | Lot #   | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|-----------------------------------|---------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | A12244 / Stearic acid, 98%, 100 g | U20E006 | 04/02/2026      | 04/02/2021 / apatel     | 04/02/2021 / apatel         | W2817          |

| Supplier         | ItemCode / ItemName         | Lot #      | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|------------------|-----------------------------|------------|-----------------|-------------------------|-----------------------------|----------------|
| Seidler Chemical | H223-57 / Hexadecane, 99.0% | 0000266903 | 05/04/2027      | 09/07/2021 / apatel     | 08/26/2021 / apatel         | W2871          |

## CHEMICAL RECEIPT LOG BOOK

| Supplier                    | ItemCode / ItemName                         | Lot #  | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|---------------------------------------------|--------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | J4296-1 / ZINC ACETATE,DIHYD,CRYS,AC S,500G | 383058 | 07/05/2027      | 07/05/2022 / ketankumar | 07/05/2022 / ketankumar     | W2926          |

| Supplier         | ItemCode / ItemName         | Lot #    | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|------------------|-----------------------------|----------|-----------------|-------------------------|-----------------------------|----------------|
| Seidler Chemical | H223-57 / Hexadecane, 99.0% | SHBP8192 | 02/27/2028      | 02/27/2023 / lwona      | 02/27/2023 / lwona          | W3009          |

| Supplier      | ItemCode / ItemName     | Lot #    | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---------------|-------------------------|----------|-----------------|-------------------------|-----------------------------|----------------|
| SIGMA ALDRICH | 270970-1L / Pyridine 1L | SHBQ2113 | 04/03/2028      | 04/03/2023 / lwona      | 04/03/2023 / lwona          | W3019          |

| Supplier                    | ItemCode / ItemName               | Lot #   | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|-----------------------------------|---------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | A12244 / Stearic acid, 98%, 100 g | U23E020 | 02/26/2029      | 02/26/2024 / lwona      | 02/26/2024 / lwona          | W3082          |

| Supplier                    | ItemCode / ItemName                     | Lot #    | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|-----------------------------------------|----------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | 566002 / BUFFER PH 7.00 GREEN 1PINT PK6 | 44001f99 | 12/31/2025      | 04/03/2024 / jignesh    | 04/02/2024 / jignesh        | W3093          |

| Supplier                    | ItemCode / ItemName                           | Lot #   | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|-----------------------------------------------|---------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | AL69870-8 / SODIUM THIOSULFATE,0.025N,4LIT RE | 4403S13 | 09/30/2025      | 04/22/2024 / lwona      | 04/22/2024 / lwona          | W3105          |

## CHEMICAL RECEIPT LOG BOOK

| Supplier         | ItemCode / ItemName | Lot #               | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|------------------|---------------------|---------------------|-----------------|-------------------------|-----------------------------|----------------|
| Seidler Chemical | DIW / DI Water      | Daily Lab-Certified | 07/03/2029      | 07/03/2024 / lwona      | 07/03/2024 / lwona          | W3112          |

| Supplier                    | ItemCode / ItemName                        | Lot #      | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|--------------------------------------------|------------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | PC19510-7 / Sodium Hydroxide Pellets 12 Kg | 23B1556310 | 12/31/2025      | 07/08/2024 / lwona      | 07/08/2024 / lwona          | W3113          |

| Supplier                  | ItemCode / ItemName                             | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---------------------------|-------------------------------------------------|-------|-----------------|-------------------------|-----------------------------|----------------|
| Environmental Express LTD | B1010 / COD Digestion Vials Low Level 0-150Mg/L | 13821 | 10/31/2027      | 06/16/2025 / lwona      | 07/25/2024 / lwona          | W3129          |

| Supplier                    | ItemCode / ItemName                             | Lot #   | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|-------------------------------------------------|---------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | PC05050-1 / EDTA, disodium salt, dihydrate 1 lb | 2ND0156 | 07/10/2026      | 07/26/2024 / lwona      | 07/26/2024 / lwona          | W3132          |

| Supplier                    | ItemCode / ItemName                         | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|---------------------------------------------|-------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | 140476 / Test Paper,PH Short Range 9.0/10.0 | L23   | 08/22/2029      | 08/22/2024 / lwona      | 08/22/2024 / lwona          | W3133          |

| Supplier                    | ItemCode / ItemName                 | Lot #    | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|-------------------------------------|----------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | JTE494-6 / CHLORAMINE-T BAKER 250GM | 10239484 | 09/09/2029      | 09/09/2024 / lwona      | 09/09/2024 / lwona          | W3139          |

## CHEMICAL RECEIPT LOG BOOK

| Supplier                    | ItemCode / ItemName             | Lot #   | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|---------------------------------|---------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | AL70850-8 / Starch Solution, 4L | 4408P62 | 08/31/2026      | 10/16/2024 / Iwona      | 10/16/2024 / Iwona          | W3149          |

| Supplier                    | ItemCode / ItemName                           | Lot #  | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|-----------------------------------------------|--------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | 140730 / TEST PAPER,POT.IOD-STRCH,P K100,CS12 | 14-860 | 12/02/2029      | 12/02/2024 / Iwona      | 12/02/2024 / Iwona          | W3155          |

| Supplier                    | ItemCode / ItemName                      | Lot #   | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|------------------------------------------|---------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | AL13850-1 / Buffer Solution, PH2 (500ml) | 2411E26 | 10/31/2026      | 12/09/2024 / Iwona      | 12/09/2024 / Iwona          | W3161          |

| Supplier                    | ItemCode / ItemName                              | Lot #      | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|--------------------------------------------------|------------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | P243-500 / Potassium Hydrogen Phthalate, 500 gms | 24H0956262 | 04/28/2026      | 01/03/2025 / Iwona      | 01/03/2025 / Iwona          | W3169          |

| Supplier                    | ItemCode / ItemName                  | Lot #   | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|--------------------------------------|---------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | J9416-1 / Sodium Hypochlorite 500 ml | 2501J28 | 07/31/2025      | 01/24/2025 / Iwona      | 01/24/2025 / Iwona          | W3174          |

| Supplier                    | ItemCode / ItemName              | Lot #   | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|----------------------------------|---------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | AL14055-3 / PH 4 BUFFER SOLUTION | 2411A93 | 10/30/2026      | 04/01/2025 / JIGNESH    | 01/27/2025 / jignesh        | W3178          |

## CHEMICAL RECEIPT LOG BOOK

| Supplier         | ItemCode / ItemName                      | Lot #      | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|------------------|------------------------------------------|------------|-----------------|-------------------------|-----------------------------|----------------|
| Seidler Chemical | BA-3382-05 / Sand, Purified (cs/4x2.5kg) | 25A2756718 | 12/31/2028      | 03/10/2025 / Iwona      | 03/10/2025 / Iwona          | W3186          |

| Supplier                    | ItemCode / ItemName                     | Lot #   | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|-----------------------------------------|---------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | 1601-1 / PH 10.01 BUFFER,COLOR CD 475ML | 2410F80 | 03/31/2026      | 04/01/2025 / JIGNESH    | 03/13/2025 / jignesh        | W3191          |

| Supplier                    | ItemCode / ItemName                    | Lot #      | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|----------------------------------------|------------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | J0660-1 / AMMONIUM CHLORIDE, ACS, 500G | 24L0356561 | 08/31/2027      | 03/19/2025 / Iwona      | 03/19/2025 / Iwona          | W3195          |

| Supplier                    | ItemCode / ItemName                    | Lot #    | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|----------------------------------------|----------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | J0660-1 / AMMONIUM CHLORIDE, ACS, 500G | MKCV1009 | 09/30/2026      | 03/19/2025 / Iwona      | 03/19/2025 / Iwona          | W3196          |

| Supplier               | ItemCode / ItemName       | Lot #   | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|------------------------|---------------------------|---------|-----------------|-------------------------|-----------------------------|----------------|
| RICCA CHEMICAL COMPANY | 1615-16 / pH 12.00 Buffer | 2504F20 | 09/30/2026      | 04/11/2025 / Iwona      | 04/11/2025 / Iwona          | W3200          |

| Supplier                    | ItemCode / ItemName              | Lot #    | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|----------------------------------|----------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | J3568-1 / Sodium Borate, 500 gms | BCCL9613 | 05/31/2029      | 04/16/2025 / Iwona      | 04/16/2025 / Iwona          | W3201          |

## CHEMICAL RECEIPT LOG BOOK

| Supplier                    | ItemCode / ItemName                    | Lot #     | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|----------------------------------------|-----------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | EM-BX0035-3 / Barbituric Acid, 100 gms | WXBF3271V | 05/16/2029      | 04/21/2025 / lwona      | 04/21/2025 / lwona          | W3203          |

| Supplier         | ItemCode / ItemName                       | Lot #      | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|------------------|-------------------------------------------|------------|-----------------|-------------------------|-----------------------------|----------------|
| Seidler Chemical | BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L) | 25c0362005 | 04/30/2026      | 04/22/2025 / jignesh    | 04/18/2025 / jignesh        | W3204          |

| Supplier                    | ItemCode / ItemName                  | Lot #      | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|--------------------------------------|------------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | AL35830-4 / IODINE SOLUTION .025N 1L | MK25A21527 | 01/20/2029      | 05/21/2025 / lwona      | 05/21/2025 / lwona          | W3213          |

| Supplier                    | ItemCode / ItemName                | Lot #   | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|------------------------------------|---------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | RC2543-4 / CYANIDE STD 1000PPM 4OZ | 1505H73 | 11/30/2025      | 05/21/2025 / lwona      | 05/21/2025 / lwona          | W3214          |

| Supplier                    | ItemCode / ItemName                     | Lot #   | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|-----------------------------------------|---------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | AL14455-3 / buffer solution pH 7 yellow | 2504D34 | 03/31/2027      | 07/02/2025 / jignesh    | 06/26/2025 / lwona          | W3217          |

| Supplier                    | ItemCode / ItemName                              | Lot #      | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|--------------------------------------------------|------------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | P243-500 / Potassium Hydrogen Phthalate, 500 gms | 2025040493 | 06/30/2030      | 06/26/2025 / lwona      | 06/26/2025 / lwona          | W3219          |

### **CHEMICAL RECEIPT LOG BOOK**

| Supplier                       | ItemCode / ItemName                                        | Lot #    | Expiration Date | Date Opened /<br>Opened By | Received Date /<br>Received By | Chemtech<br>Lot # |
|--------------------------------|------------------------------------------------------------|----------|-----------------|----------------------------|--------------------------------|-------------------|
| PCI Scientific<br>Supply, Inc. | LC135457 / Cyanide<br>Standard, 1000 PPM,<br>Second Source | 45060288 | 12/24/2025      | 07/07/2025 /<br>lwona      | 07/07/2025 /<br>lwona          | W3224             |



# Certificate Of Analysis

|                   |                                     |                  |         |
|-------------------|-------------------------------------|------------------|---------|
| Item Number       | P1060                               | Lot Number       | 2HD0179 |
| Item              | Phenol, Loose Crystal, Reagent, ACS |                  |         |
| CAS Number        | 108-95-2                            |                  |         |
| Molecular Formula | C <sub>6</sub> H <sub>6</sub> O     | Molecular Weight | 94.11   |

| Test                                     | Specification |        | Result      |
|------------------------------------------|---------------|--------|-------------|
|                                          | min           | max    |             |
| ASSAY (C <sub>6</sub> H <sub>5</sub> OH) | 99.0 %        |        | 100.02 %    |
| FREEZING POINT (DRY)                     | 40.5 C        |        | 40.5°C      |
| CLARITY OF SOLUTION                      | TO PASS TEST  |        | PASSES TEST |
| RESIDUE AFTER EVAPORATION                |               | 0.05 % | <0.05 %     |
| WATER                                    |               | 0.5 %  | 0.0087 %    |
| DATE OF MANUFACTURE                      |               |        | 06-MAR-2018 |

Spectrum Chemical Mfg Corp  
755 Jersey Avenue  
New Brunswick 08901 NJ



Certificate Of Analysis Results Certified by

Ibad Tirmizi  
Director of Quality  
Spectrum Chemical Mfg. Corp.

All pharmaceutical ingredients are tested using current edition of applicable pharmacopeia.

Read and understand label and SDS before handling any chemicals. All Spectrum's chemicals are for manufacturing, processing, repacking or research purposes by experienced personnel only. It is the customer's responsibility to provide adequate hazardous material training and ensure that appropriate Personal Protective Equipment (PPE) is used before handling any chemical.



Hexadecane, 99.0%



Material No.: H223-57  
Batch No.: 0000266903  
Manufactured Date: 2020/05/05  
Retest Date: 2027/05/04  
Revision No: 1

## Certificate of Analysis

| Test                                                         | Specification  | Result |
|--------------------------------------------------------------|----------------|--------|
| Assay ( $\text{CH}_3(\text{CH}_2)_{14}\text{CH}_3$ ) (by GC) | $\geq 99.0 \%$ | 99.3   |
| Infrared Spectrum                                            | Passes Test    | PT     |

For Laboratory, Research or Manufacturing Use

Country of Origin: US  
Packaging Site: Paris Mfg Ctr & DC

  
Jamie Ethier  
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700  
Avantor Performance Materials, LLC  
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

W2666 Recived on 02/10/2020 by AP

Product No.: 87683

Product: Sodium pentacyanonitrosylferrate(III) dihydrate, ACS,  
99.0-102.0%

Lot No.: W12F013

| Test                  | Limits         | Results      |
|-----------------------|----------------|--------------|
| Assay                 | 99.0 - 102.0 % | 99.67 %      |
| Insoluble             | 0.01 % max     | 0.0079 %     |
| Chloride              | 0.02 % max     | Not detected |
| Sulfate               | To pass test   | Passes test  |
| Aqueous solubility    | To pass test   | Passes test  |
| Limit on Ferricyanide | To pass test   | Passes test  |
| Limit on Ferrocyanide | To pass test   | Passes test  |

Order our products online [alfa.com](https://www.alfa.com)

**This document has been electronically generated and does not require a signature.**

This is to certify that units of the lot number above were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the purchaser, formulator or those performing further manufacturing to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The above information is the actual analytical results obtained.

**Product Name:** Stearic acid, 98%, Thermo Scientific Chemicals  
**Catalog Number:** A12244.14

**CAS Number:** 57-11-4  
**Molecular Formula:** C<sub>18</sub>H<sub>36</sub>O<sub>2</sub>  
**Molecular Weight:** 284.48  
**InChI Key:** QIQXTHQIDYTRH-UHFFFAOYSA-N  
**SMILES:** CCCCCCCCCCCCCCCC(O)=O  
**Synonym:** stearic acid acide stearique hydrofol acid 1855 hydrofol acid 1655 industrene 5016  
stearic acid, ion(1-) (8Cl) glycon TP glycon DP acidum stearinicum hydrofol acid 150

### Product Specification

**Appearance (Color):** White  
**Form:** Crystals or powder or crystalline powder or flakes or waxy solid  
**Assay (Silylated GC):** ≥97.5%  
**Melting Point (clear melt):** 67.0-74.0°C

**Date Of Print:** 11/30/2023

*Product Specifications are subject to amendment and may change over time. Data contained is accurate as of the date printed.*

W3009  
rec. 2/27/2023 12

Product Name:

Hexadecane - ReagentPlus®, 99%

## Certificate of Analysis

Product Number:

H6703

Batch Number:

SHBP8192

 $\text{CH}_3(\text{CH}_2)_{14}\text{CH}_3$ 

Brand:

SIAL

CAS Number:

544-76-3

MDL Number:

MFCD00008998

Formula:

C16H34

Formula Weight:

226.44 g/mol

Quality Release Date:

04 AUG 2022

| Test                       | Specification         | Result    |
|----------------------------|-----------------------|-----------|
| Appearance (Color)         | Colorless or White    | Colorless |
| Appearance (Form)          | Liquid or Solid       | Liquid    |
| Infrared Spectrum          | Conforms to Structure | Conforms  |
| Refractive index at 20 ° C | 1.432 - 1.436         | 1.435     |
| Purity (GC)                | ≥ 98.5 %              | 99.3 %    |
| Color Test                 | ≤ 20 APHA             | < 5 APHA  |

  
Larry Coers, Director

Quality Control

Sheboygan Falls, WI US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at [Sigma-Aldrich.com](http://Sigma-Aldrich.com). For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.



W3019  
rec 4/3/23

3050 Spruce Street, Saint Louis, MO 63103, USA

Website: [www.sigmaaldrich.com](http://www.sigmaaldrich.com)Email USA: [techserv@sial.com](mailto:techserv@sial.com)Outside USA: [eurtechserv@sial.com](mailto:eurtechserv@sial.com)

## Certificate of Analysis

Product Name:

Pyridine - anhydrous, 99.8%

Product Number:

270970

Batch Number:

SHBQ2113

Brand:

SIAL

CAS Number:

110-86-1

MDL Number:

MFCD00011732

Formula:


C<sub>5</sub>H<sub>5</sub>N

Formula Weight:

79.10 g/mol

Quality Release Date:

15 DEC 2022



| Test                    | Specification         | Result     |
|-------------------------|-----------------------|------------|
| Appearance (Color)      | Colorless             | Colorless  |
| Appearance (Form)       | Liquid                | Liquid     |
| Infrared Spectrum       | Conforms to Structure | Conforms   |
| Purity (GC)             | ≥ 99.75 %             | 99.99 %    |
| Water (by Karl Fischer) | ≤ 0.003 %             | 0.002 %    |
| Residue on Evaporation  | ≤ 0.0005 %            | < 0.0001 % |

  
Larry Coers, Director  
Quality Control  
Sheboygan Falls, WI US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at [Sigma-Aldrich.com](http://Sigma-Aldrich.com). For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.





## Certificate of Analysis

Date of Release: 2/26/2020

Name: Formaldehyde Solution  
GR ACS  
Meets ACS Specifications

Item No: FX0410 all size codes

Lot / Batch No: 60045

Country of Origin: USA

| Characteristic             | Requirement |       | Results     | Units |
|----------------------------|-------------|-------|-------------|-------|
|                            | Min.        | Max.  |             |       |
| Assay                      | 36.5        | 38.0  | 36.71       | %     |
| Chloride (Cl)              |             | 5     | <5          | ppm   |
| Color (APHA)               |             | 10    | <10         |       |
| Form                       |             |       | Passes test |       |
| Heavy metals (as Pb)       |             | 5     | <5          | ppm   |
| Iron (Fe)                  |             | 5     | 0.6         | ppm   |
| Residue after ignition     |             | 0.005 | <0.0050     | %     |
| Sulfate (SO <sub>4</sub> ) |             | 0.002 | <0.0020     | %     |
| Titrate acid               |             | 0.006 | <0.0060     | meq/g |

Heather Sinn,

-----  
Quality Control Manager

This document has been produced electronically and is valid without a signature.

EMD Millipore Corporation, an affiliate of Merck KGaA, Darmstadt, Germany  
290 Concord Road  
Billerica, MA 01821  
U.S.A

The life science business of Merck KGaA, Darmstadt, Germany operates as MilliporeSigma in the U.S. and Canada.

Certificate of Analysis

**ThermoFisher**  
SCIENTIFIC

## Certificate of Analysis

1 Reagent Lane

Fair Lawn, NJ 07410

201.796.7100 tel

201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System  
Standard ISO9001:2015 by SAI Global Certificate Number CERT – 0120632

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

|                   |                                                                                                                                                                                                   |                             |            |
|-------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------------------|------------|
| Catalog Number    | P243                                                                                                                                                                                              | Quality Test / Release Date | 06/19/2020 |
| Lot Number        | 201089                                                                                                                                                                                            |                             |            |
| Description       | POTASSIUM HYDROGEN PHTHALATE, ACIDIMETRIC STANDARD, A.C.S.                                                                                                                                        |                             |            |
| Country of Origin | Spain                                                                                                                                                                                             | Suggested Retest Date       | Jun/2025   |
| Chemical Origin   | Organic - non animal                                                                                                                                                                              |                             |            |
| BSE/TSE Comment   | No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product. |                             |            |

| N/A                                |                     |                                  |                |
|------------------------------------|---------------------|----------------------------------|----------------|
| Result Name                        | Units               | Specifications                   | Test Value     |
| APPEARANCE                         |                     | REPORT                           | WHITE CRYSTALS |
| ASSAY POTASSIUM HYDROGEN PHTHALATE | %                   | Inclusive Between 99.95 - 100.05 | 100.03         |
| CHLORINE COMPOUNDS                 | %                   | <= 0.003                         | <0.003         |
| HEAVY METALS (as Pb)               | ppm                 | <= 5                             | <5             |
| IDENTIFICATION                     | PASS/FAIL           | = PASS TEST                      | PASS TEST      |
| INSOLUBLE MATTER                   | %                   | <= 0.005                         | <0.005         |
| IRON (Fe)                          | ppm                 | <= 5                             | <5             |
| PH OF 0.05M SOLUTION               |                     | Inclusive Between 4.00 - 4.02    | 4.00           |
| SODIUM (Na)                        | %                   | <= 0.005                         | <0.005         |
| SULFUR COMPOUNDS                   | %                   | <= 0.002                         | <0.002%        |
| TRACEABLE TO NIST                  | SOD CARBONATE       | = LOT 351a                       | 351a           |
| TRACEABLE TO NIST KHP STD          | POT. ACID PHTHALATE | = LOT 84L                        | 84L            |



Julian Burton - Quality Control Manager – Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.  
If there are any questions with this certificate, please call at (800) 227-6701.

\*Based on suggested storage condition.



**PRODUCTOS  
QUÍMICOS  
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR  
MONTERREY, N.L. MEXICO  
CP 64070  
TEL +52 81 13 52 57 57  
www.pqm.com.mx

## CERTIFICATE OF ANALYSIS

|                        |                                   |               |                                 |
|------------------------|-----------------------------------|---------------|---------------------------------|
| PRODUCT :              | SODIUM SULFATE CRYSTALS ANHYDROUS |               |                                 |
| QUALITY :              | ACS (CODE RMB3375)                | FORMULA :     | Na <sub>2</sub> SO <sub>4</sub> |
| SPECIFICATION NUMBER : | 6399                              | RELEASE DATE: | ABR/21/2023                     |
| LOT NUMBER :           | 313201                            |               |                                 |

| TEST                                     | SPECIFICATIONS | LOT VALUES  |
|------------------------------------------|----------------|-------------|
| Assay (Na <sub>2</sub> SO <sub>4</sub> ) | Min. 99.0%     | 99.7 %      |
| pH of a 5% solution at 25°C              | 5.2 - 9.2      | 6.1         |
| Insoluble matter                         | Max. 0.01%     | 0.005 %     |
| Loss on ignition                         | Max. 0.5%      | 0.1 %       |
| Chloride (Cl)                            | Max. 0.001%    | <0.001 %    |
| Nitrogen compounds (as N)                | Max. 5 ppm     | <5 ppm      |
| Phosphate (PO <sub>4</sub> )             | Max. 0.001%    | <0.001 %    |
| Heavy metals (as Pb)                     | Max. 5 ppm     | <5 ppm      |
| Iron (Fe)                                | Max. 0.001%    | <0.001 %    |
| Calcium (Ca)                             | Max. 0.01%     | 0.002 %     |
| Magnesium (Mg)                           | Max. 0.005%    | 0.001 %     |
| Potassium (K)                            | Max. 0.008%    | 0.003 %     |
| Extraction-concentration suitability     | Passes test    | Passes test |
| Appearance                               | Passes test    | Passes test |
| Identification                           | Passes test    | Passes test |
| Solubility and foreign matter            | Passes test    | Passes test |
| Retained on US Standard No. 10 sieve     | Max. 1%        | 0.1 %       |
| Retained on US Standard No. 60 sieve     | Min. 94%       | 97.3 %      |
| Through US Standard No. 60 sieve         | Max. 5%        | 2.5 %       |
| Through US Standard No. 100 sieve        | Max. 10%       | 0.1 %       |

### COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

Recd. by R3 on 7/24/23 E 3551

RC-02-01, Ed. 3



Acetone

BAKER RESI-ANALYZED® Reagent  
For Organic Residue Analysis

avantor™



Material No.: 9254-03

Batch No.: 24H2762008

Manufactured Date: 2024-04-18

Expiration Date: 2027-04-18

Revision No.: 0

## Certificate of Analysis

| Test                                                                    | Specification | Result      |
|-------------------------------------------------------------------------|---------------|-------------|
| Assay ((CH <sub>3</sub> ) <sub>2</sub> CO) (by GC, corrected for water) | >= 99.4 %     | 100.0 %     |
| Color (APHA)                                                            | <= 10         | 5           |
| Residue after Evaporation                                               | <= 1.0 ppm    | 0.0 ppm     |
| Substances Reducing Permanganate                                        | Passes Test   | Passes Test |
| Titration Acid (μeq/g)                                                  | <= 0.3        | 0.2         |
| Titration Base (μeq/g)                                                  | <= 0.6        | <0.1        |
| Water (H <sub>2</sub> O)                                                | <= 0.5 %      | <0.1 %      |
| FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)    | <= 5          | 1           |
| ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)    | <= 10         | 1           |

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

Recd by RP on 03/31/25

E3917

Jamie Croak  
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Sulfuric Acid  
BAKER INSTRA-ANALYZED® Reagent  
For Trace Metal Analysis  
Low Selenium

avantor™



M 6041-4b  
MS

Material No.: 9673-33  
Batch No.: 23D2462010  
Manufactured Date: 2023-03-22  
Retest Date: 2028-03-20  
Revision No.: 0

## Certificate of Analysis

| Test                                                         | Specification | Result      |
|--------------------------------------------------------------|---------------|-------------|
| ACS – Assay (H <sub>2</sub> SO <sub>4</sub> )                | 95.0 – 98.0 % | 96.1 %      |
| Appearance                                                   | Passes Test   | Passes Test |
| ACS – Color (APHA)                                           | ≤ 10          | 5           |
| ACS – Residue after Ignition                                 | ≤ 3 ppm       | < 1 ppm     |
| ACS – Substances Reducing Permanganate (as SO <sub>2</sub> ) | ≤ 2 ppm       | < 2 ppm     |
| Ammonium (NH <sub>4</sub> )                                  | ≤ 1 ppm       | 1 ppm       |
| Chloride (Cl)                                                | ≤ 0.1 ppm     | < 0.1 ppm   |
| Nitrate (NO <sub>3</sub> )                                   | ≤ 0.2 ppm     | < 0.1 ppm   |
| Phosphate (PO <sub>4</sub> )                                 | ≤ 0.5 ppm     | < 0.1 ppm   |
| Trace Impurities – Aluminum (Al)                             | ≤ 30.0 ppb    | < 5.0 ppb   |
| Arsenic and Antimony (as As)                                 | ≤ 4.0 ppb     | < 2.0 ppb   |
| Trace Impurities – Boron (B)                                 | ≤ 10.0 ppb    | 8.5 ppb     |
| Trace Impurities – Cadmium (Cd)                              | ≤ 2.0 ppb     | < 0.3 ppb   |
| Trace Impurities – Chromium (Cr)                             | ≤ 6.0 ppb     | < 0.4 ppb   |
| Trace Impurities – Cobalt (Co)                               | ≤ 0.5 ppb     | < 0.3 ppb   |
| Trace Impurities – Copper (Cu)                               | ≤ 1.0 ppb     | < 0.1 ppb   |
| Trace Impurities – Gold (Au)                                 | ≤ 10.0 ppb    | 0.5 ppb     |
| Heavy Metals (as Pb)                                         | ≤ 500.0 ppb   | < 100.0 ppb |
| Trace Impurities – Iron (Fe)                                 | ≤ 50.0 ppb    | 1.3 ppb     |
| Trace Impurities – Lead (Pb)                                 | ≤ 0.5 ppb     | < 0.5 ppb   |
| Trace Impurities – Magnesium (Mg)                            | ≤ 7.0 ppb     | 0.8 ppb     |
| Trace Impurities – Manganese (Mn)                            | ≤ 1.0 ppb     | < 0.4 ppb   |
| Trace Impurities – Mercury (Hg)                              | ≤ 0.5 ppb     | < 0.1 ppb   |
| Trace Impurities – Nickel (Ni)                               | ≤ 2.0 ppb     | 0.3 ppb     |
| Trace Impurities – Potassium (K)                             | ≤ 500.0 ppb   | < 2.0 ppb   |
| Trace Impurities – Selenium (Se)                             | ≤ 50.0 ppb    | < 0.1 ppb   |
| Trace Impurities – Silicon (Si)                              | ≤ 100.0 ppb   | 31.5 ppb    |
| Trace Impurities – Silver (Ag)                               | ≤ 1.0 ppb     | < 0.3 ppb   |

>>> Continued on page 2 >>>

Sulfuric Acid  
BAKER INSTRA-ANALYZED® Reagent  
For Trace Metal Analysis  
Low Selenium

 **avantor™**




Material No.: 9673-33  
Batch No.: 23D2462010

| Test                              | Specification    | Result    |
|-----------------------------------|------------------|-----------|
| Trace Impurities – Sodium (Na)    | $\leq 500.0$ ppb | 5.4 ppb   |
| Trace Impurities – Strontium (Sr) | $\leq 5.0$ ppb   | < 0.2 ppb |
| Trace Impurities – Tin (Sn)       | $\leq 5.0$ ppb   | < 0.8 ppb |
| Trace Impurities – Zinc (Zn)      | $\leq 5.0$ ppb   | 0.4 ppb   |

For Laboratory, Research, or Manufacturing Use

Country of Origin: USA  
Packaging Site: Phillipsburg Mfg Ctr & DC

  
Jamie Ethier  
Vice President Global Quality



## Certificate of Analysis

### Product information

|                      |                                |
|----------------------|--------------------------------|
| Product              | pH-Fix 0.3-2.3                 |
| REF                  | 92180                          |
| LOT                  | 80A0441                        |
| Expiration date:     | 29.02.2028                     |
| Date of examination: | 23.01.2024                     |
| Gradation:           | pH 0.3-0.7-1.0-1.3-1.6-1.9-2.3 |

### Confirmation

Hereby we confirm, that the above mentioned product has successfully passed our quality control system in accordance with ISO 9001 and meets the specific quality criteria.

This document has been produced electronically and is valid without a signature.



Hydrochloric Acid, 36.5–38.0%  
BAKER INSTRA-ANALYZED® Reagent  
For Trace Metal Analysis

 **avantor™**



M6151

R → 11/15/25

Material No.: 9530-33  
Batch No.: 22G2862015  
Manufactured Date: 2022-06-15  
Retest Date: 2027-06-14  
Revision No.: 0

## Certificate of Analysis

| Test                                      | Specification | Result      |
|-------------------------------------------|---------------|-------------|
| ACS – Assay (as HCl) (by acid–base titrn) | 36.5 – 38.0 % | 37.9 %      |
| ACS – Color (APHA)                        | ≤ 10          | 5           |
| ACS – Residue after Ignition              | ≤ 3 ppm       | < 1 ppm     |
| ACS – Specific Gravity at 60°/60°F        | 1.185 – 1.192 | 1.191       |
| ACS – Bromide (Br)                        | ≤ 0.005 %     | < 0.005 %   |
| ACS – Extractable Organic Substances      | ≤ 5 ppm       | < 1 ppm     |
| ACS – Free Chlorine (as Cl <sub>2</sub> ) | ≤ 0.5 ppm     | < 0.5 ppm   |
| Phosphate (PO <sub>4</sub> )              | ≤ 0.05 ppm    | < 0.03 ppm  |
| Sulfate (SO <sub>4</sub> )                | ≤ 0.5 ppm     | < 0.3 ppm   |
| Sulfite (SO <sub>3</sub> )                | ≤ 0.8 ppm     | 0.3 ppm     |
| Ammonium (NH <sub>4</sub> )               | ≤ 3 ppm       | < 1 ppm     |
| Trace Impurities – Arsenic (As)           | ≤ 0.010 ppm   | < 0.003 ppm |
| Trace Impurities – Aluminum (Al)          | ≤ 10.0 ppb    | 1.3 ppb     |
| Arsenic and Antimony (as As)              | ≤ 5.0 ppb     | < 3.0 ppb   |
| Trace Impurities – Barium (Ba)            | ≤ 1.0 ppb     | 0.2 ppb     |
| Trace Impurities – Beryllium (Be)         | ≤ 1.0 ppb     | < 0.2 ppb   |
| Trace Impurities – Bismuth (Bi)           | ≤ 10.0 ppb    | < 1.0 ppb   |
| Trace Impurities – Boron (B)              | ≤ 20.0 ppb    | < 5.0 ppb   |
| Trace Impurities – Cadmium (Cd)           | ≤ 1.0 ppb     | < 0.3 ppb   |
| Trace Impurities – Calcium (Ca)           | ≤ 50.0 ppb    | 163.0 ppb   |
| Trace Impurities – Chromium (Cr)          | ≤ 1.0 ppb     | 0.7 ppb     |
| Trace Impurities – Cobalt (Co)            | ≤ 1.0 ppb     | < 0.3 ppb   |
| Trace Impurities – Copper (Cu)            | ≤ 1.0 ppb     | < 0.1 ppb   |
| Trace Impurities – Gallium (Ga)           | ≤ 1.0 ppb     | < 0.2 ppb   |
| Trace Impurities – Germanium (Ge)         | ≤ 3.0 ppb     | < 2.0 ppb   |
| Trace Impurities – Gold (Au)              | ≤ 4.0 ppb     | 0.6 ppb     |
| Heavy Metals (as Pb)                      | ≤ 100 ppb     | < 50 ppb    |
| Trace Impurities – Iron (Fe)              | ≤ 15 ppb      | 6 ppb       |

>>> Continued on page 2 >>>

Hydrochloric Acid, 36.5–38.0%  
BAKER INSTRA-ANALYZED® Reagent  
For Trace Metal Analysis

 **avantorsm**



Material No.: 9530-33  
Batch No.: 22G2862015

| Test                                                   | Specification | Result     |
|--------------------------------------------------------|---------------|------------|
| Trace Impurities – Lead (Pb)                           | ≤ 1.0 ppb     | < 0.5 ppb  |
| Trace Impurities – Lithium (Li)                        | ≤ 1.0 ppb     | < 0.2 ppb  |
| Trace Impurities – Magnesium (Mg)                      | ≤ 10.0 ppb    | 2.9 ppb    |
| Trace Impurities – Manganese (Mn)                      | ≤ 1.0 ppb     | < 0.4 ppb  |
| Trace Impurities – Mercury (Hg)                        | ≤ 0.5 ppb     | 0.1 ppb    |
| Trace Impurities – Molybdenum (Mo)                     | ≤ 10.0 ppb    | < 3.0 ppb  |
| Trace Impurities – Nickel (Ni)                         | ≤ 4.0 ppb     | < 0.3 ppb  |
| Trace Impurities – Niobium (Nb)                        | ≤ 1.0 ppb     | 0.8 ppb    |
| Trace Impurities – Potassium (K)                       | ≤ 9.0 ppb     | < 2.0 ppb  |
| Trace Impurities – Selenium (Se), For Information Only |               | < 1.0 ppb  |
| Trace Impurities – Silicon (Si)                        | ≤ 100.0 ppb   | < 10.0 ppb |
| Trace Impurities – Silver (Ag)                         | ≤ 1.0 ppb     | 0.5 ppb    |
| Trace Impurities – Sodium (Na)                         | ≤ 100.0 ppb   | 2.3 ppb    |
| Trace Impurities – Strontium (Sr)                      | ≤ 1.0 ppb     | < 0.2 ppb  |
| Trace Impurities – Tantalum (Ta)                       | ≤ 1.0 ppb     | 1.6 ppb    |
| Trace Impurities – Thallium (Tl)                       | ≤ 5.0 ppb     | < 2.0 ppb  |
| Trace Impurities – Tin (Sn)                            | ≤ 5.0 ppb     | 4.0 ppb    |
| Trace Impurities – Titanium (Ti)                       | ≤ 1.0 ppb     | 1.5 ppb    |
| Trace Impurities – Vanadium (V)                        | ≤ 1.0 ppb     | < 0.2 ppb  |
| Trace Impurities – Zinc (Zn)                           | ≤ 5.0 ppb     | 0.8 ppb    |
| Trace Impurities – Zirconium (Zr)                      | ≤ 1.0 ppb     | 0.3 ppb    |

>>> Continued on page 3 >>>

Hydrochloric Acid, 36.5–38.0%  
BAKER INSTRA–ANALYZED® Reagent  
For Trace Metal Analysis



Material No.: 9530-33  
Batch No.: 22G2862015

| Test | Specification | Result |
|------|---------------|--------|
|------|---------------|--------|

For Laboratory, Research, or Manufacturing Use  
Product Information (not specifications):  
Appearance (clear, fuming liquid)  
Meets ACS Specifications  
Storage Condition: Store below 25 °C.

Country of Origin: USA  
Packaging Site: Phillipsburg Mfg Ctr & DC

A handwritten signature in cursive script that reads 'Jamie Ethier'.  
Jamie Ethier  
Vice President Global Quality

Sodium Phosphate, Monobasic, Monohydrate,  
Crystal  
BAKER ANALYZED® A.C.S. Reagent

(sodium dihydrogen phosphate, monohydrate)



Material No.: 3818-05  
Batch No.: 0000225799  
Manufactured Date: 2018/12/05  
Retest Date: 2025/12/03  
Revision No: 1

## Certificate of Analysis

Meets ACS Reagent Chemical Requirements,

| Test                                                         | Specification  | Result    |
|--------------------------------------------------------------|----------------|-----------|
| Assay ( $\text{NaH}_2\text{PO}_4 \cdot \text{H}_2\text{O}$ ) | 98.0 – 102.0 % | 99.5      |
| pH of 5% Solution at 25°C                                    | 4.1 – 4.5      | 4.3       |
| Insoluble Matter                                             | $\leq 0.01$ %  | $< 0.01$  |
| Chloride (Cl)                                                | $\leq 5$ ppm   | $< 5$     |
| ACS – Sulfate ( $\text{SO}_4$ )                              | $\leq 0.003$ % | $< 0.003$ |
| Calcium (Ca)                                                 | $\leq 0.005$ % | $< 0.005$ |
| Potassium (K)                                                | $\leq 0.01$ %  | $< 0.01$  |
| Heavy Metals (as Pb)                                         | $\leq 0.001$ % | $< 0.001$ |
| Trace Impurities – Iron (Fe)                                 | $\leq 0.001$ % | $< 0.001$ |

For Laboratory, Research or Manufacturing Use  
Meets Reagent Specifications for testing USP/NF monographs

Country of Origin: IN  
Packaging Site: Paris Mfg Ctr & DC

  
Jamie Ethier  
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700  
Avantor Performance Materials, LLC  
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700




## Certificate of Analysis

Product Name:

Zinc acetate dihydrate - ACS reagent,  $\geq 98\%$ 

Product Number: 383058  
Batch Number: MKCQ9159  
Brand: SIGALD  
CAS Number: 5970-45-6  
MDL Number: MFCD00066961  
Formula:  $C_4H_6O_4Zn \cdot 2H_2O$   
Formula Weight: 219.51 g/mol  
Quality Release Date: 06 JAN 2022



W2926  
Open 7/5/22  
received  
on  
7/5/22

| Test                       | Specification                 | Result             |
|----------------------------|-------------------------------|--------------------|
| Appearance (Color)         | White                         | White              |
| Appearance (Form)          | Powder or Crystal or Chunk(s) | Powder             |
| Infrared Spectrum          | Conforms to Structure         | Conforms           |
| Insoluble Matter           | $\leq 0.005\%$                | 0.003 %            |
| Calcium (Ca)               | $\leq 0.005\%$                | 0.003 %            |
| Chloride (Cl)              | $\leq 5$ ppm                  | $< 5$ ppm          |
| Iron (Fe)                  | $\leq 5$ ppm                  | $< 5$ ppm          |
| Potassium (K)              | $\leq 0.01\%$                 | 0.00 %             |
| Magnesium (Mg)             | $\leq 0.005\%$                | 0.003 %            |
| Sodium (Na)                | $\leq 0.05\%$                 | 0.03 %             |
| Lead (Pb)                  | $\leq 0.002\%$                | $< 0.001\%$        |
| pH                         | 6.0 - 7.0                     | 6.1                |
| Sulfate (SO <sub>4</sub> ) | $\leq 0.005\%$                | $< 0.005\%$        |
| Complexometric EDTA        | 98.0 - 101.0 %                | 100.3 %            |
| Meets ACS Requirements     | Meets Requirements            | Meets Requirements |

  
Larry Coers, Director  
Quality Control  
Milwaukee, WI US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at [Sigma-Aldrich.com](http://Sigma-Aldrich.com). For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.



# Certificate of analysis

W3082 Received on 2/26/2026 by IZ

Product No.: A12244  
Product: Stearic acid, 98%  
Lot No.: U23E020

Appearance White flakes  
Assay 98.7 %

This document has been electronically generated and does not require a signature.

Order our products online [www.alfa.com](http://www.alfa.com)

**ThermoFisher**  
S C I E N T I F I C

**RICCA CHEMICAL COMPANY®**

1490 Lammers Pike

Batesville, IN 47006

<http://www.riccachemical.com>

1-888-GO-RICCA

customerservice@riccachemical.com

# Certificate of Analysis

W3093  
004121  
04/03/2024  
16

**Buffer, Reference Standard, pH 7.00 ± 0.01 at 25°C (Color Coded Yellow)****Lot Number:** 4401F99**Product Number:** 1551**Manufacture Date:** JAN 08, 2024**Expiration Date:** DEC 2025

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

|    |      |      |      |      |      |      |      |      |      |      |      |
|----|------|------|------|------|------|------|------|------|------|------|------|
| °C | 0    | 5    | 10   | 15   | 20   | 25   | 30   | 35   | 40   | 45   | 50   |
| pH | 7.12 | 7.09 | 7.06 | 7.04 | 7.02 | 7.00 | 6.99 | 6.98 | 6.98 | 6.97 | 6.97 |

| Name                           | CAS#        | Grade           |
|--------------------------------|-------------|-----------------|
| Water                          | 7732-18-5   | ACS/ASTM/USP/EP |
| Sodium Phosphate Dibasic       | 7558-79-4   | ACS             |
| Potassium Dihydrogen Phosphate | 7778-77-0   | ACS             |
| Preservative                   | Proprietary |                 |
| Yellow Dye                     | Proprietary |                 |
| Sodium Hydroxide               | 1310-73-2   |                 |

| Test       | Specification | Result |
|------------|---------------|--------|
| Appearance | Yellow liquid | Passed |

\*Not a certified value.

| Test                                  | Certified Value | Uncertainty | NIST SRM#               |
|---------------------------------------|-----------------|-------------|-------------------------|
| pH at 25°C (Method: SQCP027, SQCP033) | 7.004           | 0.02        | 186-I-g, 186-II-g, 191d |

| Specification               | Reference       |
|-----------------------------|-----------------|
| Commercial Buffer Solutions | ASTM (D 1293 B) |
| Buffer A                    | ASTM (D 5464)   |
| Buffer A                    | ASTM (D 5128)   |

pH measurements were performed in our Batesville, IN laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.02) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number | Size / Package Type | Shelf Life (Unopened Container) |
|-------------|---------------------|---------------------------------|
| 1551-1      | 4 L natural poly    | 24 months                       |
| 1551-1CT    | 4 L Cubitainer®     | 24 months                       |
| 1551-2.5    | 10 L Cubitainer®    | 24 months                       |
| 1551-5      | 20 L Cubitainer®    | 24 months                       |

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)



Paul Brandon (01/08/2024)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --  
Contents of Certificates and Labels."

**This product was tested in an ISO 17025 Accredited Laboratory**

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

# Certificate of Analysis

## Sodium Thiosulfate, 0.0250 Normal (N/40)

**Lot Number:** 4403S13

**Product Number:** 7900

**Manufacture Date:** MAR 29, 2024

**Expiration Date:** SEP 2025

This product is specially formulated to increase its stability. A preservative is added to prevent bacterial contamination. However, all Sodium Thiosulfate solutions are subject to slow chemical deterioration and should be restandardized periodically.

| Name                            | CAS#        | Grade           |
|---------------------------------|-------------|-----------------|
| Water                           | 7732-18-5   | ACS/ASTM/USP/EP |
| Sodium Thiosulfate Pentahydrate | 10102-17-7  | ACS             |
| Organic Preservative            | Proprietary |                 |
| Sodium Carbonate                | 497-19-8    | ACS             |

| Test                                | Specification             | Result            | NIST SRM# |
|-------------------------------------|---------------------------|-------------------|-----------|
| Appearance                          | Colorless liquid          | Passed            |           |
| Assay (vs. Potassium Iodate/Starch) | 0.02499-0.02501 N at 20°C | 0.02501 N at 20°C | 136       |

| Specification                                  | Reference           |
|------------------------------------------------|---------------------|
| Standard Sodium Thiosulfate Solution, 0.0250 N | APHA (4500-S2- F)   |
| Standard Sodium Thiosulfate Titrant            | APHA (4500-O D)     |
| Standard Sodium Thiosulfate Titrant            | APHA (4500-O E)     |
| Standard Sodium Thiosulfate Titrant            | APHA (4500-O F)     |
| Standard Sodium Thiosulfate Titrant, 0.025 N   | APHA (4500-CI B)    |
| Standard Sodium Thiosulfate Titrant            | APHA (4500-O C)     |
| Standard Sodium Thiosulfate Titrant, 0.025 M   | APHA (5530 C)       |
| Standard Sodium Thiosulfate Solution (0.025 N) | EPA (SW-846) (9031) |
| Standard Sodium Thiosulfate solution (0.025 N) | EPA (SW-846) (9034) |

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number | Size / Package Type | Shelf Life (Unopened Container) |
|-------------|---------------------|---------------------------------|
| 7900-1      | 4 L natural poly    | 18 months                       |
| 7900-16     | 500 mL natural poly | 18 months                       |
| 7900-1CT    | 4 L Cubitainer®     | 18 months                       |
| 7900-32     | 1 L natural poly    | 18 months                       |

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)



Paul Brandon (03/29/2024)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --  
Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.



# Certificate of Analysis



## Sodium Hydroxide (Pellets)

**Material:** 0583  
**Grade:** ACS GRADE  
**Batch Number:** 23B1556310

Chemical Formula: NaOH  
Molecular Weight: 40  
CAS #: 1310-73-2  
Appearance:

Manufacture Date: 12/14/2022  
Expiration Date: 12/31/2025

Storage: Room Temperature

Pellets

| TEST               | SPECIFICATION | ANALYSIS | DISPOSITION |
|--------------------|---------------|----------|-------------|
| Calcium            | <= 0.005 %    | <0.005 % | PASS        |
| Chloride           | <= 0.005 %    | 0.002 %  | PASS        |
| Heavy Metals       | <= 0.002 %    | <0.002 % | PASS        |
| Iron               | <= 0.001 %    | <0.001 % | PASS        |
| Magnesium          | <= 0.002 %    | <0.002 % | PASS        |
| Mercury            | <= 0.1 ppm    | <0.1 ppm | PASS        |
| Nickel             | <= 0.001 %    | <0.001 % | PASS        |
| Nitrogen Compounds | <= 0.001 %    | <0.001 % | PASS        |
| Phosphate          | <= 0.001 %    | <0.001 % | PASS        |
| Potassium          | <= 0.02 %     | <0.02 %  | PASS        |
| Purity             | >= 97.0 %     | 99.2 %   | PASS        |
| Sodium Carbonate   | <= 1.0 %      | 0.5 %    | PASS        |
| Sulfate            | <= 0.003 %    | <0.003 % | PASS        |

Internal ID #: 710

### Signature

We certify that this batch conforms to the specifications listed.

This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon  
VWR Chemicals, LLC.  
28600 Fountain Parkway, Solon OH 44139 USA

### Additional Information

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.



## Sodium Hydroxide (Pellets)

**Material:** 0583  
**Grade:** ACS GRADE  
**Batch Number:** 23B1556310

Chemical Formula: NaOH  
Molecular Weight: 40  
CAS #: 1310-73-2  
Appearance:

Manufacture Date: 12/14/2022  
Expiration Date: 12/31/2025

Storage: Room Temperature

Pellets

Spec Set: 0583ACS

Internal ID #: 710

### Signature

We certify that this batch conforms to the specifications listed.

This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon  
VWR Chemicals, LLC.  
28600 Fountain Parkway, Solon OH 44139 USA

### Additional Information

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.



W3127 rec. 7/25/24 12  
W3128 exp. 10/31/27  
W3129

**ENVIRONMENTAL EXPRESS**  
Charleston, SC USA  
[www.envexp.com](http://www.envexp.com)  
(800) 343-5319

October 27, 2022

**CERTIFICATE OF ANALYSIS**

Environmental Express certifies that the following COD Reagent Vials have been rigorously checked against NIST Traceable standards and also compared for conformance to another major brand name product. Environmental Express COD Vial performance is evaluated using bench top spectrophotometers. Acceptance guidelines are strict and ensure dependable, quality results.

Environmental Express further certifies that the COD products listed below are recognized by the United States Environmental Protection Agency (USEPA) as equivalent to an approved Water Pollutant Testing Procedure for COD (Federal Register, Vol. 45, No. 78, Monday, April 20<sup>th</sup>, 1980, page 26811) and as such can be used for National Pollution Discharge Elimination System (NPDES) reporting.

| <u>Cat. No.</u> | <u>Lot No.</u> | <u>Product Description</u>     |
|-----------------|----------------|--------------------------------|
| B1010           | 13821          | COD Reagent Vials, 0 - 150 ppm |

|                   |                                      |                  |           |
|-------------------|--------------------------------------|------------------|-----------|
| Item Number       | ED150                                | Lot Number       | 2ND0156   |
| Item              | Edetate Disodium, Dihydrate, USP     | CAS Number       | 6381-92-6 |
| Molecular Formula | $C_{10}H_{14}N_2Na_2O_8 \cdot 2H_2O$ | Molecular Weight | 372.24    |

| TEST                                                           | SPECIFICATION                                         |             | RESULT                                                |
|----------------------------------------------------------------|-------------------------------------------------------|-------------|-------------------------------------------------------|
|                                                                | MIN                                                   | MAX         |                                                       |
| ASSAY (DRIED BASIS)                                            | 99.0                                                  | 101.0 %     | 99.5 %                                                |
| pH OF A 5% SOLUTION @ 25°C                                     | 4.0                                                   | 6.0         | 4.6                                                   |
| LOSS ON DRYING                                                 | 8.7                                                   | 11.4 %      | 8.90 %                                                |
| CALCIUM (Ca)                                                   | NO PRECIPITATE IS FORMED                              |             | NO PRECIPITATE IS FORMED                              |
| ELEMENTAL IMPURITIES:                                          |                                                       |             | .                                                     |
| NICKEL (Ni)                                                    | AS REPORTED                                           |             | <0.3 ppm                                              |
| CHROMIUM (Cr)                                                  | AS REPORTED                                           |             | <0.3 ppm                                              |
| NITRILOTRIACETIC ACID[n[(HOCOCH <sub>2</sub> ) <sub>3</sub> N] |                                                       | 0.1 %       | <0.10 %                                               |
| IDENTIFICATION A                                               | MATCHES REFERENCE                                     |             | MATCHES REFERENCE                                     |
| IDENTIFICATION B                                               | RED COLOR IS DISCHARGED, LEAVING A YELLOWISH SOLUTION |             | RED COLOR IS DISCHARGED, LEAVING A YELLOWISH SOLUTION |
| IDENTIFICATION C                                               | MEETS THE REQUIREMENTS FOR SODIUM                     |             | MEETS THE REQUIREMENTS FOR SODIUM                     |
| CERTIFIED HALAL                                                |                                                       |             | CERTIFIED HALAL                                       |
| EXPIRATION DATE                                                |                                                       |             | 10-JUL-2026                                           |
| DATE OF MANUFACTURE                                            |                                                       |             | 11-JUL-2023                                           |
| APPEARANCE                                                     |                                                       |             | WHITE CRYSTALLINE POWDER                              |
| RESIDUAL SOLVENTS                                              |                                                       | AS REPORTED | NO RESIDUAL SOLVENTS PRESENT                          |
| MONOGRAPH EDITION                                              |                                                       |             | USP 2024                                              |

Certificate of Analysis Results Entered By:

CACEVEDO  
Charmian Acevedo  
22-MAY-24 08:12:30

Certificate of Analysis Results Approved By:

GHERRERA  
Genaro Herrera  
22-MAY-24 12:32:01

Spectrum Chemical Mfg Corp  
755 Jersey Avenue  
New Brunswick 08901 NJ



**All pharmaceutical ingredients are tested using current edition of applicable pharmacopeia.**

**Read and understand label and SDS before handling any chemicals. All Spectrum's chemicals are for manufacturing, processing, repacking or research purposes by experienced personnel only. It is the customer's responsibility to provide adequate hazardous material training and ensure that appropriate Personal Protective Equipment (PPE) is used before handling any chemical.**

The Elemental Impurities standards implemented by USP and other Pharmaceutical Compendia reflect a growing understanding of the toxicology of trace levels of elemental impurities that can remain in drug substances originating from either raw materials or manufacturing processes. Identifying and quantifying impurities can be critical to predicting the best possible patient outcomes. Elemental Impurities has been a requirement of all products meeting USP/NF, EP and BP monographs since January 1, 2018. More information can be found in USP sections <232> Elemental Impurities – Limits and <233> Elemental Impurities – Procedures. Data for drug substances furnished by Spectrum Chemical Mfg. Corp can be used to ensure that patient daily exposures by oral administration to the selected elements are not exceeded in the formulation of pharmaceutical products.

W3139 Received on 9/9/24 by IZ

Product No.: A12044  
Product: Chloramine-T trihydrate, 98%  
Lot No.: 10239484

|                               |              |
|-------------------------------|--------------|
| Appearance:                   | White powder |
| Melting Point:                | 166°C(dec)   |
| Assay (Iodometric titration): | 100.5%       |
| Identification (FTIR):        | Conforms     |

Order our products online [thermofisher.com/chemicals](https://thermofisher.com/chemicals)

**This document has been electronically generated and does not require a signature.**

Products are processed under ISO 9001:2015 quality management systems and samples are tested for conformance to the noted specifications. Certain data may have been supplied by third parties. We disclaim the implied warranties of merchantability and fitness for a particular purpose, and the accuracy of third party data or information associated with the product. Products are for research and development use only. Products are not for direct administration to humans or animals. It is the responsibility of the final formulator or end user to determine suitability, and to qualify and/or validate each product for its intended use.



# Certificate of Analysis

## Starch Indicator, 0.5% (w/v), Mercury Free, for Iodometric Titrations

Lot Number: 4408P62

Product Number: 8000

Manufacture Date: AUG 28, 2024

Expiration Date: AUG 2026

This product is Mercury-free.

| Name            | CAS#      | Grade           |
|-----------------|-----------|-----------------|
| Water           | 7732-18-5 | ACS/ASTM/USP/EP |
| Starch, soluble | 9005-84-9 | ACS             |
| Salicylic Acid  | 69-72-7   | ACS             |

| Test                | Specification                                     | Result |
|---------------------|---------------------------------------------------|--------|
| Appearance          | White translucent liquid                          | Passed |
| Suitability for Use | Colorless (Iodine absent) - Blue (Iodine present) | Passed |

| Specification             | Reference           |
|---------------------------|---------------------|
| Starch Solution           | APHA (4500-S2- F)   |
| Starch Indicator Solution | APHA (4500-CI B)    |
| Starch Indicator          | APHA (4500-SO32- B) |
| Starch indicator solution | APHA (2350 B)       |
| Starch indicator solution | APHA (2350 E)       |
| Starch Solution           | APHA (510 B)        |
| Starch Solution           | APHA (5530 C)       |
| Starch Indicator          | APHA (4500-CI C)    |
| Starch Indicator          | EPA (345.1)         |

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number | Size / Package Type | Shelf Life (Unopened Container) |
|-------------|---------------------|---------------------------------|
| 8000-1      | 4 L natural poly    | 24 months                       |
| 8000-16     | 500 mL natural poly | 24 months                       |
| 8000-32     | 1 L natural poly    | 24 months                       |

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)

A handwritten signature in blue ink that reads "Paul Brandon". The signature is fluid and cursive, with the first name "Paul" and last name "Brandon" clearly distinguishable.

Paul Brandon (08/28/2024)  
Production Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.



# Certificate of Analysis

**Buffer, Reference Standard, pH 2.00 ± 0.01 at 25°C****Lot Number:** 2411E26**Product Number:** 1493**Manufacture Date:** NOV 11, 2024**Expiration Date:** OCT 2026

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

|    |      |      |      |      |      |      |      |      |      |
|----|------|------|------|------|------|------|------|------|------|
| °C | 10   | 15   | 20   | 25   | 30   | 35   | 40   | 45   | 50   |
| pH | 1.93 | 1.98 | 1.98 | 2.00 | 2.01 | 2.03 | 2.03 | 2.04 | 2.04 |

| Name               | CAS#      | Grade           |
|--------------------|-----------|-----------------|
| Water              | 7732-18-5 | ACS/ASTM/USP/EP |
| Potassium Chloride | 7447-40-7 | ACS             |
| Hydrochloric Acid  | 7647-01-0 | ACS             |

| Test       | Specification    | Result |
|------------|------------------|--------|
| Appearance | Colorless liquid | Passed |

\*Not a certified value.

| Test                                  | Certified Value | Uncertainty | NIST SRM#               |
|---------------------------------------|-----------------|-------------|-------------------------|
| pH at 25°C (Method: SQCP027, SQCP033) | 1.994           | 0.02        | 185i, 186-I-g, 186-II-g |

pH measurements were performed in our Pocomoke City, MD laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.01) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number | Size / Package Type | Shelf Life (Unopened Container) |
|-------------|---------------------|---------------------------------|
| 1493-1      | 4 L natural poly    | 24 months                       |
| 1493-16     | 500 mL natural poly | 24 months                       |
| 1493-1CT    | 4 L Cubitainer®     | 24 months                       |
| 1493-2.5    | 10 L Cubitainer®    | 24 months                       |
| 1493-32     | 1 L natural poly    | 24 months                       |

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)



Jose Pena (11/11/2024)  
Operations Manager

**This product was tested in an ISO 17025 Accredited Laboratory**

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.





|                      |                                |
|----------------------|--------------------------------|
| Material             | BDH9260-500G                   |
| Material Description | BDH POTASS HYDRGN PHTHLTE 500G |
| Grade                | ACS GRADE                      |
| Batch                | 24H0956262                     |
| Reassay Date         | 04/28/2026                     |
| CAS Number           | 877-24-7                       |
| Molecular Formula    | HOCC6H4COOK                    |
| Molecular Mass       | 204.22                         |
| Date of Manufacture  | 04/29/2023                     |
| Storage              | Room Temperature               |

| Characteristics        | Specifications   | Measured Values |
|------------------------|------------------|-----------------|
| Appearance             | White crystals.  | White crystals. |
| Assay (dried basis)    | 99.95 - 100.05 % | 99.98 %         |
| Chlorine Compounds     | <= 0.003 %       | <0.003 %        |
| Heavy Metals (as Pb)   | <= 5 ppm         | <5 ppm          |
| Insoluble Matter       | <= 0.005 %       | 0.003 %         |
| Iron                   | <= 5 ppm         | <5 ppm          |
| pH (0.05M, Water) @25C | 4.00 - 4.02      | 4.00            |
| Sodium                 | <= 0.005 %       | <0.005 %        |
| Sulfur Compounds       | <= 0.002 %       | <0.002 %        |

Internal ID #: 322

| Signature                                                                                                                                                                                                                                                                                                 | Additional Information                                                                                                                                   |
|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------|
| <p>We certify that this batch conforms to the specifications listed above.</p> <p>This document has been electronically produced and is valid without a signature.</p> <p>Leona Edwardson, Quality Control Sr. Manager - Solon<br/>VWR Chemicals, LLC.<br/>28600 Fountain Parkway, Solon OH 44139 USA</p> | <p>Analysis may have been rounded to significant digits in specification limits</p> <p>Product meets analytical specifications of the grades listed.</p> |

# Certificate of Analysis

## Sodium Hypochlorite Solution, 5% available Chlorine

**Lot Number:** 2501J28**Product Number:** 7495.5**Manufacture Date:** JAN 17, 2025**Expiration Date:** JUL 2025

This solution is subject to slow decomposition upon exposure to air. Keep container tightly capped. Refrigeration may improve stability.  
When used in the Phenate method for Ammonia, APHA recommends replacing this solution about every 2 months.

| Name                | CAS#      | Grade      |
|---------------------|-----------|------------|
| Water               | 7732-18-5 | Commercial |
| Sodium Hypochlorite | 7681-52-9 | Commercial |

| Test                                  | Specification                       | Result                     | NIST SRM# |
|---------------------------------------|-------------------------------------|----------------------------|-----------|
| Appearance                            | Colorless to greenish-yellow liquid | Passed                     |           |
| Assay (vs. Sodium Thiosulfate/Starch) | 4.75-5.25 % (w/w) $\text{Cl}_2$     | 5.17 % (w/w) $\text{Cl}_2$ | 136       |

| Specification           | Reference         |
|-------------------------|-------------------|
| Sodium Hypochlorite, 5% | APHA (4500-NH3 F) |
| Sodium Hypochlorite     | ASTM (D 4785)     |

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number | Size / Package Type | Shelf Life (Unopened Container) |
|-------------|---------------------|---------------------------------|
| 7495.5-1    | 4 L black poly      | 6 months                        |
| 7495.5-16   | 500 mL amber poly   | 6 months                        |
| 7495.5-32   | 1 L amber poly      | 6 months                        |
| 7495.5-8    | 250 mL amber poly   | 6 months                        |

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)

Jose Pena (01/17/2025)  
Operations Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

**RICCA CHEMICAL COMPANY®**

1841 Broad Street  
Pocomoke City, MD 21851  
<http://www.riccachemical.com>  
1-888-GO-RICCA  
[customerservice@riccachemical.com](mailto:customerservice@riccachemical.com)

# Certificate of Analysis

W21758 58

Buffer, Reference Standard, pH 4.00 ± 0.01 at 25°C (Color Coded Red)

Lot Number: 2411A93

Product Number: 1501

Manufacture Date: NOV 04, 2024

Expiration Date: OCT 2026

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST Traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

|    |      |      |      |      |      |      |      |      |      |      |      |
|----|------|------|------|------|------|------|------|------|------|------|------|
| °C | 0    | 5    | 10   | 15   | 20   | 25   | 30   | 35   | 40   | 45   | 50   |
| pH | 4.00 | 4.00 | 4.00 | 4.00 | 4.00 | 4.00 | 4.01 | 4.02 | 4.03 | 4.04 | 4.06 |

| Name                     | CAS#        | Grade           |
|--------------------------|-------------|-----------------|
| Water                    | 7732-18-5   | ACS/ASTM/USP/EP |
| Potassium Acid Phthalate | 877-24-7    | Buffer          |
| Preservative             | Proprietary | Commercial      |
| Red Dye                  | Proprietary | Purified        |

| Test       | Specification | Result |
|------------|---------------|--------|
| Appearance | Red liquid    | Passed |

\*Not a certified value.

| Test                                  | Certified Value | Uncertainty | NIST SRM#               |
|---------------------------------------|-----------------|-------------|-------------------------|
| pH at 25°C (Method: SQCP027, SQCP033) | 4.008           | 0.02        | 185i, 186-I-g, 186-II-g |

| Specification               | Reference       |
|-----------------------------|-----------------|
| Commercial Buffer Solutions |                 |
| Buffer B                    | ASTM (D 1293 B) |
| Buffer B                    | ASTM (D 5464)   |
| Buffer B                    | ASTM (D 5128)   |

pH measurements were performed in our Pocomoke City, MD laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.01) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number | Size / Package Type | Shelf Life (Unopened Container) |
|-------------|---------------------|---------------------------------|
| 1501-16     | 500 mL natural poly | 24 months                       |
| 1501-2.5    | 10 L Cubitainer®    | 24 months                       |
| 1501-5      | 20 L Cubitainer®    | 24 months                       |

Recommended Storage: 15°C - 30°C (59°F - 86°F)



# Certificate of Analysis



|                      |                                |
|----------------------|--------------------------------|
| Material             | BDH9274-2.5KG                  |
| Material Description | BDH SAND STDD OTTAWA W+I 2.5KG |
| Grade                | NOT APPLICABLE                 |
| Batch                | 25A2756718                     |
| Reassay Date         | 12/31/2028                     |
| CAS Number           | 14808-60-7                     |
| Molecular Formula    | SiO <sub>2</sub>               |
| Molecular Mass       | 60.09                          |
| Date of Manufacture  | 12/05/2024                     |
| Storage              | Room Temperature               |

| Characteristics               | Specifications  | Measured Values |
|-------------------------------|-----------------|-----------------|
| Appearance                    | Beige granules. | Beige granules. |
| Moisture                      | <= 0.1 %        | 0.1 %           |
| Particle Size 30-40 mesh      | >= 80 %         | 99 %            |
| CUSTOMER PART # BDH9274-2.5KG |                 |                 |

Internal ID #: 793

| Signature                                                                                                                                                                                                                                                                                        | Additional Information                                                                                                                            |
|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------|
| We certify that this batch conforms to the specifications listed above.<br><br>This document has been electronically produced and is valid without a signature.<br><br>Leona Edwardson, Quality Control Sr. Manager - Solon<br>VWR Chemicals, LLC.<br>28600 Fountain Parkway, Solon OH 44139 USA | Analysis may have been rounded to significant digits in specification limits<br><br>Product meets analytical specifications of the grades listed. |



RICCA CHEMICAL COMPANY®

1841 Broad Street  
Pocomoke City, MD 21851  
<http://www.riccachemical.com>  
1-888-GO-RICCA  
[customerservice@riccachemical.com](mailto:customerservice@riccachemical.com)

Certificate of Analysis

Buffer, Reference Standard, pH 10.00 ± 0.01 at 25°C (Color Coded Blue)

Lot Number: 2410F80

Product Number: 1601

Manufacture Date: OCT 09, 2024

Expiration Date: MAR 2026

The certified value for this product is confirmed in independent testing by a second qualified chemist.  
The NIST traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

| °C | 0     | 5     | 10    | 15    | 20    | 25    | 30   | 35   | 40   | 50   |
|----|-------|-------|-------|-------|-------|-------|------|------|------|------|
| pH | 10.31 | 10.23 | 10.17 | 10.11 | 10.05 | 10.00 | 9.95 | 9.91 | 9.87 | 9.81 |

| Name               | CAS#        | Grade           |
|--------------------|-------------|-----------------|
| Water              | 7732-18-5   | ACS/ASTM/USP/EP |
| Sodium Carbonate   | 497-19-8    | ACS             |
| Sodium Bicarbonate | 144-55-8    | ACS             |
| Sodium Hydroxide   | 1310-73-2   | Reagent         |
| Preservative       | Proprietary |                 |
| Blue Dye           | Proprietary |                 |

| Test       | Specification | Result |
|------------|---------------|--------|
| Appearance | Blue liquid   | Passed |

| Test                                  | Certified Value | Uncertainty | NIST SRM#               |
|---------------------------------------|-----------------|-------------|-------------------------|
| pH at 25°C (Method: SQCP027, SQCP033) | 10.009          | 0.02        | 186-I-g, 186-II-g, 191d |

| Specification               | Reference       |
|-----------------------------|-----------------|
| Commercial Buffer Solutions |                 |
| Buffer C                    | ASTM (D 1293 B) |
| Buffer C                    | ASTM (D 5464)   |
|                             | ASTM (D 5128)   |

pH measurements were performed in our Pocomoke City, MD laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.01) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number | Size / Package Type | Shelf Life (Unopened Container) |
|-------------|---------------------|---------------------------------|
| 1601-1      | 4 L natural poly    | 18 months                       |
| 1601-16     | 500 mL natural poly | 18 months                       |
| 1601-1CT    | 4 L Cubitainer®     | 18 months                       |
| 1601-2.5    | 10 L Cubitainer®    | 18 months                       |
| 1601-32     | 1 L natural poly    | 18 months                       |
| 1601-5      | 20 L Cubitainer®    | 18 months                       |

Version: 1.3

Lot Number: 2410F80

Product Number: 1601

Page 1 of 2



W3195 Received on 03/19/2025 by IZ

# Certificate of Analysis



|                      |                                |
|----------------------|--------------------------------|
| Material             | BDH9208-500G                   |
| Material Description | BDH AMMONIUM CHLORIDE ACS 500G |
| Grade                | U S P REAGENT (ACS GRADE)      |
| Batch                | 24L0356561                     |
| Reassay Date         | 08/31/2027                     |
| CAS Number           | 12125-02-9                     |
| Molecular Formula    | NH <sub>4</sub> Cl             |
| Molecular Mass       | 53.49                          |
| Date of Manufacture  | 08/01/2024                     |
| Storage              | Room Temperature               |

| Characteristics      | Specifications                                             | Measured Values       |
|----------------------|------------------------------------------------------------|-----------------------|
| Appearance           | White granular powder                                      | White granular powder |
| Calcium              | <= 0.001 %                                                 | 0.001 %               |
| Heavy Metals (as Pb) | <= 0.0005 %                                                | <0.0002 %             |
| Insolubles           | <= 0.005 %                                                 | 0.001 %               |
| Iron                 | <= 0.0002 %                                                | <0.0002 %             |
| Magnesium            | <= 0.0005 %                                                | 0.0001 %              |
| pH (5%, Water) @25C  | 4.5 - 5.5                                                  | 4.8                   |
| Phosphate            | <= 0.0002 %                                                | <0.0002 %             |
| Purity               | >= 99.5 %                                                  | 99.8 %                |
| Residue on Ignition  | <= 0.01 %                                                  | 0.003 %               |
| Sulfate              | <= 0.002 %                                                 | <0.002 %              |
| Extra Description:   | Meets Reagent Specifications for testing USP/NF monographs |                       |

Internal ID #: 710


| Signature                                                                                                                                                                                                                                                                                        | Additional Information                                                                                                                            |
|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------|
| We certify that this batch conforms to the specifications listed above.<br><br>This document has been electronically produced and is valid without a signature.<br><br>Leona Edwardson, Quality Control Sr. Manager - Solon<br>VWR Chemicals, LLC.<br>28600 Fountain Parkway, Solon OH 44139 USA | Analysis may have been rounded to significant digits in specification limits<br><br>Product meets analytical specifications of the grades listed. |

W3196 Received on 03/19/2025 by IZ

## Certificate of Analysis

Product Name:

Ammonium chloride - ACS reagent, ≥99.5%



**Product Number:** 213330  
**Batch Number:** MKCV1009  
**Brand:** SIGALD  
**CAS Number:** 12125-02-9  
**MDL Number:** MFCD00011420  
**Formula:** H4ClN  
**Formula Weight:** 53.49 g/mol  
**Quality Release Date:** 23 OCT 2023  
**Recommended Retest Date:** SEP 2026

| Test                           | Specification                  | Result    |
|--------------------------------|--------------------------------|-----------|
| Appearance (Color)             | White                          | White     |
| Appearance (Form)              | Powder or Crystals or Chunk(s) | Crystals  |
| Titration by AgNO <sub>3</sub> | ≥ 99.5 %                       | 100.2 %   |
| pH                             | 4.5 - 5.5                      | 4.9       |
| @ 25 Deg c (5% Solution)       |                                |           |
| Insoluble Matter               | ≤ 0.005 %                      | 0.001 %   |
| 10%, H <sub>2</sub> O          |                                |           |
| Residue on ignition (Ash)      | ≤ 0.01 %                       | < 0.01 %  |
| Calcium (Ca)                   | ≤ 0.001 %                      | < 0.001 % |
| Magnesium (Mg)                 | ≤ 5 ppm                        | 1 ppm     |
| Heavy Metals                   | ≤ 5 ppm                        | < 1 ppm   |
| by ICP                         |                                |           |
| Iron (Fe)                      | ≤ 2 ppm                        | < 1 ppm   |
| Phosphate (PO <sub>4</sub> )   | ≤ 2 ppm                        | < 2 ppm   |
| Sulfate (SO <sub>4</sub> )     | ≤ 0.002 %                      | < 0.002 % |
| Meets ACS Requirements         | Current ACS Specification      | Conforms  |
| Recommended Retest Period      | -----                          | -----     |
| 3 Years                        |                                |           |



Larry Coers, Director

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at [Sigma-Aldrich.com](http://Sigma-Aldrich.com). For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.



## Certificate of Analysis

**Product Number:** 213330  
**Batch Number:** MKCV1009

---

Quality Control  
Milwaukee, WI US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at [Sigma-Aldrich.com](http://Sigma-Aldrich.com). For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.







# Certificate of Analysis

**Buffer, Reference Standard, pH 12.00 ± 0.01 at 25°C****Lot Number:** 2504F20**Product Number:** 1615**Manufacture Date:** APR 08, 2025**Expiration Date:** SEP 2026

The certified value for this product is confirmed in independent testing by a second qualified chemist.

|    |       |       |       |       |       |       |
|----|-------|-------|-------|-------|-------|-------|
| °C | 15    | 20    | 25    | 30    | 35    | 40    |
| pH | 12.35 | 12.17 | 11.99 | 11.78 | 11.62 | 11.46 |

| Name               | CAS#      | Grade              |
|--------------------|-----------|--------------------|
| Water              | 7732-18-5 | ACS/ASTM/USP/EP    |
| Potassium Chloride | 7447-40-7 | ACS                |
| Sodium Hydroxide   | 1310-73-2 | Reagent (from ACS) |

| Test       | Specification    | Result |
|------------|------------------|--------|
| Appearance | Colorless liquid | Passed |

\*Not a certified value.

| Test                                  | Certified Value | Uncertainty | NIST SRM#               |
|---------------------------------------|-----------------|-------------|-------------------------|
| pH at 25°C (Method: SQCP027, SQCP033) | 12.009          | 0.02        | 186-I-g, 186-II-g, 191d |

pH measurements were performed in our Pocomoke City, MD laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.01) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number | Size / Package Type | Shelf Life (Unopened Container) |
|-------------|---------------------|---------------------------------|
| 1615-1      | 4 L natural poly    | 18 months                       |
| 1615-16     | 500 mL clear PET-G  | 18 months                       |
| 1615-5      | 20 L Cubitainer®    | 18 months                       |

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)



Jose Pena (04/08/2025)  
Operations Manager

**This product was tested in an ISO 17025 Accredited Laboratory**

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.


W3201 Received on 4/16/25 by IZ

## Certificate of Analysis

Product Name:

Sodium tetraborate decahydrate - ACS reagent, ≥99.5%

**Product Number:** S9640  
**Batch Number:** BCCL9613  
**Brand:** SIGALD  
**CAS Number:** 1303-96-4  
**Formula:** B<sub>4</sub>Na<sub>2</sub>O<sub>7</sub> · 10H<sub>2</sub>O  
**Formula Weight:** 381,37 g/mol  
**Quality Release Date:** 05 JUL 2024  
**Recommended Retest Date:** MAY 2029



| Test                              | Specification               | Result      |
|-----------------------------------|-----------------------------|-------------|
| Appearance (Color)                | White                       | White       |
| Appearance (Form)                 | Powder or Crystals          | Powder      |
| Titration with NaOH               | 99.5 - 105.0 %              | 100.7 %     |
| pH                                | 9.15 - 9.20                 | 9.20        |
| 0.01 m Solution at 25 Deg C       |                             |             |
| Meets ACS Requirements            | Corresponds to Requirements | Corresponds |
| ACS Specifications                | Corresponds to Requirements | Corresponds |
| Insoluble Matter ≤ 0.005% / Heavy |                             |             |
| Metals (As Pb) ≤ 0.001%           |                             |             |
| Calcium (Ca)                      | ≤ 50 mg/kg                  | < 50 mg/kg  |
| Iron (Fe)                         | ≤ 5 mg/kg                   | < 5 mg/kg   |
| Total Sulfur                      | ≤ 50 mg/kg                  | < 50 mg/kg  |
| as SO <sub>4</sub> (ICP)          |                             |             |
| Chloride (Cl)                     | ≤ 10 mg/kg                  | < 10 mg/kg  |
| Phosphate (PO <sub>4</sub> )      | ≤ 10 mg/kg                  | < 10 mg/kg  |

Dr. Reinhold Schwenninger  
Quality Assurance  
Buchs, Switzerland CH

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at [Sigma-Aldrich.com](http://Sigma-Aldrich.com). For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.




## Certificate of Analysis

Product Name:

Barbituric acid - ReagentPlus®, 99%

Product Number: 185698  
Batch Number: WXBFB3271V  
Brand: SIAL  
CAS Number: 67-52-7  
Formula: C<sub>4</sub>H<sub>4</sub>N<sub>2</sub>O<sub>3</sub>  
Formula Weight: 128.09 g/mol  
Quality Release Date: 16 MAY 2024



| Test                       | Specification         | Result   |
|----------------------------|-----------------------|----------|
| Appearance (Colour)        | White to Off-White    | White    |
| Appearance (Form)          | Powder                | Powder   |
| Infrared spectrum          | Conforms to Structure | Conforms |
| Purity (Titration by NaOH) | 98.5 - 101.5 %        | 100.4 %  |
| GC (area %)                | ≥ 98 %                | 100 %    |
| VPCT                       |                       |          |



Kang Chen  
Quality Manager  
Wuxi, China CN

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at [Sigma-Aldrich.com](http://Sigma-Aldrich.com). For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.



n-Hexane 95%  
ULTRA RESI-ANALYZED  
For Organic Residue Analysis

avantor™



W3204  
084K: 09/22/2025  
38

Material No.: 9262-03  
Batch No.: 25C0362005  
Manufactured Date: 2025-01-29  
Expiration Date: 2026-04-30  
Revision No.: 0

## Certificate of Analysis

| Test                                                                           | Specification          | Result      |
|--------------------------------------------------------------------------------|------------------------|-------------|
| FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)           | $\leq 5$               | 1           |
| ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)            | $\leq 10$              | 6           |
| ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL) | $\leq 5$               | 5           |
| Assay (Total Saturated C <sub>6</sub> Isomers) (by GC, corrected for water)    | $\geq 99.5 \%$         | 100.0 %     |
| Assay (as n-Hexane) (by GC, corrected for water)                               | $\geq 95 \%$           | 100 %       |
| Color (APHA)                                                                   | $\leq 10$              | 10          |
| Residue after Evaporation                                                      | $\leq 1.0 \text{ ppm}$ | 0.1 ppm     |
| Substances Darkened by H <sub>2</sub> SO <sub>4</sub>                          | Passes Test            | Passes Test |
| Water (by KF, coulometric)                                                     | $\leq 0.05 \%$         | $< 0.01 \%$ |

For Laboratory, Research, or Manufacturing Use  
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States  
Packaging Site: Phillipsburg Mfg Ctr & DC

*J. Croak*

Jamie Croak  
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700



|                           |                        |
|---------------------------|------------------------|
| Material                  | BDHVBDH7206-1          |
| Material Description      | IODINE SOLUTION 0.025N |
| Lot                       | 25A2461008             |
| Expires end of            | 2029-Jan-20            |
| Molecular mass            | 0                      |
| Last Quality Control      | 2025-Jan-24            |
| Date of manufacture       | 2025-Jan-21            |
| Made in                   | United States          |
| Manufacturer Source Batch | MK25A21527             |

Additional information

| Characteristics                 | Specifications  | Measured values |
|---------------------------------|-----------------|-----------------|
| Prepared to formulation on file | Confirmed       | Confirmed       |
| Appearance                      | Passes Test     | Passes Test     |
| Normality, N                    | 0.0200 - 0.0300 | 0.0268          |

Signature

We certify that this batch conforms to the specifications listed above.

This document has been produced electronically and is valid without a signature.

Michelle Bales - Sr. Manager Quality Assurance  
 Avantor Performance Materials, LLC

For Professional use in Laboratory or Manufacturing. Not for use as an Active Pharmaceutical Ingredient or Food or Animal Feed. Suitability and intended use of the product remains the responsibility of the user.

VWR International LLC, Radnor Corporate Center, Building One, Suite 200, 100 Matsonford Road, Radnor, PA 19087, USA.

VWR International bv, Haasrode Research Park Zone 2020, Geldenaaksebaan 464, 3001 Leuven, Belgium

BDHVBDH72 25A2461008 Page 1 / 1

# Certificate of Analysis

## Cyanide Standard, 1000 ppm CN<sup>-</sup>

**Lot Number:** 1505H73

**Product Number:** 2543

**Manufacture Date:** MAY 08, 2025

**Expiration Date:** NOV 2025

This standard is prepared using accurate volumetric techniques from material that has been assayed against Silver Nitrate solution certified traceable to NIST Standard Reference Material 999. The certified value reported is the prepared value based upon the method of preparation of the material. The uncertainty in the prepared value is the combined uncertainty based on the stability of the assayed Potassium Cyanide, and the uncertainty in the mass and volume measurements.

Use 0.16% (w/v) (0.04 N) Sodium Hydroxide or 0.225 % (w/v) (0.04 N) Potassium Hydroxide to make dilutions of this standard. Restandardize weekly if extreme accuracy is required.

| Name              | CAS#      | Grade              |
|-------------------|-----------|--------------------|
| Water             | 7732-18-5 | ACS/ASTM/USP/EP    |
| Potassium Cyanide | 151-50-8  | ACS                |
| Sodium Hydroxide  | 1310-73-2 | Reagent (from ACS) |

| Test                       | Specification    | Result   |
|----------------------------|------------------|----------|
| Appearance                 | Colorless liquid | Passed   |
| Cyanide (CN <sup>-</sup> ) | 995-1005 ppm     | 1000 ppm |

| Specification                                                    | Reference              |
|------------------------------------------------------------------|------------------------|
| Stock Standard Cyanide Solution                                  | APHA (4500-CN- F)      |
| Stock Cyanide Solution                                           | APHA (4500-CN- E)      |
| Stock Cyanide Solution                                           | APHA (4500-CN- K)      |
| Stock Cyanide Solution                                           | APHA (4500-CN- H)      |
| Cyanide Reference Solution (1000 mg/L)                           | EPA (SW-846) (7.3.3.2) |
| Cyanide Calibration Stock Solution (1,000 mg/L CN <sup>-</sup> ) | EPA (SW-846) (9213)    |
| Stock Cyanide Solution                                           | EPA (335.3)            |
| Stock Cyanide Solution                                           | EPA (335.2)            |
| Cyanide Solution Stock                                           | ASTM (D 4282)          |
| Simple Cyanide Solution, Stock (1.0 g/L CN <sup>-</sup> )        | ASTM (D 4374)          |

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number | Size / Package Type | Shelf Life (Unopened Container) |
|-------------|---------------------|---------------------------------|
| 2543-16     | 500 mL amber poly   | 6 months                        |
| 2543-32     | 1 L amber poly      | 6 months                        |
| 2543-4      | 120 mL amber poly   | 6 months                        |

**Recommended Storage:** 2°C - 8°C (36°F - 46°F)



Ernest Mahan (05/08/2025)  
Plant Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.



# Certificate of Analysis

**Buffer, Reference Standard, pH 7.00 ± 0.01 at 25°C (Color Coded Yellow)**

**Lot Number:** 2504D34

**Product Number:** 1551

**Manufacture Date:** APR 03, 2025

**Expiration Date:** MAR 2027

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

|    |      |      |      |      |      |      |      |      |      |      |      |
|----|------|------|------|------|------|------|------|------|------|------|------|
| °C | 0    | 5    | 10   | 15   | 20   | 25   | 30   | 35   | 40   | 45   | 50   |
| pH | 7.12 | 7.09 | 7.06 | 7.04 | 7.02 | 7.00 | 6.99 | 6.98 | 6.98 | 6.97 | 6.97 |

| Name                           | CAS#        | Grade              |
|--------------------------------|-------------|--------------------|
| Water                          | 7732-18-5   | ACS/ASTM/USP/EP    |
| Sodium Phosphate Dibasic       | 7558-79-4   | ACS                |
| Potassium Dihydrogen Phosphate | 7778-77-0   | ACS                |
| Preservative                   | Proprietary |                    |
| Yellow Dye                     | Proprietary |                    |
| Sodium Hydroxide               | 1310-73-2   | Reagent (from ACS) |

| Test       | Specification | Result                         |
|------------|---------------|--------------------------------|
| Appearance | Yellow liquid | Passed *Not a certified value. |

| Test                                  | Certified Value | Uncertainty | NIST SRM#               |
|---------------------------------------|-----------------|-------------|-------------------------|
| pH at 25°C (Method: SQCP027, SQCP033) | 7.003           | 0.02        | 186-I-g, 186-II-g, 191d |

| Specification               | Reference       |
|-----------------------------|-----------------|
| Commercial Buffer Solutions | ASTM (D 1293 B) |
| Buffer A                    | ASTM (D 5464)   |
| Buffer A                    | ASTM (D 5128)   |

pH measurements were performed in our Pocomoke City, MD laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.01) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number | Size / Package Type | Shelf Life (Unopened Container) |
|-------------|---------------------|---------------------------------|
| 1551-2.5    | 10 L Cubitainer®    | 24 months                       |
| 1551-20     | 20 x 20 mL pack     | 24 months                       |
| 1551-32     | 1 L natural poly    | 24 months                       |
| 1551-5      | 20 L Cubitainer®    | 24 months                       |

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)



Jose Pena (04/03/2025)  
Operations Manager

**This product was tested in an ISO 17025 Accredited Laboratory**

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

# Certificate Of Analysis



Date of Release: 4/8/2025

Name: **Potassium Hydrogen Phthalate**

ACS

Item No: **PX1476 All Sizes**

Lot / Batch No: **2025040493**

Country of Origin: **USA**

| Item                           | Specifications | Analysis    |
|--------------------------------|----------------|-------------|
| Assay (Dried Basis)            | 99.95-100.05%  | 99.98%      |
| Chlorine compounds (as Cl)     | 0.003% max.    | <0.003%     |
| Color                          | White          | Passes Test |
| Form                           | Crystals       | Passes Test |
| Heavy metals (by ICP-OES)      | 5 ppm max.     | <5 ppm      |
| Insoluble Matter               | 0.005% max.    | <0.005%     |
| Iron (Fe)                      | 5 ppm max.     | <5 ppm      |
| pH of a 0.05m solution @ 25.0C | 4.00-4.02      | 4.00        |
| Sodium (Na)                    | 0.005% max.    | <0.005%     |
| Sulfur compounds (as S)        | 0.002% max.    | <0.002%     |

Joe Schoellkopf

-----  
Quality Control Manager

This document has been produced electronically and is valid without a signature.

EMD Millipore is a division of Merck KGaA, Darmstadt, Germany

EMD Millipore Corporation

400 Summit Drive  
Burlington, MA 01803  
U.S.A.

Form number: 00005624CA, Rev. 2.0



Part of TCP Analytical Group

Jackson's Pointe Commerce Park- Building 1000  
1010 Jackson's Pointe Court, Zelienople, PA 16063

## Certificate of Analysis

### Cyanide Standard 1000 ppm (1ml = 1mg CN)

Product Code: **LC13545**

Manufacture Date: June 25, 2025

Lot Number: **45060288**

Expiration Date: December 24, 2025

| Test                  | Specification      | Result         |
|-----------------------|--------------------|----------------|
| Appearance (clarity)  | clear solution     | clear solution |
| Appearance (color)    | colorless          | colorless      |
| Concentration (CN)    | 0.990 - 1.010mg/mL | 1.000mg/mL     |
| Concentration (CN)    | 990 - 1,010ppm     | 1,000ppm       |
| Traceable to NIST SRM | Report             | 999b           |

**Intended Use** - Product is intended for use in manufacturing procedures and laboratory procedures and protocols.

**Storage Information** - Unless noted on the product label, store the product under normal lab conditions in its tightly closed, original container. Do not pipet directly from the container or return unused portions to the container.

**Instructions for Handling and Use** - Please refer to the associated product label and Safety Data Sheet (SDS) for information regarding safety and handling of this product.

**Preparation** - All products are manufactured and tested according to established, documented procedures and methodology. Production documentation records manufacturing data, raw material traceability and testing history on a per lot basis. Balances, thermometers, and glassware are calibrated before first use and on a regular schedule with references traceable to NIST

\*The suffix of the product code may differ from what is on your product label. The suffix will designate the size and be associated with a numeric digit(s). Visit [LabChem.com](http://LabChem.com) for more information\*

| Suffix | 1         | 2         | 3/3S/36/36S                           | 4/4C | 5   | 6   | 7     | 8   | 9    | 20      | 44   | 200  | 246    | 486    |
|--------|-----------|-----------|---------------------------------------|------|-----|-----|-------|-----|------|---------|------|------|--------|--------|
| Size   | 500mL org | 1L or 1kg | 2.5L/2.5L Coated/6x2.5L/6x2.5L Coated | 4L   | 20L | 10L | 125mL | 25g | 100g | 20x20mL | 4x4L | 200L | 24x6mL | 48x6mL |

*Michael Monteleone*

Michael Monteleone  
Chemistry Supervisor - Quality Control  
20250703 15:30:45ahoffman-0-0

ISO9001:2015 Registration #0306-01



PERCENT SOLID

Supervisor: Iwona  
Analyst: jignesh  
Date: 7/14/2025

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:10  
In Date: 07/11/2025  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104  
Time OUT: 08:25  
Out Date: 07/12/2025  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: % SOLID-OVEN

QC:LB136437

| Lab ID   | Client SampleID       | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments   |
|----------|-----------------------|--------|----------------|--------------|-------------------------|---------------------------|---------|------------|
| Q2561-01 | AUD-25-0115           | 1      | 1.15           | 10.37        | 11.52                   | 11.33                     | 98.2    |            |
| Q2561-02 | AUD-25-0116           | 2      | 1.13           | 10.33        | 11.46                   | 11.22                     | 97.7    |            |
| Q2578-02 | WC-A5-01-C            | 13     | 1.15           | 10.23        | 11.38                   | 9.84                      | 84.9    |            |
| Q2578-06 | WC-A2-09-C            | 14     | 1.18           | 10.33        | 11.51                   | 8.94                      | 75.1    |            |
| Q2578-10 | WC-A2-10-C            | 15     | 1.15           | 10.33        | 11.48                   | 8.58                      | 71.9    |            |
| Q2578-14 | WC-A2-11-C            | 16     | 1.16           | 10.44        | 11.6                    | 9.81                      | 82.9    |            |
| Q2578-18 | WC-A2-12-C            | 17     | 1.18           | 10.29        | 11.47                   | 7.91                      | 65.4    |            |
| Q2579-02 | WC-A2-13-C            | 18     | 1.19           | 10.53        | 11.72                   | 8.78                      | 72.1    |            |
| Q2579-06 | WC-A2-14-C            | 19     | 1.12           | 10.54        | 11.66                   | 9.82                      | 82.5    |            |
| Q2580-01 | WC-URBAN-FILL-B1      | 20     | 1.18           | 10.04        | 11.22                   | 10.6                      | 93.8    |            |
| Q2580-02 | WC-URBAN-FILL-B2      | 21     | 1.17           | 10.82        | 11.99                   | 11.4                      | 94.5    |            |
| Q2581-05 | SVOC-GPC-BLANK        | 3      | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   |            |
| Q2581-06 | PEST-GPC-BLANK        | 4      | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   |            |
| Q2581-07 | PEST-GPC-BLANK-SPIKE  | 5      | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   |            |
| Q2581-08 | PCB-GPC-BLANK         | 6      | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   |            |
| Q2581-09 | PCB-GPC-BLANK-SPIKE   | 7      | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   |            |
| Q2581-10 | SVOC-GPC2-BLANK       | 8      | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   |            |
| Q2581-11 | PEST-GPC2-BLANK       | 9      | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   |            |
| Q2581-12 | PEST-GPC2-BLANK-SPIKE | 10     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   |            |
| Q2581-13 | PCB-GPC2-BLANK        | 11     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   |            |
| Q2581-14 | PCB-GCP2-BLANK-SPIKE  | 12     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   |            |
| Q2586-01 | TP-16                 | 22     | 1.18           | 10.21        | 11.39                   | 10.36                     | 89.9    |            |
| Q2586-02 | TP-16-EPH             | 23     | 1.15           | 10.84        | 11.99                   | 11.34                     | 94.0    |            |
| Q2586-03 | TP-16-VOC             | 24     | 1.13           | 10.44        | 11.57                   | 10.45                     | 89.3    |            |
| Q2587-01 | E-2257                | 25     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | oil sample |
| Q2587-02 | F-2258                | 26     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | oil sample |
| Q2587-03 | O-2267                | 27     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | oil sample |
| Q2589-01 | AU-6-071125           | 28     | 1.15           | 10.43        | 11.58                   | 10.35                     | 88.2    |            |



PERCENT SOLID

Supervisor: Iwona  
Analyst: jignesh  
Date: 7/14/2025

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:10  
In Date: 07/11/2025  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104  
Time OUT: 08:25  
Out Date: 07/12/2025  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: % SOLID-OVEN

QC:LB136437

| Lab ID   | Client SampleID | Dish # | Dish Wt (g) (A) | Sample Wt (g) | Dish + Sample Wt (g) (B) | Dish+Dry Sample Wt (g) (C) | % Solid | Comments    |
|----------|-----------------|--------|-----------------|---------------|--------------------------|----------------------------|---------|-------------|
| Q2590-01 | NB-7-071125     | 29     | 1.14            | 10.44         | 11.58                    | 11.23                      | 96.6    |             |
| Q2590-02 | NB-7-071125-E2  | 30     | 1.13            | 10.74         | 11.87                    | 10.71                      | 89.2    |             |
| Q2591-01 | 72-12000-1      | 31     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | WIPE SAMPLE |
| Q2591-02 | 72-12000-2      | 32     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | WIPE SAMPLE |
| Q2591-03 | NWB-2188        | 33     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | WIPE SAMPLE |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

# WORKLIST(Hardcopy Internal Chain)

136437

WorkList Name : %1-071125 WorkList ID : 190653 Department : Wet-Chemistry Date : 07-11-2025 07:59:07

| Sample   | Customer Sample       | Matrix | Test           | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method       |
|----------|-----------------------|--------|----------------|--------------|----------|-----------------------------|--------------|--------------|
| Q2581-05 | SVOC-GPC-BLANK        | Solid  | Percent Solids | Cool 4 deg C | CHEM02   | D31                         | 07/04/2025   | Chemtech -SO |
| Q2581-06 | PEST-GPC-BLANK        | Solid  | Percent Solids | Cool 4 deg C | CHEM02   | D31                         | 07/04/2025   | Chemtech -SO |
| Q2581-07 | PEST-GPC-BLANK-SPIKE  | Solid  | Percent Solids | Cool 4 deg C | CHEM02   | D31                         | 07/04/2025   | Chemtech -SO |
| Q2581-08 | PCB-GPC-BLANK         | Solid  | Percent Solids | Cool 4 deg C | CHEM02   | D31                         | 07/04/2025   | Chemtech -SO |
| Q2581-09 | PCB-GPC-BLANK-SPIKE   | Solid  | Percent Solids | Cool 4 deg C | CHEM02   | D31                         | 07/04/2025   | Chemtech -SO |
| Q2581-10 | SVOC-GPC2-BLANK       | Solid  | Percent Solids | Cool 4 deg C | CHEM02   | D31                         | 07/04/2025   | Chemtech -SO |
| Q2581-11 | PEST-GPC2-BLANK       | Solid  | Percent Solids | Cool 4 deg C | CHEM02   | D31                         | 07/04/2025   | Chemtech -SO |
| Q2581-12 | PEST-GPC2-BLANK-SPIKE | Solid  | Percent Solids | Cool 4 deg C | CHEM02   | D31                         | 07/04/2025   | Chemtech -SO |
| Q2581-13 | PCB-GPC2-BLANK        | Solid  | Percent Solids | Cool 4 deg C | CHEM02   | D31                         | 07/04/2025   | Chemtech -SO |
| Q2581-14 | PCB-GPC2-BLANK-SPIKE  | Solid  | Percent Solids | Cool 4 deg C | CHEM02   | D31                         | 07/04/2025   | Chemtech -SO |
| Q2561-01 | AUD-25-0115           | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | O21                         | 07/10/2025   | Chemtech -SO |
| Q2561-02 | AUD-25-0116           | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | O21                         | 07/10/2025   | Chemtech -SO |
| Q2586-01 | TP-16                 | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | D41                         | 07/11/2025   | Chemtech -SO |
| Q2586-02 | TP-16-EPH             | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | D41                         | 07/11/2025   | Chemtech -SO |
| Q2586-03 | TP-16-VOC             | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | D41                         | 07/11/2025   | Chemtech -SO |
| Q2587-01 | E-2257                | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | D41                         | 07/11/2025   | Chemtech -SO |
| Q2587-02 | F-2258                | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | D41                         | 07/11/2025   | Chemtech -SO |
| Q2587-03 | O-2267                | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | D41                         | 07/11/2025   | Chemtech -SO |
| Q2591-01 | 72-12000-1            | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | D41                         | 07/11/2025   | Chemtech -SO |
| Q2591-02 | 72-12000-2            | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | D41                         | 07/11/2025   | Chemtech -SO |
| Q2591-03 | NWB-2188              | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | D41                         | 07/11/2025   | Chemtech -SO |

Date/Time 07-11-25 13:00 Date/Time 07-11-25 17:30  
Raw Sample Received by: [Signature] Raw Sample Received by: [Signature]  
Raw Sample Relinquished by: [Signature] Raw Sample Relinquished by: [Signature]

# WORKLIST(Hardcopy Internal Chain)

136437

WorkList Name : %1-071125      WorkList ID : 190653      Department : Wet-Chemistry      Date : 07-11-2025 07:59:07

| Sample   | Customer Sample  | Matrix | Test           | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method       |
|----------|------------------|--------|----------------|--------------|----------|-----------------------------|--------------|--------------|
| Q2590-02 | NB-07-071125-E2  | Solid  | Percent Solids | Cool 4 deg C | PSEG05   | D51                         | 07/11/2025   | Chemtech -SO |
| Q2589-01 | AU-06-071125     | Solid  | Percent Solids | Cool 4 deg C | PSEG05   | D31                         | 07/11/2025   | Chemtech -SO |
| Q2590-01 | NB-07-071125     | Solid  | Percent Solids | Cool 4 deg C | PSEG05   | D51                         | 07/11/2025   | Chemtech -SO |
| Q2578-02 | WC-A5-01-C       | Solid  | Percent Solids | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | Chemtech -SO |
| Q2578-06 | WC-A2-09-C       | Solid  | Percent Solids | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | Chemtech -SO |
| Q2578-10 | WC-A2-10-C       | Solid  | Percent Solids | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | Chemtech -SO |
| Q2578-14 | WC-A2-11-C       | Solid  | Percent Solids | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | Chemtech -SO |
| Q2578-18 | WC-A2-12-C       | Solid  | Percent Solids | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | Chemtech -SO |
| Q2579-02 | WC-A2-13-C       | Solid  | Percent Solids | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | Chemtech -SO |
| Q2579-06 | WC-A2-14-C       | Solid  | Percent Solids | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | Chemtech -SO |
| Q2580-01 | WC-URBAN-FILL-B1 | Solid  | Percent Solids | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | Chemtech -SO |
| Q2580-02 | WC-URBAN-FILL-B2 | Solid  | Percent Solids | Cool 4 deg C | ENTA05   | D41                         | 07/10/2025   | Chemtech -SO |

Date/Time 07-11-25 15:00  
Raw Sample Received by: SS  
Raw Sample Relinquished by: SS

Date/Time 07-11-25 17:30  
Raw Sample Received by: SS  
Raw Sample Relinquished by: SS





# SHIPPING DOCUMENTS



284 Sheffield Street, Mountainside, NJ 07092

(908) 789-8900 Fax: (908) 788-9222

www.chemtech.net

## CHAIN OF CUSTODY RECORD

Alliance Project Number:

Q 2579

COC Number: 2042113

Page 1 of 2

## CLIENT INFORMATION

COMPANY: ENTACT, LLC

ADDRESS: 150 Bay Street, Suite 806

CITY: Jersey City STATE: NJ ZIP: 07302

ATTENTION: Austin Farmerie

PHONE: 412-716-1366

FAX:

## PROJECT INFORMATION

PROJECT NAME: 540 Degraw St Brooklyn, NY

PROJECT #: E9309 LOCATION: Brooklyn, NY

PROJECT MANAGER: Austin Farmerie

E-MAIL: afarmerie@entact.com

PHONE: 412-716-1366

FAX:

## BILLING INFORMATION

BILL TO: ENTACT, LLC

PO# E9309

ADDRESS: 999 Oakmont Plaza Drive, Suite 300

CITY: Westmont

STATE: IL ZIP: 60559

ATTENTION: Wendy Murray

PHONE: 800-936-8228

## ANALYSIS

| TCLP VOCs | TCLP ICP Metals +<br>Cu, Ni, Zn | TCLP Herb | TCLP Pest | TCLP SVOCs | TCLP pH * | I/C/R | PCBs | Oil & Grease |
|-----------|---------------------------------|-----------|-----------|------------|-----------|-------|------|--------------|
| 1         | 2                               | 3         | 4         | 5          | 6         | 7     | 8    | 9            |

\* For TCLP pH -  
include preparatory  
information for TCLP  
leachate

## PRESERVATIVES

## COMMENTS

<-- Specify Preservatives  
A-HCl B-HNO3  
C-H2SO4 D-NaOH  
E-ICE F-Other

## DATA TURNAROUND INFORMATION

FAX: 3 DAYS\*

HARD COPY: 3 DAYS\*

EDD 3 DAYS\*

\* TO BE APPROVED BY ALLIANCE

STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS

## DATA DELIVERABLE INFORMATION

- ☐ RESEULTS ONLY ☐ USEPA CLP  
☐ RESULTS + QC ☐ New York State ASP "B"  
☐ New Jersey REDUCED ☐ New York State ASP "A"  
☐ New Jersey CLP ☐ Other \_\_\_\_\_  
☐ EDD Format \_\_\_\_\_

| CHEMTECH<br>SAMPLE<br>ID | PROJECT<br>SAMPLE IDENTIFICATION | SAMPLE<br>MATRIX | SAMPLE<br>TYPE |      | SAMPLE<br>COLLECTION |       | # of Bottles | PRESERVATIVES |   |   |   |   |   |   |   |   | COMMENTS |
|--------------------------|----------------------------------|------------------|----------------|------|----------------------|-------|--------------|---------------|---|---|---|---|---|---|---|---|----------|
|                          |                                  |                  | COMP           | GRAB | DATE                 | TIME  |              | E             | E | E | E | E | E | E | E | E |          |
| 1.                       | WC-A2-13-G                       | Soil             |                | X    | 7/10                 | 12:00 | 1            | X             |   |   |   |   |   |   |   |   |          |
| 2.                       | WC-A2-13-C                       | Soil             | X              |      | 7/10                 | 12:00 | 11           |               | X | X | X | X | X | X | X | X |          |
| 3.                       | WC-A2-14-G                       | Soil             |                | X    | 7/10                 | 12:00 | 1            | X             |   |   |   |   |   |   |   |   |          |
| 4.                       | WC-A2-14-C                       | Soil             | X              |      | 7/10                 | 12:00 | 11           |               | X | X | X | X | X | X | X | X |          |
| 5.                       |                                  |                  |                |      |                      |       |              |               |   |   |   |   |   |   |   |   |          |
| 6.                       |                                  |                  |                |      |                      |       |              |               |   |   |   |   |   |   |   |   |          |
| 7.                       |                                  |                  |                |      |                      |       |              |               |   |   |   |   |   |   |   |   |          |
| 8.                       |                                  |                  |                |      |                      |       |              |               |   |   |   |   |   |   |   |   |          |
| 9.                       |                                  |                  |                |      |                      |       |              |               |   |   |   |   |   |   |   |   |          |
| 10.                      |                                  |                  |                |      |                      |       |              |               |   |   |   |   |   |   |   |   |          |

## SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE PROSESSION INCLUDING COURIER DELIVERY

|                                               |                           |                     |                 |                                                                                                                                                                                                             |
|-----------------------------------------------|---------------------------|---------------------|-----------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| RELINQUISHED BY SAMPLER<br>1. Austin Farmerie | DATE/TIME<br>7/10/25 1350 | RECEIVED BY<br>     | 1350<br>7-10-25 | Conditions of bottles or coolers at receipt: <input type="checkbox"/> Compliant <input type="checkbox"/> Non Compliant <input type="checkbox"/> Cooler Temp. 41.4° <input type="checkbox"/> Ice in Cooler?: |
| RELINQUISHED BY                               | DATE/TIME                 | RECEIVED BY         |                 | Comments:                                                                                                                                                                                                   |
| 2.                                            |                           |                     |                 |                                                                                                                                                                                                             |
| RELINQUISHED BY                               | DATE/TIME                 | RECEIVED FOR LAB BY |                 |                                                                                                                                                                                                             |
| 3.                                            | 7-10-25 1520              |                     |                 |                                                                                                                                                                                                             |

Page \_\_\_\_\_ of \_\_\_\_\_

SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ Overnight  
ALLIANCE: ☐ Picked Up ☐ Overnight

Shipment Complete  
☐ YES ☐ NO

WHITE - ALLIANCE COPY FOR RETURN TO CLIENT

YELLOW - ALLIANCE COPY

PINK - SAMPLER COPY



284 Sheffield Street, Mountainside, NJ 07092  
(908) 789-8900 Fax: (908) 788-9222  
www.chemtech.net

### CHAIN OF CUSTODY RECORD

Alliance Project Number:

Q2579

COC Number: 2042113

Page 2 of 2

| CLIENT INFORMATION                                                                                                                |  | PROJECT INFORMATION                                                                                                                                                                                              |  | BILLING INFORMATION                                                                                                                                                        |  |                                                                                                   |  |                   |                      |              |  |                   |  |                                            |  |               |  |                   |  |                        |  |  |  |  |  |                     |  |  |  |  |  |
|-----------------------------------------------------------------------------------------------------------------------------------|--|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--|---------------------------------------------------------------------------------------------------|--|-------------------|----------------------|--------------|--|-------------------|--|--------------------------------------------|--|---------------|--|-------------------|--|------------------------|--|--|--|--|--|---------------------|--|--|--|--|--|
| COMPANY: ENTACT, LLC                                                                                                              |  | PROJECT NAME: 540 Degraw St Brooklyn, NY                                                                                                                                                                         |  | BILL TO: ENTACT, LLC                                                                                                                                                       |  |                                                                                                   |  |                   | PO# E9309            |              |  |                   |  |                                            |  |               |  |                   |  |                        |  |  |  |  |  |                     |  |  |  |  |  |
| ADDRESS: 150 Bay Street, Suite 806                                                                                                |  | PROJECT #: E9309 LOCATION: Brooklyn, NY                                                                                                                                                                          |  | ADDRESS: 999 Oakmont Plaza Drive, Suite 300                                                                                                                                |  |                                                                                                   |  |                   |                      |              |  |                   |  |                                            |  |               |  |                   |  |                        |  |  |  |  |  |                     |  |  |  |  |  |
| CITY Jersey City STATE: NJ ZIP: 07302                                                                                             |  | PROJECT MANAGER: Austin Farmerie                                                                                                                                                                                 |  | CITY: Westmont                                                                                                                                                             |  |                                                                                                   |  |                   | STATE: IL ZIP: 60559 |              |  |                   |  |                                            |  |               |  |                   |  |                        |  |  |  |  |  |                     |  |  |  |  |  |
| ATTENTION: Austin Farmerie                                                                                                        |  | E-MAIL: afarmerie@entact.com                                                                                                                                                                                     |  | ATTENTION: Wendy Murray                                                                                                                                                    |  |                                                                                                   |  |                   | PHONE: 800-936-8228  |              |  |                   |  |                                            |  |               |  |                   |  |                        |  |  |  |  |  |                     |  |  |  |  |  |
| PHONE: 412-716-1366 FAX:                                                                                                          |  | PHONE: 412-716-1366 FAX:                                                                                                                                                                                         |  | ANALYSIS                                                                                                                                                                   |  |                                                                                                   |  |                   |                      |              |  |                   |  |                                            |  |               |  |                   |  |                        |  |  |  |  |  |                     |  |  |  |  |  |
| DATA TURNAROUND INFORMATION                                                                                                       |  | DATA DELIVERABLE INFORMATION                                                                                                                                                                                     |  | PRESERVATIVES                                                                                                                                                              |  |                                                                                                   |  |                   |                      |              |  |                   |  |                                            |  |               |  |                   |  |                        |  |  |  |  |  |                     |  |  |  |  |  |
| FAX: 3 DAYS*<br>HARD COPY: 3 DAYS*<br>EDD 3 DAYS*<br>* TO BE APPROVED BY ALLIANCE<br>STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS |  | <input type="checkbox"/> RESEULTS ONLY<br><input type="checkbox"/> RESULTS + QC<br><input type="checkbox"/> New Jersey REDUCED<br><input type="checkbox"/> New Jersey CLP<br><input type="checkbox"/> EDD Format |  | <input type="checkbox"/> USEPA CLP<br><input type="checkbox"/> New York State ASP "B"<br><input type="checkbox"/> New York State ASP "A"<br><input type="checkbox"/> Other |  | ASTM COD<br>ASTM Ammonia Nitrogen<br>ASTM O&G<br>ASTM TS<br>TS, TVS<br>pH (9045D)<br>Paint Filter |  |                   |                      |              |  |                   |  |                                            |  |               |  |                   |  |                        |  |  |  |  |  |                     |  |  |  |  |  |
|                                                                                                                                   |  |                                                                                                                                                                                                                  |  | COMMENTS                                                                                                                                                                   |  |                                                                                                   |  |                   |                      |              |  |                   |  |                                            |  |               |  |                   |  |                        |  |  |  |  |  |                     |  |  |  |  |  |
| CHEMTECH SAMPLE ID                                                                                                                |  | PROJECT SAMPLE IDENTIFICATION                                                                                                                                                                                    |  | SAMPLE MATRIX                                                                                                                                                              |  | SAMPLE TYPE                                                                                       |  | SAMPLE COLLECTION |                      | # of Bottles |  | E E E E E E E     |  |                                            |  |               |  |                   |  |                        |  |  |  |  |  |                     |  |  |  |  |  |
|                                                                                                                                   |  |                                                                                                                                                                                                                  |  |                                                                                                                                                                            |  | COMP GRAB                                                                                         |  | DATE TIME         |                      |              |  | 1 2 3 4 5 6 7 8 9 |  |                                            |  |               |  |                   |  |                        |  |  |  |  |  |                     |  |  |  |  |  |
| 1.                                                                                                                                |  | WC-A2-13-G                                                                                                                                                                                                       |  | Soil                                                                                                                                                                       |  |                                                                                                   |  | X                 |                      | 7/10 12:00   |  | 1                 |  |                                            |  |               |  |                   |  |                        |  |  |  |  |  |                     |  |  |  |  |  |
| 2.                                                                                                                                |  | WC-A2-13-C                                                                                                                                                                                                       |  | Soil                                                                                                                                                                       |  | X                                                                                                 |  |                   |                      | 7/10 12:00   |  | 11                |  | X X X X X X X                              |  |               |  |                   |  |                        |  |  |  |  |  |                     |  |  |  |  |  |
| 3.                                                                                                                                |  | WC-A2-14-G                                                                                                                                                                                                       |  | Soil                                                                                                                                                                       |  |                                                                                                   |  | X                 |                      | 7/10 12:00   |  | 1                 |  |                                            |  |               |  |                   |  |                        |  |  |  |  |  |                     |  |  |  |  |  |
| 4.                                                                                                                                |  | WC-A2-14-C                                                                                                                                                                                                       |  | Soil                                                                                                                                                                       |  | X                                                                                                 |  |                   |                      | 7/10 12:00   |  | 11                |  | X X X X X X X                              |  |               |  |                   |  |                        |  |  |  |  |  |                     |  |  |  |  |  |
| 5.                                                                                                                                |  |                                                                                                                                                                                                                  |  |                                                                                                                                                                            |  |                                                                                                   |  |                   |                      |              |  |                   |  |                                            |  |               |  |                   |  |                        |  |  |  |  |  |                     |  |  |  |  |  |
| 6.                                                                                                                                |  |                                                                                                                                                                                                                  |  |                                                                                                                                                                            |  |                                                                                                   |  |                   |                      |              |  |                   |  |                                            |  |               |  |                   |  |                        |  |  |  |  |  |                     |  |  |  |  |  |
| 7.                                                                                                                                |  |                                                                                                                                                                                                                  |  |                                                                                                                                                                            |  |                                                                                                   |  |                   |                      |              |  |                   |  |                                            |  |               |  |                   |  |                        |  |  |  |  |  |                     |  |  |  |  |  |
| 8.                                                                                                                                |  |                                                                                                                                                                                                                  |  |                                                                                                                                                                            |  |                                                                                                   |  |                   |                      |              |  |                   |  |                                            |  |               |  |                   |  |                        |  |  |  |  |  |                     |  |  |  |  |  |
| 9.                                                                                                                                |  |                                                                                                                                                                                                                  |  |                                                                                                                                                                            |  |                                                                                                   |  |                   |                      |              |  |                   |  |                                            |  |               |  |                   |  |                        |  |  |  |  |  |                     |  |  |  |  |  |
| 10.                                                                                                                               |  |                                                                                                                                                                                                                  |  |                                                                                                                                                                            |  |                                                                                                   |  |                   |                      |              |  |                   |  |                                            |  |               |  |                   |  |                        |  |  |  |  |  |                     |  |  |  |  |  |
| SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE PROSESSION INCLUDING COURIER DELIVERY.                           |  |                                                                                                                                                                                                                  |  |                                                                                                                                                                            |  |                                                                                                   |  |                   |                      |              |  |                   |  |                                            |  |               |  |                   |  |                        |  |  |  |  |  |                     |  |  |  |  |  |
| RELINQUISHED BY SAMPLER                                                                                                           |  | DATE/TIME                                                                                                                                                                                                        |  | RECEIVED BY                                                                                                                                                                |  | Conditions of bottles or coolers at receipt:                                                      |  |                   |                      |              |  |                   |  | Compliant                                  |  | Non Compliant |  | Cooler Temp       |  | Ice in Cooler?         |  |  |  |  |  |                     |  |  |  |  |  |
| 1. Austin Farmerie                                                                                                                |  | 7/10/25                                                                                                                                                                                                          |  | 1350                                                                                                                                                                       |  | Comments:                                                                                         |  |                   |                      |              |  |                   |  |                                            |  |               |  | 4.4 C             |  |                        |  |  |  |  |  |                     |  |  |  |  |  |
| RELINQUISHED BY                                                                                                                   |  | DATE/TIME                                                                                                                                                                                                        |  | RECEIVED BY                                                                                                                                                                |  |                                                                                                   |  |                   |                      |              |  |                   |  |                                            |  |               |  |                   |  |                        |  |  |  |  |  |                     |  |  |  |  |  |
| 2.                                                                                                                                |  |                                                                                                                                                                                                                  |  |                                                                                                                                                                            |  |                                                                                                   |  |                   |                      |              |  |                   |  |                                            |  |               |  |                   |  |                        |  |  |  |  |  |                     |  |  |  |  |  |
| RELINQUISHED BY                                                                                                                   |  | DATE/TIME                                                                                                                                                                                                        |  | RECEIVED FOR LAB BY                                                                                                                                                        |  | SHIPPED VIA: CLIENT:                                                                              |  |                   |                      |              |  |                   |  | Hand Delivered                             |  | Overnight     |  | Shipment Complete |  | YES NO                 |  |  |  |  |  |                     |  |  |  |  |  |
| 3.                                                                                                                                |  | 7-10-25                                                                                                                                                                                                          |  | 1527                                                                                                                                                                       |  | ALLIANCE:                                                                                         |  |                   |                      |              |  |                   |  | Picked Up                                  |  | Overnight     |  |                   |  |                        |  |  |  |  |  |                     |  |  |  |  |  |
| Page _____ of _____                                                                                                               |  |                                                                                                                                                                                                                  |  |                                                                                                                                                                            |  |                                                                                                   |  |                   |                      |              |  |                   |  | WHITE - ALLIANCE COPY FOR RETURN TO CLIENT |  |               |  |                   |  | YELLOW - ALLIANCE COPY |  |  |  |  |  | PINK - SAMPLER COPY |  |  |  |  |  |

### Laboratory Certification

| Certified By         | License No.      |
|----------------------|------------------|
|                      |                  |
| CAS EPA CLP Contract | 68HERH20D0011    |
|                      |                  |
| Connecticut          | PH-0830          |
|                      |                  |
| DOD ELAP (ANAB)      | L2219            |
|                      |                  |
| Maine                | 2024021          |
|                      |                  |
| Maryland             | 296              |
|                      |                  |
| New Hampshire        | 255424 Rev 1     |
|                      |                  |
| New Jersey           | 20012            |
|                      |                  |
| New York             | 11376            |
|                      |                  |
| Pennsylvania         | 68-00548         |
|                      |                  |
| Soil Permit          | 525-24-234-08441 |
|                      |                  |
| Texas                | T104704488       |