

Preparation Blank Summary

Client: ENTACT

SDG No.: Q2579

Project: 540 Degraw St, Brooklyn, NY - E9309

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB136448BL TS	%	< 2.5000	2.5000	U	1	5	07/11/2025
Sample ID: LB136449BL TVS	%	< 5.0000	5.0000	U	1	10	07/11/2025
Sample ID: LB136455BL ASTM TS	mg/L	< 2.5000	2.5000	U	1	5	07/14/2025
Sample ID: LB136456BL ASTM Oil and Grease	mg/L	< 2.5000	2.5000	U	0.29	5.0	07/14/2025
Sample ID: LB136474BL Oil and Grease	mg/Kg	< 12.5000	12.5000	U	5.8	25	07/15/2025
Sample ID: LB136478BL ASTM COD	mg/L	< 5.0000	5.0000	U	1.5	10.0	07/15/2025
Sample ID: PB168800BL Reactive Sulfide	mg/Kg	< 5.0000	5.0000	U	0.201	10	07/11/2025
Sample ID: PB168801BL Reactive Cyanide	mg/Kg	< 0.0250	0.0250	U	0.0084	0.05	07/11/2025
Sample ID: PB168836BL ASTM Ammonia	mg/L	< 0.0500	0.0500	U	0.03	0.1	07/14/2025