

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q2579

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB136456

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136456BS							
ASTM Oil and Grease	mg/L	20.0	16.7		84	1	78-114	07/14/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q2579

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB136474

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136474BS							
Oil and Grease	mg/Kg	100	94.9		95	1	80-120	07/15/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q2579

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB136478

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136478BS							
ASTM COD	mg/L	50	51.0		102	1	90-110	07/15/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q2579

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB136468

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168836BS							
ASTM Ammonia	mg/L	1	1.00		100	1	90-110	07/14/2025