

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>ENTACT</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q2579</u>
Contract:	<u>ENTA05</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q2586-04</u>	Client ID:	<u>TP-16DUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q2586-04DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	1170		1110		5		P
Cadmium	ug/L	20	30.0	U	30.0	U			P
Chromium	ug/L	20	50.0	U	50.0	U			P
Copper	ug/L	20	44.7	J	42.3	J	6		P
Lead	ug/L	20	60.0	U	60.0	U			P
Mercury	ug/L	20	2.00	U	2.00	U			CV
Nickel	ug/L	20	200	U	200	U			P
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P
Zinc	ug/L	20	188	J	179	J	5		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>ENTACT</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q2579</u>
Contract:	<u>ENTA05</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q2586-04MS</u>	Client ID:	<u>TP-16MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q2586-04MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	4310		4300		0		P
Barium	ug/L	20	2210		2220		0		P
Cadmium	ug/L	20	1000		1000		0		P
Chromium	ug/L	20	1920		1920		0		P
Copper	ug/L	20	1580		1570		1		P
Lead	ug/L	20	4840		4810		1		P
Mercury	ug/L	20	46.9		46.8		0		CV
Nickel	ug/L	20	2580		2570		0		P
Selenium	ug/L	20	11200		11200		0		P
Silver	ug/L	20	317		288		10		P
Zinc	ug/L	20	1360		1350		1		P