

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance
Contract: ENTA05

Lab Code: ACE
SDG NO.: Q2579

Instrument ID: ECD_P
Date(s) Analyzed: 07/07/2025 07/08/2025

GC Column: ZB-MR1
ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.69	4.59	4.79	17866400
		2	4.78	4.68	4.88	13504500
		3	4.85	4.75	4.95	41604800
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.85	4.75	4.95	32650800
		2	5.37	5.27	5.47	16254700
		3	5.66	5.56	5.76	32993400
		4	5.82	5.72	5.92	16445200
		5	5.91	5.81	6.01	10570400
Aroclor-1262	500	1	8.07	7.97	8.17	87027200
		2	8.38	8.28	8.48	198028000
		3	8.70	8.60	8.80	125725000
		4	8.78	8.68	8.88	93262800
		5	9.43	9.33	9.53	63389400

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance
Contract: ENTA05

Lab Code: ACE
SDG NO.: Q2579

Instrument ID: ECD_P
Date(s) Analyzed: 07/07/2025 07/08/2025

GC Column: ZB-MR2
ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	3.99	3.89	4.09	27045000
		2	4.08	3.98	4.18	20360400
		3	4.15	4.05	4.25	61448600
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.15	4.05	4.25	46614400
		2	4.88	4.78	4.98	47659800
		3	5.05	4.95	5.15	25029000
		4	5.14	5.04	5.24	21656200
		5	5.31	5.21	5.41	22563400
Aroclor-1262	500	1	6.89	6.79	6.99	147692000
		2	7.15	7.05	7.25	127829000
		3	7.67	7.57	7.77	114019000
		4	7.74	7.64	7.84	184049000
		5	8.23	8.13	8.33	84435600