

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q2579

Continuing Calib Date: 07/14/2025

Initial Calibration Date(s): 07/08/2025

07/08/2025

Continuing Calib Time: 17:01

Initial Calibration Time(s): 14:06

22:16

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1254-1 (1)   | 5.56       | 5.57      | 5.47      | 5.67 | 0.01       |
| Aroclor-1254-2 (2)   | 5.71       | 5.72      | 5.62      | 5.82 | 0.01       |
| Aroclor-1254-3 (3)   | 6.12       | 6.12      | 6.02      | 6.22 | 0.00       |
| Aroclor-1254-4 (4)   | 6.35       | 6.35      | 6.25      | 6.45 | 0.00       |
| Aroclor-1254-5 (5)   | 6.76       | 6.77      | 6.67      | 6.87 | 0.01       |
| Tetrachloro-m-xylene | 3.67       | 3.67      | 3.57      | 3.77 | 0.00       |
| Decachlorobiphenyl   | 8.70       | 8.70      | 8.60      | 8.80 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q2579

Continuing Calib Date: 07/14/2025

Initial Calibration Date(s): 07/08/2025

07/08/2025

Continuing Calib Time: 17:01

Initial Calibration Time(s): 14:06

22:16

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1254-1 (1)   | 5.54       | 5.54      | 5.44      | 5.64 | 0.00       |
| Aroclor-1254-2 (2)   | 5.69       | 5.69      | 5.59      | 5.79 | 0.00       |
| Aroclor-1254-3 (3)   | 6.09       | 6.09      | 5.99      | 6.19 | 0.00       |
| Aroclor-1254-4 (4)   | 6.32       | 6.32      | 6.22      | 6.42 | 0.00       |
| Aroclor-1254-5 (5)   | 6.73       | 6.73      | 6.63      | 6.83 | 0.00       |
| Tetrachloro-m-xylene | 3.67       | 3.67      | 3.57      | 3.77 | 0.01       |
| Decachlorobiphenyl   | 8.65       | 8.65      | 8.55      | 8.75 | 0.00       |