

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** ENTA05  
**Lab Code:** ACE **SDG NO.:** Q2579  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

**Client Sample No.:** CCAL02 **Date Analyzed:** 07/14/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO112270.D **Time Analyzed:** 20:16

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Aroclor-1016-1       | 4.758 | 4.661     | 4.861 | 566.300            | 500.000           | 13.3  |
| Aroclor-1016-2       | 4.776 | 4.680     | 4.880 | 570.420            | 500.000           | 14.1  |
| Aroclor-1016-3       | 4.833 | 4.736     | 4.936 | 552.280            | 500.000           | 10.5  |
| Aroclor-1016-4       | 4.953 | 4.856     | 5.056 | 566.420            | 500.000           | 13.3  |
| Aroclor-1016-5       | 5.210 | 5.113     | 5.313 | 597.680            | 500.000           | 19.5  |
| Aroclor-1260-1       | 6.247 | 6.151     | 6.351 | 449.180            | 500.000           | -10.2 |
| Aroclor-1260-2       | 6.437 | 6.341     | 6.541 | 439.890            | 500.000           | -12.0 |
| Aroclor-1260-3       | 6.803 | 6.707     | 6.907 | 558.730            | 500.000           | 11.7  |
| Aroclor-1260-4       | 7.062 | 6.967     | 7.167 | 544.230            | 500.000           | 8.8   |
| Aroclor-1260-5       | 7.304 | 7.209     | 7.409 | 444.700            | 500.000           | -11.1 |
| Decachlorobiphenyl   | 8.697 | 8.603     | 8.803 | 43.550             | 50.000            | -12.9 |
| Tetrachloro-m-xylene | 3.670 | 3.573     | 3.773 | 59.970             | 50.000            | 19.9  |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** ENTA05  
**Lab Code:** ACE **SDG NO.:** Q2579  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

**Client Sample No.:** CCAL02 **Date Analyzed:** 07/14/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO112270.D **Time Analyzed:** 20:16

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.738 | 4.661     | 4.861 | 524.110            | 500.000           | 4.8  |
| Aroclor-1016-2       | 4.758 | 4.679     | 4.879 | 548.630            | 500.000           | 9.7  |
| Aroclor-1016-3       | 4.932 | 4.854     | 5.054 | 544.810            | 500.000           | 9.0  |
| Aroclor-1016-4       | 4.975 | 4.897     | 5.097 | 530.620            | 500.000           | 6.1  |
| Aroclor-1016-5       | 5.187 | 5.109     | 5.309 | 550.020            | 500.000           | 10.0 |
| Aroclor-1260-1       | 6.216 | 6.138     | 6.338 | 507.230            | 500.000           | 1.4  |
| Aroclor-1260-2       | 6.404 | 6.326     | 6.526 | 505.440            | 500.000           | 1.1  |
| Aroclor-1260-3       | 6.555 | 6.478     | 6.678 | 496.400            | 500.000           | -0.7 |
| Aroclor-1260-4       | 7.026 | 6.948     | 7.148 | 460.080            | 500.000           | -8.0 |
| Aroclor-1260-5       | 7.267 | 7.190     | 7.390 | 466.320            | 500.000           | -6.7 |
| Decachlorobiphenyl   | 8.645 | 8.568     | 8.768 | 47.560             | 50.000            | -4.9 |
| Tetrachloro-m-xylene | 3.663 | 3.584     | 3.784 | 57.170             | 50.000            | 14.3 |