

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2579
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089561.D **Time Analyzed:** 09:31

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.079	8.971	9.171	49.340	50.000	-1.3
Endrin	6.580	6.472	6.672	51.200	50.000	2.4
gamma-BHC (Lindane)	4.336	4.228	4.428	51.290	50.000	2.6
Heptachlor	4.935	4.827	5.027	51.880	50.000	3.8
Heptachlor epoxide	5.696	5.588	5.788	50.590	50.000	1.2
Methoxychlor	7.499	7.392	7.592	50.410	50.000	0.8
Tetrachloro-m-xylene	3.556	3.448	3.648	50.160	50.000	0.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2579
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089561.D **Time Analyzed:** 09:31

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.070	7.970	8.170	50.180	50.000	0.4
Endrin	5.787	5.687	5.887	50.870	50.000	1.7
gamma-BHC (Lindane)	3.727	3.627	3.827	49.690	50.000	-0.6
Heptachlor	4.081	3.980	4.180	48.980	50.000	-2.0
Heptachlor epoxide	4.871	4.770	4.970	49.620	50.000	-0.8
Methoxychlor	6.752	6.651	6.851	48.240	50.000	-3.5
Tetrachloro-m-xylene	2.878	2.779	2.979	49.030	50.000	-1.9