

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance
Contract: ENTA05
Lab Code: ACE
SDG NO.: Q2579
GC Column: ZB-MR1
ID: 0.32 (mm)
 Initi. Calib. Date(s): 07/21/2025 07/21/2025

Client Sample No.: CCAL02
Date Analyzed: 07/22/2025
Lab Sample No.: PSTDCCC050
Data File : PD089575.D
Time Analyzed: 13:31

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.076	8.971	9.171	49.360	50.000	-1.3
Endrin	6.578	6.472	6.672	50.990	50.000	2.0
gamma-BHC (Lindane)	4.335	4.228	4.428	51.470	50.000	2.9
Heptachlor	4.933	4.827	5.027	51.230	50.000	2.5
Heptachlor epoxide	5.695	5.588	5.788	50.270	50.000	0.5
Methoxychlor	7.497	7.392	7.592	49.720	50.000	-0.6
Tetrachloro-m-xylene	3.554	3.448	3.648	49.540	50.000	-0.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2579
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089575.D **Time Analyzed:** 13:31

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.070	7.970	8.170	50.950	50.000	1.9
Endrin	5.788	5.687	5.887	53.120	50.000	6.2
gamma-BHC (Lindane)	3.728	3.627	3.827	50.930	50.000	1.9
Heptachlor	4.081	3.980	4.180	50.660	50.000	1.3
Heptachlor epoxide	4.871	4.770	4.970	50.600	50.000	1.2
Methoxychlor	6.752	6.651	6.851	50.350	50.000	0.7
Tetrachloro-m-xylene	2.879	2.779	2.979	50.720	50.000	1.4