

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2579</u>	Analytical Method:	<u>8081B</u>
Client:	<u>ENTACT</u>	DataFile :	<u>PD089567.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec	RPD		Limits		
			Result				Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q2578-03MSD	Client Sample ID:		WC-A5-01-CMSD								
	(Column 1)											
	gamma-BHC (Lindane)	5	0	4.90	ug/L	98		2		30 (60)	150 (152)	20 (20)
	Heptachlor	5	0	4.60	ug/L	92		0		30 (56)	150 (147)	20 (20)
	Heptachlor epoxide	5	0	4.70	ug/L	94		2		30 (77)	150 (143)	20 (20)
	Endrin	5	0	4.80	ug/L	96		0		30 (76)	150 (144)	20 (20)
	Methoxychlor	5	0	4.80	ug/L	96		0		30 (70)	150 (142)	20 (20)
Lab Sample ID:	Q2578-03MSD	Client Sample ID:		WC-A5-01-CMSD								
	(Column 2)											
	gamma-BHC (Lindane)	5	0	4.40	ug/L	88		2		30 (60)	150 (152)	20 (20)
	Heptachlor	5	0	4.50	ug/L	90		2		30 (56)	150 (147)	20 (20)
	Heptachlor epoxide	5	0	4.60	ug/L	92		0		30 (77)	150 (143)	20 (20)
	Endrin	5	0	4.80	ug/L	96		0		30 (76)	150 (144)	20 (20)
	Methoxychlor	5	0	4.80	ug/L	96		2		30 (70)	150 (142)	20 (20)