

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>ENTA05</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2579</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>07/14/2025</u> <u>09:16</u>
Lab File ID:	<u>VX046961.D</u>	Init. Calib. Date(s):	<u>07/02/2025</u> <u>07/02/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>12:11</u> <u>14:31</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.569	0.565		-0.7	20
1,1-Dichloroethene	0.542	0.553		2.03	20
2-Butanone	0.274	0.340		24.09	20
Carbon Tetrachloride	0.484	0.506		4.55	20
Chloroform	1.087	1.141		4.97	20
Benzene	1.385	1.391		0.43	20
1,2-Dichloroethane	0.470	0.497		5.74	20
Trichloroethene	0.361	0.355		-1.66	20
Tetrachloroethene	0.333	0.320		-3.9	20
Chlorobenzene	1.081	1.090	0.3	0.83	20
1,2-Dichloroethane-d4	0.661	0.674		1.97	20
Dibromofluoromethane	0.339	0.336		-0.88	20
Toluene-d8	1.197	1.136		-5.1	20
4-Bromofluorobenzene	0.455	0.448		-1.54	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.