

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX070225

Review By	John Carlone	Review On	7/3/2025 8:44:30 AM
Supervise By	Maresh Dadoda	Supervise On	7/3/2025 3:09:59 PM
SubDirectory	VX070225	HP Acquire Method	HP Processing Method 82X070225W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134642 VP134633,VP134634,VP134635,VP134638,VP134639,VP134640 VP134641

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046859.D	02 Jul 2025 11:12		JC/MD	Ok
2	VSTDICC001	VSTDICC001	VX046860.D	02 Jul 2025 12:11		JC/MD	Ok,M
3	VSTDICC005	VSTDICC005	VX046861.D	02 Jul 2025 12:37	%D Failed for com.#13 in 5 ppb	JC/MD	Ok,M
4	VSTDICC020	VSTDICC020	VX046862.D	02 Jul 2025 13:18		JC/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VX046863.D	02 Jul 2025 13:39	LR-13	JC/MD	Ok,M
6	VSTDICC100	VSTDICC100	VX046864.D	02 Jul 2025 14:10		JC/MD	Ok,M
7	VSTDICC150	VSTDICC150	VX046865.D	02 Jul 2025 14:31		JC/MD	Ok,M
8	VIBLK	VIBLK	VX046866.D	02 Jul 2025 14:53		JC/MD	Ok
9	VSTDICV050	ICVVX070225	VX046867.D	02 Jul 2025 15:14		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX071425

Review By	Maresh Dadoda	Review On	7/15/2025 4:54:26 AM
Supervise By	Semsettin Yesilyurt	Supervise On	7/15/2025 4:55:39 AM
SubDirectory	VX071425	HP Acquire Method	HP Processing Method 82X070225W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134751
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134752,VP134753

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046960.D	14 Jul 2025 08:10		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX046961.D	14 Jul 2025 09:16	pH#Lot#V12668	JC/MD	Ok,M
3	VX0714MBL01	VX0714MBL01	VX046962.D	14 Jul 2025 10:54		JC/MD	Ok
4	VX0714WBL01	VX0714WBL01	VX046963.D	14 Jul 2025 11:18		JC/MD	Ok
5	VX0714MBS01	VX0714MBS01	VX046964.D	14 Jul 2025 11:39		JC/MD	Ok,M
6	VX0714MBSD01	VX0714MBSD01	VX046965.D	14 Jul 2025 12:06		JC/MD	Ok,M
7	Q2480-04ME	GPX4ME	VX046966.D	14 Jul 2025 12:27		JC/MD	Ok
8	Q2480-02ME	GPX2ME	VX046967.D	14 Jul 2025 12:49		JC/MD	Ok,M
9	VX0714WBS01	VX0714WBS01	VX046968.D	14 Jul 2025 13:10		JC/MD	Ok,M
10	VX0714WBSD01	VX0714WBSD01	VX046969.D	14 Jul 2025 13:37		JC/MD	Ok,M
11	Q2523-01	MOO-25-0190	VX046970.D	14 Jul 2025 13:58	Need Soil Run	JC/MD	Not Ok
12	Q2526-02	LAW-25-0101	VX046971.D	14 Jul 2025 14:19	wipe sample	JC/MD	Ok
13	Q2565-03	MOO-25-0191	VX046972.D	14 Jul 2025 14:41	wipe sample	JC/MD	Ok
14	Q2565-04	MOO-25-0196	VX046973.D	14 Jul 2025 15:02	wipe sample	JC/MD	Ok,M
15	Q2565-05	MOO-25-0180	VX046974.D	14 Jul 2025 15:23	Need Soil Run	JC/MD	Not Ok
16	IBLK	IBLK	VX046975.D	14 Jul 2025 15:45		JC/MD	Ok
17	Q2578-01	WC-A5-01-G	VX046976.D	14 Jul 2025 16:06	vial A pH#5.0 Surrogate fail	JC/MD	ReRun
18	Q2578-05	WC-A2-09-G	VX046977.D	14 Jul 2025 16:28	vial A pH#5.0 Surrogate fail	JC/MD	ReRun

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX071425

Review By	Maresh Dadoda	Review On	7/15/2025 4:54:26 AM
Supervise By	Semsettin Yesilyurt	Supervise On	7/15/2025 4:55:39 AM
SubDirectory	VX071425	HP Acquire Method	HP Processing Method 82X070225W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134751		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134752,VP134753		

19	Q2578-09	WC-A2-10-G	VX046978.D	14 Jul 2025 16:49	vial A pH#5.0 Surrogate fail	JC/MD	ReRun
20	Q2578-13	WC-A2-11-G	VX046979.D	14 Jul 2025 17:10	vial A pH#5.0 Surrogate fail	JC/MD	ReRun
21	Q2578-17	WC-A2-12-G	VX046980.D	14 Jul 2025 17:31	vial A pH#5.0 Surrogate fail	JC/MD	ReRun
22	Q2579-01	WC-A2-13-G	VX046981.D	14 Jul 2025 17:53	vial A pH#5.0 Surrogate fail	JC/MD	ReRun
23	Q2579-05	WC-A2-14-G	VX046982.D	14 Jul 2025 18:14	vial A pH#5.0 Surrogate fail	JC/MD	ReRun
24	IBLK	IBLK	VX046983.D	14 Jul 2025 18:35		JC/MD	Ok
25	Q2553-05	FB	VX046984.D	14 Jul 2025 18:56	vial B pH<2 FB	JC/MD	Ok
26	Q2595-01	CHRT-24849	VX046985.D	14 Jul 2025 19:17	jelly like sample	JC/MD	Ok
27	VSTDCCC050	VSTDCCC050EC	VX046986.D	14 Jul 2025 19:38		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX071525

Review By	John Carlone	Review On	7/16/2025 8:51:22 AM
Supervise By		Supervise On	
SubDirectory	VX071525	HP Acquire Method	HP Processing Method 82X070225W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134768
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134769,VP134770

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046987.D	15 Jul 2025 08:29		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX046988.D	15 Jul 2025 10:18	CCC fail High com. #25	JC/MD	Ok,NS
3	VX0715MBL01	VX0715MBL01	VX046989.D	15 Jul 2025 10:39		JC/MD	Ok
4	VX0715WBL01	VX0715WBL01	VX046990.D	15 Jul 2025 11:00		JC/MD	Ok
5	VX0715WBS01	VX0715WBS01	VX046991.D	15 Jul 2025 11:21		JC/MD	Ok,NS
6	VX0715WBSD01	VX0715WBSD01	VX046992.D	15 Jul 2025 11:47		JC/MD	Ok,NS
7	Q2578-01RE	WC-A5-01-GRE	VX046993.D	15 Jul 2025 12:08	vial B pH#5.0 Surrogate Fail	JC/MD	Confirms
8	Q2578-05RE	WC-A2-09-GRE	VX046994.D	15 Jul 2025 12:30	vial B pH#5.0 Surrogate Fail	JC/MD	Confirms
9	Q2578-09RE	WC-A2-10-GRE	VX046995.D	15 Jul 2025 12:51	vial B pH#5.0 Surrogate Fail	JC/MD	Confirms
10	Q2578-13RE	WC-A2-11-GRE	VX046996.D	15 Jul 2025 13:12	vial B pH#5.0 Surrogate Fail	JC/MD	Confirms
11	Q2578-17RE	WC-A2-12-GRE	VX046997.D	15 Jul 2025 13:33	vial B pH#5.0 Surrogate Fail	JC/MD	Confirms
12	Q2579-01RE	WC-A2-13-GRE	VX046998.D	15 Jul 2025 13:54	vial B pH#5.0 Surrogate Fail	JC/MD	Confirms
13	Q2579-05RE	WC-A2-14-GRE	VX046999.D	15 Jul 2025 14:16	vial B pH#5.0 Surrogate Fail	JC/MD	Confirms
14	Q2565-01	MOO-25-0192-0193	VX047000.D	15 Jul 2025 14:37	vial A pH<2 foamy sample	JC/MD	Ok
15	Q2602-01	FRAC-TANK-266380	VX047001.D	15 Jul 2025 14:58	vial A pH<2 CCC Failed High,BS,BSD Failed High	JC/MD	ReRun
16	Q2571-04	TP-18	VX047002.D	15 Jul 2025 15:20	vial A pH#5.0	JC/MD	Ok,NS
17	Q2571-08	TP-17	VX047003.D	15 Jul 2025 15:41	vial A pH#5.0	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX071525

Review By	John Carlone	Review On	7/16/2025 8:51:22 AM
Supervise By		Supervise On	
SubDirectory	VX071525	HP Acquire Method	HP Processing Method 82X070225W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134768		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134769,VP134770		

18	Q2586-04	TP-16	VX047004.D	15 Jul 2025 16:02	vial A pH#5.0	JC/MD	Ok,NS
19	Q2561-07	AUD-25-0117	VX047005.D	15 Jul 2025 16:24	vial A pH#5.0	JC/MD	Ok
20	PB168804TB	PB168804TB	VX047006.D	15 Jul 2025 16:45		JC/MD	Ok,NS
21	VX0715MBS01	VX0715MBS01	VX047007.D	15 Jul 2025 17:06		JC/MD	Ok,NS
22	VX0715MBSD01	VX0715MBSD01	VX047008.D	15 Jul 2025 17:27		JC/MD	Ok,NS
23	Q2503-03ME	GCAP2ME	VX047009.D	15 Jul 2025 17:49		JC/MD	Ok
24	Q2503-04ME	GCAP3ME	VX047010.D	15 Jul 2025 18:10		JC/MD	Ok
25	Q2600-03	TRENCH	VX047011.D	15 Jul 2025 18:31	Need Soil Run	JC/MD	Not Ok
26	Q2600-07	STOCK-PILE	VX047012.D	15 Jul 2025 18:52	Need Soil Run	JC/MD	Not Ok
27	Q2600-11	END-OF-TRENCH	VX047013.D	15 Jul 2025 19:13	Need 20X	JC/MD	Dilution
28	VSTDCCC050	VSTDCCC050EC	VX047014.D	15 Jul 2025 19:34		JC/MD	Ok,NS

M : Manual Integration